

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG012615\
 Data File : BG015939.D
 Acq On : 27 Jan 2015 9:44
 Operator : TP/IZ
 Sample : G1139-03
 Misc : S
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 121B2B

Quant Time: Jan 27 10:32:11 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 20 17:47:40 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	92052	20.00	ng	0.00
21) Naphthalene-d8	10.47	136	415862	20.00	ng	0.00
38) Acenaphthene-d10	14.34	164	340153	20.00	ng	0.00
63) Phenanthrene-d10	17.09	188	715517	20.00	ng	0.01
75) Chrysene-d12	21.27	240	962850	20.00	ng	0.00
86) Perylene-d12	23.52	264	990591	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.28	112	440638	71.28	ng	0.00
7) Phenol-d6	6.87	99	682600	68.52	ng	0.00
23) Nitrobenzene-d5	8.84	82	633407	43.91	ng	0.00
41) 2,4,6-Tribromophenol	15.84	330	379935	61.01	ng	0.01
44) 2-Fluorobiphenyl	12.96	172	1096157	50.56	ng	0.00
78) Terphenyl-d14	19.72	244	2223165	59.00	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	10.53	128	1051901	49.42	ng	# 92
37) 2-Methylnaphthalene	12.16	142	3343164	197.78	ng	97
45) 1,1'-Biphenyl	13.18	154	726324	33.83	ng	93
51) Acenaphthene	14.41	154	271357	15.69	ng	# 82
57) Fluorene	15.40	166	680429	26.43	ng	# 92
70) Phenanthrene	17.14	178	2283461	63.61	ng	# 93
74) Fluoranthene	19.16	202	147485	3.08	ng	73
77) Pyrene	19.51	202	730149	15.96	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG012615\
Data File : BG015939.D
Acq On : 27 Jan 2015 9:44
Operator : TP/IZ
Sample : G1139-03
Misc : S
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
121B2B

Quant Time: Jan 27 10:32:11 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 20 17:47:40 2015
Response via : Initial Calibration

