

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG012615\
 Data File : BG015940.D
 Acq On : 27 Jan 2015 10:18
 Operator : TP/IZ
 Sample : G1139-04
 Misc : S
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 121B2C

Quant Time: Jan 27 11:48:01 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 20 17:47:40 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	96424	20.00	ng	0.00
21) Naphthalene-d8	10.47	136	397258	20.00	ng	0.00
38) Acenaphthene-d10	14.34	164	336169	20.00	ng	0.00
63) Phenanthrene-d10	17.10	188	708582	20.00	ng	0.01
75) Chrysene-d12	21.27	240	1009559	20.00	ng	0.00
86) Perylene-d12	23.52	264	1015630	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.28	112	613208	94.70	ng	0.00
7) Phenol-d6	6.87	99	946810	90.74	ng	0.00
23) Nitrobenzene-d5	8.85	82	827296	60.03	ng	0.00
41) 2,4,6-Tribromophenol	15.84	330	555069	90.19	ng	0.01
44) 2-Fluorobiphenyl	12.96	172	1444376	67.41	ng	0.00
78) Terphenyl-d14	19.72	244	2626326	69.36	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	10.53	128	1358679	66.82	ng	# 92
37) 2-Methylnaphthalene	12.16	142	3733892	231.23	ng	97
45) 1,1'-Biphenyl	13.17	154	783609	36.93	ng	94
49) Dimethylphthalate	13.80	163	244530	10.11	ng	# 90
51) Acenaphthene	14.40	154	242763	14.20	ng	# 89
57) Fluorene	15.39	166	603345	23.71	ng	# 86
70) Phenanthrene	17.14	178	1926573	54.19	ng	# 95
77) Pyrene	19.52	202	550027	11.46	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG012615\
Data File : BG015940.D
Acq On : 27 Jan 2015 10:18
Operator : TP/IZ
Sample : G1139-04
Misc : S
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
121B2C

Quant Time: Jan 27 11:48:01 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 20 17:47:40 2015
Response via : Initial Calibration

