

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG012617\
 Data File : BG025661.D
 Acq On : 26 Jan 2017 10:53
 Operator : SJ/MA
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleID :
 SSTD02020

Quant Time: Jan 27 03:16:18 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG011817.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 25 18:08:33 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.18	152	71381	20.00	ng/ul	0.00
18) Naphthalene-d8	11.00	136	321042	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.81	164	266828	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.65	188	533361	20.00	ng/ul	0.10
75) Chrysene-d12	21.85	240	679996	20.00	ng/ul	0.00
83) Perylene-d12	25.24	264	692023	20.00	ng/ul	0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.52	96	10055	7.05	ng/uL	0.00
5) Phenol-d5	7.34	99	131441	21.48	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.49	67	87965	22.26	ng/ul	0.00
9) 2-Chlorophenol-d4	7.70	132	94704	21.95	ng/ul	0.00
13) 4-Methylphenol-d8	8.90	113	108441	20.85	ng/ul	0.00
19) Nitrobenzene-d5	9.36	128	48890	21.48	ng/ul	0.00
22) 2-Nitrophenol-d4	10.09	143	58127	21.41	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.64	165	113464	21.18	ng/ul	0.00
29) 4-Chloroaniline-d4	11.15	131	128643	21.48	ng/ul	0.00
43) Dimethylphthalate-d6	14.20	166	380841	20.21	ng/ul	0.00
46) Acenaphthylene-d8	14.51	160	449287	22.07	ng/ul	0.00
51) 4-Nitrophenol-d4	15.04	143	47903	18.55	ng/ul	0.00
57) Fluorene-d10	15.80	176	360627	20.30	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.94	200	68949	20.96	ng/ul	0.00
70) Anthracene-d10	17.65	188	533361	23.49	ng/ul	0.00
76) Pyrene-d10	19.93	212	657742	23.25	ng/ul	0.00
87) Benzo(a)pyrene-d12	25.00	264	641133	22.04	ng/ul	0.00

Target Compounds

					Qvalue
2) 1,4-Dioxane	3.55	88	12896	7.65 ng/uL#	85
4) Benzaldehyde	7.32	77	106494	25.28 ng/ul	93
6) Phenol	7.36	94	140242	21.82 ng/ul	93
8) Bis(2-Chloroethyl)ether	7.59	93	106174	23.13 ng/ul	97
10) 2-Chlorophenol	7.74	128	95720	22.40 ng/ul	94
11) 2-Methylphenol	8.63	108	101146	21.13 ng/ul	96
12) 2,2'-oxybis(1-Chloropropan	8.70	45	206075	23.63 ng/ul	98
14) Acetophenone	9.01	105	168859	20.79 ng/ul	97
15) N-Nitroso-di-n-propylamine	8.98	70	106264	21.86 ng/ul#	89
16) 4-Methylphenol	8.96	108	113675	21.66 ng/ul	92
17) Hexachloroethane	9.26	117	46833	23.53 ng/ul	93
20) Nitrobenzene	9.40	77	158324	21.73 ng/ul#	91
21) Isophorone	9.92	82	283538	20.72 ng/ul	96
23) 2-Nitrophenol	10.12	139	63024	22.72 ng/ul	89
24) 2,4-Dimethylphenol	10.17	107	143345	21.44 ng/ul	96
25) Bis(2-Chloroethoxy)methane	10.40	93	151331	21.92 ng/ul	99
28) Naphthalene	11.05	128	318861	21.51 ng/ul	96
30) 4-Chloroaniline	11.18	127	131374	21.78 ng/ul	91
31) Hexachlorobutadiene	11.31	225	95053	20.91 ng/ul	96
32) Caprolactam	11.94	113	43029	19.68 ng/ul	87
33) 4-Chloro-3-methylphenol	12.29	107	138177	21.87 ng/ul	93
34) 2-Methylnaphthalene	12.65	142	253183	20.92 ng/ul	91
36) 1,2,4,5-Tetrachlorobenzene	13.01	216	175064	21.79 ng/ul	96

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG012617\
 Data File : BG025661.D
 Acq On : 26 Jan 2017 10:53
 Operator : SJ/MA
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SSTD02020

Quant Time: Jan 27 03:16:18 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG011817.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 25 18:08:33 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)

37) Hexachlorocyclopentadiene	12.97	237	66222	20.88	ng/ul#	86
38) 2,4,6-Trichlorophenol	13.35	196	112255	23.08	ng/ul	98
39) 2,4,5-Trichlorophenol	13.35	196	112255	22.01	ng/ul	93
40) 1,1'-Biphenyl	13.64	154	356483	21.56	ng/ul	99
41) 2-Chloronaphthalene	13.69	162	279377	21.77	ng/ul	97
42) 2-Nitroaniline	13.92	65	116212	20.89	ng/ul	94
44) Dimethylphthalate	14.25	163	364383	19.67	ng/ul#	97
45) 2,6-Dinitrotoluene	14.40	165	77925	20.43	ng/ul	94
47) Acenaphthylene	14.54	152	427388	21.57	ng/ul	97
48) 3-Nitroaniline	14.73	138	76327	22.12	ng/ul#	97
49) Acenaphthene	14.87	153	299115	21.19	ng/ul	96
50) 2,4-Dinitrophenol	14.97	184	48344	21.56	ng/ul#	77
52) 4-Nitrophenol	15.06	109	59523	16.64	ng/ul	89
53) Dibenzofuran	15.21	168	445654	20.54	ng/ul	98
54) 2,4-Dinitrotoluene	15.20	165	121944	20.96	ng/ul#	85
55) 2,3,4,6-Tetrachlorophenol	15.44	232	106474	19.73	ng/ul#	95
56) Diethylphthalate	15.60	149	395817	19.76	ng/ul	98
58) Fluorene	15.85	166	368937	20.82	ng/ul	98
59) 4-Chlorophenyl-phenylether	15.83	204	195129	19.52	ng/ul	95
60) 4-Nitroaniline	15.90	138	71674	21.46	ng/ul	94
64) N-Nitrosodiphenylamine	16.05	169	348226	25.33	ng/ul	87
65) 4-Bromophenyl-phenylether	16.73	248	147211	23.78	ng/ul	98
66) Hexachlorobenzene	16.85	284	160879	22.87	ng/ul	94
67) Atrazine	16.99	200	146095	22.62	ng/ul	98
69) Phenanthrene	17.69	178	625292	24.12	ng/ul	99
71) Anthracene	17.69	178	625292	23.99	ng/ul	100
72) Carbazole	17.96	167	545250	25.22	ng/ul	99
73) Di-n-butylphthalate	18.47	149	694823	24.86	ng/ul	98
74) Fluoranthene	19.60	202	758158	24.78	ng/ul	98
77) Pyrene	19.96	202	774075	23.15	ng/ul	99
78) Butylbenzylphthalate	20.81	149	308938	23.76	ng/ul	97
79) 3,3'-Dichlorobenzidine	21.73	252	309539	24.81	ng/ul	96
80) Benzo(a)anthracene	21.83	228	814959	22.68	ng/ul	99
81) Bis(2-ethylhexyl)phthalate	21.67	149	438561	24.16	ng/ul	97
82) Chrysene	21.89	228	755912	22.37	ng/ul	99
84) Di-n-octyl phthalate	22.92	149	739029	24.88	ng/ul	100
85) Benzo(b)fluoranthene	24.15	252	789483	22.04	ng/ul	98
86) Benzo(k)fluoranthene	24.22	252	755381	22.48	ng/ul	95
88) Benzo(a)pyrene	25.08	252	762917	22.52	ng/ul	97
89) Indeno(1,2,3-cd)pyrene	29.16	276	913423	21.29	ng/ul#	95
90) Dibenzo(a,h)anthracene	29.19	278	785553	21.98	ng/ul	95
91) Benzo(g,h,i)perylene	30.38	276	762353	21.64	ng/ul#	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 7.65 ng/uL

RT: 3.55 min Scan# 121

Delta R.T. 0.01 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53

Instrument :

BNA_G

ClientSampleId :

SSTD02020


Tgt Ion: 88 Resp: 12896

Ion Ratio Lower Upper




88 100

43 49.2 39.2 58.8

58 108.3 69.4 104.0#



Abundance Ion 88.00 (87.70 to 88.70): BGC
Ion 43.00 (42.70 to 43.70): BGC
Ion 58.00 (57.70 to 58.70): BGC



#4

Benzaldehyde

Concen: 25.28 ng/uL

RT: 7.32 min Scan# 762

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53


Tgt Ion: 77 Resp: 106494

Ion Ratio Lower Upper



77 100


105 85.9 62.0 93.0

106 79.7 60.2 90.4



Abundance Ion 77.00 (76.70 to 77.70): BGC
Ion 105.00 (104.70 to 105.70): BGC
Ion 106.00 (105.70 to 106.70): BGC





#6

Phenol

Concen: 21.82 ng/ul

RT: 7.36 min Scan# 770

Delta R.T. -0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53

Instrument :

BNA_G

ClientSampleId :

SSTD02020


Tgt Ion: 94 Resp: 140242

Ion Ratio Lower Upper




94 100

65 34.9 26.8 40.2

66 48.6 44.4 66.6



Abundance Ion 94.00 (93.70 to 94.70): BGC
Ion 65.00 (64.70 to 65.70): BGC
Ion 66.00 (65.70 to 66.70): BGC



#8

Bis(2-Chloroethyl)ether

Concen: 23.13 ng/ul

RT: 7.59 min Scan# 808

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53


Tgt Ion: 93 Resp: 106174

Ion Ratio Lower Upper



93 100


63 96.2 75.5 113.3

95 33.5 29.2 43.8



Abundance Ion 93.00 (92.70 to 93.70): BGC
Ion 63.00 (62.70 to 63.70): BGC
Ion 95.00 (94.70 to 95.70): BGC









#10
2-Chlorophenol
Concen: 22.40 ng/ul
RT: 7.74 min Scan# 834
Delta R.T. 0.00 min
Lab File: BG025661.D
Acq: 26 Jan 2017 10:53




Instrument :
BNA_G
ClientSampleId :
SSTD02020


Tgt Ion:128 Resp: 95720
Ion Ratio Lower Upper
128 100
64 64.9 56.3 84.5
130 31.4 26.7 40.1



#11
2-Methylphenol
Concen: 21.13 ng/ul
RT: 8.63 min Scan# 985
Delta R.T. 0.00 min
Lab File: BG025661.D
Acq: 26 Jan 2017 10:53

Tgt Ion:108 Resp: 101146
Ion Ratio Lower Upper
108 100
107 91.7 76.7 115.1









#12
 2,2'-oxybis(1-chloropropane)
 Concen: 23.63 ng/ul
 RT: 8.70 min Scan# 998
 Delta R.T. 0.00 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02020


Tgt Ion: 45 Resp: 206075
 Ion Ratio Lower Upper
 45 100
 77 12.3 10.0 15.0
 79 7.9 7.3 10.9



#14
 Acetophenone
 Concen: 20.79 ng/ul
 RT: 9.01 min Scan# 1050
 Delta R.T. 0.00 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53

Tgt Ion: 105 Resp: 168859
 Ion Ratio Lower Upper
 105 100
 77 94.7 76.0 114.0
 51 38.8 34.8 52.2









#15
 N-Nitroso-di-n-propylamine
 Concen: 21.86 ng/ul
 RT: 8.98 min Scan# 1045
 Delta R.T. 0.00 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02020


Tgt Ion: 70 Resp: 106264
 Ion Ratio Lower Upper
 70 100
 42 64.9 59.0 88.6
 101 11.5 6.6 9.8#
 130 14.2 15.6 23.4#



#16
 4-Methylphenol
 Concen: 21.66 ng/ul
 RT: 8.96 min Scan# 1042
 Delta R.T. 0.01 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53

Tgt Ion: 108 Resp: 113675
 Ion Ratio Lower Upper
 108 100
 107 106.0 91.7 137.5





#17

Hexachloroethane

Concen: 23.53 ng/ul

RT: 9.26 min Scan# 1092

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53

Instrument :

BNA_G

ClientSampleId :

SSTD02020


Tgt Ion:117 Resp: 46833

Ion Ratio Lower Upper

117 100

201 123.6 91.9 137.9




199 75.4 64.3 96.5



Abundance Ion 117.00 (116.70 to 117.70): E

Ion 201.00 (200.70 to 201.70): E

Ion 199.00 (198.70 to 199.70): E



#20

Nitrobenzene

Concen: 21.73 ng/ul

RT: 9.40 min Scan# 1117

Delta R.T. -0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53


Tgt Ion: 77 Resp: 158324

Ion Ratio Lower Upper

77 100

123 28.6 27.4 41.0



65 13.2 13.3 19.9#




Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 123.00 (122.70 to 123.70): E

Ion 65.00 (64.70 to 65.70): BGC









#21
Isophorone
Concen: 20.72 ng/ul
RT: 9.92 min Scan# 1205
Delta R.T. 0.01 min
Lab File: BG025661.D
Acq: 26 Jan 2017 10:53




Instrument :
BNA_G
ClientSampleId :
SSTD02020


Tgt Ion: 82 Resp: 283538
Ion Ratio Lower Upper
82 100
95 9.8 7.8 11.6
138 12.4 12.2 18.2



#23
2-Nitrophenol
Concen: 22.72 ng/ul
RT: 10.12 min Scan# 1239
Delta R.T. 0.00 min
Lab File: BG025661.D
Acq: 26 Jan 2017 10:53

Tgt Ion: 139 Resp: 63024
Ion Ratio Lower Upper
139 100
65 73.9 67.0 100.6
109 41.9 39.6 59.4









#24
 2,4-Dimethylphenol
 Concen: 21.44 ng/ul
 RT: 10.17 min Scan# 1247
 Delta R.T. 0.00 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02020


Tgt Ion: 107 Resp: 143345
 Ion Ratio Lower Upper
 107 100
 121 44.7 32.3 48.5
 122 78.6 61.2 91.8



#25
 Bis(2-Chloroethoxy)methane
 Concen: 21.92 ng/ul
 RT: 10.40 min Scan# 1286
 Delta R.T. 0.00 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53

Tgt Ion: 93 Resp: 151331
 Ion Ratio Lower Upper
 93 100
 95 29.0 23.4 35.0
 123 8.7 7.8 11.6





#28

Naphthalene

Concen: 21.51 ng/ul

RT: 11.05 min Scan# 1398

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53

Instrument :

BNA_G

ClientSampleId :

SSTD02020


Tgt Ion:128 Resp: 318861

Ion Ratio Lower Upper

128 100

129 10.9 9.4 14.2




127 15.1 10.3 15.5



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#30

4-Chloroaniline

Concen: 21.78 ng/ul

RT: 11.18 min Scan# 1419

Delta R.T. 0.00 min

Lab File: BG025661.D


Acq: 26 Jan 2017 10:53

Tgt Ion:127 Resp: 131374

Ion Ratio Lower Upper



127 100


129 30.8 28.9 43.3



Abundance Ion 127.00 (126.70 to 127.70): E

Ion 129.00 (128.70 to 129.70): E





#31

Hexachlorobutadiene

Concen: 20.91 ng/ul

RT: 11.31 min Scan# 1442

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53

Instrument :

BNA_G

ClientSampleId :

SSTD02020


Tgt Ion:225 Resp: 95053

Ion Ratio Lower Upper

225 100

223 62.3 45.9 68.9




227 56.2 44.7 67.1



Abundance Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E



#32

Caprolactam

Concen: 19.68 ng/ul

RT: 11.94 min Scan# 1549

Delta R.T. 0.01 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53


Tgt Ion:113 Resp: 43029

Ion Ratio Lower Upper

113 100

55 225.5 168.6 252.8



56 183.3 128.5 192.7




Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BGC

Ion 56.00 (55.70 to 56.70): BGC









#33
 4-Chloro-3-methylphenol
 Concen: 21.87 ng/ul
 RT: 12.29 min Scan# 1609
 Delta R.T. 0.00 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02020


Tgt Ion:107 Resp: 138177
 Ion Ratio Lower Upper
 107 100
 144 24.6 18.2 27.4
 142 62.2 55.0 82.4



#34
 2-Methylnaphthalene
 Concen: 20.92 ng/ul
 RT: 12.65 min Scan# 1669
 Delta R.T. 0.00 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53

Tgt Ion:142 Resp: 253183
 Ion Ratio Lower Upper
 142 100
 141 93.6 68.6 102.8









#36
 1,2,4,5-Tetrachlorobenzene
 Concen: 21.79 ng/ul
 RT: 13.01 min Scan# 1731
 Delta R.T. 0.00 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02020

Tgt Ion:	216	Resp:	175064
Ion	Ratio	Lower	Upper
216	100		
214	80.1	67.0	100.6
179	20.5	18.1	27.1
108	14.7	14.1	21.1




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#37
 Hexachlorocyclopentadiene
 Concen: 20.88 ng/ul
 RT: 12.97 min Scan# 1724
 Delta R.T. 0.00 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53

Tgt Ion:	237	Resp:	66222
Ion	Ratio	Lower	Upper
237	100		
235	51.1	50.2	75.4
272	9.6	10.6	16.0#



Abundance Ion 237.00 (236.70 to 237.70): E
 Ion 235.00 (234.70 to 235.70): E
 Ion 272.00 (271.70 to 272.70): E









#38
 2,4,6-Trichlorophenol
 Concen: 23.08 ng/ul
 RT: 13.35 min Scan# 1788
 Delta R.T. 0.08 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02020


Tgt Ion:196 Resp: 112255
 Ion Ratio Lower Upper
 196 100
 198 89.0 71.4 107.2
 200 34.4 24.6 37.0



#39
 2,4,5-Trichlorophenol
 Concen: 22.01 ng/ul
 RT: 13.35 min Scan# 1788
 Delta R.T. 0.00 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53

Tgt Ion:196 Resp: 112255
 Ion Ratio Lower Upper
 196 100
 198 89.0 78.6 118.0
 200 34.4 27.3 40.9









#40
 1,1'-Biphenyl
 Concen: 21.56 ng/ul
 RT: 13.64 min Scan# 1838
 Delta R.T. 0.00 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53




Instrument :
 BNA_G
 ClientSampleId :
 SST02020


Tgt Ion	Ratio	Lower	Upper
154	100		
153	40.6	32.2	48.2
76	12.6	9.6	14.4



#41
 2-Chloronaphthalene
 Concen: 21.77 ng/ul
 RT: 13.69 min Scan# 1847
 Delta R.T. 0.00 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53

Tgt Ion	Ratio	Lower	Upper
162	100		
164	32.4	25.0	37.6
127	35.9	30.7	46.1









#42
2-Nitroaniline
Concen: 20.89 ng/ul
RT: 13.92 min Scan# 1885
Delta R.T. 0.00 min
Lab File: BG025661.D
Acq: 26 Jan 2017 10:53




Instrument :
BNA_G
ClientSampleId :
SSTD02020


Tgt Ion: 65 Resp: 116212
Ion Ratio Lower Upper
65 100
92 58.1 50.9 76.3
138 73.4 61.8 92.6



#44
Dimethylphthalate
Concen: 19.67 ng/ul
RT: 14.25 min Scan# 1942
Delta R.T. 0.00 min
Lab File: BG025661.D
Acq: 26 Jan 2017 10:53

Tgt Ion: 163 Resp: 364383
Ion Ratio Lower Upper
163 100
194 6.3 3.4 5.0#
164 10.6 9.0 13.4









#45
 2,6-Dinitrotoluene
 Concen: 20.43 ng/ul
 RT: 14.40 min Scan# 1967
 Delta R.T. 0.00 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02020


Tgt Ion:165 Resp: 77925
 Ion Ratio Lower Upper
 165 100
 89 59.4 52.9 79.3
 121 28.5 22.2 33.4



#47
 Acenaphthylene
 Concen: 21.57 ng/ul
 RT: 14.54 min Scan# 1991
 Delta R.T. 0.01 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53

Tgt Ion:152 Resp: 427388
 Ion Ratio Lower Upper
 152 100
 151 21.7 16.7 25.1
 153 12.3 11.4 17.2









#48
3-Nitroaniline
Concen: 22.12 ng/ul
RT: 14.73 min Scan# 2024
Delta R.T. -0.00 min
Lab File: BG025661.D
Acq: 26 Jan 2017 10:53




Instrument :
BNA_G
ClientSampleId :
SSTD02020


Tgt Ion:138 Resp: 76327
Ion Ratio Lower Upper
138 100
108 12.3 6.3 9.5#
92 130.5 106.4 159.6



#49
Acenaphthene
Concen: 21.19 ng/ul
RT: 14.87 min Scan# 2048
Delta R.T. 0.00 min
Lab File: BG025661.D
Acq: 26 Jan 2017 10:53

Tgt Ion:153 Resp: 299115
Ion Ratio Lower Upper
153 100
152 45.3 36.5 54.7
154 85.3 72.3 108.5









#50
2,4-Dinitrophenol
Concen: 21.56 ng/ul
RT: 14.97 min Scan# 2065
Delta R.T. 0.01 min
Lab File: BG025661.D
Acq: 26 Jan 2017 10:53

Instrument :
BNA_G
ClientSampleId :
SSTD02020

Tgt Ion:184 Resp: 48344
Ion Ratio Lower Upper
184 100
63 63.0 65.1 97.7#
154 61.7 68.2 102.4#




Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BG
Ion 154.00 (153.70 to 154.70): E






#52
4-Nitrophenol
Concen: 16.64 ng/ul
RT: 15.06 min Scan# 2080
Delta R.T. 0.01 min
Lab File: BG025661.D
Acq: 26 Jan 2017 10:53

Tgt Ion:109 Resp: 59523
Ion Ratio Lower Upper
109 100
139 89.6 62.6 93.8
65 123.7 90.4 135.6



Abundance Ion 109.00 (108.70 to 109.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 65.00 (64.70 to 65.70): BG





#53

Dibenzofuran

Concen: 20.54 ng/ul

RT: 15.21 min Scan# 2105

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53

Instrument :

BNA_G

ClientSampleId :


SSTD02020

Tgt Ion:168 Resp: 445654

Ion Ratio Lower Upper




168 100

139 41.2 33.8 50.8



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E



#54

2,4-Dinitrotoluene

Concen: 20.96 ng/ul

RT: 15.20 min Scan# 2103

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53


Tgt Ion:165 Resp: 121944

Ion Ratio Lower Upper

165 100

63 70.2 46.8 70.2



182 2.7 1.8 2.6#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BG0

Ion 182.00 (181.70 to 182.70): E





#55

2,3,4,6-Tetrachlorophenol

Concen: 19.73 ng/ul

RT: 15.44 min Scan# 2144

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53

Instrument :

BNA_G

ClientSampleId :

SSTD02020

Tgt Ion:232 Resp: 106474


Ion Ratio Lower Upper

232 100

131 37.3 33.3 49.9

130 3.3 1.8 2.6#

166 28.3 21.9 32.9






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E



#56

Diethylphthalate

Concen: 19.76 ng/ul

RT: 15.60 min Scan# 2171

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53


Tgt Ion:149 Resp: 395817

Ion Ratio Lower Upper

149 100

177 22.8 17.8 26.8



150 14.1 10.2 15.4




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E





#58

Fluorene

Concen: 20.82 ng/ul

RT: 15.85 min Scan# 2215

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53

Instrument :

BNA_G

ClientSampleId :

SSTD02020





Tgt Ion:166 Resp: 368937

Ion Ratio Lower Upper

166 100

165 100.8 79.7 119.5

167 12.6 11.4 17.2



#59

4-Chlorophenyl-phenylether

Concen: 19.52 ng/ul

RT: 15.83 min Scan# 2211

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53




Tgt Ion:204 Resp: 195129


Ion Ratio Lower Upper

204 100

206 38.8 32.3 48.5

141 66.2 57.5 86.3









#60
4-Nitroaniline
Concen: 21.46 ng/ul
RT: 15.90 min Scan# 2223
Delta R.T. 0.00 min
Lab File: BG025661.D
Acq: 26 Jan 2017 10:53




Instrument :
BNA_G
ClientSampleId :
SSTD02020


Tgt Ion:138 Resp: 71674
Ion Ratio Lower Upper
138 100
92 67.5 56.9 85.3
108 138.5 104.9 157.3



#64
N-Nitrosodiphenylamine
Concen: 25.33 ng/ul
RT: 16.05 min Scan# 2249
Delta R.T. 0.00 min
Lab File: BG025661.D
Acq: 26 Jan 2017 10:53

Tgt Ion:169 Resp: 348226
Ion Ratio Lower Upper
169 100
168 61.2 59.1 88.7
167 31.5 29.7 44.5







#65
4-Bromophenyl-phenylether
Concen: 23.78 ng/ul
RT: 16.73 min Scan# 2364
Delta R.T. 0.01 min
Lab File: BG025661.D
Acq: 26 Jan 2017 10:53

Instrument :
BNA_G
ClientSampleId :
SSTD02020



Tgt Ion:248 Resp: 147211
Ion Ratio Lower Upper
248 100
250 91.4 71.9 107.9
141 66.0 53.8 80.8



Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E




Time--> 16.65 16.70 16.75 16.80




#66
Hexachlorobenzene
Concen: 22.87 ng/ul
RT: 16.85 min Scan# 2385
Delta R.T. 0.00 min
Lab File: BG025661.D
Acq: 26 Jan 2017 10:53


Tgt Ion:284 Resp: 160879
Ion Ratio Lower Upper
284 100
142 28.3 26.1 39.1
249 30.4 25.9 38.9




Abundance Ion 284.00 (283.70 to 284.70): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E



Time--> 16.80 16.85 16.90









#67
 Atrazine
 Concen: 22.62 ng/ul
 RT: 16.99 min Scan# 2408
 Delta R.T. 0.00 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02020


Tgt Ion	Ratio	Lower	Upper
200	100		
173	21.3	19.2	28.8
215	50.3	40.3	60.5



#69
 Phenanthrene
 Concen: 24.12 ng/ul
 RT: 17.69 min Scan# 2528
 Delta R.T. 0.10 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53

Tgt Ion	Ratio	Lower	Upper
178	100		
179	15.8	12.4	18.6
176	20.2	16.5	24.7





#71

Anthracene

Concen: 23.99 ng/ul

RT: 17.69 min Scan# 2528

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53

Instrument :

BNA_G

ClientSampleId :

SSTD02020


Tgt Ion:178 Resp: 625292

Ion Ratio Lower Upper

178 100

179 15.8 12.6 19.0




176 20.2 15.9 23.9



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#72

Carbazole

Concen: 25.22 ng/ul

RT: 17.96 min Scan# 2574

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53


Tgt Ion:167 Resp: 545250

Ion Ratio Lower Upper

167 100

166 22.3 17.9 26.9



139 14.3 10.5 15.7




Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E





#73

Di-n-butylphthalate

Concen: 24.86 ng/ul

RT: 18.47 min Scan# 2661

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53

Instrument :

BNA_G

ClientSampleId :

SSTD02020


Tgt Ion:149 Resp: 694823

Ion Ratio Lower Upper

149 100

150 9.3 7.9 11.9




104 6.3 5.8 8.8



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 24.78 ng/ul

RT: 19.60 min Scan# 2852

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53


Tgt Ion:202 Resp: 758158

Ion Ratio Lower Upper

202 100

101 8.5 6.2 9.2



100 6.7 5.2 7.8




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): E





#77

Pyrene

Concen: 23.15 ng/ul

RT: 19.96 min Scan# 2914

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53

Instrument :

BNA_G

ClientSampleId :

SSTD02020


Tgt Ion:202 Resp: 774075

Ion Ratio Lower Upper

202 100

101 9.5 6.9 10.3




100 8.3 6.6 10.0



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B



#78

Butylbenzylphthalate

Concen: 23.76 ng/ul

RT: 20.81 min Scan# 3058

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53


Tgt Ion:149 Resp: 308938

Ion Ratio Lower Upper

149 100

91 81.1 65.8 98.6



206 20.4 20.1 30.1




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): B

Ion 206.00 (205.70 to 206.70): E









#79
 3,3'-Dichlorobenzidine
 Concen: 24.81 ng/ul
 RT: 21.73 min Scan# 3215
 Delta R.T. 0.00 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02020


Tgt Ion:252 Resp: 309539
 Ion Ratio Lower Upper
 252 100
 254 64.2 49.0 73.6
 126 7.5 5.1 7.7



#80
 Benzo(a)anthracene
 Concen: 22.68 ng/ul
 RT: 21.83 min Scan# 3232
 Delta R.T. 0.01 min
 Lab File: BG025661.D
 Acq: 26 Jan 2017 10:53

Tgt Ion:228 Resp: 814959
 Ion Ratio Lower Upper
 228 100
 229 19.4 16.3 24.5
 226 27.6 21.9 32.9





#81

Bis(2-ethylhexyl)phthalate

Concen: 24.16 ng/ul

RT: 21.67 min Scan# 3204

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53

Instrument :

BNA_G

ClientSampleId :

SSTD02020


Tgt Ion:149 Resp: 438561

Ion Ratio Lower Upper

149 100

167 28.8 21.8 32.8




279 5.1 4.5 6.7



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#82

Chrysene

Concen: 22.37 ng/ul

RT: 21.89 min Scan# 3243

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53


Tgt Ion:228 Resp: 755912

Ion Ratio Lower Upper

228 100

226 30.3 24.0 36.0



229 20.2 16.9 25.3




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#84

Di-n-octyl phthalate

Concen: 24.88 ng/ul

RT: 22.92 min Scan# 3417

Delta R.T. 0.00 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53


Instrument :

BNA_G




ClientSampleId :

SSTD02020

Tgt Ion:149 Resp: 739029



Abundance Ion 149.00 (148.70 to 149.70): E



#85

Benzo(b)fluoranthene

Concen: 22.04 ng/ul

RT: 24.15 min Scan# 3627

Delta R.T. 0.01 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53


Tgt Ion:252 Resp: 789483

Ion Ratio Lower Upper



252 100


253 22.9 17.7 26.5

125 5.3 4.0 6.0



Abundance Ion 252.00 (251.70 to 252.70): E





#86

Benzo(k)fluoranthene

Concen: 22.48 ng/ul

RT: 24.22 min Scan# 3639

Delta R.T. 0.01 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53

Instrument :

BNA_G

ClientSampleId :

SSTD02020


Tgt Ion:252 Resp: 755381

Ion Ratio Lower Upper

252 100

253 20.7 18.5 27.7




125 4.3 4.1 6.1



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(a)pyrene

Concen: 22.52 ng/ul

RT: 25.08 min Scan# 3785

Delta R.T. 0.01 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53


Tgt Ion:252 Resp: 762917

Ion Ratio Lower Upper

252 100

253 20.7 17.8 26.6



125 6.1 5.4 8.0




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#89

Indeno(1,2,3-cd)pyrene

Concen: 21.29 ng/ul

RT: 29.16 min Scan# 4480

Delta R.T. 0.03 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53

Instrument :

BNA_G

ClientSampleId :

SSTD02020


Tgt Ion:276 Resp: 913423

Ion Ratio Lower Upper

276 100

138 13.5 8.2 12.4#




277 24.9 18.6 28.0



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#90

Dibenzo(a,h)anthracene

Concen: 21.98 ng/ul

RT: 29.19 min Scan# 4485

Delta R.T. 0.02 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53


Tgt Ion:278 Resp: 785553

Ion Ratio Lower Upper

278 100

139 9.7 6.7 10.1



279 23.3 20.9 31.3




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 21.64 ng/ul

RT: 30.38 min Scan# 4687

Delta R.T. 0.01 min

Lab File: BG025661.D

Acq: 26 Jan 2017 10:53

Instrument :

BNA_G

ClientSampleId :

SSTD02020


Tgt Ion:276 Resp: 762353

Ion Ratio Lower Upper

276 100

138 13.8 8.6 12.8#



277 24.0 17.6 26.4



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



Time-->