

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG012617\
 Data File : BG025672.D
 Acq On : 26 Jan 2017 18:34
 Operator : SJ/MA
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02021

Quant Time: Jan 27 03:39:17 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG011817.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jan 27 03:32:12 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	69174	20.00	ng/ul	0.00
18) Naphthalene-d8	11.00	136	330167	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.81	164	258558	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.56	188	574247	20.00	ng/ul	0.00
75) Chrysene-d12	21.84	240	683400	20.00	ng/ul	0.00
83) Perylene-d12	25.23	264	686990	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.51	96	11511	8.33	ng/uL	0.00
5) Phenol-d5	7.34	99	136098	22.96	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.49	67	90774	23.71	ng/ul	0.00
9) 2-Chlorophenol-d4	7.71	132	98135	23.47	ng/ul	0.00
13) 4-Methylphenol-d8	8.89	113	109686	21.76	ng/ul	0.00
19) Nitrobenzene-d5	9.36	128	49257	21.05	ng/ul	0.00
22) 2-Nitrophenol-d4	10.09	143	61209	21.92	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.63	165	114127	20.72	ng/ul	0.00
29) 4-Chloroaniline-d4	11.15	131	136319	22.14	ng/ul	0.00
43) Dimethylphthalate-d6	14.20	166	372626	20.41	ng/ul	0.00
46) Acenaphthylene-d8	14.51	160	446343	22.63	ng/ul	0.00
51) 4-Nitrophenol-d4	15.05	143	42839	17.12	ng/ul	0.00
57) Fluorene-d10	15.80	176	362200	21.04	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.95	200	69846	19.72	ng/ul	0.00
70) Anthracene-d10	17.66	188	535808	21.91	ng/ul	0.00
76) Pyrene-d10	19.93	212	641285	22.55	ng/ul	0.00
87) Benzo(a)pyrene-d12	25.01	264	644515	22.32	ng/ul	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.54	88	13782	8.44	ng/uL	96
4) Benzaldehyde	7.31	77	107399	26.31	ng/ul	97
6) Phenol	7.37	94	139403	22.38	ng/ul#	91
8) Bis(2-Chloroethyl)ether	7.58	93	106796	24.01	ng/ul	95
10) 2-Chlorophenol	7.74	128	97976	23.66	ng/ul	95
11) 2-Methylphenol	8.63	108	105402	22.72	ng/ul	95
12) 2,2'-oxybis(1-Chloropropan	8.70	45	201160	23.80	ng/ul	99
14) Acetophenone	9.01	105	171740	21.82	ng/ul	95
15) N-Nitroso-di-n-propylamine	8.98	70	104514	22.19	ng/ul#	92
16) 4-Methylphenol	8.96	108	117299	23.06	ng/ul	97
17) Hexachloroethane	9.25	117	45078	23.37	ng/ul	94
20) Nitrobenzene	9.41	77	155628	20.77	ng/ul	96
21) Isophorone	9.92	82	285269	20.27	ng/ul	96
23) 2-Nitrophenol	10.12	139	62604	21.95	ng/ul#	84
24) 2,4-Dimethylphenol	10.16	107	139054	20.23	ng/ul	95
25) Bis(2-Chloroethoxy)methane	10.39	93	148769	20.95	ng/ul	95
27) 2,4-Dichlorophenol	10.66	162	112962	21.11	ng/ul	95
28) Naphthalene	11.06	128	329341	21.61	ng/ul#	96
30) 4-Chloroaniline	11.17	127	138927	22.39	ng/ul	88
31) Hexachlorobutadiene	11.31	225	96616	20.67	ng/ul	95
32) Caprolactam	11.94	113	42737	19.00	ng/ul	85
33) 4-Chloro-3-methylphenol	12.29	107	139914	21.53	ng/ul#	92
34) 2-Methylnaphthalene	12.64	142	260694	20.95	ng/ul	98

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG012617\
 Data File : BG025672.D
 Acq On : 26 Jan 2017 18:34
 Operator : SJ/MA
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02021

Quant Time: Jan 27 03:39:17 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG011817.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jan 27 03:32:12 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
36) 1,2,4,5-Tetrachlorobenzene	13.01	216	177892	22.85	ng/ul	93
37) Hexachlorocyclopentadiene	12.97	237	69745	22.70	ng/ul	100
38) 2,4,6-Trichlorophenol	13.25	196	103125	21.88	ng/ul	89
39) 2,4,5-Trichlorophenol	13.35	196	109948	22.24	ng/ul	94
40) 1,1'-Biphenyl	13.64	154	366456	22.87	ng/ul	95
41) 2-Chloronaphthalene	13.70	162	289759	23.30	ng/ul	97
42) 2-Nitroaniline	13.91	65	115665	21.45	ng/ul	95
44) Dimethylphthalate	14.25	163	373572	20.81	ng/ul	99
45) 2,6-Dinitrotoluene	14.39	165	83694	22.64	ng/ul#	93
47) Acenaphthylene	14.54	152	426439	22.21	ng/ul	99
48) 3-Nitroaniline	14.74	138	75594	22.61	ng/ul#	95
49) Acenaphthene	14.87	153	304335	22.25	ng/ul	99
50) 2,4-Dinitrophenol	14.97	184	44990	20.71	ng/ul#	84
52) 4-Nitrophenol	15.06	109	58349	16.83	ng/ul	91
53) Dibenzofuran	15.21	168	452375	21.52	ng/ul	98
54) 2,4-Dinitrotoluene	15.19	165	120357	21.35	ng/ul#	73
55) 2,3,4,6-Tetrachlorophenol	15.44	232	108552	20.76	ng/ul#	96
56) Diethylphthalate	15.60	149	397963	20.50	ng/ul	99
58) Fluorene	15.85	166	367740	21.42	ng/ul	99
59) 4-Chlorophenyl-phenylether	15.83	204	201963	20.86	ng/ul	91
63) 4,6-Dinitro-2-methylphenol	15.96	198	73928	20.25	ng/ul#	93
64) N-Nitrosodiphenylamine	16.05	169	339268	22.92	ng/ul	95
65) 4-Bromophenyl-phenylether	16.73	248	144839	21.73	ng/ul	91
66) Hexachlorobenzene	16.86	284	164887	21.77	ng/ul	95
67) Atrazine	16.99	200	151091	21.72	ng/ul	92
68) Pentachlorophenol	17.21	266	55734	17.69	ng/ul	90
69) Phenanthrene	17.60	178	626106	22.43	ng/ul	98
71) Anthracene	17.69	178	628660	22.40	ng/ul	99
72) Carbazole	17.97	167	539873	23.20	ng/ul	99
73) Di-n-butylphthalate	18.48	149	661693	21.99	ng/ul	99
74) Fluoranthene	19.60	202	757021	22.98	ng/ul	99
77) Pyrene	19.96	202	767467	22.84	ng/ul	99
78) Butylbenzylphthalate	20.81	149	309687	23.70	ng/ul	95
79) 3,3'-Dichlorobenzidine	21.73	252	293135	23.38	ng/ul	94
80) Benzo(a)anthracene	21.83	228	805609	22.31	ng/ul	98
81) Bis(2-ethylhexyl)phthalate	21.67	149	439644	24.10	ng/ul	99
82) Chrysene	21.89	228	751609	22.13	ng/ul	98
84) Di-n-octyl phthalate	22.91	149	741882	25.16	ng/ul	100
85) Benzo(b)fluoranthene	24.22	252	770660	21.67	ng/ul	99
86) Benzo(k)fluoranthene	24.22	252	772366	23.16	ng/ul	99
88) Benzo(a)pyrene	25.08	252	758074	22.54	ng/ul	98
90) Dibenzo(a,h)anthracene	29.19	278	768998	21.67	ng/ul	97
91) Benzo(g,h,i)perylene	30.38	276	718919	20.55	ng/ul	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG012617\
Data File : BG025672.D
Acq On : 26 Jan 2017 18:34
Operator : SJ/MA
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleID :
SSTD02021

Quant Time: Jan 27 03:39:17 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG011817.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jan 27 03:32:12 2017
Response via : Initial Calibration





#2

1,4-Dioxane

Concen: 8.44 ng/uL

RT: 3.54 min Scan# 120

Delta R.T. -0.01 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34

Instrument :

BNA_G

ClientSampleId :

SSTD02021





Tgt Ion: 88 Resp: 13782

Ion Ratio Lower Upper

88 100

43 53.3 39.2 58.8

58 84.5 69.4 104.0



#4

Benzaldehyde

Concen: 26.31 ng/uL

RT: 7.31 min Scan# 762

Delta R.T. -0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34




Tgt Ion: 77 Resp: 107399


Ion Ratio Lower Upper

77 100

105 80.8 62.0 93.0

106 77.1 60.2 90.4





#6

Phenol

Concen: 22.38 ng/ul

RT: 7.37 min Scan# 771

Delta R.T. 0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34

Instrument :

BNA_G

ClientSampleId :

SSTD02021


Tgt Ion: 94 Resp: 139403

Ion Ratio Lower Upper




94 100

65 41.9 26.8 40.2#

66 51.8 44.4 66.6



Abundance Ion 94.00 (93.70 to 94.70): BGC
Ion 65.00 (64.70 to 65.70): BGC
Ion 66.00 (65.70 to 66.70): BGC



#8

Bis(2-Chloroethyl)ether

Concen: 24.01 ng/ul

RT: 7.58 min Scan# 808

Delta R.T. -0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34


Tgt Ion: 93 Resp: 106796

Ion Ratio Lower Upper



93 100


63 90.7 75.5 113.3

95 31.7 29.2 43.8



Abundance Ion 93.00 (92.70 to 93.70): BGC
Ion 63.00 (62.70 to 63.70): BGC
Ion 95.00 (94.70 to 95.70): BGC









#10
2-Chlorophenol
Concen: 23.66 ng/ul
RT: 7.74 min Scan# 835
Delta R.T. 0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34




Instrument :
BNA_G
ClientSampleId :
SSTD02021


Tgt Ion:128 Resp: 97976
Ion Ratio Lower Upper
128 100
64 68.3 56.3 84.5
130 27.1 26.7 40.1



#11
2-Methylphenol
Concen: 22.72 ng/ul
RT: 8.63 min Scan# 986
Delta R.T. 0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34

Tgt Ion:108 Resp: 105402
Ion Ratio Lower Upper
108 100
107 91.4 76.7 115.1









#12
2,2'-oxybis(1-Chloropropane)
Concen: 23.80 ng/ul
RT: 8.70 min Scan# 998
Delta R.T. -0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34




Instrument :
BNA_G
ClientSampleId :
SSTD02021


Tgt Ion: 45 Resp: 201160
Ion Ratio Lower Upper
45 100
77 12.9 10.0 15.0
79 9.2 7.3 10.9



#14
Acetophenone
Concen: 21.82 ng/ul
RT: 9.01 min Scan# 1051
Delta R.T. 0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34

Tgt Ion: 105 Resp: 171740
Ion Ratio Lower Upper
105 100
77 92.4 76.0 114.0
51 36.7 34.8 52.2









#15
N-Nitroso-di-n-propylamine
Concen: 22.19 ng/ul
RT: 8.98 min Scan# 1045
Delta R.T. -0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34




Instrument :
BNA_G
ClientSampleId :
SSTD02021


Tgt Ion: 70 Resp: 104514
Ion Ratio Lower Upper
70 100
42 67.5 59.0 88.6
101 10.6 6.6 9.8#
130 15.4 15.6 23.4#



#16
4-Methylphenol
Concen: 23.06 ng/ul
RT: 8.96 min Scan# 1042
Delta R.T. -0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34

Tgt Ion: 108 Resp: 117299
Ion Ratio Lower Upper
108 100
107 111.8 91.7 137.5





#17

Hexachloroethane

Concen: 23.37 ng/ul

RT: 9.25 min Scan# 1092

Delta R.T. -0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34

Instrument :

BNA_G

ClientSampleId :

SSTD02021


Tgt Ion: 117 Resp: 45078

Ion Ratio Lower Upper

117 100

201 113.5 91.9 137.9




199 69.7 64.3 96.5



Abundance Ion 117.00 (116.70 to 117.70): E

Ion 201.00 (200.70 to 201.70): E

Ion 199.00 (198.70 to 199.70): E



#20

Nitrobenzene

Concen: 20.77 ng/ul

RT: 9.41 min Scan# 1118

Delta R.T. 0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34


Tgt Ion: 77 Resp: 155628

Ion Ratio Lower Upper

77 100

123 31.9 27.4 41.0



65 15.1 13.3 19.9




Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 123.00 (122.70 to 123.70): E

Ion 65.00 (64.70 to 65.70): BGC









#21
Isophorone
Concen: 20.27 ng/ul
RT: 9.92 min Scan# 1205
Delta R.T. -0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34




Instrument :
BNA_G
ClientSampleId :
SSTD02021


Tgt Ion: 82 Resp: 285269
Ion Ratio Lower Upper
82 100
95 8.1 7.8 11.6
138 13.9 12.2 18.2



#23
2-Nitrophenol
Concen: 21.95 ng/ul
RT: 10.12 min Scan# 1240
Delta R.T. 0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34

Tgt Ion: 139 Resp: 62604
Ion Ratio Lower Upper
139 100
65 63.6 67.0 100.6#
109 52.3 39.6 59.4









#24
 2,4-Dimethylphenol
 Concen: 20.23 ng/ul
 RT: 10.16 min Scan# 1247
 Delta R.T. -0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34




Instrument :
 BNA_G
 ClientSampled :
 SSTD02021


Tgt Ion: 107 Resp: 139054
 Ion Ratio Lower Upper
 107 100
 121 41.4 32.3 48.5
 122 71.0 61.2 91.8



#25
 Bis(2-Chloroethoxy)methane
 Concen: 20.95 ng/ul
 RT: 10.39 min Scan# 1286
 Delta R.T. -0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34

Tgt Ion: 93 Resp: 148769
 Ion Ratio Lower Upper
 93 100
 95 32.3 23.4 35.0
 123 8.7 7.8 11.6









#27
2,4-Dichlorophenol
Concen: 21.11 ng/ul
RT: 10.66 min Scan# 1332
Delta R.T. -0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34




Instrument :
BNA_G
ClientSampleId :
SSTD02021


Tgt Ion	Ratio	Lower	Upper
162	100		
164	65.3	50.8	76.2
98	34.6	32.1	48.1



#28
Naphthalene
Concen: 21.61 ng/ul
RT: 11.06 min Scan# 1399
Delta R.T. 0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34

Tgt Ion	Ratio	Lower	Upper
128	100		
129	9.4	9.4	14.2#
127	11.8	10.3	15.5









#30
4-Chloroaniline
Concen: 22.39 ng/ul
RT: 11.17 min Scan# 1419
Delta R.T. -0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34




Instrument :
BNA_G
ClientSampleId :
SSTD02021


Tgt Ion:127 Resp: 138927
Ion Ratio Lower Upper
127 100
129 29.1 28.9 43.3



#31
Hexachlorobutadiene
Concen: 20.67 ng/ul
RT: 11.31 min Scan# 1442
Delta R.T. -0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34

Tgt Ion:225 Resp: 96616
Ion Ratio Lower Upper
225 100
223 63.2 45.9 68.9
227 53.9 44.7 67.1









#32
 Caprolactam
 Concen: 19.00 ng/ul
 RT: 11.94 min Scan# 1550
 Delta R.T. 0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02021


Tgt Ion: 113 Resp: 42737
 Ion Ratio Lower Upper
 113 100
 55 171.6 168.6 252.8
 56 163.0 128.5 192.7



#33
 4-Chloro-3-methylphenol
 Concen: 21.53 ng/ul
 RT: 12.29 min Scan# 1609
 Delta R.T. -0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34

Tgt Ion: 107 Resp: 139914
 Ion Ratio Lower Upper
 107 100
 144 18.0 18.2 27.4#
 142 62.6 55.0 82.4









#34
 2-Methylnaphthalene
 Concen: 20.95 ng/ul
 RT: 12.64 min Scan# 1669
 Delta R.T. -0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02021


Tgt Ion:142 Resp: 260694
 Ion Ratio Lower Upper
 142 100
 141 87.8 68.6 102.8



#36
 1,2,4,5-Tetrachlorobenzene
 Concen: 22.85 ng/ul
 RT: 13.01 min Scan# 1731
 Delta R.T. -0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34

Tgt Ion:216 Resp: 177892
 Ion Ratio Lower Upper
 216 100
 214 76.9 67.0 100.6
 179 20.7 18.1 27.1
 108 15.5 14.1 21.1









#37
Hexachlorocyclopentadiene
Concen: 22.70 ng/ul
RT: 12.97 min Scan# 1725
Delta R.T. 0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34




Instrument :
BNA_G
ClientSampleId :
SSTD02021


Tgt Ion:237 Resp: 69745
Ion Ratio Lower Upper
237 100
235 62.6 50.2 75.4
272 13.4 10.6 16.0



#38
2,4,6-Trichlorophenol
Concen: 21.88 ng/ul
RT: 13.25 min Scan# 1773
Delta R.T. -0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34

Tgt Ion:196 Resp: 103125
Ion Ratio Lower Upper
196 100
198 102.8 71.4 107.2
200 30.5 24.6 37.0









#39
 2,4,5-Trichlorophenol
 Concen: 22.24 ng/ul
 RT: 13.35 min Scan# 1789
 Delta R.T. 0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34




Instrument :
 BNA_G
 ClientSampleId :
 SST02021


Tgt Ion:196 Resp: 109948
 Ion Ratio Lower Upper
 196 100
 198 93.4 78.6 118.0
 200 29.6 27.3 40.9



#40
 1,1'-Biphenyl
 Concen: 22.87 ng/ul
 RT: 13.64 min Scan# 1839
 Delta R.T. 0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34

Tgt Ion:154 Resp: 366456
 Ion Ratio Lower Upper
 154 100
 153 36.5 32.2 48.2
 76 10.9 9.6 14.4









#41
 2-Chloronaphthalene
 Concen: 23.30 ng/ul
 RT: 13.70 min Scan# 1848
 Delta R.T. 0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02021


Tgt Ion	Ratio	Lower	Upper
162	100		
164	30.3	25.0	37.6
127	36.1	30.7	46.1



#42
 2-Nitroaniline
 Concen: 21.45 ng/ul
 RT: 13.91 min Scan# 1885
 Delta R.T. -0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34

Tgt Ion	Ratio	Lower	Upper
65	100		
92	58.9	50.9	76.3
138	73.7	61.8	92.6





#44

Dimethylphthalate

Concen: 20.81 ng/ul

RT: 14.25 min Scan# 1942

Delta R.T. -0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34

Instrument :

BNA_G

ClientSampleId :

SSTD02021


Tgt Ion:163 Resp: 373572

Ion Ratio Lower Upper

163 100

194 4.5 3.4 5.0




164 11.9 9.0 13.4



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#45

2,6-Dinitrotoluene

Concen: 22.64 ng/ul

RT: 14.39 min Scan# 1967

Delta R.T. -0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34


Tgt Ion:165 Resp: 83694

Ion Ratio Lower Upper

165 100

89 69.7 52.9 79.3



121 21.5 22.2 33.4#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 89.00 (88.70 to 89.70): BG

Ion 121.00 (120.70 to 121.70): E





#47

Acenaphthylene

Concen: 22.21 ng/ul

RT: 14.54 min Scan# 1991

Delta R.T. -0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34

Instrument :

BNA_G

ClientSampleId :

SSTD02021


Tgt Ion:152 Resp: 426439

Ion Ratio Lower Upper

152 100

151 21.1 16.7 25.1




153 14.6 11.4 17.2



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#48

3-Nitroaniline

Concen: 22.61 ng/ul

RT: 14.74 min Scan# 2026

Delta R.T. 0.01 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34


Tgt Ion:138 Resp: 75594

Ion Ratio Lower Upper

138 100

108 14.1 6.3 9.5#



92 128.2 106.4 159.6




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BGC





#49

Acenaphthene

Concen: 22.25 ng/ul

RT: 14.87 min Scan# 2048

Delta R.T. -0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34

Instrument :

BNA_G

ClientSampleId :

SSTD02021


Tgt Ion:153 Resp: 304335

Ion Ratio Lower Upper

153 100

152 47.1 36.5 54.7




154 90.1 72.3 108.5



Abundance Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E

Ion 154.00 (153.70 to 154.70): E



#50

2,4-Dinitrophenol

Concen: 20.71 ng/ul

RT: 14.97 min Scan# 2065

Delta R.T. -0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34


Tgt Ion:184 Resp: 44990

Ion Ratio Lower Upper

184 100

63 72.3 65.1 97.7



154 65.8 68.2 102.4#




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 63.00 (62.70 to 63.70): BG

Ion 154.00 (153.70 to 154.70): E









#52
4-Nitrophenol
Concen: 16.83 ng/ul
RT: 15.06 min Scan# 2081
Delta R.T. 0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34




Instrument :
BNA_G
ClientSampleId :
SSTD02021


Tgt Ion:109 Resp: 58349
Ion Ratio Lower Upper
109 100
139 85.5 62.6 93.8
65 123.2 90.4 135.6



#53
Dibenzofuran
Concen: 21.52 ng/ul
RT: 15.21 min Scan# 2105
Delta R.T. -0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34

Tgt Ion:168 Resp: 452375
Ion Ratio Lower Upper
168 100
139 43.6 33.8 50.8





#54

2,4-Dinitrotoluene

Concen: 21.35 ng/ul

RT: 15.19 min Scan# 2103

Delta R.T. -0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34

Instrument :

BNA_G

ClientSampleId :

SSTD02021


Tgt Ion:165 Resp: 120357

Ion Ratio Lower Upper

165 100

63 79.6 46.8 70.2#




182 3.3 1.8 2.6#



Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BG

Ion 182.00 (181.70 to 182.70): E



#55

2,3,4,6-Tetrachlorophenol

Concen: 20.76 ng/ul

RT: 15.44 min Scan# 2145

Delta R.T. 0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34

Tgt Ion:232 Resp: 108552


Ion Ratio Lower Upper

232 100

131 38.3 33.3 49.9

130 3.6 1.8 2.6#

166 26.4 21.9 32.9






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E









#56
Diethylphthalate
Concen: 20.50 ng/ul
RT: 15.60 min Scan# 2172
Delta R.T. 0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34




Instrument :
BNA_G
ClientSampleId :
SSTD02021


Tgt Ion:149 Resp: 397963
Ion Ratio Lower Upper
149 100
177 22.5 17.8 26.8
150 13.0 10.2 15.4



#58
Fluorene
Concen: 21.42 ng/ul
RT: 15.85 min Scan# 2215
Delta R.T. -0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34

Tgt Ion:166 Resp: 367740
Ion Ratio Lower Upper
166 100
165 100.9 79.7 119.5
167 14.8 11.4 17.2









#59
 4-Chlorophenyl-phenylether
 Concen: 20.86 ng/ul
 RT: 15.83 min Scan# 2211
 Delta R.T. -0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02021

Tgt Ion: 204 Resp: 201963
 Ion Ratio Lower Upper
 204 100
 206 34.6 32.3 48.5
 141 64.5 57.5 86.3




Abundance Ion 204.00 (203.70 to 204.70): E
 Ion 206.00 (205.70 to 206.70): E
 Ion 141.00 (140.70 to 141.70): E






#63
 4,6-Dinitro-2-methylphenol
 Concen: 20.25 ng/ul
 RT: 15.96 min Scan# 2233
 Delta R.T. -0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34

Tgt Ion: 198 Resp: 73928
 Ion Ratio Lower Upper
 198 100
 182 4.1 5.8 6.4#
 77 32.1 22.1 33.9



Abundance Ion 198.00 (197.70 to 198.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 77.00 (76.70 to 77.70): BG





#64

N-Nitrosodiphenylamine

Concen: 22.92 ng/ul

RT: 16.05 min Scan# 2249

Delta R.T. -0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34

Instrument :

BNA_G

ClientSampleId :

SSTD02021


Tgt Ion:169 Resp: 339268

Ion Ratio Lower Upper

169 100

168 69.2 59.1 88.7




167 38.9 29.7 44.5



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#65

4-Bromophenyl-phenylether

Concen: 21.73 ng/ul

RT: 16.73 min Scan# 2364

Delta R.T. -0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34


Tgt Ion:248 Resp: 144839

Ion Ratio Lower Upper

248 100

250 97.3 71.9 107.9



141 76.3 53.8 80.8




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E









#66
Hexachlorobenzene
Concen: 21.77 ng/ul
RT: 16.86 min Scan# 2386
Delta R.T. 0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34




Instrument :
BNA_G
ClientSampled :
SSTD02021


Tgt Ion	Ratio	Lower	Upper
284	100		
142	27.0	26.1	39.1
249	32.2	25.9	38.9



#67
Atrazine
Concen: 21.72 ng/ul
RT: 16.99 min Scan# 2408
Delta R.T. -0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34

Tgt Ion	Ratio	Lower	Upper
200	100		
173	20.4	19.2	28.8
215	44.8	40.3	60.5









#68
 Pentachlorophenol
 Concen: 17.69 ng/ul
 RT: 17.21 min Scan# 2447
 Delta R.T. -0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02021


Tgt Ion:266 Resp: 55734
 Ion Ratio Lower Upper
 266 100
 264 52.8 47.1 70.7
 268 60.3 56.2 84.4



#69
 Phenanthrene
 Concen: 22.43 ng/ul
 RT: 17.60 min Scan# 2512
 Delta R.T. -0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34

Tgt Ion:178 Resp: 626106
 Ion Ratio Lower Upper
 178 100
 179 15.6 12.4 18.6
 176 19.4 16.5 24.7





#71

Anthracene

Concen: 22.40 ng/ul

RT: 17.69 min Scan# 2528

Delta R.T. -0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34

Instrument :

BNA_G

ClientSampleId :

SSTD02021


Tgt Ion:178 Resp: 628660

Ion Ratio Lower Upper

178 100

179 16.1 12.6 19.0




176 20.4 15.9 23.9



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#72

Carbazole

Concen: 23.20 ng/ul

RT: 17.97 min Scan# 2575

Delta R.T. 0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34


Tgt Ion:167 Resp: 539873

Ion Ratio Lower Upper

167 100

166 22.4 17.9 26.9



139 14.4 10.5 15.7




Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E





#73

Di-n-butylphthalate

Concen: 21.99 ng/ul

RT: 18.48 min Scan# 2662

Delta R.T. 0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34

Instrument :

BNA_G

ClientSampleId :

SSTD02021


Tgt Ion:149 Resp: 661693

Ion Ratio Lower Upper

149 100

150 9.7 7.9 11.9




104 6.9 5.8 8.8



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 22.98 ng/ul

RT: 19.60 min Scan# 2853

Delta R.T. 0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34


Tgt Ion:202 Resp: 757021

Ion Ratio Lower Upper

202 100

101 8.0 6.2 9.2



100 6.0 5.2 7.8




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B





#77

Pyrene

Concen: 22.84 ng/ul

RT: 19.96 min Scan# 2915

Delta R.T. 0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34

Instrument :

BNA_G

ClientSampleId :

SSTD02021


Tgt Ion:202 Resp: 767467

Ion Ratio Lower Upper

202 100

101 9.1 6.9 10.3




100 7.8 6.6 10.0



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): E



#78

Butylbenzylphthalate

Concen: 23.70 ng/ul

RT: 20.81 min Scan# 3059

Delta R.T. 0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34


Tgt Ion:149 Resp: 309687

Ion Ratio Lower Upper

149 100

91 76.7 65.8 98.6



206 24.2 20.1 30.1




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): E

Ion 206.00 (205.70 to 206.70): E









#79
 3,3'-Dichlorobenzidine
 Concen: 23.38 ng/ul
 RT: 21.73 min Scan# 3216
 Delta R.T. 0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34




Instrument :
 BNA_G
 ClientSampleId :
 SST02021


Tgt Ion:252 Resp: 293135
 Ion Ratio Lower Upper
 252 100
 254 65.8 49.0 73.6
 126 7.2 5.1 7.7



#80
 Benzo(a)anthracene
 Concen: 22.31 ng/ul
 RT: 21.83 min Scan# 3232
 Delta R.T. -0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34

Tgt Ion:228 Resp: 805609
 Ion Ratio Lower Upper
 228 100
 229 20.6 16.3 24.5
 226 29.0 21.9 32.9









#81
 Bis(2-ethylhexyl)phthalate
 Concen: 24.10 ng/ul
 RT: 21.67 min Scan# 3205
 Delta R.T. 0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34




Instrument :
 BNA_G
 ClientSampleId :
 SST02021


Tgt Ion:149 Resp: 439644
 Ion Ratio Lower Upper
 149 100
 167 27.0 21.8 32.8
 279 5.5 4.5 6.7



#82
 Chrysene
 Concen: 22.13 ng/ul
 RT: 21.89 min Scan# 3243
 Delta R.T. -0.00 min
 Lab File: BG025672.D
 Acq: 26 Jan 2017 18:34

Tgt Ion:228 Resp: 751609
 Ion Ratio Lower Upper
 228 100
 226 29.5 24.0 36.0
 229 19.9 16.9 25.3





#84

Di-n-octyl phthalate

Concen: 25.16 ng/ul

RT: 22.91 min Scan# 3417

Delta R.T. -0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34


Instrument :

BNA_G




ClientSampleId :

SSTD02021

Tgt Ion:149 Resp: 741882



Abundance Ion 149.00 (148.70 to 149.70): E



#85

Benzo(b)fluoranthene

Concen: 21.67 ng/ul

RT: 24.22 min Scan# 3639

Delta R.T. 0.07 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34


Tgt Ion:252 Resp: 770660

Ion Ratio Lower Upper



252 100


253 22.6 17.7 26.5

125 5.3 4.0 6.0



Abundance Ion 252.00 (251.70 to 252.70): E









#86
Benzo(k)fluoranthene
Concen: 23.16 ng/ul
RT: 24.22 min Scan# 3639
Delta R.T. -0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34




Instrument :
BNA_G
ClientSampleId :
SSTD02021


Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.6	18.5	27.7
125	5.3	4.1	6.1



#88
Benzo(a)pyrene
Concen: 22.54 ng/ul
RT: 25.08 min Scan# 3786
Delta R.T. 0.00 min
Lab File: BG025672.D
Acq: 26 Jan 2017 18:34

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.3	17.8	26.6
125	5.7	5.4	8.0





#90

Dibenzo(a,h)anthracene

Concen: 21.67 ng/ul

RT: 29.19 min Scan# 4485

Delta R.T. -0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34

Instrument :

BNA_G

ClientSampleId :

SSTD02021


Tgt Ion:278 Resp: 768998

Ion Ratio Lower Upper

278 100

139 8.4 6.7 10.1




279 24.1 20.9 31.3



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#91

Benzo(g,h,i)perylene

Concen: 20.55 ng/ul

RT: 30.38 min Scan# 4688

Delta R.T. 0.00 min

Lab File: BG025672.D

Acq: 26 Jan 2017 18:34


Tgt Ion:276 Resp: 718919

Ion Ratio Lower Upper

276 100

138 9.3 8.6 12.8

277 24.4 17.6 26.4



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E

