

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012621\
 Data File : BG048015.D
 Acq On : 26 Jan 2021 20:34
 Operator : CG/JU
 Sample : M1188-06
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP-SWITCH-GEAR

Quant Time: Jan 26 23:33:59 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 25 19:39:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.134	152	37332	20.00	ng	# 0.00
21) Naphthalene-d8	10.943	136	154088	20.00	ng	# 0.00
39) Acenaphthene-d10	14.750	164	123311	20.00	ng	0.00
64) Phenanthrene-d10	17.494	188	344374	20.00	ng	0.00
76) Chrysene-d12	21.765	240	348458	20.00	ng	# 0.00
86) Perylene-d12	25.056	264	337276	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.690	112	335433	138.09	ng	0.00
7) Phenol-d6	7.276	99	450047	121.85	ng	0.00
23) Nitrobenzene-d5	9.292	82	332078	85.27	ng	0.00
42) 2,4,6-Tribromophenol	16.236	330	308457	150.19	ng	0.00
45) 2-Fluorobiphenyl	13.381	172	779889	81.77	ng	0.00
79) Terphenyl-d14	20.097	244	1515969	75.43	ng	0.00
Target Compounds						Qvalue
32) Benzoic acid	10.167	122	16156	7.74	ng	88
70) Pentachlorophenol	17.135	266	37946	12.49	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012621\
Data File : BG048015.D
Acq On : 26 Jan 2021 20:34
Operator : CG/JU
Sample : M1188-06
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP-SWITCH-GEAR

Quant Time: Jan 26 23:33:59 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 25 19:39:42 2021
Response via : Initial Calibration

