

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012621\
 Data File : BG048000.D
 Acq On : 26 Jan 2021 9:14
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 26 10:57:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 25 19:39:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
2	1,4-Dioxane	0.580	0.602	-3.8	99	0.00
3	Pyridine	1.741	1.822	-4.7	89	0.00
4	n-Nitrosodimethylamine	0.805	0.826	-2.6	88	0.00
5 S	2-Fluorophenol	1.301	1.321	-1.5	90	0.00
6	Aniline	2.390	2.388	0.1	86	0.00
7 S	Phenol-d6	1.979	2.048	-3.5	90	0.00
8	2-Chlorophenol	1.362	1.407	-3.3	90	0.00
9	Benzaldehyde	1.015	1.063	-4.7	90	0.00
10 C	Phenol	1.897	1.929	-1.7	88	0.00
11	bis(2-Chloroethyl)ether	1.485	1.498	-0.9	87	0.00
12	1,3-Dichlorobenzene	1.644	1.704	-3.6	92	0.00
13 C	1,4-Dichlorobenzene	1.648	1.712	-3.9	91	0.00
14	1,2-Dichlorobenzene	1.573	1.595	-1.4	88	0.00
15	Benzyl Alcohol	1.674	1.709	-2.1	87	0.00
16	2,2'-oxybis(1-Chloropropane	2.085	2.033	2.5	85	-0.01
17	2-Methylphenol	1.282	1.277	0.4	88	0.00
18	Hexachloroethane	0.643	0.656	-2.0	87	0.00
19 P	n-Nitroso-di-n-propylamine	1.444	1.423	1.5	86	0.00
20	3+4-Methylphenols	1.773	1.757	0.9	85	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
22	Acetophenone	0.616	0.620	-0.6	85	0.00
23 S	Nitrobenzene-d5	0.505	0.518	-2.6	87	0.00
24	Nitrobenzene	0.506	0.523	-3.4	87	0.00
25	Isophorone	0.907	0.935	-3.1	87	0.00
26 C	2-Nitrophenol	0.208	0.213	-2.4	88	0.00
27	2,4-Dimethylphenol	0.303	0.305	-0.7	87	0.00
28	bis(2-Chloroethoxy)methane	0.546	0.559	-2.4	88	0.00
29 C	2,4-Dichlorophenol	0.389	0.409	-5.1	89	0.00
30	1,2,4-Trichlorobenzene	0.471	0.495	-5.1	88	0.00
31	Naphthalene	1.156	1.184	-2.4	87	0.00
32	Benzoic acid	0.271	0.310	-14.4	91	0.00
33	4-Chloroaniline	0.529	0.557	-5.3	89	0.00
34 C	Hexachlorobutadiene	0.334	0.343	-2.7	87	0.00
35	Caprolactam	0.148	0.162	-9.5	92	0.00
36 C	4-Chloro-3-methylphenol	0.437	0.462	-5.7	90	0.00
37	2-Methylnaphthalene	0.873	0.918	-5.2	89	0.00
38	1-Methylnaphthalene	0.833	0.864	-3.7	87	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
40	1,2,4,5-Tetrachlorobenzene	0.762	0.772	-1.3	90	0.00
41 P	Hexachlorocyclopentadiene	0.432	0.434	-0.5	87	0.00
42 S	2,4,6-Tribromophenol	0.333	0.369	-10.8	97	0.00
43 C	2,4,6-Trichlorophenol	0.532	0.556	-4.5	92	0.00
44	2,4,5-Trichlorophenol	0.551	0.578	-4.9	90	0.00
45 S	2-Fluorobiphenyl	1.547	1.583	-2.3	91	0.00
46	1,1'-Biphenyl	1.546	1.570	-1.6	90	0.00
47	2-Chloronaphthalene	1.346	1.363	-1.3	90	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012621\
 Data File : BG048000.D
 Acq On : 26 Jan 2021 9:14
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 26 10:57:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 25 19:39:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.488	0.510	-4.5	89	0.00
49	Acenaphthylene	1.986	2.060	-3.7	92	0.00
50	Dimethylphthalate	1.810	1.906	-5.3	93	0.00
51	2,6-Dinitrotoluene	0.376	0.400	-6.4	93	0.00
52 C	Acenaphthene	1.344	1.385	-3.1	91	0.00
53	3-Nitroaniline	0.413	0.453	-9.7	96	0.00
54 P	2,4-Dinitrophenol	0.235	0.268	-14.0	95	0.00
55	Dibenzofuran	2.039	2.132	-4.6	93	0.00
56 P	4-Nitrophenol	0.322	0.375	-16.5	100	0.00
57	2,4-Dinitrotoluene	0.543	0.597	-9.9	94	0.00
58	Fluorene	1.639	1.738	-6.0	94	0.00
59	2,3,4,6-Tetrachlorophenol	0.551	0.605	-9.8	97	0.00
60	Diethylphthalate	1.861	1.967	-5.7	93	0.00
61	4-Chlorophenyl-phenylether	1.004	1.063	-5.9	93	0.00
62	4-Nitroaniline	0.450	0.502	-11.6	97	0.00
63	Azobenzene	1.752	1.835	-4.7	93	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
65	4,6-Dinitro-2-methylphenol	0.129	0.133	-3.1	94	0.00
66 c	n-Nitrosodiphenylamine	0.601	0.601	0.0	94	0.00
67	4-Bromophenyl-phenylether	0.267	0.271	-1.5	96	0.00
68	Hexachlorobenzene	0.291	0.295	-1.4	97	0.00
69	Atrazine	0.228	0.240	-5.3	98	0.00
70 C	Pentachlorophenol	0.176	0.189	-7.4	97	0.00
71	Phenanthrene	1.154	1.185	-2.7	97	0.00
72	Anthracene	1.135	1.160	-2.2	97	0.00
73	Carbazole	1.059	1.106	-4.4	99	0.00
74	Di-n-butylphthalate	1.293	1.330	-2.9	97	0.00
75 C	Fluoranthene	1.513	1.599	-5.7	101	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
77	Benzydine	0.577	0.638	-10.6	101	0.00
78	Pyrene	1.465	1.498	-2.3	101	0.00
79 S	Terphenyl-d14	1.154	1.164	-0.9	100	0.00
80	Butylbenzylphthalate	0.559	0.576	-3.0	101	0.00
81	Benzo(a)anthracene	1.454	1.486	-2.2	102	0.00
82	3,3'-Dichlorobenzidine	0.490	0.509	-3.9	102	0.00
83	Chrysene	1.381	1.413	-2.3	102	0.00
84	Bis(2-ethylhexyl)phthalate	0.766	0.780	-1.8	102	0.00
85 c	Di-n-octyl phthalate	1.282	1.301	-1.5	100	0.00
86 I	Perylene-d12	1.000	1.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	1.609	1.632	-1.4	101	-0.01
88	Benzo(b)fluoranthene	1.421	1.437	-1.1	100	0.00
89	Benzo(k)fluoranthene	1.381	1.407	-1.9	101	0.00
90 C	Benzo(a)pyrene	1.335	1.361	-1.9	101	0.00
91	Dibenzo(a,h)anthracene	1.328	1.334	-0.5	99	0.00
92	Benzo(g,h,i)perylene	1.279	1.299	-1.6	101	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012621\
Data File : BG048000.D
Acq On : 26 Jan 2021 9:14
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jan 26 10:57:29 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 25 19:39:42 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0