

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012623\
 Data File : BG056442.D
 Acq On : 26 Jan 2023 14:24
 Operator : CG/JU
 Sample : 01283-02
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP-H

Quant Time: Jan 27 00:58:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 25 04:54:56 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.293	152	32909	20.000	ng	0.00
21) Naphthalene-d8	11.125	136	118520	20.000	ng	0.00
39) Acenaphthene-d10	14.909	164	93985	20.000	ng	0.00
64) Phenanthrene-d10	17.647	188	265155	20.000	ng	0.00
76) Chrysene-d12	21.936	240	272022	20.000	ng	# 0.00
86) Perylene-d12	25.373	264	293134	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.820	112	256824	139.341	ng	0.00
7) Phenol-d6	7.441	99	311032	109.885	ng	0.00
23) Nitrobenzene-d5	9.468	82	192156	82.932	ng	0.00
42) 2,4,6-Tribromophenol	16.395	330	198677	166.913	ng	0.00
45) 2-Fluorobiphenyl	13.534	172	638769	93.853	ng	0.00
79) Terphenyl-d14	20.220	244	1492069	98.241	ng	0.00


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012623\
Data File : BG056442.D
Acq On : 26 Jan 2023 14:24
Operator : CG/JU
Sample : 01283-02
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP-H

Quant Time: Jan 27 00:58:39 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 25 04:54:56 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.293 min Scan# 81
Delta R.T. 0.000 min
Lab File: BG056442.D
Acq: 26 Jan 2023 14:24




Instrument :
BNA_G
ClientSampleId :
TP-H


Tgt Ion:152 Resp: 32909
Ion Ratio Lower Upper
152 100
150 156.8 127.9 191.9
115 54.5 44.2 66.2



#5
2-Fluorophenol
Concen: 139.341 ng
RT: 5.820 min Scan# 431
Delta R.T. 0.006 min
Lab File: BG056442.D
Acq: 26 Jan 2023 14:24

Tgt Ion:112 Resp: 256824
Ion Ratio Lower Upper
112 100
64 40.0 33.3 49.9
63 23.0 19.0 28.4









#7
Phenol-d6
Concen: 109.885 ng
RT: 7.441 min Scan# 70
Delta R.T. 0.006 min
Lab File: BG056442.D
Acq: 26 Jan 2023 14:24




Instrument :
BNA_G
ClientSampleId :
TP-H


Tgt Ion: 99 Resp: 311032
Ion Ratio Lower Upper
99 100
42 7.7 5.9 8.9
71 29.4 25.4 38.2



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.125 min Scan# 1334
Delta R.T. -0.006 min
Lab File: BG056442.D
Acq: 26 Jan 2023 14:24

Tgt Ion: 136 Resp: 118520
Ion Ratio Lower Upper
136 100
137 11.8 9.4 14.0
54 2.4 2.2 3.2
68 4.3 3.4 5.0









#23
Nitrobenzene-d5
Concen: 82.932 ng
RT: 9.468 min Scan# 1052
Delta R.T. -0.006 min
Lab File: BG056442.D
Acq: 26 Jan 2023 14:24




Instrument :
BNA_G
ClientSampleId :
TP-H


Tgt Ion: 82 Resp: 192156
Ion Ratio Lower Upper
82 100
128 48.8 33.4 50.2
54 26.3 20.3 30.5



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.909 min Scan# 1978
Delta R.T. -0.006 min
Lab File: BG056442.D
Acq: 26 Jan 2023 14:24

Tgt Ion: 164 Resp: 93985
Ion Ratio Lower Upper
164 100
162 104.2 84.7 127.1
160 41.6 36.3 54.5





#42

2,4,6-Tribromophenol

Concen: 166.913 ng

RT: 16.395 min Scan# 21

Delta R.T. -0.000 min

Lab File: BG056442.D

Acq: 26 Jan 2023 14:24

Instrument :

BNA_G

ClientSampleId :

TP-H





Tgt Ion:330 Resp: 198677

Ion Ratio Lower Upper

330 100

332 97.5 77.3 115.9

141 30.3 27.9 41.9



#45

2-Fluorobiphenyl

Concen: 93.853 ng

RT: 13.534 min Scan# 1744

Delta R.T. -0.006 min

Lab File: BG056442.D

Acq: 26 Jan 2023 14:24




Tgt Ion:172 Resp: 638769


Ion Ratio Lower Upper

172 100

171 35.2 29.6 44.4

170 23.4 19.4 29.2









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.647 min Scan# 2444
Delta R.T. -0.006 min
Lab File: BG056442.D
Acq: 26 Jan 2023 14:24




Instrument :
BNA_G
ClientSampleId :
TP-H


Tgt Ion:188 Resp: 265155
Ion Ratio Lower Upper
188 100
94 4.9 4.2 6.2
80 5.9 5.0 7.4



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.936 min Scan# 3174
Delta R.T. -0.006 min
Lab File: BG056442.D
Acq: 26 Jan 2023 14:24

Tgt Ion:240 Resp: 272022
Ion Ratio Lower Upper
240 100
120 4.5 2.6 4.0#
236 26.2 20.8 31.2









#79
 Terphenyl-d14
 Concen: 98.241 ng
 RT: 20.220 min Scan# 2882
 Delta R.T. -0.006 min
 Lab File: BG056442.D
 Acq: 26 Jan 2023 14:24

Instrument :
 BNA_G
 ClientSampleId :
 TP-H

Tgt Ion:244 Resp: 1492069
 Ion Ratio Lower Upper
 244 100
 212 7.1 5.7 8.5
 122 5.5 3.5 5.3#



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 25.373 min Scan# 3759
 Delta R.T. -0.000 min
 Lab File: BG056442.D
 Acq: 26 Jan 2023 14:24

Tgt Ion:264 Resp: 293134
 Ion Ratio Lower Upper
 264 100
 260 23.1 18.6 28.0
 265 21.5 16.7 25.1

