


Data Path : U:\HPCHEM1\BNA G\Data\BG012718\
 Data File : BG032365.D
 Acq On : 27 Jan 2018 21:10
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG011218.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jan 24 14:02:41 2018



AutoFind: Scans 2582, 2583, 2584; Background Corrected with Scan 2576

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	20.2	10778	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	23.0	12287	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	33.2	17737	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	53429	PASS
199	198	5	9	6.8	3631	PASS
275	198	10	60	35.0	18721	PASS
365	198	1	100	3.7	1958	PASS
441	442	0.01	24	15.8	6622	PASS
442	198	50	100	78.2	41800	PASS
443	442	15	24	19.7	8216	PASS