

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012721\
 Data File : BG048034.D
 Acq On : 27 Jan 2021 18:52
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 27 23:29:39 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 25 19:39:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.580	0.584	-0.7	108	0.00
3	Pyridine	1.741	1.776	-2.0	98	0.00
4	n-Nitrosodimethylamine	0.805	0.835	-3.7	101	0.00
5 S	2-Fluorophenol	1.301	1.289	0.9	99	0.00
6	Aniline	2.390	2.304	3.6	94	0.00
7 S	Phenol-d6	1.979	1.982	-0.2	98	0.00
8	2-Chlorophenol	1.362	1.328	2.5	96	0.00
9	Benzaldehyde	1.015	1.024	-0.9	97	0.00
10 C	Phenol	1.897	1.874	1.2	97	0.00
11	bis(2-Chloroethyl)ether	1.485	1.426	4.0	93	0.00
12	1,3-Dichlorobenzene	1.644	1.634	0.6	99	-0.01
13 C	1,4-Dichlorobenzene	1.648	1.619	1.8	97	-0.01
14	1,2-Dichlorobenzene	1.573	1.525	3.1	95	0.00
15	Benzyl Alcohol	1.674	1.648	1.6	94	0.00
16	2,2'-oxybis(1-Chloropropane	2.085	1.864	10.6	88	-0.01
17	2-Methylphenol	1.282	1.216	5.1	94	0.00
18	Hexachloroethane	0.643	0.603	6.2	91	0.00
19 P	n-Nitroso-di-n-propylamine	1.444	1.354	6.2	92	0.00
20	3+4-Methylphenols	1.773	1.688	4.8	93	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
22	Acetophenone	0.616	0.626	-1.6	92	0.00
23 S	Nitrobenzene-d5	0.505	0.527	-4.4	95	0.00
24	Nitrobenzene	0.506	0.527	-4.2	95	-0.01
25	Isophorone	0.907	0.878	3.2	89	0.00
26 C	2-Nitrophenol	0.208	0.221	-6.3	98	0.00
27	2,4-Dimethylphenol	0.303	0.306	-1.0	94	0.00
28	bis(2-Chloroethoxy)methane	0.546	0.545	0.2	92	0.00
29 C	2,4-Dichlorophenol	0.389	0.420	-8.0	98	0.00
30	1,2,4-Trichlorobenzene	0.471	0.495	-5.1	95	0.00
31	Naphthalene	1.156	1.188	-2.8	94	0.00
32	Benzoic acid	0.271	0.291	-7.4	92	0.00
33	4-Chloroaniline	0.529	0.539	-1.9	92	0.00
34 C	Hexachlorobutadiene	0.334	0.352	-5.4	96	-0.01
35	Caprolactam	0.148	0.150	-1.4	92	0.00
36 C	4-Chloro-3-methylphenol	0.437	0.446	-2.1	94	0.00
37	2-Methylnaphthalene	0.873	0.884	-1.3	93	0.00
38	1-Methylnaphthalene	0.833	0.841	-1.0	92	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
40	1,2,4,5-Tetrachlorobenzene	0.762	0.810	-6.3	98	0.00
41 P	Hexachlorocyclopentadiene	0.432	0.472	-9.3	97	0.00
42 S	2,4,6-Tribromophenol	0.333	0.366	-9.9	99	0.00
43 C	2,4,6-Trichlorophenol	0.532	0.569	-7.0	98	0.00
44	2,4,5-Trichlorophenol	0.551	0.594	-7.8	96	0.00
45 S	2-Fluorobiphenyl	1.547	1.606	-3.8	95	0.00
46	1,1'-Biphenyl	1.546	1.585	-2.5	94	0.00
47	2-Chloronaphthalene	1.346	1.384	-2.8	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012721\
 Data File : BG048034.D
 Acq On : 27 Jan 2021 18:52
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 27 23:29:39 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 25 19:39:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.488	0.506	-3.7	91	0.00
49	Acenaphthylene	1.986	2.051	-3.3	94	0.00
50	Dimethylphthalate	1.810	1.867	-3.1	94	0.00
51	2,6-Dinitrotoluene	0.376	0.388	-3.2	93	0.00
52 C	Acenaphthene	1.344	1.400	-4.2	95	0.00
53	3-Nitroaniline	0.413	0.429	-3.9	94	0.00
54 P	2,4-Dinitrophenol	0.235	0.278	-18.3	101	0.00
55	Dibenzofuran	2.039	2.090	-2.5	94	0.00
56 P	4-Nitrophenol	0.322	0.345	-7.1	95	0.00
57	2,4-Dinitrotoluene	0.543	0.572	-5.3	93	0.00
58	Fluorene	1.639	1.695	-3.4	95	0.00
59	2,3,4,6-Tetrachlorophenol	0.551	0.583	-5.8	97	0.00
60	Diethylphthalate	1.861	1.890	-1.6	92	0.00
61	4-Chlorophenyl-phenylether	1.004	1.041	-3.7	94	0.00
62	4-Nitroaniline	0.450	0.468	-4.0	93	0.00
63	Azobenzene	1.752	1.740	0.7	91	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
65	4,6-Dinitro-2-methylphenol	0.129	0.141	-9.3	98	0.00
66 c	n-Nitrosodiphenylamine	0.601	0.603	-0.3	94	0.00
67	4-Bromophenyl-phenylether	0.267	0.274	-2.6	96	0.00
68	Hexachlorobenzene	0.291	0.303	-4.1	99	0.00
69	Atrazine	0.228	0.232	-1.8	93	0.00
70 C	Pentachlorophenol	0.176	0.197	-11.9	100	0.00
71	Phenanthrene	1.154	1.172	-1.6	94	0.00
72	Anthracene	1.135	1.149	-1.2	95	0.00
73	Carbazole	1.059	1.079	-1.9	95	0.00
74	Di-n-butylphthalate	1.293	1.280	1.0	93	0.00
75 C	Fluoranthene	1.513	1.581	-4.5	99	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
77	Benzidine	0.577	0.609	-5.5	98	0.00
78	Pyrene	1.465	1.458	0.5	99	0.00
79 S	Terphenyl-d14	1.154	1.140	1.2	99	0.00
80	Butylbenzylphthalate	0.559	0.552	1.3	98	0.00
81	Benzo(a)anthracene	1.454	1.477	-1.6	102	-0.01
82	3,3'-Dichlorobenzidine	0.490	0.517	-5.5	104	0.00
83	Chrysene	1.381	1.406	-1.8	102	0.00
84	Bis(2-ethylhexyl)phthalate	0.766	0.766	0.0	101	0.00
85 c	Di-n-octyl phthalate	1.282	1.315	-2.6	102	-0.01
86 I	Perylene-d12	1.000	1.000	0.0	104	-0.01
87	Indeno(1,2,3-cd)pyrene	1.609	1.671	-3.9	106	-0.03
88	Benzo(b)fluoranthene	1.421	1.466	-3.2	106	-0.01
89	Benzo(k)fluoranthene	1.381	1.422	-3.0	106	-0.01
90 C	Benzo(a)pyrene	1.335	1.378	-3.2	106	0.00
91	Dibenzo(a,h)anthracene	1.328	1.367	-2.9	105	-0.03
92	Benzo(g,h,i)perylene	1.279	1.335	-4.4	107	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012721\
Data File : BG048034.D
Acq On : 27 Jan 2021 18:52
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jan 27 23:29:39 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 25 19:39:42 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0