

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012721\
 Data File : BG048034.D
 Acq On : 27 Jan 2021 18:52
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 27 23:29:39 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 25 19:39:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	99	0.00
2	1,4-Dioxane	40.000	40.255	-0.6	108	0.00
3	Pyridine	40.000	40.788	-2.0	98	0.00
4	n-Nitrosodimethylamine	40.000	41.535	-3.8	101	0.00
5 S	2-Fluorophenol	80.000	79.226	1.0	99	0.00
6	Aniline	40.000	38.554	3.6	94	0.00
7 S	Phenol-d6	80.000	80.147	-0.2	98	0.00
8	2-Chlorophenol	40.000	38.998	2.5	96	0.00
9	Benzaldehyde	40.000	40.348	-0.9	97	0.00
10 C	Phenol	40.000	39.526	1.2	97	0.00
11	bis(2-Chloroethyl)ether	40.000	38.410	4.0	93	0.00
12	1,3-Dichlorobenzene	40.000	39.761	0.6	99	-0.01
13 C	1,4-Dichlorobenzene	40.000	39.302	1.7	97	-0.01
14	1,2-Dichlorobenzene	40.000	38.762	3.1	95	0.00
15	Benzyl Alcohol	40.000	39.387	1.5	94	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.774	10.6	88	-0.01
17	2-Methylphenol	40.000	37.936	5.2	94	0.00
18	Hexachloroethane	40.000	37.533	6.2	91	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.511	6.2	92	0.00
20	3+4-Methylphenols	40.000	38.069	4.8	93	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	94	0.00
22	Acetophenone	40.000	40.649	-1.6	92	0.00
23 S	Nitrobenzene-d5	80.000	83.453	-4.3	95	0.00
24	Nitrobenzene	40.000	41.636	-4.1	95	-0.01
25	Isophorone	40.000	38.726	3.2	89	0.00
26 C	2-Nitrophenol	40.000	42.459	-6.1	98	0.00
27	2,4-Dimethylphenol	40.000	40.472	-1.2	94	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.976	0.1	92	0.00
29 C	2,4-Dichlorophenol	40.000	43.111	-7.8	98	0.00
30	1,2,4-Trichlorobenzene	40.000	42.059	-5.1	95	0.00
31	Naphthalene	40.000	41.098	-2.7	94	0.00
32	Benzoic acid	40.000	43.004	-7.5	92	0.00
33	4-Chloroaniline	40.000	40.772	-1.9	92	0.00
34 C	Hexachlorobutadiene	40.000	42.163	-5.4	96	-0.01
35	Caprolactam	40.000	40.375	-0.9	92	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.836	-2.1	94	0.00
37	2-Methylnaphthalene	40.000	40.466	-1.2	93	0.00
38	1-Methylnaphthalene	40.000	40.394	-1.0	92	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	94	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.484	-6.2	98	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.737	-9.3	97	0.00
42 S	2,4,6-Tribromophenol	80.000	87.919	-9.9	99	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.813	-7.0	98	0.00
44	2,4,5-Trichlorophenol	40.000	43.103	-7.8	96	0.00
45 S	2-Fluorobiphenyl	80.000	83.065	-3.8	95	0.00
46	1,1'-Biphenyl	40.000	41.015	-2.5	94	0.00
47	2-Chloronaphthalene	40.000	41.136	-2.8	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012721\
 Data File : BG048034.D
 Acq On : 27 Jan 2021 18:52
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 27 23:29:39 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 25 19:39:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.460	-3.7	91	0.00
49	Acenaphthylene	40.000	41.310	-3.3	94	0.00
50	Dimethylphthalate	40.000	41.249	-3.1	94	0.00
51	2,6-Dinitrotoluene	40.000	41.236	-3.1	93	0.00
52 C	Acenaphthene	40.000	41.661	-4.2	95	0.00
53	3-Nitroaniline	40.000	41.567	-3.9	94	0.00
54 P	2,4-Dinitrophenol	40.000	47.188	-18.0	101	0.00
55	Dibenzofuran	40.000	40.990	-2.5	94	0.00
56 P	4-Nitrophenol	40.000	42.854	-7.1	95	0.00
57	2,4-Dinitrotoluene	40.000	42.124	-5.3	93	0.00
58	Fluorene	40.000	41.354	-3.4	95	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.288	-5.7	97	0.00
60	Diethylphthalate	40.000	40.619	-1.5	92	0.00
61	4-Chlorophenyl-phenylether	40.000	41.487	-3.7	94	0.00
62	4-Nitroaniline	40.000	41.601	-4.0	93	0.00
63	Azobenzene	40.000	39.719	0.7	91	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	95	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.620	-9.0	98	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.124	-0.3	94	0.00
67	4-Bromophenyl-phenylether	40.000	41.025	-2.6	96	0.00
68	Hexachlorobenzene	40.000	41.726	-4.3	99	0.00
69	Atrazine	40.000	40.569	-1.4	93	0.00
70 C	Pentachlorophenol	40.000	44.543	-11.4	100	0.00
71	Phenanthrene	40.000	40.627	-1.6	94	0.00
72	Anthracene	40.000	40.493	-1.2	95	0.00
73	Carbazole	40.000	40.764	-1.9	95	0.00
74	Di-n-butylphthalate	40.000	39.591	1.0	93	0.00
75 C	Fluoranthene	40.000	41.791	-4.5	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	103	0.00
77	Benzydine	40.000	42.210	-5.5	98	0.00
78	Pyrene	40.000	39.815	0.5	99	0.00
79 S	Terphenyl-d14	80.000	79.039	1.2	99	0.00
80	Butylbenzylphthalate	40.000	39.500	1.3	98	0.00
81	Benzo(a)anthracene	40.000	40.643	-1.6	102	-0.01
82	3,3'-Dichlorobenzidine	40.000	42.215	-5.5	104	0.00
83	Chrysene	40.000	40.721	-1.8	102	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.016	-0.0	101	0.00
85 c	Di-n-octyl phthalate	40.000	41.050	-2.6	102	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	104	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	41.539	-3.8	106	-0.03
88	Benzo(b)fluoranthene	40.000	41.251	-3.1	106	-0.01
89	Benzo(k)fluoranthene	40.000	41.178	-2.9	106	-0.01
90 C	Benzo(a)pyrene	40.000	41.292	-3.2	106	0.00
91	Dibenzo(a,h)anthracene	40.000	41.179	-2.9	105	-0.03
92	Benzo(g,h,i)perylene	40.000	41.754	-4.4	107	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012721\
Data File : BG048034.D
Acq On : 27 Jan 2021 18:52
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jan 27 23:29:39 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 25 19:39:42 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0