

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012723\
 Data File : BG056461.D
 Acq On : 27 Jan 2023 17:47
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG012723

Quant Time: Jan 28 00:41:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG012723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 28 00:39:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	91	0.00
2	1,4-Dioxane	40.000	39.784	0.5	92	0.00
3	Pyridine	40.000	41.666	-4.2	93	0.00
4	n-Nitrosodimethylamine	40.000	40.587	-1.5	92	0.00
5 S	2-Fluorophenol	80.000	83.718	-4.6	95	0.00
6	Aniline	40.000	42.110	-5.3	94	0.00
7 S	Phenol-d6	80.000	84.943	-6.2	95	0.00
8	2-Chlorophenol	40.000	41.343	-3.4	94	0.00
9	Benzaldehyde	40.000	44.756	-11.9	101	0.00
10 C	Phenol	40.000	41.774	-4.4	95	0.00
11	bis(2-Chloroethyl)ether	40.000	41.732	-4.3	94	0.00
12	1,3-Dichlorobenzene	40.000	41.690	-4.2	96	0.00
13 C	1,4-Dichlorobenzene	40.000	41.505	-3.8	96	0.00
14	1,2-Dichlorobenzene	40.000	41.570	-3.9	94	0.00
15	Benzyl Alcohol	40.000	43.036	-7.6	94	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.172	-2.9	91	0.00
17	2-Methylphenol	40.000	43.123	-7.8	96	0.00
18	Hexachloroethane	40.000	40.176	-0.4	92	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.687	-6.7	94	0.00
20	3+4-Methylphenols	40.000	43.439	-8.6	95	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	96	0.00
22	Acetophenone	40.000	40.295	-0.7	94	0.00
23 S	Nitrobenzene-d5	80.000	81.018	-1.3	96	0.00
24	Nitrobenzene	40.000	40.513	-1.3	95	0.00
25	Isophorone	40.000	40.662	-1.7	94	0.00
26 C	2-Nitrophenol	40.000	41.201	-3.0	95	0.00
27	2,4-Dimethylphenol	40.000	41.202	-3.0	96	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.020	-0.1	94	0.00
29 C	2,4-Dichlorophenol	40.000	41.331	-3.3	95	0.00
30	1,2,4-Trichlorobenzene	40.000	40.956	-2.4	95	0.00
31	Naphthalene	40.000	40.680	-1.7	95	0.00
32	Benzoic acid	40.000	39.378	1.6	89	0.00
33	4-Chloroaniline	40.000	40.827	-2.1	93	0.00
34 C	Hexachlorobutadiene	40.000	41.281	-3.2	97	0.00
35	Caprolactam	40.000	40.128	-0.3	92	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.482	-3.7	95	0.00
37	2-Methylnaphthalene	40.000	41.242	-3.1	95	0.00
38	1-Methylnaphthalene	40.000	40.770	-1.9	94	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	94	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.019	-0.0	94	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.308	1.7	96	0.00
42 S	2,4,6-Tribromophenol	80.000	81.747	-2.2	95	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.923	-2.3	93	0.00
44	2,4,5-Trichlorophenol	40.000	40.911	-2.3	95	0.00
45 S	2-Fluorobiphenyl	80.000	80.863	-1.1	95	0.00
46	1,1'-Biphenyl	40.000	40.631	-1.6	95	0.00
47	2-Chloronaphthalene	40.000	40.286	-0.7	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012723\
 Data File : BG056461.D
 Acq On : 27 Jan 2023 17:47
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG012723

Quant Time: Jan 28 00:41:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG012723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 28 00:39:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.047	-2.6	94	0.00
49	Acenaphthylene	40.000	40.659	-1.6	94	0.00
50	Dimethylphthalate	40.000	40.337	-0.8	94	0.00
51	2,6-Dinitrotoluene	40.000	40.818	-2.0	96	0.00
52 C	Acenaphthene	40.000	40.160	-0.4	93	0.00
53	3-Nitroaniline	40.000	40.429	-1.1	93	0.00
54 P	2,4-Dinitrophenol	40.000	42.659	-6.6	94	0.00
55	Dibenzofuran	40.000	40.512	-1.3	94	0.00
56 P	4-Nitrophenol	40.000	41.755	-4.4	93	0.00
57	2,4-Dinitrotoluene	40.000	41.163	-2.9	94	0.00
58	Fluorene	40.000	40.377	-0.9	94	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.747	-6.9	95	0.00
60	Diethylphthalate	40.000	40.257	-0.6	94	0.00
61	4-Chlorophenyl-phenylether	40.000	40.889	-2.2	95	0.00
62	4-Nitroaniline	40.000	41.544	-3.9	96	0.00
63	Azobenzene	40.000	40.271	-0.7	93	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	96	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.367	-3.4	94	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.836	-2.1	96	0.00
67	4-Bromophenyl-phenylether	40.000	41.167	-2.9	95	0.00
68	Hexachlorobenzene	40.000	40.821	-2.1	95	0.00
69	Atrazine	40.000	41.717	-4.3	95	0.00
70 C	Pentachlorophenol	40.000	41.146	-2.9	93	0.00
71	Phenanthrene	40.000	40.309	-0.8	95	0.00
72	Anthracene	40.000	40.681	-1.7	95	0.00
73	Carbazole	40.000	40.608	-1.5	96	0.00
74	Di-n-butylphthalate	40.000	40.090	-0.2	95	0.00
75 C	Fluoranthene	40.000	40.040	-0.1	94	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	97	0.00
77	Benzidine	40.000	39.461	1.3	94	0.00
78	Pyrene	40.000	39.517	1.2	95	0.00
79 S	Terphenyl-d14	80.000	80.000	0.0	96	0.00
80	Butylbenzylphthalate	40.000	39.452	1.4	94	0.00
81	Benzo(a)anthracene	40.000	40.538	-1.3	96	0.00
82	3,3'-Dichlorobenzidine	40.000	39.104	2.2	93	0.00
83	Chrysene	40.000	40.810	-2.0	98	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.282	-0.7	97	0.00
85 c	Di-n-octyl phthalate	40.000	40.402	-1.0	96	0.00
86 I	Perylene-d12	20.000	20.000	0.0	96	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.034	-2.6	97	0.00
88	Benzo(b)fluoranthene	40.000	41.397	-3.5	98	0.00
89	Benzo(k)fluoranthene	40.000	41.239	-3.1	98	-0.01
90 C	Benzo(a)pyrene	40.000	41.075	-2.7	98	0.00
91	Dibenzo(a,h)anthracene	40.000	41.044	-2.6	98	0.00
92	Benzo(g,h,i)perylene	40.000	41.151	-2.9	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012723\
Data File : BG056461.D
Acq On : 27 Jan 2023 17:47
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG012723

Quant Time: Jan 28 00:41:31 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG012723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 28 00:39:13 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0