

Data Path : Z:\HPCHEM1\BNA G\DATA\BG012818\
 Data File : BG032405.D
 Acq On : 29 Jan 2018 1:19
 Operator : SJ/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02086

Manual Integrations
 APPROVED

Sohil
 1/29/2018 5:50:04 PM

Quant Time: Jan 29 06:37:50 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG012218.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 29 06:15:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.72	152	31124	20.00	ng/ul	0.00
18) Naphthalene-d8	11.60	136	123489	20.00	ng/ul	0.00
35) Acenaphthene-d10	15.35	164	89496	20.00	ng/ul	0.00
61) Phenanthrene-d10	18.09	188	247085	20.00	ng/ul	0.00
77) Chrysene-d12	22.42	240	323963	20.00	ng/ul	0.00
85) Perylene-d12	26.09	264	350756	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.77	96	4378	8.51	ng/uL	0.00
5) Phenol-d5	7.83	99	39616	18.52	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	8.01	67	17486	17.33	ng/ul	0.00
9) 2-Chlorophenol-d4	8.23	132	38583	19.34	ng/ul	0.00
13) 4-Methylphenol-d8	9.42	113	35604	18.33	ng/ul	0.00
19) Nitrobenzene-d5	9.91	128	17846	19.83	ng/ul	0.00
22) 2-Nitrophenol-d4	10.65	143	22882	21.21	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	11.21	165	46902	19.81	ng/ul	0.00
29) 4-Chloroaniline-d4	11.72	131	40732	24.67	ng/ul	0.00
43) Dimethylphthalate-d6	14.74	166	154127	20.63	ng/ul	0.00
46) Acenaphthylene-d8	15.05	160	180934	19.75	ng/ul	0.00
51) 4-Nitrophenol-d4	15.51	143	24161	24.27	ng/ul	0.00
57) Fluorene-d10	16.33	176	143860	20.68	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	16.43	200	34211	22.44	ng/ul	0.00
70) Anthracene-d10	18.19	188	235576	19.83	ng/ul	0.00
78) Pyrene-d10	20.42	212	299670	19.93	ng/ul	0.00
89) Benzo(a)pyrene-d12	25.83	264	317404	19.58	ng/ul	0.00

Target Compounds

				Ovalue	
2) 1,4-Dioxane	3.82	88	3556	6.089	ng/uL# 85
4) Benzaldehyde	7.82	77	24856	23.350	ng/ul 86
6) Phenol	7.86	94	39149	18.677	ng/ul# 82
8) Bis(2-Chloroethyl)ether	8.10	93	28476	19.321	ng/ul 89
10) 2-Chlorophenol	8.27	128	37062	19.661	ng/ul# 85
11) 2-Methylphenol	9.15	108	30635	18.363	ng/ul 97
12) 2,2'-oxybis(1-Chloropropan	9.25	45	24851	16.933	ng/ul# 88
14) Acetophenone	9.56	105	51399	18.894	ng/ul 96
15) N-Nitroso-di-n-propylamine	9.54	70	22687	17.868	ng/ul 99
16) 4-Methylphenol	9.49	108	33369	18.273	ng/ul 96
17) Hexachloroethane	9.84	117	14574	19.586	ng/ul# 75
20) Nitrobenzene	9.96	77	35061	19.355	ng/ul# 83
21) Isophorone	10.48	82	64299	19.708	ng/ul# 93
23) 2-Nitrophenol	10.69	139	21817	19.717	ng/ul# 89
24) 2,4-Dimethylphenol	10.72	107	42275	19.068	ng/ul 98
25) Bis(2-Chloroethoxy)methane	10.96	93	37129	17.945	ng/ul# 96
27) 2,4-Dichlorophenol	11.23	162	44513	19.968	ng/ul# 92
28) Naphthalene	11.65	128	112414	19.954	ng/ul 98
30) 4-Chloroaniline	11.75	127	40320	25.172	ng/ul 100
31) Hexachlorobutadiene	11.92	225	40344	20.022	ng/ul 98
32) Caprolactam	12.47	113	12302	21.339	ng/ul# 71
33) 4-Chloro-3-methylphenol	12.82	107	40650	20.150	ng/ul 94
34) 2-Methylnaphthalene	13.21	142	93560	19.527	ng/ul 100

Data Path : Z:\HPCHEM1\BNA G\DATA\BG012818\
 Data File : BG032405.D
 Acq On : 29 Jan 2018 1:19
 Operator : SJ/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02086


Manual Integrations
 APPROVED

Sohil
 1/29/2018 5:50:04 PM

Quant Time: Jan 29 06:37:50 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG012218.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 29 06:15:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
36) 1,2,4,5-Tetrachlorobenzene	13.57	216	75082	18.815	ng/ul	93
37) Hexachlorocyclopentadiene	13.55	237	38129	22.146	ng/ul	99
38) 2,4,6-Trichlorophenol	13.80	196	44440	19.278	ng/ul	92
39) 2,4,5-Trichlorophenol	13.87	196	47825	20.098	ng/ul	95
40) 1,1'-Biphenyl	14.20	154	132305	18.732	ng/ul#	96
41) 2-Chloronaphthalene	14.25	162	108381	19.138	ng/ul	98
42) 2-Nitroaniline	14.44	65	23553	20.826	ng/ul	89
44) Dimethylphthalate	14.78	163	147076	20.508	ng/ul	98
45) 2,6-Dinitrotoluene	14.91	165	30941	21.132	ng/ul	94
47) Acenaphthylene	15.08	152	155377	19.452	ng/ul	97
48) 3-Nitroaniline	15.24	138	25933	24.263	ng/ul#	94
49) Acenaphthene	15.42	153	112851	19.357	ng/ul	97
50) 2,4-Dinitrophenol	15.44	184	18728	24.746	ng/ul#	55
52) 4-Nitrophenol	15.52	109	24337	26.505	ng/ul	87
53) Dibenzofuran	15.75	168	176829	20.137	ng/ul	98
54) 2,4-Dinitrotoluene	15.69	165	48928	23.110	ng/ul	90
55) 2,3,4,6-Tetrachlorophenol	15.96	232	51259	22.513	ng/ul#	95
56) Diethylphthalate	16.12	149	146015	21.512	ng/ul	96
58) Fluorene	16.39	166	142460	20.464	ng/ul	96
59) 4-Chlorophenyl-phenylether	16.37	204	88430	20.121	ng/ul	96
60) 4-Nitroaniline	16.40	138	30074	24.408	ng/ul	94
63) 4,6-Dinitro-2-methylphenol	16.45	198	34897	21.963	ng/ul#	88
64) N-Nitrosodiphenylamine	16.58	169	128082	18.231	ng/ul	98
65) 4-Bromophenyl-phenylether	17.26	248	67656	19.015	ng/ul	93
66) Hexachlorobenzene	17.39	284	72296	19.565	ng/ul	96
67) Atrazine	17.50	200	67220	21.280	ng/ul	92
68) Pentachlorophenol	17.73	266	38688	20.994	ng/ul	95
69) Phenanthrene	18.13	178	255599	19.657	ng/ul	100
71) Anthracene	18.22	178	262578	19.680	ng/ul	99
72) 1,2,3,4-Tetrachlorobenzene	14.17	216	77358	16.439	ng/uL	96
73) Pentachlorobenzene	15.66	250	86055	17.396	ng/uL	98
74) Carbazole	18.48	167	218664	21.148	ng/ul	97
75) Di-n-butylphthalate	18.99	149	251923	20.662	ng/ul	100
76) Fluoranthene	20.09	202	356642	22.237	ng/ul#	92
79) Pvrene	20.45	202	362957	19.847	ng/ul#	90
80) Butylbenzylphthalate	21.31	149	119135	19.998	ng/ul#	85
81) 3,3'-Dichlorobenzidine	22.29	252	129673	25.036	ng/ul#	98
82) Benzo(a)anthracene	22.39	228	386022	19.712	ng/ul	99
83) Bis(2-ethylhexyl)phthalate	22.23	149	166328	19.674	ng/ul#	96
84) Chrysene	22.47	228	349779	19.679	ng/ul	99
86) Di-n-octyl phthalate	23.60	149	284790	19.468	ng/ul	100
87) Benzo(b)fluoranthene	24.91	252	403166	19.456	ng/ul#	98
88) Benzo(k)fluoranthene	24.99	252	390825	19.199	ng/ul#	95
90) Benzo(a)pyrene	25.92	252	386833	19.340	ng/ul#	96
91) Indeno(1,2,3-cd)pyrene	30.35	276	473403	20.916	ng/ul#	93
92) Dibenzo(a,h)anthracene	30.42	278	388563	20.984	ng/ul#	95
93) Benzo(g,h,i)perylene	31.68	276	395602m	21.957	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 6.089 ng/uL

RT: 3.82 min Scan# 90

Delta R.T. 0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Instrument :

BNA_G





ClientSampled :

SSTD02086

Tot Ion:	88	Resp:	3556
Ion	Ratio	Lower	Upper
88	100		
43	32.2	29.0	43.4
58	71.7	45.9	68.9#

Manual Integrations
APPROVED

Sohil
1/29/2018 5:50:04 PM



#4

Benzaldehyde

Concen: 23.350 ng/uL




RT: 7.82 min Scan# 771


Delta R.T. -0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Tgt Ion:	77	Resp:	24856
Ion	Ratio	Lower	Upper
77	100		
105	106.5	71.0	106.6
106	95.0	69.5	104.3





#6

Phenol

Concen: 18.677 ng/ul

RT: 7.86 min Scan# 777

Delta R.T. -0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Instrument :

BNA_G





ClientSampleId :

SSTD02086

Tot Ion:	94	Resp:	39149
Ion	Ratio	Lower	Upper
94	100		
65	34.3	23.0	34.6
66	52.2	30.4	45.6

Manual Integrations
APPROVED

Sohil
1/29/2018 5:50:04 PM



#8

Bis(2-Chloroethyl)ether

Concen: 19.321 ng/ul




RT: 8.10 min Scan# 819


Delta R.T. -0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Tgt Ion:	93	Resp:	28476
Ion	Ratio	Lower	Upper
93	100		
63	63.4	59.4	89.2
95	36.0	26.4	39.6









#10
2-Chlorophenol
Concen: 19.661 ng/ul
RT: 8.27 min Scan# 847
Delta R.T. 0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Instrument :
BNA_G
ClientSampleId :
SSTD02086

Tot Ion	Ratio	Lower	Upper
128	100		
64	33.9	39.8	59.6#
130	30.4	25.8	38.8




Manual Integrations
APPROVED


Sohil
1/29/2018 5:50:04 PM



#11
2-Methylphenol
Concen: 18.363 ng/ul
RT: 9.15 min Scan# 998
Delta R.T. -0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Lower	Upper
108	100		
107	95.4	74.1	111.1





#12

2,2'-oxvbis(1-Chloropropane)

Concen: 16.933 ng/ul

RT: 9.25 min Scan# 1015

Delta R.T. -0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Instrument :

BNA_G





ClientSampleId :

SSTD02086

Tot Ion:	45	Resp:	24851
Ion Ratio	Lower	Upper	
45	100		
77	22.2	12.5	18.7#
79	14.8	9.4	14.2#

Manual Integrations
APPROVED

Sohil
1/29/2018 5:50:04 PM



#14

Acetophenone

Concen: 18.894 ng/ul




RT: 9.56 min Scan# 1067




Delta R.T. -0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Tgt Ion:	105	Resp:	51399
Ion Ratio	Lower	Upper	
105	100		
77	81.6	63.2	94.8
51	28.7	19.8	29.8





#15

N-Nitroso-di-n-propylamine

Concen: 17.868 ng/ul

RT: 9.54 min Scan# 1063

Delta R.T. 0.01 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Tot Ion:	70	Resp:	22687
Ion Ratio	Lower	Upper	
70	100		
42	51.0	40.2	60.4
101	10.2	7.7	11.5
130	22.0	17.8	26.6

Instrument :





BNA_G

ClientSampleId :

SSTD02086

Manual Integrations
APPROVED

Sohil
1/29/2018 5:50:04 PM



#16

4-Methylphenol

Concen: 18.273 ng/ul


RT: 9.49 min Scan# 1055


Delta R.T. 0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Tgt Ion:	108	Resp:	33369
Ion Ratio	Lower	Upper	
108	100		
107	113.2	93.6	140.4









#17
Hexachloroethane
Concen: 19.586 ng/ul
RT: 9.84 min Scan# 1115
Delta R.T. -0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Instrument :
BNA_G
ClientSampleId :
SSTD02086

Tot Ion	Ratio	Lower	Upper
117	100		
201	146.4	96.1	144.1#
199	97.6	59.7	89.5#




Manual Integrations
APPROVED


Sohil
1/29/2018 5:50:04 PM



#20
Nitrobenzene
Concen: 19.355 ng/ul
RT: 9.96 min Scan# 1135
Delta R.T. 0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Lower	Upper
77	100		
123	53.2	31.8	47.8#
65	15.0	10.6	16.0









#21
Isophorone
Concen: 19.708 ng/ul
RT: 10.48 min Scan# 1224
Delta R.T. 0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Instrument :
BNA_G
ClientSampled :
SSTD02086

Tot Ion	82	Resp	64299
Ion	Ratio	Lower	Upper
82	100		
95	8.3	6.1	9.1
138	20.8	13.5	20.3#




Manual Integrations
APPROVED


Sohil
1/29/2018 5:50:04 PM



#23
2-Nitrophenol
Concen: 19.717 ng/ul
RT: 10.69 min Scan# 1259
Delta R.T. 0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Tgt Ion	139	Resp	21817
Ion	Ratio	Lower	Upper
139	100		
65	54.0	45.3	67.9
109	47.5	27.7	41.5#





#24

2,4-Dimethylphenol

Concen: 19.068 ng/ul

RT: 10.72 min Scan# 1265

Delta R.T. 0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Instrument :

BNA_G

ClientSampleId :

SSTD02086

Tot Ion: 107 Resp: 42275

Ion Ratio Lower Upper

107 100


121 49.8 38.1 57.1

122 78.9 63.8 95.8

Manual Integrations
APPROVED

Sohil




1/29/2018 5:50:04 PM



Abundance Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E

Ion 122.00 (121.70 to 122.70): E



#25

Bis(2-Chloroethoxy)methane

Concen: 17.945 ng/ul

RT: 10.96 min Scan# 1306

Delta R.T. 0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19


Tgt Ion: 93 Resp: 37129

Ion Ratio Lower Upper

93 100

95 32.6 24.7 37.1



123 12.8 8.2 12.2#




Abundance Ion 93.00 (92.70 to 93.70): BGC

Ion 95.00 (94.70 to 95.70): BGC

Ion 123.00 (122.70 to 123.70): E









#27
 2,4-Dichlorophenol
 Concen: 19.968 ng/ul
 RT: 11.23 min Scan# 1351
 Delta R.T. -0.00 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Instrument :
 BNA_G
 ClientSampleId :
 SST02086

Tot Ion	Ratio	Lower	Upper
162	100		
164	68.3	52.0	78.0
98	25.2	26.6	39.8




Manual Integrations
 APPROVED


Sohil
 1/29/2018 5:50:04 PM



#28
 Naphthalene
 Concen: 19.954 ng/ul
 RT: 11.65 min Scan# 1423
 Delta R.T. -0.00 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.3	12.5
127	14.3	10.8	16.2









#30
4-Chloroaniline
Concen: 25.172 ng/ul
RT: 11.75 min Scan# 1439
Delta R.T. 0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Instrument :
BNA_G
ClientSampleId :
SSTD02086

Tot Ion:127 Resp: 40320
Ion Ratio Lower Upper
127 100
129 30.6 24.6 37.0




Manual Integrations
APPROVED


Sohil
1/29/2018 5:50:04 PM



#31
Hexachlorobutadiene
Concen: 20.022 ng/ul
RT: 11.92 min Scan# 1469
Delta R.T. 0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Tgt Ion:225 Resp: 40344
Ion Ratio Lower Upper
225 100
223 63.4 48.1 72.1
227 65.3 52.1 78.1









#32
 Caprolactam
 Concen: 21.339 ng/ul
 RT: 12.47 min Scan# 1563
 Delta R.T. 0.01 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Instrument :
 BNA_G
 ClientSampleId :
 SST02086

Tot Ion	Ratio	Lower	Upper
113	100		
55	108.5	131.7	197.5#
56	103.0	92.7	139.1




Manual Integrations
 APPROVED


Sohil
 1/29/2018 5:50:04 PM



#33
 4-Chloro-3-methylphenol
 Concen: 20.150 ng/ul
 RT: 12.82 min Scan# 1622
 Delta R.T. 0.00 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Lower	Upper
107	100		
144	28.2	21.1	31.7
142	81.5	60.1	90.1









#34
 2-Methylnaphthalene
 Concen: 19.527 ng/ul
 RT: 13.21 min Scan# 1689
 Delta R.T. -0.00 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Instrument :
 BNA_G
 ClientSampleId :
 SST02086

Tot Ion:142 Resp: 93560
 Ion Ratio Lower Upper
 142 100
 141 88.9 71.4 107.2




Manual Integrations
 APPROVED


Sohil
 1/29/2018 5:50:04 PM



#36
 1,2,4,5-Tetrachlorobenzene
 Concen: 18.815 ng/ul
 RT: 13.57 min Scan# 1750
 Delta R.T. 0.00 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Tgt Ion:216 Resp: 75082
 Ion Ratio Lower Upper
 216 100
 214 73.1 64.1 96.1
 179 17.2 14.5 21.7
 108 10.5 10.4 15.6









#37
Hexachlorocyclopentadiene
Concen: 22.146 ng/ul
RT: 13.55 min Scan# 1746
Delta R.T. 0.01 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Instrument :
BNA_G
ClientSampleId :
SSTD02086

Tot Ion	Ratio	Resp	Lower	Upper
237	100			
235	61.0	49.5	74.3	
272	10.4	8.6	12.8	




Manual Integrations
APPROVED


Sohil
1/29/2018 5:50:04 PM



#38
2,4,6-Trichlorophenol
Concen: 19.278 ng/ul
RT: 13.80 min Scan# 1788
Delta R.T. -0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Resp	Lower	Upper
196	100			
198	105.8	75.9	113.9	
200	32.1	25.8	38.6	






#39
 2,4,5-Trichlorophenol
 Concen: 20.098 ng/ul
 RT: 13.87 min Scan# 1800
 Delta R.T. 0.00 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Instrument :
 BNA_G
 ClientSampleId :
 SST02086

Tot Ion	Ratio	Resp	Lower	Upper
196	100	47825		
198	99.6	74.7	112.1	
200	30.2	25.4	38.0	

Manual Integrations
 APPROVED

Sohil
 1/29/2018 5:50:04 PM






Abundance

Ion 196.00 (195.70 to 196.70): E


Ion 198.00 (197.70 to 198.70): E

Ion 200.00 (199.70 to 200.70): E



#40
 1,1'-Biphenyl
 Concen: 18.732 ng/ul
 RT: 14.20 min Scan# 1856
 Delta R.T. 0.00 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Resp	Lower	Upper
154	100	132305		
153	43.9	34.0	51.0	
76	6.4	8.4	12.6	






Abundance

Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BGC









#41
 2-Chloronaphthalene
 Concen: 19.138 ng/ul
 RT: 14.25 min Scan# 1865
 Delta R.T. 0.00 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02086

Tot Ion	Ion	Ratio	Resp	Lower	Upper
162	100				
164	33.8	26.2	39.4		
127	35.4	29.5	44.3		




Manual Integrations
 APPROVED


Sohil
 1/29/2018 5:50:04 PM



#42
 2-Nitroaniline
 Concen: 20.826 ng/ul
 RT: 14.44 min Scan# 1897
 Delta R.T. 0.00 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
65	100				
92	73.1	57.0	85.4		
138	126.3	87.2	130.8		





#44

Dimethylphthalate

Concen: 20.508 ng/ul

RT: 14.78 min Scan# 1956

Delta R.T. 0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Instrument :

BNA_G





ClientSampleId :

SSTD02086

Tot Ion:	163	Resp:	147076
Ion Ratio	Lower	Upper	
163	100		
194	5.3	4.0	6.0
164	9.4	8.4	12.6

Manual Integrations
APPROVED

Sohil
1/29/2018 5:50:04 PM



#45

2,6-Dinitrotoluene

Concen: 21.132 ng/ul




RT: 14.91 min Scan# 1978


Delta R.T. 0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Tgt Ion:	165	Resp:	30941
Ion Ratio	Lower	Upper	
165	100		
89	48.6	43.8	65.8
121	22.4	17.8	26.6





#47

Acenaphthylene

Concen: 19.452 ng/ul

RT: 15.08 min Scan# 2007

Delta R.T. 0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Instrument :

BNA_G





ClientSampleId :

SSTD02086

Tot Ion:	152	Resp:	155377
Ion Ratio	Lower	Upper	
152	100		
151	22.3	16.1	24.1
153	13.9	10.8	16.2

Manual Integrations
APPROVED

Sohil
1/29/2018 5:50:04 PM



#48

3-Nitroaniline

Concen: 24.263 ng/ul




RT: 15.24 min Scan# 2034


Delta R.T. -0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Tgt Ion:	138	Resp:	25933
Ion Ratio	Lower	Upper	
138	100		
108	11.7	7.6	11.4#
92	110.7	94.2	141.2





#49

Acenaphthene

Concen: 19.357 ng/ul

RT: 15.42 min Scan# 2064

Delta R.T. 0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Instrument :

BNA_G





ClientSampleId :

SSTD02086

Tot Ion:	153	Resp:	112851
Ion Ratio	Lower	Upper	
153	100		
152	47.5	37.9	56.9
154	83.6	69.8	104.6

Manual Integrations
APPROVED

Sohil
1/29/2018 5:50:04 PM



#50

2,4-Dinitrophenol

Concen: 24.746 ng/ul




RT: 15.44 min Scan# 2068


Delta R.T. -0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Tgt Ion:	184	Resp:	18728
Ion Ratio	Lower	Upper	
184	100		
63	48.2	39.7	59.5
154	117.3	45.0	67.6#






#52
4-Nitrophenol
Concen: 26.505 ng/ul
RT: 15.52 min Scan# 2082
Delta R.T. 0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Instrument :
BNA_G
ClientSampleId :
SSTD02086




Tot Ion	Ratio	Lower	Upper
109	100		
139	103.1	91.2	136.8
65	97.7	92.6	139.0

Manual Integrations
APPROVED

Sohil
1/29/2018 5:50:04 PM




Abundance Ion 109.00 (108.70 to 109.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 65.00 (64.70 to 65.70): BGC






#53
Dibenzofuran
Concen: 20.137 ng/ul
RT: 15.75 min Scan# 2120
Delta R.T. 0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Lower	Upper
168	100		
139	37.7	31.0	46.4



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E









#54
 2,4-Dinitrotoluene
 Concen: 23.110 ng/ul
 RT: 15.69 min Scan# 2110
 Delta R.T. -0.00 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Instrument :
 BNA_G
 ClientSampleId :
 SST02086

Tot Ion	Ratio	Resp	Lower	Upper
165	100	48928		
63	32.6		31.4	47.0
182	3.2		2.7	4.1




Manual Integrations
 APPROVED


Sohil
 1/29/2018 5:50:04 PM



#55
 2,3,4,6-Tetrachlorophenol
 Concen: 22.513 ng/ul
 RT: 15.96 min Scan# 2156
 Delta R.T. -0.00 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Resp	Lower	Upper
232	100	51259		
131	31.1		27.1	40.7
130	1.8		0.0	0.0#
166	22.4		20.0	30.0





#56

Diethylphthalate

Concen: 21.512 ng/ul

RT: 16.12 min Scan# 2184

Delta R.T. -0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Instrument :

BNA_G





ClientSampleId :

SSTD02086

Tot Ion	Ratio	Resp	Lower	Upper
149	100	146015		
177	26.1	19.3	28.9	
150	12.0	10.6	15.8	

Manual Integrations
APPROVED

Sohil
1/29/2018 5:50:04 PM



#58

Fluorene

Concen: 20.464 ng/ul




RT: 16.39 min Scan# 2230


Delta R.T. 0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Resp	Lower	Upper
166	100	142460		
165	101.8	78.4	117.6	
167	14.6	11.1	16.7	









#59
 4-Chlorophenyl-phenylether
 Concen: 20.121 ng/ul
 RT: 16.37 min Scan# 2226
 Delta R.T. 0.00 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Instrument :
 BNA_G
 ClientSampled :
 SST02086

Tot Ion	Ratio	Lower	Upper
204	100		
206	33.5	26.6	40.0
141	55.8	40.7	61.1




Manual Integrations
 APPROVED


Sohil
 1/29/2018 5:50:04 PM



#60
 4-Nitroaniline
 Concen: 24.408 ng/ul
 RT: 16.40 min Scan# 2231
 Delta R.T. 0.01 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Lower	Upper
138	100		
92	45.3	38.2	57.4
108	94.5	70.0	105.0









#63
 4,6-Dinitro-2-methylphenol
 Concen: 21.963 ng/ul
 RT: 16.45 min Scan# 2239
 Delta R.T. -0.00 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Instrument :
 BNA_G
 ClientSampleId :
 SST02086

Tot Ion	Ratio	Resp	Lower	Upper
198	100	34897		
182	3.2		3.5	5.3#
77	16.1		18.2	27.4#




Manual Integrations
 APPROVED


Sohil
 1/29/2018 5:50:04 PM



#64
 N-Nitrosodiphenylamine
 Concen: 18.231 ng/ul
 RT: 16.58 min Scan# 2262
 Delta R.T. 0.00 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Resp	Lower	Upper
169	100	128082		
168	69.2		53.8	80.8
167	37.4		29.5	44.3





#65

4-Bromophenyl-phenylether

Concen: 19.015 ng/ul

RT: 17.26 min Scan# 2378

Delta R.T. 0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Instrument :

BNA_G





ClientSampleId :

SSTD02086

Tot Ion:	248	Resp:	67656
Ion Ratio	Lower	Upper	
248	100		
250	101.9	79.0	118.4
141	54.3	51.9	77.9

Manual Integrations
APPROVED

Sohil
1/29/2018 5:50:04 PM



#66

Hexachlorobenzene

Concen: 19.565 ng/ul




RT: 17.39 min Scan# 2400


Delta R.T. 0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Tgt Ion:	284	Resp:	72296
Ion Ratio	Lower	Upper	
284	100		
142	22.8	21.3	31.9
249	33.7	26.3	39.5









#67
Atrazine
Concen: 21.280 ng/ul
RT: 17.50 min Scan# 2419
Delta R.T. 0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Instrument :
BNA_G
ClientSampleId :
SSTD02086

Tot Ion	Ratio	Lower	Upper
200	100		
173	23.6	20.6	31.0
215	56.6	39.8	59.6




Manual Integrations
APPROVED


Sohil
1/29/2018 5:50:04 PM



#68
Pentachlorophenol
Concen: 20.994 ng/ul
RT: 17.73 min Scan# 2457
Delta R.T. 0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Lower	Upper
266	100		
264	68.2	48.2	72.4
268	66.1	52.3	78.5





#69

Phenanthrene

Concen: 19.657 ng/ul

RT: 18.13 min Scan# 2526

Delta R.T. 0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Instrument :

BNA_G





ClientSampleId :

SSTD02086

Tot Ion:	178	Resp:	255599
Ion Ratio	Lower	Upper	
178	100		
179	15.3	12.3	18.5
176	18.9	15.4	23.0

Manual Integrations
APPROVED

Sohil
1/29/2018 5:50:04 PM



#71

Anthracene

Concen: 19.680 ng/ul




RT: 18.22 min Scan# 2541


Delta R.T. -0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Tgt Ion:	178	Resp:	262578
Ion Ratio	Lower	Upper	
178	100		
179	16.0	13.4	20.2
176	18.9	15.4	23.0





#72

1,2,3,4-Tetrachlorobenzene

Concen: 16.439 ng/uL

RT: 14.17 min Scan# 1851

Delta R.T. 0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Instrument :

BNA_G


ClientSampleId :

SSTD02086




Tot Ion:	216	Resp:	77358
Ion Ratio	Lower	Upper	
216	100		
214	74.4	64.1	96.1
218	49.7	39.4	59.0

Manual Integrations
APPROVED

Sohil
1/29/2018 5:50:04 PM



Abundance Ion 216.00 (215.70 to 216.70): E
Ion 214.00 (213.70 to 214.70): E
Ion 218.00 (217.70 to 218.70): E



#73

Pentachlorobenzene

Concen: 17.396 ng/uL


RT: 15.66 min Scan# 2106

Delta R.T. -0.00 min



Lab File: BG032405.D


Acq: 29 Jan 2018 1:19

Tgt Ion:	250	Resp:	86055
Ion Ratio	Lower	Upper	
250	100		
248	64.5	51.4	77.2
252	64.8	49.8	74.6



Abundance Ion 250.00 (249.70 to 250.70): E
Ion 248.00 (247.70 to 248.70): E
Ion 252.00 (251.70 to 252.70): E









#74
 Carbazole
 Concen: 21.148 ng/ul
 RT: 18.48 min Scan# 2585
 Delta R.T. 0.00 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02086

Tot Ion	Ratio	Resp	Lower	Upper
167	100	218664		
166	22.9	17.0	25.4	
139	11.1	9.1	13.7	




Manual Integrations
 APPROVED


Sohil
 1/29/2018 5:50:04 PM



#75
 Di-n-butylphthalate
 Concen: 20.662 ng/ul
 RT: 18.99 min Scan# 2672
 Delta R.T. 0.00 min
 Lab File: BG032405.D
 Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Resp	Lower	Upper
149	100	251923		
150	9.3	7.5	11.3	
104	5.1	4.1	6.1	





#76

Fluoranthene

Concen: 22.237 ng/ul

RT: 20.09 min Scan# 2860

Delta R.T. 0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Instrument :

BNA_G





ClientSampleId :

SSTD02086

Tot Ion	Ion	Ratio	Resp Lower	Resp Upper
202	100			
101	4.7	6.1	9.1#	
100	3.2	4.6	7.0#	

Manual Integrations
APPROVED

Sohil
1/29/2018 5:50:04 PM



#79

Pyrene

Concen: 19.847 ng/ul




RT: 20.45 min Scan# 2921


Delta R.T. 0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Tgt Ion	Ion	Ratio	Resp Lower	Resp Upper
202	100			
101	4.1	6.9	10.3#	
100	4.3	5.6	8.4#	









#80
Butylbenzylphthalate
Concen: 19.998 ng/ul
RT: 21.31 min Scan# 3066
Delta R.T. -0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Instrument :
BNA_G
ClientSampleId :
SSTD02086

Tot Ion	Ratio	Resp	Lower	Upper
149	100	119135		
91	53.5	53.4	80.0	
206	29.5	19.3	28.9	




Manual Integrations
APPROVED


Sohil
1/29/2018 5:50:04 PM



#81
3,3'-Dichlorobenzidine
Concen: 25.036 ng/ul
RT: 22.29 min Scan# 3233
Delta R.T. 0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Resp	Lower	Upper
252	100	129673		
254	65.0	52.8	79.2	
126	5.3	6.9	10.3	









#82
Benzo(a)anthracene
Concen: 19.712 ng/ul
RT: 22.39 min Scan# 3251
Delta R.T. -0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Instrument :
BNA_G
ClientSampleId :
SSTD02086

Tot Ion	Ratio	Lower	Upper
228	100		
229	20.8	16.0	24.0
226	26.9	21.8	32.8




Manual Integrations
APPROVED


Sohil
1/29/2018 5:50:04 PM



#83
Bis(2-ethylhexyl)phthalate
Concen: 19.674 ng/ul
RT: 22.23 min Scan# 3224
Delta R.T. 0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Lower	Upper
149	100		
167	30.9	23.0	34.6
279	7.5	4.3	6.5





#84

Chrysene

Concen: 19.679 ng/ul

RT: 22.47 min Scan# 3264

Delta R.T. 0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Instrument :

BNA_G





ClientSampleId :

SSTD02086

Tot Ion	Ratio	Lower	Upper
228	100		
226	29.7	24.1	36.1
229	19.8	16.0	24.0

Manual Integrations
APPROVED

Sohil
1/29/2018 5:50:04 PM



#86

Di-n-octyl phthalate

Concen: 19.468 ng/ul




RT: 23.60 min Scan# 3457


Delta R.T. -0.00 min

Lab File: BG032405.D

Acq: 29 Jan 2018 1:19

Tgt Ion:149 Resp: 284790









#87
Benzo(b)fluoranthene
Concen: 19.456 ng/ul
RT: 24.91 min Scan# 3679
Delta R.T. -0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Instrument :
BNA_G
ClientSampleId :
SSTD02086

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.4	26.0
125	3.5	4.9	7.3




Manual Integrations
APPROVED


Sohil
1/29/2018 5:50:04 PM



#88
Benzo(k)fluoranthene
Concen: 19.199 ng/ul
RT: 24.99 min Scan# 3693
Delta R.T. 0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.0	17.1	25.7
125	3.8	5.4	8.0









#90
Benzo(a)pyrene
Concen: 19.340 ng/u1
RT: 25.92 min Scan# 3851
Delta R.T. 0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Instrument :
BNA_G
ClientSampleId :
SSTD02086

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.6	17.2	25.8
125	4.4	5.8	8.6#




Manual Integrations
APPROVED


Sohil
1/29/2018 5:50:04 PM



#91
Indeno(1,2,3-cd)pyrene
Concen: 20.916 ng/u1
RT: 30.35 min Scan# 4605
Delta R.T. 0.01 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Lower	Upper
276	100		
138	7.8	10.9	16.3#
277	24.7	18.9	28.3









#92
Dibenzo(a,h)anthracene
Concen: 20.984 ng/ul
RT: 30.42 min Scan# 4617
Delta R.T. -0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Instrument :
BNA_G
ClientSampleId :
SSTD02086

Tot Ion	Ratio	Lower	Upper
278	100		
139	7.4	9.0	13.4#
279	22.9	19.1	28.7

Manual Integrations
APPROVED

Sohil
1/29/2018 5:50:04 PM



#93
Benzo(g,h,i)perylene
Concen: 21.957 ng/ul m
RT: 31.68 min Scan# 4832
Delta R.T. -0.00 min
Lab File: BG032405.D
Acq: 29 Jan 2018 1:19

Tgt Ion	Ratio	Lower	Upper
276	100		
138	10.2	11.2	16.8#
277	25.0	19.0	28.6

