

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_G\DATA\BG012918\
 Data File : BG032419.D
 Acq On : 29 Jan 2018 18:03
 Operator : SJ/JU
 Sample : J1196-22
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 1035

Quant Time: Jan 30 04:23:59 2018
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 29 15:32:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.72	152	24233	20.00	ng	0.00
21) Naphthalene-d8	11.60	136	103565	20.00	ng	0.00
38) Acenaphthene-d10	15.36	164	80829	20.00	ng	0.00
63) Phenanthrene-d10	18.09	188	220932	20.00	ng	0.00
75) Chrysene-d12	22.41	240	272557	20.00	ng	0.00
86) Perylene-d12	26.10	264	286401	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	6.16	112	196940	163.23	ng	0.00
7) Phenol-d6	7.84	99	254851	158.30	ng	0.00
23) Nitrobenzene-d5	9.92	82	160957	97.07	ng	0.00
41) 2,4,6-Tribromophenol	16.84	330	238889	155.27	ng	0.00
44) 2-Fluorobiphenyl	13.99	172	606126	89.13	ng	0.00
78) Terphenyl-d14	20.63	244	1385885	108.92	ng	0.00
Target Compounds						
8) 2-Chlorophenol	8.27	128	93343	65.451	ng	Qvalue # 80
17) 2-Methylphenol	9.16	107	96361	77.602	ng	95
20) 3+4-Methylphenols	9.49	107	254644	146.197	ng	96
26) 2-Nitrophenol	10.69	139	102244	105.492	ng	# 79
27) 2,4-Dimethylphenol	10.73	122	130584	93.869	ng	94
29) 2,4-Dichlorophenol	11.23	162	204061	110.731	ng	89
36) 4-Chloro-3-methylphenol	12.82	107	308495	176.481	ng	88
42) 2,4,6-Trichlorophenol	13.81	196	343342	166.925	ng	99
43) 2,4,5-Trichlorophenol	13.87	196	106915	44.590	ng	# 89
53) 2,4-Dinitrophenol	15.45	184	183287	182.186	ng	# 55
55) 4-Nitrophenol	15.54	139	175218	171.296	ng	# 74
64) 4,6-Dinitro-2-methylphenol	16.45	198	162638	98.538	ng	# 68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA_G\DATA\BG012918\
Data File : BG032419.D
Acq On : 29 Jan 2018 18:03
Operator : SJ/JU
Sample : J1196-22
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
1035

Quant Time: Jan 30 04:23:59 2018
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 29 15:32:55 2018
Response via : Initial Calibration

