

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_G\DATA\BG012918\
 Data File : BG032432.D
 Acq On : 30 Jan 2018 4:41
 Operator : SJ/JU
 Sample : J1303-02
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 BUS-WASH-ZIPPER-DRAINS

Quant Time: Jan 30 06:22:48 2018
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 29 15:32:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.75	152	25500	20.00	ng	0.03
21) Naphthalene-d8	11.62	136	107178	20.00	ng	0.02
38) Acenaphthene-d10	15.37	164	86135	20.00	ng	0.02
63) Phenanthrene-d10	18.10	188	248087	20.00	ng	0.02
75) Chrysene-d12	22.43	240	296742	20.00	ng	0.01
86) Perylene-d12	26.12	264	316153	20.00	ng	0.04
System Monitoring Compounds						
5) 2-Fluorophenol	6.18	112	153498	120.91	ng	0.03
7) Phenol-d6	7.86	99	179437	105.92	ng	0.04
23) Nitrobenzene-d5	9.94	82	154639	90.12	ng	0.02
41) 2,4,6-Tribromophenol	16.85	330	189683	115.69	ng	0.01
44) 2-Fluorobiphenyl	14.00	172	607016	83.76	ng	0.02
78) Terphenyl-d14	20.64	244	1328571	95.91	ng	0.01
Target Compounds						
10) Phenol	7.89	94	3644	2.265	ng	Qvalue # 54
31) Naphthalene	11.68	128	21141	4.039	ng	97
32) Benzoic acid	10.82	122	2222	9.038	ng	# 55
49) Dimethylphthalate	14.80	163	24701	3.110	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_G\DATA\BG012918\
Data File : BG032432.D
Acq On : 30 Jan 2018 4:41
Operator : SJ/JU
Sample : J1303-02
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
BUS-WASH-ZIPPER-DRAINS

Quant Time: Jan 30 06:22:48 2018
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 29 15:32:55 2018
Response via : Initial Calibration

