

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_G\Data\BG012918\
 Data File : BG032414.D
 Acq On : 29 Jan 2018 14:29
 Operator : SJ/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG012918

Quant Time: Jan 30 03:45:28 2018
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 29 15:27:12 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	111	0.01
2	1,4-Dioxane	40.000	38.876	2.8	119	0.00
3	Pyridine	40.000	41.716	-4.3	114	0.00
4	n-Nitrosodimethylamine	40.000	39.363	1.6	103	0.01
5 S	2-Fluorophenol	80.000	78.107	2.4	106	0.01
6	Aniline	40.000	39.262	1.8	105	0.01
7 S	Phenol-d6	80.000	79.131	1.1	109	0.02
8	2-Chlorophenol	40.000	40.914	-2.3	111	0.01
9	Benzaldehyde	40.000	38.009	5.0	106	0.01
10 C	Phenol	40.000	39.950	0.1	108	0.01
11	bis(2-Chloroethyl)ether	40.000	39.051	2.4	108	0.01
12	1,3-Dichlorobenzene	40.000	38.468	3.8	108	0.02
13 C	1,4-Dichlorobenzene	40.000	39.612	1.0	111	0.01
14	1,2-Dichlorobenzene	40.000	39.020	2.4	108	0.01
15	Benzyl Alcohol	40.000	40.101	-0.3	107	0.01
16	2,2'-oxybis(1-Chloropropane)	40.000	37.022	7.4	102	0.02
17	2-Methylphenol	40.000	38.968	2.6	106	0.01
18	Hexachloroethane	40.000	37.011	7.5	105	0.01
19 P	n-Nitroso-di-n-propylamine	40.000	38.732	3.2	104	0.01
20	3+4-Methylphenols	40.000	38.322	4.2	100	0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	107	0.01
22	Acetophenone	40.000	38.535	3.7	104	0.01
23 S	Nitrobenzene-d5	80.000	78.447	1.9	103	0.01
24	Nitrobenzene	40.000	39.127	2.2	105	0.01
25	Isophorone	40.000	38.129	4.7	102	0.01
26 C	2-Nitrophenol	40.000	39.621	0.9	101	0.01
27	2,4-Dimethylphenol	40.000	38.707	3.2	103	0.01
28	bis(2-Chloroethoxy)methane	40.000	40.011	-0.0	107	0.01
29 C	2,4-Dichlorophenol	40.000	39.252	1.9	105	0.00
30	1,2,4-Trichlorobenzene	40.000	38.651	3.4	106	0.00
31	Naphthalene	40.000	38.843	2.9	105	0.01
32	Benzoic acid	40.000	36.707	8.2	99	0.00
33	4-Chloroaniline	40.000	38.655	3.4	105	0.01
34 C	Hexachlorobutadiene	40.000	38.359	4.1	104	0.00
35	Caprolactam	40.000	38.266	4.3	100	0.01
36 C	4-Chloro-3-methylphenol	40.000	38.719	3.2	106	0.01
37	2-Methylnaphthalene	40.000	38.305	4.2	103	0.01
38 I	Acenaphthene-d10	20.000	20.000	0.0	99	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	40.850	-2.1	104	0.00
40 P	Hexachlorocyclopentadiene	40.000	41.384	-3.5	104	0.00
41 S	2,4,6-Tribromophenol	80.000	82.140	-2.7	104	0.00
42 C	2,4,6-Trichlorophenol	40.000	41.051	-2.6	105	0.00
43	2,4,5-Trichlorophenol	40.000	41.283	-3.2	104	0.00
44 S	2-Fluorobiphenyl	80.000	79.498	0.6	102	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_G\Data\BG012918\
 Data File : BG032414.D
 Acq On : 29 Jan 2018 14:29
 Operator : SJ/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG012918

Quant Time: Jan 30 03:45:28 2018
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 29 15:27:12 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	40.514	-1.3	103	0.01
46	2-Chloronaphthalene	40.000	40.418	-1.0	103	0.00
47	2-Nitroaniline	40.000	40.409	-1.0	102	0.00
48	Acenaphthylene	40.000	40.408	-1.0	103	0.00
49	Dimethylphthalate	40.000	39.303	1.7	103	0.00
50	2,6-Dinitrotoluene	40.000	40.400	-1.0	105	0.00
51 C	Acenaphthene	40.000	40.596	-1.5	103	0.00
52	3-Nitroaniline	40.000	40.329	-0.8	102	0.00
53 P	2,4-Dinitrophenol	40.000	38.931	2.7	103	0.00
54	Dibenzofuran	40.000	39.600	1.0	103	0.00
55 P	4-Nitrophenol	40.000	41.070	-2.7	105	0.00
56	2,4-Dinitrotoluene	40.000	40.266	-0.7	102	0.00
57	Fluorene	40.000	39.710	0.7	104	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	40.176	-0.4	101	0.00
59	Diethylphthalate	40.000	39.474	1.3	104	0.00
60	4-Chlorophenyl-phenylether	40.000	39.971	0.1	103	0.00
61	4-Nitroaniline	40.000	39.580	1.1	104	0.00
62	Azobenzene	40.000	39.481	1.3	104	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
64	4,6-Dinitro-2-methylphenol	40.000	40.772	-1.9	103	0.00
65 c	n-Nitrosodiphenylamine	40.000	38.975	2.6	102	0.00
66	4-Bromophenyl-phenylether	40.000	39.433	1.4	102	0.00
67	Hexachlorobenzene	40.000	39.255	1.9	103	0.00
68	Atrazine	40.000	40.482	-1.2	103	0.00
69 C	Pentachlorophenol	40.000	41.608	-4.0	109	0.00
70	Phenanthrene	40.000	38.567	3.6	102	0.00
71	Anthracene	40.000	39.265	1.8	102	0.00
72	Carbazole	40.000	39.005	2.5	102	0.00
73	Di-n-butylphthalate	40.000	39.784	0.5	103	0.00
74 C	Fluoranthene	40.000	38.185	4.5	102	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	102	0.00
76	Benzidine	40.000	42.558	-6.4	98	0.00
77	Pyrene	40.000	39.362	1.6	102	0.00
78 S	Terphenyl-d14	80.000	78.553	1.8	103	0.00
79	Butylbenzylphthalate	40.000	40.553	-1.4	103	0.00
80	Benzo(a)anthracene	40.000	39.036	2.4	101	0.00
81	3,3'-Dichlorobenzidine	40.000	38.426	3.9	98	0.00
82	Chrysene	40.000	38.719	3.2	102	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	40.084	-0.2	101	0.00
84 c	Di-n-octyl phthalate	40.000	40.353	-0.9	103	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	39.142	2.1	101	0.02
86 I	Perylene-d12	20.000	20.000	0.0	101	0.01
87	Benzo(b)fluoranthene	40.000	39.086	2.3	101	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_G\Data\BG012918\
Data File : BG032414.D
Acq On : 29 Jan 2018 14:29
Operator : SJ/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG012918

Quant Time: Jan 30 03:45:28 2018
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 29 15:27:12 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	39.501	1.2	103	0.00
89 C	Benzo(a)pyrene	40.000	39.205	2.0	102	0.01
90	Dibenzo(a,h)anthracene	40.000	39.334	1.7	101	0.00
91	Benzo(g,h,i)perylene	40.000	39.270	1.8	101	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0