

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_G\Data\BG012918\
 Data File : BG032423.D
 Acq On : 29 Jan 2018 20:34
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 30 03:28:27 2018
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 29 15:32:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	39.393	1.5	108	0.00
3	Pyridine	40.000	42.902	-7.3	105	0.00
4	n-Nitrosodimethylamine	40.000	41.229	-3.1	97	0.00
5 S	2-Fluorophenol	80.000	80.443	-0.6	98	0.00
6	Aniline	40.000	39.889	0.3	95	0.00
7 S	Phenol-d6	80.000	81.247	-1.6	100	0.00
8	2-Chlorophenol	40.000	40.217	-0.5	98	0.00
9	Benzaldehyde	40.000	39.307	1.7	98	0.00
10 C	Phenol	40.000	40.752	-1.9	99	0.00
11	bis(2-Chloroethyl)ether	40.000	40.730	-1.8	101	0.00
12	1,3-Dichlorobenzene	40.000	39.660	0.9	100	0.00
13 C	1,4-Dichlorobenzene	40.000	39.591	1.0	100	0.00
14	1,2-Dichlorobenzene	40.000	38.896	2.8	96	0.00
15	Benzyl Alcohol	40.000	40.700	-1.8	97	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	37.969	5.1	94	0.00
17	2-Methylphenol	40.000	40.073	-0.2	97	0.00
18	Hexachloroethane	40.000	37.425	6.4	96	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.554	1.1	95	0.00
20	3+4-Methylphenols	40.000	40.471	-1.2	95	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	94	0.00
22	Acetophenone	40.000	39.789	0.5	95	0.00
23 S	Nitrobenzene-d5	80.000	80.214	-0.3	93	0.00
24	Nitrobenzene	40.000	41.046	-2.6	97	0.00
25	Isophorone	40.000	40.840	-2.1	97	0.00
26 C	2-Nitrophenol	40.000	41.626	-4.1	93	0.00
27	2,4-Dimethylphenol	40.000	40.344	-0.9	95	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.383	-3.5	98	0.00
29 C	2,4-Dichlorophenol	40.000	42.098	-5.2	99	0.00
30	1,2,4-Trichlorobenzene	40.000	40.367	-0.9	98	0.00
31	Naphthalene	40.000	41.001	-2.5	98	0.00
32	Benzoic acid	40.000	36.215	9.5	86	0.00
33	4-Chloroaniline	40.000	39.643	0.9	95	0.00
34 C	Hexachlorobutadiene	40.000	39.669	0.8	95	0.00
35	Caprolactam	40.000	42.521	-6.3	98	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.442	-1.1	97	0.00
37	2-Methylnaphthalene	40.000	39.868	0.3	95	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	95	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	39.169	2.1	96	0.00
40 P	Hexachlorocyclopentadiene	40.000	39.029	2.4	94	0.00
41 S	2,4,6-Tribromophenol	80.000	81.289	-1.6	99	0.00
42 C	2,4,6-Trichlorophenol	40.000	39.944	0.1	99	0.00
43	2,4,5-Trichlorophenol	40.000	39.957	0.1	97	0.00
44 S	2-Fluorobiphenyl	80.000	77.422	3.2	95	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_G\Data\BG012918\
 Data File : BG032423.D
 Acq On : 29 Jan 2018 20:34
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 30 03:28:27 2018
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 29 15:32:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	39.099	2.3	96	0.00
46	2-Chloronaphthalene	40.000	39.193	2.0	96	0.00
47	2-Nitroaniline	40.000	39.713	0.7	96	0.00
48	Acenaphthylene	40.000	39.013	2.5	96	0.00
49	Dimethylphthalate	40.000	39.338	1.7	99	0.00
50	2,6-Dinitrotoluene	40.000	40.153	-0.4	100	0.00
51 C	Acenaphthene	40.000	39.449	1.4	96	0.00
52	3-Nitroaniline	40.000	40.417	-1.0	98	0.00
53 P	2,4-Dinitrophenol	40.000	37.846	5.4	96	0.00
54	Dibenzofuran	40.000	38.923	2.7	97	0.00
55 P	4-Nitrophenol	40.000	41.320	-3.3	101	0.00
56	2,4-Dinitrotoluene	40.000	40.277	-0.7	98	0.00
57	Fluorene	40.000	38.498	3.8	97	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	40.272	-0.7	97	0.00
59	Diethylphthalate	40.000	39.129	2.2	99	0.00
60	4-Chlorophenyl-phenylether	40.000	39.522	1.2	98	0.00
61	4-Nitroaniline	40.000	40.682	-1.7	103	0.00
62	Azobenzene	40.000	39.065	2.3	99	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	99	0.00
64	4,6-Dinitro-2-methylphenol	40.000	39.250	1.9	96	0.00
65 c	n-Nitrosodiphenylamine	40.000	39.067	2.3	99	0.00
66	4-Bromophenyl-phenylether	40.000	38.854	2.9	98	0.00
67	Hexachlorobenzene	40.000	38.703	3.2	99	0.00
68	Atrazine	40.000	39.301	1.7	97	0.00
69 C	Pentachlorophenol	40.000	38.783	3.0	99	0.00
70	Phenanthrene	40.000	38.478	3.8	98	0.00
71	Anthracene	40.000	39.162	2.1	99	0.00
72	Carbazole	40.000	38.647	3.4	98	0.00
73	Di-n-butylphthalate	40.000	39.537	1.2	99	0.00
74 C	Fluoranthene	40.000	38.399	4.0	100	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
76	Benzidine	40.000	44.804	-12.0	100	0.00
77	Pyrene	40.000	39.230	1.9	99	0.00
78 S	Terphenyl-d14	80.000	77.962	2.5	100	0.00
79	Butylbenzylphthalate	40.000	39.846	0.4	98	0.00
80	Benzo(a)anthracene	40.000	39.045	2.4	99	0.00
81	3,3'-Dichlorobenzidine	40.000	39.242	1.9	98	0.00
82	Chrysene	40.000	38.700	3.2	99	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	39.963	0.1	98	0.00
84 c	Di-n-octyl phthalate	40.000	40.119	-0.3	99	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	39.702	0.7	99	0.00
86 I	Perylene-d12	20.000	20.000	0.0	98	0.01
87	Benzo(b)fluoranthene	40.000	39.790	0.5	99	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_G\Data\BG012918\
Data File : BG032423.D
Acq On : 29 Jan 2018 20:34
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jan 30 03:28:27 2018
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 29 15:32:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	39.386	1.5	99	0.00
89 C	Benzo(a)pyrene	40.000	39.605	1.0	99	0.00
90	Dibenzo(a,h)anthracene	40.000	39.740	0.6	98	0.00
91	Benzo(g,h,i)perylene	40.000	39.720	0.7	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0