

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_G\Data\BG013018\
 Data File : BG032442.D
 Acq On : 30 Jan 2018 15:07
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 31 00:41:49 2018
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 29 15:32:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	96	0.00
2	1,4-Dioxane	40.000	41.001	-2.5	108	0.00
3	Pyridine	40.000	41.649	-4.1	98	0.00
4	n-Nitrosodimethylamine	40.000	40.384	-1.0	91	0.00
5 S	2-Fluorophenol	80.000	81.780	-2.2	96	0.00
6	Aniline	40.000	39.967	0.1	92	0.00
7 S	Phenol-d6	80.000	79.737	0.3	95	0.01
8	2-Chlorophenol	40.000	40.941	-2.4	96	0.00
9	Benzaldehyde	40.000	38.621	3.4	93	0.00
10 C	Phenol	40.000	40.449	-1.1	94	0.00
11	bis(2-Chloroethyl)ether	40.000	39.835	0.4	95	0.00
12	1,3-Dichlorobenzene	40.000	40.973	-2.4	100	0.00
13 C	1,4-Dichlorobenzene	40.000	40.947	-2.4	99	0.00
14	1,2-Dichlorobenzene	40.000	39.812	0.5	95	0.00
15	Benzyl Alcohol	40.000	40.631	-1.6	94	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.690	3.3	92	0.00
17	2-Methylphenol	40.000	38.519	3.7	90	0.00
18	Hexachloroethane	40.000	38.212	4.5	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.335	-0.8	93	0.00
20	3+4-Methylphenols	40.000	39.105	2.2	89	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	90	0.00
22	Acetophenone	40.000	38.038	4.9	87	0.00
23 S	Nitrobenzene-d5	80.000	78.436	2.0	87	0.00
24	Nitrobenzene	40.000	40.365	-0.9	92	0.00
25	Isophorone	40.000	40.999	-2.5	93	0.00
26 C	2-Nitrophenol	40.000	40.275	-0.7	87	0.00
27	2,4-Dimethylphenol	40.000	39.778	0.6	90	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.792	0.5	90	0.00
29 C	2,4-Dichlorophenol	40.000	40.397	-1.0	92	0.00
30	1,2,4-Trichlorobenzene	40.000	38.597	3.5	90	0.00
31	Naphthalene	40.000	39.653	0.9	91	0.00
32	Benzoic acid	40.000	42.713	-6.8	102	0.01
33	4-Chloroaniline	40.000	38.757	3.1	89	0.00
34 C	Hexachlorobutadiene	40.000	38.945	2.6	90	0.00
35	Caprolactam	40.000	46.265	-15.7	103	0.01
36 C	4-Chloro-3-methylphenol	40.000	40.252	-0.6	93	0.00
37	2-Methylnaphthalene	40.000	38.621	3.4	88	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	92	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	36.737	8.2	87	0.00
40 P	Hexachlorocyclopentadiene	40.000	33.758	15.6	79	0.00
41 S	2,4,6-Tribromophenol	80.000	83.228	-4.0	98	0.00
42 C	2,4,6-Trichlorophenol	40.000	39.715	0.7	95	0.00
43	2,4,5-Trichlorophenol	40.000	39.431	1.4	93	0.00
44 S	2-Fluorobiphenyl	80.000	73.530	8.1	88	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_G\Data\BG013018\
 Data File : BG032442.D
 Acq On : 30 Jan 2018 15:07
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 31 00:41:49 2018
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 29 15:32:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	37.452	6.4	89	0.00
46	2-Chloronaphthalene	40.000	38.001	5.0	90	0.00
47	2-Nitroaniline	40.000	41.128	-2.8	97	0.00
48	Acenaphthylene	40.000	38.595	3.5	92	0.00
49	Dimethylphthalate	40.000	38.968	2.6	95	0.00
50	2,6-Dinitrotoluene	40.000	39.702	0.7	96	0.00
51 C	Acenaphthene	40.000	38.944	2.6	92	0.00
52	3-Nitroaniline	40.000	40.466	-1.2	95	0.00
53 P	2,4-Dinitrophenol	40.000	34.134	14.7	83	0.00
54	Dibenzofuran	40.000	38.208	4.5	93	0.00
55 P	4-Nitrophenol	40.000	39.195	2.0	93	0.02
56	2,4-Dinitrotoluene	40.000	41.023	-2.6	97	0.00
57	Fluorene	40.000	39.003	2.5	95	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	39.591	1.0	92	0.00
59	Diethylphthalate	40.000	40.218	-0.5	99	0.00
60	4-Chlorophenyl-phenylether	40.000	38.872	2.8	93	0.00
61	4-Nitroaniline	40.000	40.358	-0.9	99	0.00
62	Azobenzene	40.000	39.135	2.2	96	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	97	0.00
64	4,6-Dinitro-2-methylphenol	40.000	38.212	4.5	91	0.00
65 c	n-Nitrosodiphenylamine	40.000	39.208	2.0	97	0.00
66	4-Bromophenyl-phenylether	40.000	39.182	2.0	96	0.00
67	Hexachlorobenzene	40.000	38.869	2.8	97	0.00
68	Atrazine	40.000	41.117	-2.8	99	0.00
69 C	Pentachlorophenol	40.000	41.301	-3.3	102	0.00
70	Phenanthrene	40.000	38.689	3.3	96	0.00
71	Anthracene	40.000	39.531	1.2	97	0.00
72	Carbazole	40.000	39.103	2.2	97	0.00
73	Di-n-butylphthalate	40.000	41.000	-2.5	100	0.00
74 C	Fluoranthene	40.000	38.104	4.7	97	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
76	Benzidine	40.000	41.140	-2.9	97	0.00
77	Pyrene	40.000	36.463	8.8	97	0.00
78 S	Terphenyl-d14	80.000	73.235	8.5	98	0.00
79	Butylbenzylphthalate	40.000	40.566	-1.4	105	0.00
80	Benzo(a)anthracene	40.000	38.974	2.6	104	0.00
81	3,3'-Dichlorobenzidine	40.000	39.185	2.0	103	0.00
82	Chrysene	40.000	37.934	5.2	102	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	41.596	-4.0	108	0.00
84 c	Di-n-octyl phthalate	40.000	41.260	-3.1	107	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	39.387	1.5	104	0.00
86 I	Perylene-d12	20.000	20.000	0.0	103	0.00
87	Benzo(b)fluoranthene	40.000	39.003	2.5	102	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_G\Data\BG013018\
Data File : BG032442.D
Acq On : 30 Jan 2018 15:07
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jan 31 00:41:49 2018
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 29 15:32:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	38.989	2.5	103	0.00
89 C	Benzo(a)pyrene	40.000	39.403	1.5	104	0.00
90	Dibenzo(a,h)anthracene	40.000	39.742	0.6	104	0.00
91	Benzo(g,h,i)perylene	40.000	39.155	2.1	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0