

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG013020\
 Data File : BG044250.D
 Acq On : 30 Jan 2020 15:59
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG013020

Quant Time: Jan 30 16:38:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG013020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 30 16:05:09 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.509	0.543	-6.7	90	0.00
3	Pyridine	1.357	1.550	-14.2	89	0.00
4	n-Nitrosodimethylamine	0.567	0.592	-4.4	90	0.00
5 S	2-Fluorophenol	1.133	1.187	-4.8	89	0.00
6	Aniline	1.889	1.959	-3.7	87	0.00
7 S	Phenol-d6	1.522	1.589	-4.4	89	0.00
8	2-Chlorophenol	1.245	1.281	-2.9	87	0.00
9	Benzaldehyde	0.867	0.873	-0.7	90	0.00
10 C	Phenol	1.506	1.579	-4.8	88	0.00
11	bis(2-Chloroethyl)ether	1.288	1.308	-1.6	86	0.00
12	1,3-Dichlorobenzene	1.487	1.531	-3.0	89	0.00
13 C	1,4-Dichlorobenzene	1.473	1.532	-4.0	88	0.00
14	1,2-Dichlorobenzene	1.392	1.436	-3.2	87	0.00
15	Benzyl Alcohol	1.077	1.123	-4.3	87	0.00
16	2,2'-oxybis(1-Chloropropane	1.743	1.777	-2.0	86	0.00
17	2-Methylphenol	1.075	1.110	-3.3	88	0.00
18	Hexachloroethane	0.553	0.570	-3.1	89	0.00
19 P	n-Nitroso-di-n-propylamine	1.014	1.040	-2.6	87	0.00
20	3+4-Methylphenols	1.481	1.536	-3.7	87	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
22	Acetophenone	0.487	0.488	-0.2	87	0.00
23 S	Nitrobenzene-d5	0.354	0.355	-0.3	87	0.00
24	Nitrobenzene	0.346	0.344	0.6	87	0.00
25	Isophorone	0.660	0.653	1.1	86	0.00
26 C	2-Nitrophenol	0.179	0.181	-1.1	86	0.00
27	2,4-Dimethylphenol	0.253	0.258	-2.0	88	0.00
28	bis(2-Chloroethoxy)methane	0.440	0.446	-1.4	87	0.00
29 C	2,4-Dichlorophenol	0.306	0.308	-0.7	87	0.01
30	1,2,4-Trichlorobenzene	0.351	0.353	-0.6	87	0.00
31	Naphthalene	0.970	0.975	-0.5	87	0.00
32	Benzoic acid	0.135	0.133	1.5	87	0.00
33	4-Chloroaniline	0.441	0.443	-0.5	87	0.00
34 C	Hexachlorobutadiene	0.216	0.215	0.5	86	0.00
35	Caprolactam	0.112	0.112	0.0	87	0.00
36 C	4-Chloro-3-methylphenol	0.321	0.318	0.9	87	0.00
37	2-Methylnaphthalene	0.720	0.718	0.3	87	0.00
38	1-Methylnaphthalene	0.679	0.677	0.3	86	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
40	1,2,4,5-Tetrachlorobenzene	0.598	0.604	-1.0	87	0.00
41 P	Hexachlorocyclopentadiene	0.287	0.291	-1.4	84	0.01
42 S	2,4,6-Tribromophenol	0.235	0.235	0.0	88	0.00
43 C	2,4,6-Trichlorophenol	0.387	0.390	-0.8	86	0.00
44	2,4,5-Trichlorophenol	0.395	0.415	-5.1	90	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG013020\
 Data File : BG044250.D
 Acq On : 30 Jan 2020 15:59
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICSVG013020

Quant Time: Jan 30 16:38:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG013020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 30 16:05:09 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.194	1.231	-3.1	88	0.00
46	1,1'-Biphenyl	1.443	1.471	-1.9	88	0.00
47	2-Chloronaphthalene	1.148	1.167	-1.7	89	0.00
48	2-Nitroaniline	0.339	0.338	0.3	87	0.00
49	Acenaphthylene	1.684	1.706	-1.3	88	0.00
50	Dimethylphthalate	1.438	1.446	-0.6	88	0.00
51	2,6-Dinitrotoluene	0.321	0.322	-0.3	89	0.00
52 C	Acenaphthene	1.091	1.100	-0.8	88	0.00
53	3-Nitroaniline	0.355	0.352	0.8	87	0.00
54 P	2,4-Dinitrophenol	0.159	0.167	-5.0	88	0.00
55	Dibenzofuran	1.652	1.677	-1.5	88	0.00
56 P	4-Nitrophenol	0.260	0.274	-5.4	91	0.00
57	2,4-Dinitrotoluene	0.449	0.454	-1.1	89	0.00
58	Fluorene	1.301	1.314	-1.0	88	0.00
59	2,3,4,6-Tetrachlorophenol	0.357	0.362	-1.4	88	0.00
60	Diethylphthalate	1.433	1.458	-1.7	89	0.00
61	4-Chlorophenyl-phenylether	0.706	0.707	-0.1	87	0.00
62	4-Nitroaniline	0.354	0.358	-1.1	90	0.00
63	Azobenzene	1.256	1.256	0.0	87	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
65	4,6-Dinitro-2-methylphenol	0.110	0.113	-2.7	89	0.00
66 c	n-Nitrosodiphenylamine	0.560	0.564	-0.7	88	0.00
67	4-Bromophenyl-phenylether	0.218	0.216	0.9	87	0.00
68	Hexachlorobenzene	0.244	0.242	0.8	88	0.00
69	Atrazine	0.192	0.197	-2.6	88	0.00
70 C	Pentachlorophenol	0.113	0.117	-3.5	86	0.00
71	Phenanthrene	0.968	0.990	-2.3	90	0.00
72	Anthracene	0.957	0.980	-2.4	91	0.00
73	Carbazole	0.883	0.907	-2.7	90	0.00
74	Di-n-butylphthalate	1.091	1.145	-4.9	90	0.00
75 C	Fluoranthene	1.120	1.163	-3.8	91	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
77	Benzidine	0.555	0.552	0.5	79	0.00
78	Pyrene	1.284	1.329	-3.5	92	0.00
79 S	Terphenyl-d14	0.972	0.947	2.6	93	0.00
80	Butylbenzylphthalate	0.585	0.613	-4.8	92	0.00
81	Benzo(a)anthracene	1.233	1.279	-3.7	92	0.00
82	3,3'-Dichlorobenzidine	0.478	0.488	-2.1	88	0.00
83	Chrysene	1.191	1.241	-4.2	92	0.00
84	Bis(2-ethylhexyl)phthalate	0.806	0.844	-4.7	92	0.00
85 c	Di-n-octyl phthalate	1.394	1.455	-4.4	92	0.00
86	Indeno(1,2,3-cd)pyrene	1.554	1.578	-1.5	92	0.01
87 I	Perylene-d12	1.000	1.000	0.0	90	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG013020\
Data File : BG044250.D
Acq On : 30 Jan 2020 15:59
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG013020

Quant Time: Jan 30 16:38:39 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG013020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 30 16:05:09 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.152	1.174	-1.9	92	0.00
89	Benzo(k)fluoranthene	1.123	1.152	-2.6	91	0.00
90 C	Benzo(a)pyrene	1.090	1.108	-1.7	91	0.00
91	Dibenzo(a,h)anthracene	1.078	1.093	-1.4	92	0.00
92	Benzo(a,h,i)perylene	1.080	1.088	-0.7	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0