


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG013023\
 Data File : BG056466.D
 Acq On : 30 Jan 2023 9:18
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG012723.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Jan 28 00:39:13 2023






AutoFind: Scans 2509, 2510, 2511; Background Corrected with Scan 2502

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	13.3	12658	PASS
68	69	0.00	2	0.5	84	PASS
69	198	0.00	100	19.5	18660	PASS
70	69	0.00	2	0.3	51	PASS
127	198	10	80	34.7	33141	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	95461	PASS
199	198	5	9	6.9	6627	PASS
275	198	10	60	32.2	30763	PASS
365	198	1	100	3.7	3576	PASS
441	198	0.01	100	8.5	8073	PASS
442	442	50	100	100.0	70661	PASS
443	442	15	24	19.9	14059	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG013023\
Data File : BG056466.D
Acq On : 30 Jan 2023 9:18
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Jan 31 06:19:09 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG012723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 28 00:39:13 2023
Response via : Initial Calibration



TIC: BG056466.D\data.ms

(70) Pentachlorophenol (C)

17.283min (-0.005) 24265.83 ng




response 107882

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.80	63.75
264.00	64.20	62.74
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG013023\
Data File : BG056466.D
Acq On : 30 Jan 2023 9:18
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Jan 31 06:19:09 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG012723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 28 00:39:13 2023
Response via : Initial Calibration



TIC: BG056466.D\data.ms

(77) Benzidine

19.844min (-19.844) 0.00 ng

response 0

Ion	Exp%	Act%
184.00	100.00	0.00
185.00	15.40	0.00#
183.00	12.30	0.00#
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
1/30/2023	BNA_G	<u>BG056466.D</u>
Compound Name	Response	Retention Time
DDT	343018	21.067
DDD	8839	20.62
DDE	958	20.115
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
9797	352815	2.78

Instrument :
BNA_G
ClientSampleId :
DFTPP