

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG013024\
 Data File : BG060488.D
 Acq On : 30 Jan 2024 18:02
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 31 01:22:15 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 09 04:07:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	78	0.00
2	1,4-Dioxane	40.000	44.606	-11.5	76	-0.01
3	Pyridine	40.000	39.966	0.1	65	-0.01
4	n-Nitrosodimethylamine	40.000	34.424	13.9	59	-0.01
5 S	2-Fluorophenol	80.000	76.763	4.0	79	0.00
6	Aniline	40.000	40.670	-1.7	73	0.00
7 S	Phenol-d6	80.000	82.040	-2.6	68	0.00
8	2-Chlorophenol	40.000	40.063	-0.2	76	0.00
9	Benzaldehyde	40.000	33.553	16.1	60	-0.01
10 C	Phenol	40.000	41.168	-2.9	69	0.00
11	bis(2-Chloroethyl)ether	40.000	37.956	5.1	73	-0.01
12	1,3-Dichlorobenzene	40.000	40.073	-0.2	77	0.00
13 C	1,4-Dichlorobenzene	40.000	39.247	1.9	77	0.00
14	1,2-Dichlorobenzene	40.000	38.175	4.6	67	0.00
15	Benzyl Alcohol	40.000	43.238	-8.1	78	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	39.593	1.0	69	0.00
17	2-Methylphenol	40.000	39.445	1.4	66	0.00
18	Hexachloroethane	40.000	39.058	2.4	68	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.878	-2.2	66	0.00
20	3+4-Methylphenols	40.000	43.498	-8.7	72	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	71	0.00
22	Acetophenone	40.000	40.052	-0.1	71	0.00
23 S	Nitrobenzene-d5	80.000	83.455	-4.3	74	0.00
24	Nitrobenzene	40.000	39.128	2.2	69	0.00
25	Isophorone	40.000	40.072	-0.2	71	0.00
26 C	2-Nitrophenol	40.000	46.587	-16.5	81	0.00
27	2,4-Dimethylphenol	40.000	42.441	-6.1	74	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.899	5.3	68	0.00
29 C	2,4-Dichlorophenol	40.000	42.936	-7.3	75	0.00
30	1,2,4-Trichlorobenzene	40.000	39.643	0.9	71	-0.01
31	Naphthalene	40.000	39.571	1.1	72	0.00
32	Benzoic acid	40.000	40.617	-1.5	77	0.01
33	4-Chloroaniline	40.000	40.816	-2.0	74	0.00
34 C	Hexachlorobutadiene	40.000	41.144	-2.9	79	-0.01
35	Caprolactam	40.000	44.759	-11.9	81	0.00
36 C	4-Chloro-3-methylphenol	40.000	44.696	-11.7	80	0.00
37	2-Methylnaphthalene	40.000	41.260	-3.1	76	0.00
38	1-Methylnaphthalene	40.000	40.813	-2.0	76	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	77	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.453	-1.1	81	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.769	10.6	66	0.00
42 S	2,4,6-Tribromophenol	80.000	92.435	-15.5	89	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.490	-3.7	81	0.00
44	2,4,5-Trichlorophenol	40.000	41.995	-5.0	81	0.00
45 S	2-Fluorobiphenyl	80.000	80.269	-0.3	83	0.00
46	1,1'-Biphenyl	40.000	38.634	3.4	77	0.00
47	2-Chloronaphthalene	40.000	38.584	3.5	77	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG013024\
 Data File : BG060488.D
 Acq On : 30 Jan 2024 18:02
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 31 01:22:15 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 09 04:07:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.059	-12.6	85	0.00
49	Acenaphthylene	40.000	40.846	-2.1	79	0.00
50	Dimethylphthalate	40.000	41.014	-2.5	81	0.00
51	2,6-Dinitrotoluene	40.000	45.197	-13.0	84	0.00
52 C	Acenaphthene	40.000	39.487	1.3	77	0.00
53	3-Nitroaniline	40.000	44.514	-11.3	84	0.00
54 P	2,4-Dinitrophenol	40.000	41.896	-4.7	86	0.00
55	Dibenzofuran	40.000	39.951	0.1	78	0.00
56 P	4-Nitrophenol	40.000	46.171	-15.4	81	0.00
57	2,4-Dinitrotoluene	40.000	46.509	-16.3	87	0.00
58	Fluorene	40.000	40.286	-0.7	80	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	44.924	-12.3	86	0.00
60	Diethylphthalate	40.000	41.849	-4.6	81	0.00
61	4-Chlorophenyl-phenylether	40.000	41.153	-2.9	83	0.00
62	4-Nitroaniline	40.000	44.406	-11.0	84	0.00
63	Azobenzene	40.000	39.139	2.2	77	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	84	0.00
65	4,6-Dinitro-2-methylphenol	40.000	45.364	-13.4	93	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.985	2.5	81	0.00
67	4-Bromophenyl-phenylether	40.000	40.916	-2.3	86	0.00
68	Hexachlorobenzene	40.000	40.522	-1.3	86	0.00
69	Atrazine	40.000	37.042	7.4	78	0.00
70 C	Pentachlorophenol	40.000	46.466	-16.2	90	0.00
71	Phenanthrene	40.000	39.462	1.3	83	0.00
72	Anthracene	40.000	39.852	0.4	83	0.00
73	Carbazole	40.000	39.322	1.7	82	0.00
74	Di-n-butylphthalate	40.000	40.636	-1.6	83	0.00
75 C	Fluoranthene	40.000	39.573	1.1	85	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	81	0.00
77	Benzidine	40.000	38.014	5.0	75	0.00
78	Pyrene	40.000	39.907	0.2	84	0.00
79 S	Terphenyl-d14	80.000	90.924	-13.7	91	0.00
80	Butylbenzylphthalate	40.000	43.346	-8.4	86	0.00
81	Benzo(a)anthracene	40.000	40.776	-1.9	84	0.00
82	3,3'-Dichlorobenzidine	40.000	40.251	-0.6	80	0.00
83	Chrysene	40.000	40.118	-0.3	84	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.783	-4.5	83	0.00
85 c	Di-n-octyl phthalate	40.000	43.514	-8.8	85	0.00
86 I	Perylene-d12	20.000	20.000	0.0	82	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.109	-0.3	82	0.00
88	Benzo(b)fluoranthene	40.000	40.867	-2.2	82	0.00
89	Benzo(k)fluoranthene	40.000	38.784	3.0	78	0.00
90 C	Benzo(a)pyrene	40.000	40.101	-0.3	82	0.00
91	Dibenzo(a,h)anthracene	40.000	40.224	-0.6	81	0.00
92	Benzo(g,h,i)perylene	40.000	40.086	-0.2	82	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG013024\
Data File : BG060488.D
Acq On : 30 Jan 2024 18:02
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jan 31 01:22:15 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 09 04:07:12 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0