

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG013122\
 Data File : BG052214.D
 Acq On : 31 Jan 2022 13:32
 Operator : CG/JU
 Sample : SST00549
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST005449

Quant Time: Jan 31 15:50:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG013122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 31 15:44:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.241	152	29494	20.000	ng/ul	0.00
20) Naphthalene-d8	11.072	136	110685	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.879	164	79798	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.728	188	45197	20.000	ng/ul	0.10
79) Chrysene-d12	22.046	240	208	20.000	ng/ul	# 0.10
88) Perylene-d12	25.418	264	197843	20.000	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.565	96	1373	1.663	ng/uL	0.00
4) Pyridine-d5	4.000	84	8823	3.439	ng/ul	0.01
7) Phenol-d5	7.377	99	11473	3.974	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.548	67	7000	3.820	ng/ul	0.00
11) 2-Chlorophenol-d4	7.765	132	8277	4.331	ng/ul	0.00
15) 4-Methylphenol-d8	8.940	113	8320	3.708	ng/ul	0.00
21) Nitrobenzene-d5	9.404	128	4136	4.729	ng/ul	0.00
24) 2-Nitrophenol-d4	10.133	143	4270	4.414	ng/ul	-0.01
28) 2,4-Dichlorophenol-d3	10.691	165	8537	4.515	ng/ul	0.00
31) 4-Chloroaniline-d4	11.196	131	10612	4.030	ng/ul	0.00
46) Dimethylphthalate-d6	14.268	166	29917	4.536	ng/ul	0.00
49) Acenaphthylene-d8	14.574	160	35673	4.477	ng/ul	0.00
54) 4-Nitrophenol-d4	15.067	143	2903	2.632	ng/ul	0.00
60) Fluorene-d10	15.866	176	25642	4.459	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.030	200	122	0.428	ng/ul	0.05
73) Anthracene-d10	17.769	188	180	0.084	ng/ul	0.04
81) Pyrene-d10	20.048	212	147	12.280	ng/ul	0.04
92) Benzo(a)pyrene-d12	25.177	264	47485	4.559	ng/ul	0.00

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.600	88	868	0.911	ng/uL#	83
5) Pyridine	4.017	79	10192	3.791	ng/ul	95
6) Benzaldehyde	7.354	77	8315	4.330	ng/ul	92
8) Phenol	7.407	94	10943	3.714	ng/ul#	86
10) Bis(2-Chloroethyl)ether	7.636	93	8403	3.896	ng/ul	94
12) 2-Chlorophenol	7.795	128	9026	4.642	ng/ul#	83
13) 2-Methylphenol	8.676	108	8694	4.058	ng/ul	96
14) 2,2'-oxybis(1-Chloropr...	8.758	45	8312	2.527	ng/ul#	90
16) Acetophenone	9.063	105	14921	4.213	ng/ul	98
17) N-Nitroso-di-n-propyla...	9.034	70	8768	4.052	ng/ul#	89
18) 4-Methylphenol	9.011	108	8996	3.876	ng/ul	81
19) Hexachloroethane	9.339	117	3560	4.241	ng/ul#	68
22) Nitrobenzene	9.451	77	11885	4.381	ng/ul	94
23) Isophorone	9.974	82	21162	4.126	ng/ul	99
25) 2-Nitrophenol	10.174	139	4932	4.639	ng/ul#	85
26) 2,4-Dimethylphenol	10.221	107	10113	4.384	ng/ul#	86
27) Bis(2-Chloroethoxy)met...	10.456	93	11842	4.485	ng/ul	92
29) 2,4-Dichlorophenol	10.714	162	8924	4.806	ng/ul	92
30) Naphthalene	11.125	128	28997	4.835	ng/ul	97
32) 4-Chloroaniline	11.225	127	11567	4.311	ng/ul	92
33) Hexachlorobutadiene	11.413	225	7172	4.845	ng/ul	92
34) Caprolactam	11.959	113	2595	3.591	ng/ul#	76
35) 4-Chloro-3-methylphenol	12.341	107	9321	4.348	ng/ul	98
36) 2-Methylnaphthalene	12.717	142	20253	4.740	ng/ul	90

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG013122\
 Data File : BG052214.D
 Acq On : 31 Jan 2022 13:32
 Operator : CG/JU
 Sample : SSTD00549
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD005449

Quant Time: Jan 31 15:50:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG013122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 31 15:44:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.935	142	21791	4.970	ng/ul	90
39) 1,2,4,5-Tetrachloroben...	13.087	216	13079	4.740	ng/ul	94
40) Hexachlorocyclopentadiene	13.070	237	4048	2.138	ng/ul#	95
41) 2,4,6-Trichlorophenol	13.322	196	7326	4.200	ng/ul	97
42) 2,4,5-Trichlorophenol	13.393	196	8310	4.420	ng/ul	86
43) 1,1'-Biphenyl	13.710	154	30142	4.893	ng/ul	90
44) 2-Chloronaphthalene	13.757	162	22724	4.763	ng/ul	98
45) 2-Nitroaniline	13.951	65	6833	3.737	ng/ul	95
47) Dimethylphthalate	14.315	163	28333	4.356	ng/ul#	94
48) 2,6-Dinitrotoluene	14.438	165	6110	4.456	ng/ul#	84
50) Acenaphthylene	14.603	152	37728	4.809	ng/ul	96
51) 3-Nitroaniline	14.773	138	5583	4.501	ng/ul#	93
52) Acenaphthene	14.944	153	23908	4.583	ng/ul	92
53) 2,4-Dinitrophenol	15.002	184	221	0.273	ng/ul	96
55) 4-Nitrophenol	15.085	109	3072	2.428	ng/ul#	77
56) Dibenzofuran	15.273	168	35535	4.690	ng/ul	97
57) 2,4-Dinitrotoluene	15.220	165	8473	4.274	ng/ul#	81
58) 2,3,4,6-Tetrachlorophenol	15.502	232	6800	3.899	ng/ul#	98
59) Diethylphthalate	15.672	149	29096	4.336	ng/ul	97
61) Fluorene	15.925	166	29891	4.735	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.907	204	15955	4.554	ng/ul#	84
63) 4-Nitroaniline	15.925	138	4854	3.809	ng/ul	86
66) 4,6-Dinitro-2-methylph...	16.025	198	72	0.264	ng/ul#	51
67) N-Nitrosodiphenylamine	16.119	169	24849	20.461	ng/ul	97
68) 4-Bromophenyl-phenylether	16.935	248	135	0.255	ng/ul#	1
69) Hexachlorobenzene	16.935	284	12627	21.517	ng/ul#	85
70) Atrazine	17.058	200	10775	19.123	ng/ul	93
71) Pentachlorophenol	17.352	266	210	0.641	ng/ul#	17
72) Phenanthrene	17.757	178	49919	21.078	ng/ul	98
74) Anthracene	17.757	178	49919	21.057	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.687	216	15044	22.163	ng/uL	95
76) Pentachlorobenzene	15.196	250	14477	21.913	ng/uL#	86
77) Carbazole	18.086	167	255	0.119	ng/ul#	59
78) Di-n-butylphthalate	18.568	149	50061	19.826	ng/ul	99
80) Fluoranthene	19.755	202	80	5.783	ng/ul#	72
82) Pyrene	20.031	202	64837	4760.374	ng/ul#	90
83) Butylbenzylphthalate	20.947	149	116	24.473	ng/ul#	83
84) 3,3'-Dichlorobenzidine	21.928	252	82	17.078	ng/ul#	24
85) Benzo(a)anthracene	21.993	228	60221	4402.431	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.917	149	481	69.388	ng/ul#	53
87) Chrysene	21.993	228	59308	4502.618	ng/ul	99
89) Di-n-octyl phthalate	23.097	149	51298	4.694	ng/ul	100
90) Benzo(b)fluoranthene	24.302	252	63450	4.839	ng/ul#	96
91) Benzo(k)fluoranthene	24.378	252	58556	4.830	ng/ul#	97
93) Benzo(a)pyrene	25.247	252	60008	4.836	ng/ul#	94
94) Indeno(1,2,3-cd)pyrene	29.412	276	23266	1.638	ng/ul#	85
95) Dibenzo(a,h)anthracene	29.489	278	24675	2.054	ng/ul#	88
96) Benzo(g,h,i)perylene	30.664	276	26045	2.181	ng/ul#	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

BNA G

ClientSampleId :

SSTD005449





[illegible]



#2
 1,4-Dioxane
 Concen: 0.911 ng/uL
 RT: 3.600 min Scan# 19
 Delta R.T. 0.001 min
 Lab File: BG052214.D
 Acq: 31 Jan 2022 13:32




Instrument :
 BNA_G
 ClientSampleId :
 SST005449


Tgt Ion: 88 Resp: 868
 Ion Ratio Lower Upper
 88 100
 43 53.2 23.0 34.4#
 58 83.3 62.4 93.6



#5
 Pyridine
 Concen: 3.791 ng/uL
 RT: 4.017 min Scan# 90
 Delta R.T. 0.007 min
 Lab File: BG052214.D
 Acq: 31 Jan 2022 13:32

Tgt Ion: 79 Resp: 10192
 Ion Ratio Lower Upper
 79 100
 52 79.7 60.8 91.2
 51 33.8 30.5 45.7









#6
Benzaldehyde
Concen: 4.330 ng/ul
RT: 7.354 min Scan# 611
Delta R.T. -0.005 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion: 77 Resp: 8315
Ion Ratio Lower Upper
77 100
105 75.4 70.4 105.6
106 78.6 61.2 91.8



#8
Phenol
Concen: 3.714 ng/ul
RT: 7.407 min Scan# 667
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion: 94 Resp: 10943
Ion Ratio Lower Upper
94 100
65 37.6 25.1 37.7
66 50.0 31.6 47.4#





#10

Bis(2-Chloroethyl)ether

Concen: 3.896 ng/ul

RT: 7.636 min Scan# 706

Delta R.T. 0.001 min

Lab File: BG052214.D

Acq: 31 Jan 2022 13:32

Instrument :

BNA_G

ClientSampleId :

SSTD005449





Tgt Ion: 93 Resp: 8403

Ion Ratio Lower Upper

93 100

63 91.1 68.6 103.0

95 35.5 26.1 39.1



#12

2-Chlorophenol

Concen: 4.642 ng/ul

RT: 7.795 min Scan# 733

Delta R.T. -0.005 min

Lab File: BG052214.D

Acq: 31 Jan 2022 13:32




Tgt Ion: 128 Resp: 9026


Ion Ratio Lower Upper

128 100

64 42.6 45.3 67.9#

130 26.2 27.1 40.7#





#13

2-Methylphenol

Concen: 4.058 ng/ul

RT: 8.676 min Scan# 883

Delta R.T. 0.001 min

Lab File: BG052214.D

Acq: 31 Jan 2022 13:32

Instrument :

BNA_G

ClientSampleId :





SSTD005449

Tgt Ion:108 Resp: 8694

Ion Ratio Lower Upper

108 100

107 85.9 71.8 107.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 2.527 ng/ul

RT: 8.758 min Scan# 897

Delta R.T. -0.005 min

Lab File: BG052214.D

Acq: 31 Jan 2022 13:32




Tgt Ion: 45 Resp: 8312


Ion Ratio Lower Upper

45 100

77 20.2 10.7 16.1#

79 11.2 8.5 12.7









#16
Acetophenone
Concen: 4.213 ng/ul
RT: 9.063 min Scan# 949
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:105 Resp: 14921
Ion Ratio Lower Upper
105 100
77 82.3 67.3 100.9
51 31.9 24.0 36.0



#17
N-Nitroso-di-n-propylamine
Concen: 4.052 ng/ul
RT: 9.034 min Scan# 944
Delta R.T. -0.005 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion: 70 Resp: 8768
Ion Ratio Lower Upper
70 100
42 49.4 46.0 69.0
101 9.5 7.0 10.4
130 9.8 13.4 20.2#









#18
4-Methylphenol
Concen: 3.876 ng/ul
RT: 9.011 min Scan# 940
Delta R.T. 0.007 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:108 Resp: 8996
Ion Ratio Lower Upper
108 100
107 97.8 95.0 142.4



#19
Hexachloroethane
Concen: 4.241 ng/ul
RT: 9.339 min Scan# 996
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:117 Resp: 3560
Ion Ratio Lower Upper
117 100
201 124.8 77.2 115.8#
199 91.0 49.9 74.9#





#22

Nitrobenzene

Concen: 4.381 ng/ul

RT: 9.451 min Scan# 1015

Delta R.T. 0.001 min

Lab File: BG052214.D

Acq: 31 Jan 2022 13:32

Instrument :

BNA_G

ClientSampleId :

SSTD005449





Tgt Ion: 77 Resp: 11885

Ion Ratio Lower Upper

77 100

123 32.9 30.1 45.1

65 12.7 9.6 14.4



#23

Isophorone

Concen: 4.126 ng/ul

RT: 9.974 min Scan# 1104

Delta R.T. 0.001 min

Lab File: BG052214.D

Acq: 31 Jan 2022 13:32




Tgt Ion: 82 Resp: 21162


Ion Ratio Lower Upper

82 100

95 8.2 5.6 8.4

138 15.0 11.8 17.8









#25
2-Nitrophenol
Concen: 4.639 ng/ul
RT: 10.174 min Scan# 1138
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:139 Resp: 4932
Ion Ratio Lower Upper
139 100
65 47.6 48.0 72.0#
109 43.8 28.6 42.8#



#26
2,4-Dimethylphenol
Concen: 4.384 ng/ul
RT: 10.221 min Scan# 1146
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:107 Resp: 10113
Ion Ratio Lower Upper
107 100
121 52.0 39.3 58.9
122 97.5 63.7 95.5#





#27

Bis(2-Chloroethoxy)methane

Concen: 4.485 ng/ul

RT: 10.456 min Scan# 11842

Delta R.T. 0.001 min

Lab File: BG052214.D

Acq: 31 Jan 2022 13:32

Instrument :

BNA_G

ClientSampleId :

SSTD005449





Tgt Ion: 93 Resp: 11842

Ion Ratio Lower Upper

93 100

95 38.4 26.5 39.7

123 8.9 8.1 12.1



#29

2,4-Dichlorophenol

Concen: 4.806 ng/ul

RT: 10.714 min Scan# 1230

Delta R.T. -0.005 min

Lab File: BG052214.D

Acq: 31 Jan 2022 13:32




Tgt Ion:162 Resp: 8924


Ion Ratio Lower Upper

162 100

164 69.3 49.4 74.0

98 41.3 30.4 45.6









#30
Naphthalene
Concen: 4.835 ng/ul
RT: 11.125 min Scan# 1112
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:128 Resp: 28997
Ion Ratio Lower Upper
128 100
129 12.0 8.5 12.7
127 15.6 11.8 17.6



#32
4-Chloroaniline
Concen: 4.311 ng/ul
RT: 11.225 min Scan# 1317
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:127 Resp: 11567
Ion Ratio Lower Upper
127 100
129 28.2 26.2 39.2









#33
Hexachlorobutadiene
Concen: 4.845 ng/ul
RT: 11.413 min Scan# 1141
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:225 Resp: 7172
Ion Ratio Lower Upper
225 100
223 71.8 50.6 75.8
227 66.6 49.8 74.8



#34
Caprolactam
Concen: 3.591 ng/ul
RT: 11.959 min Scan# 1442
Delta R.T. -0.011 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:113 Resp: 2595
Ion Ratio Lower Upper
113 100
55 140.1 147.0 220.6#
56 117.5 109.2 163.8





#35

4-Chloro-3-methylphenol

Concen: 4.348 ng/ul

RT: 12.341 min Scan# 1

Delta R.T. 0.007 min

Lab File: BG052214.D

Acq: 31 Jan 2022 13:32

Instrument :

BNA_G

ClientSampleId :

SSTD005449





Tgt Ion:107 Resp: 9321

Ion Ratio Lower Upper

107 100

144 23.4 19.8 29.6

142 69.2 56.2 84.4



#36

2-Methylnaphthalene

Concen: 4.740 ng/ul

RT: 12.717 min Scan# 1571

Delta R.T. -0.005 min

Lab File: BG052214.D




Acq: 31 Jan 2022 13:32


Tgt Ion:142 Resp: 20253

Ion Ratio Lower Upper

142 100

141 82.2 73.3 109.9





#37

1-Methylnaphthalene

Concen: 4.970 ng/ul

RT: 12.935 min Scan# 1

Delta R.T. -0.005 min

Lab File: BG052214.D

Acq: 31 Jan 2022 13:32

Instrument :

BNA_G

ClientSampleId :

SSTD005449





Tgt Ion:142 Resp: 21791

Ion Ratio Lower Upper

142 100

141 86.7 77.8 116.8

116 3.7 3.3 4.9



#39

1,2,4,5-Tetrachlorobenzene

Concen: 4.740 ng/ul

RT: 13.087 min Scan# 1634

Delta R.T. 0.001 min

Lab File: BG052214.D

Acq: 31 Jan 2022 13:32

Tgt Ion:216 Resp: 13079




Ion Ratio Lower Upper


216 100

214 82.3 61.0 91.4

179 17.1 16.2 24.2

108 16.6 13.1 19.7









#40
Hexachlorocyclopentadiene
Concen: 2.138 ng/ul
RT: 13.070 min Scan# 1631
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:237 Resp: 4048
Ion Ratio Lower Upper
237 100
235 58.7 49.4 74.0
272 17.9 11.1 16.7#



#41
2,4,6-Trichlorophenol
Concen: 4.200 ng/ul
RT: 13.322 min Scan# 1674
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:196 Resp: 7326
Ion Ratio Lower Upper
196 100
198 88.5 74.3 111.5
200 29.8 23.8 35.6









#42
2,4,5-Trichlorophenol
Concen: 4.420 ng/ul
RT: 13.393 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:196 Resp: 8310
Ion Ratio Lower Upper
196 100
198 79.1 77.0 115.4
200 30.0 25.4 38.2



#43
1,1'-Biphenyl
Concen: 4.893 ng/ul
RT: 13.710 min Scan# 1740
Delta R.T. -0.005 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:154 Resp: 30142
Ion Ratio Lower Upper
154 100
153 35.5 34.5 51.7
76 13.7 13.1 19.7









#44
2-Chloronaphthalene
Concen: 4.763 ng/ul
RT: 13.757 min Scan# 1748
Delta R.T. -0.005 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:162 Resp: 22724
Ion Ratio Lower Upper
162 100
164 31.1 26.1 39.1
127 37.3 31.0 46.4



#45
2-Nitroaniline
Concen: 3.737 ng/ul
RT: 13.951 min Scan# 1781
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion: 65 Resp: 6833
Ion Ratio Lower Upper
65 100
92 62.0 53.0 79.4
138 88.3 67.4 101.0









#47
Dimethylphthalate
Concen: 4.356 ng/ul
RT: 14.315 min Scan# 1843
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:163 Resp: 28333
Ion Ratio Lower Upper
163 100
194 4.3 3.0 4.4
164 7.4 8.2 12.4#



#48
2,6-Dinitrotoluene
Concen: 4.456 ng/ul
RT: 14.438 min Scan# 1864
Delta R.T. -0.005 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:165 Resp: 6110
Ion Ratio Lower Upper
165 100
89 56.9 57.4 86.0#
121 18.2 18.8 28.2#









#50
Acenaphthylene
Concen: 4.809 ng/ul
RT: 14.603 min Scan# 1892
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:152 Resp: 37728
Ion Ratio Lower Upper
152 100
151 19.2 17.4 26.0
153 13.2 10.3 15.5



#51
3-Nitroaniline
Concen: 4.501 ng/ul
RT: 14.773 min Scan# 1921
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:138 Resp: 5583
Ion Ratio Lower Upper
138 100
108 14.6 8.8 13.2#
92 115.5 98.0 147.0









#52
Acenaphthene
Concen: 4.583 ng/ul
RT: 14.944 min Scan# 1950
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:153 Resp: 23908
Ion Ratio Lower Upper
153 100
152 45.2 36.5 54.7
154 93.3 65.8 98.6



#53
2,4-Dinitrophenol
Concen: 0.273 ng/ul
RT: 15.002 min Scan# 1960
Delta R.T. 0.025 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:184 Resp: 221
Ion Ratio Lower Upper
184 100
63 82.4 66.2 99.2
154 60.6 53.6 80.4









#55
4-Nitrophenol
Concen: 2.428 ng/ul
RT: 15.085 min Scan# 1973
Delta R.T. 0.007 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:109 Resp: 3072
Ion Ratio Lower Upper
109 100
139 89.7 88.7 133.1
65 111.9 113.6 170.4#



#56
Dibenzofuran
Concen: 4.690 ng/ul
RT: 15.273 min Scan# 2006
Delta R.T. -0.005 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:168 Resp: 35535
Ion Ratio Lower Upper
168 100
139 37.2 31.2 46.8





#57

2,4-Dinitrotoluene

Concen: 4.274 ng/ul

RT: 15.220 min Scan# 1998

Delta R.T. -0.005 min

Lab File: BG052214.D

Acq: 31 Jan 2022 13:32

Instrument :

BNA_G

ClientSampleId :

SSTD005449





Tgt Ion:165 Resp: 8473

Ion Ratio Lower Upper

165 100

63 38.5 42.1 63.1#

182 0.0 1.8 2.8#



#58

2,3,4,6-Tetrachlorophenol

Concen: 3.899 ng/ul

RT: 15.502 min Scan# 2045

Delta R.T. 0.001 min

Lab File: BG052214.D

Acq: 31 Jan 2022 13:32

Tgt Ion:232 Resp: 6800




Ion Ratio Lower Upper


232 100

131 48.9 38.9 58.3

130 0.0 2.4 3.6#

166 30.5 25.8 38.8









#59
Diethylphthalate
Concen: 4.336 ng/ul
RT: 15.672 min Scan# 2074
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:149 Resp: 29096
Ion Ratio Lower Upper
149 100
177 21.8 16.8 25.2
150 14.3 10.1 15.1



#61
Fluorene
Concen: 4.735 ng/ul
RT: 15.925 min Scan# 2117
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:166 Resp: 29891
Ion Ratio Lower Upper
166 100
165 102.7 81.0 121.6
167 13.1 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 4.554 ng/ul
RT: 15.907 min Scan# 2114
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:204 Resp: 15955
Ion Ratio Lower Upper
204 100
206 30.6 27.4 41.2
141 44.8 48.6 73.0#



#63
4-Nitroaniline
Concen: 3.809 ng/ul
RT: 15.925 min Scan# 2117
Delta R.T. -0.005 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:138 Resp: 4854
Ion Ratio Lower Upper
138 100
92 51.9 49.3 73.9
108 104.1 72.6 108.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 0.264 ng/ul
 RT: 16.025 min Scan# 2129
 Delta R.T. 0.030 min
 Lab File: BG052214.D
 Acq: 31 Jan 2022 13:32




Instrument :
 BNA_G
 ClientSampleId :
 SSTD005449


Tgt Ion:198 Resp: 72
 Ion Ratio Lower Upper
 198 100
 182 0.0 3.8 5.8#
 77 0.0 23.2 34.8#



#67
 N-Nitrosodiphenylamine
 Concen: 20.461 ng/ul
 RT: 16.119 min Scan# 2150
 Delta R.T. 0.001 min
 Lab File: BG052214.D
 Acq: 31 Jan 2022 13:32

Tgt Ion:169 Resp: 24849
 Ion Ratio Lower Upper
 169 100
 168 69.1 58.5 87.7
 167 38.7 31.3 46.9









#68
4-Bromophenyl-phenylether
Concen: 0.255 ng/ul
RT: 16.935 min Scan# 21
Delta R.T. 0.130 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:248 Resp: 135
Ion Ratio Lower Upper
248 100
250 0.0 75.5 113.3#
141 268.0 67.6 101.4#



#69
Hexachlorobenzene
Concen: 21.517 ng/ul
RT: 16.935 min Scan# 2289
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:284 Resp: 12627
Ion Ratio Lower Upper
284 100
142 26.2 31.9 47.9#
249 29.1 25.8 38.6





#70

Atrazine

Concen: 19.123 ng/ul

RT: 17.058 min Scan# 21

Delta R.T. -0.005 min

Lab File: BG052214.D

Acq: 31 Jan 2022 13:32

Instrument :

BNA_G

ClientSampleId :

SSTD005449





Tgt Ion:200 Resp: 10775

Ion Ratio Lower Upper

200 100

173 26.9 20.2 30.2

215 53.6 38.1 57.1



#71

Pentachlorophenol

Concen: 0.641 ng/ul

RT: 17.352 min Scan# 2360

Delta R.T. 0.072 min

Lab File: BG052214.D

Acq: 31 Jan 2022 13:32




Tgt Ion:266 Resp: 210


Ion Ratio Lower Upper

266 100

264 0.0 54.3 81.5#

268 0.0 51.0 76.6#









#72
Phenanthrene
Concen: 21.078 ng/u1
RT: 17.757 min Scan# 2414
Delta R.T. 0.089 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:178 Resp: 49919
Ion Ratio Lower Upper
178 100
179 16.7 12.3 18.5
176 20.1 15.5 23.3



#74
Anthracene
Concen: 21.057 ng/u1
RT: 17.757 min Scan# 2429
Delta R.T. -0.005 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:178 Resp: 49919
Ion Ratio Lower Upper
178 100
179 16.7 13.0 19.6
176 20.1 15.6 23.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 22.163 ng/uL
RT: 13.687 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:216 Resp: 15044
Ion Ratio Lower Upper
216 100
214 83.8 61.3 91.9
218 46.7 37.0 55.6



#76
Pentachlorobenzene
Concen: 21.913 ng/uL
RT: 15.196 min Scan# 1993
Delta R.T. -0.005 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:250 Resp: 14477
Ion Ratio Lower Upper
250 100
248 56.1 50.5 75.7
252 48.9 51.2 76.8#









#77
 Carbazole
 Concen: 0.119 ng/ul
 RT: 18.086 min Scan# 2475
 Delta R.T. 0.060 min
 Lab File: BG052214.D
 Acq: 31 Jan 2022 13:32




Instrument :
 BNA_G
 ClientSampleId :
 SSTD005449


Tgt Ion:167 Resp: 255
 Ion Ratio Lower Upper
 167 100
 166 0.0 17.3 25.9#
 139 0.0 10.6 16.0#



#78
 Di-n-butylphthalate
 Concen: 19.826 ng/ul
 RT: 18.568 min Scan# 2567
 Delta R.T. 0.001 min
 Lab File: BG052214.D
 Acq: 31 Jan 2022 13:32

Tgt Ion:149 Resp: 50061
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.5 11.3
 104 6.4 5.0 7.6









#80
Fluoranthene
Concen: 5.783 ng/ul
RT: 19.755 min Scan# 21
Delta R.T. 0.083 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:202 Resp: 80
Ion Ratio Lower Upper
202 100
101 0.0 9.3 13.9#
100 0.0 7.0 10.6#



#82
Pyrene
Concen: 4760.374 ng/ul
RT: 20.031 min Scan# 2816
Delta R.T. -0.005 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:202 Resp: 64837
Ion Ratio Lower Upper
202 100
101 9.0 10.7 16.1#
100 7.8 9.0 13.6#









#83
Butylbenzylphthalate
Concen: 24.473 ng/ul
RT: 20.947 min Scan# 2965
Delta R.T. 0.042 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:149 Resp: 116
Ion Ratio Lower Upper
149 100
91 90.9 65.2 97.8
206 0.0 16.2 24.2#



#84
3,3'-Dichlorobenzidine
Concen: 17.078 ng/ul
RT: 21.928 min Scan# 3139
Delta R.T. 0.101 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:252 Resp: 82
Ion Ratio Lower Upper
252 100
254 85.0 50.8 76.2#
126 145.3 9.0 13.4#









#85
Benzo(a)anthracene
Concen: 4402.431 ng/u1
RT: 21.993 min Scan# 3137
Delta R.T. 0.066 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:228 Resp: 60221
Ion Ratio Lower Upper
228 100
229 20.6 15.7 23.5
226 31.0 22.3 33.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 69.388 ng/u1
RT: 21.917 min Scan# 3137
Delta R.T. 0.113 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:149 Resp: 481
Ion Ratio Lower Upper
149 100
167 0.0 21.1 31.7#
279 0.0 2.7 4.1#





#87

Chrysene

Concen: 4502.618 ng/ul

RT: 21.993 min Scan# 31

Delta R.T. -0.005 min

Lab File: BG052214.D

Acq: 31 Jan 2022 13:32

Instrument :

BNA_G

ClientSampleId :

SSTD005449





Tgt Ion:228 Resp: 59308

Ion Ratio Lower Upper

228 100

226 31.0 24.8 37.2

229 20.6 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 4.694 ng/ul




RT: 23.097 min Scan# 3338


Delta R.T. -0.005 min

Lab File: BG052214.D

Acq: 31 Jan 2022 13:32

Tgt Ion:149 Resp: 51298









#90
Benzo(b)fluoranthene
Concen: 4.839 ng/ul
RT: 24.302 min Scan# 31
Delta R.T. -0.011 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:252 Resp: 63450
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.6
125 6.6 8.5 12.7#



#91
Benzo(k)fluoranthene
Concen: 4.830 ng/ul
RT: 24.378 min Scan# 3556
Delta R.T. -0.005 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:252 Resp: 58556
Ion Ratio Lower Upper
252 100
253 21.5 17.3 25.9
125 6.1 7.8 11.6#









#93
Benzo(a)pyrene
Concen: 4.836 ng/ul
RT: 25.247 min Scan# 31
Delta R.T. -0.005 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:252 Resp: 60008
Ion Ratio Lower Upper
252 100
253 20.0 17.4 26.2
125 7.3 9.1 13.7#



#94
Indeno(1,2,3-cd)pyrene
Concen: 1.638 ng/ul
RT: 29.412 min Scan# 4413
Delta R.T. -0.011 min
Lab File: BG052214.D
Acq: 31 Jan 2022 13:32

Tgt Ion:276 Resp: 23266
Ion Ratio Lower Upper
276 100
138 10.6 15.5 23.3#
277 19.8 20.5 30.7#









#95
 Dibenzo(a,h)anthracene
 Concen: 2.054 ng/ul
 RT: 29.489 min Scan# 4428
 Delta R.T. -0.011 min
 Lab File: BG052214.D
 Acq: 31 Jan 2022 13:32

Instrument :
 BNA_G
 ClientSampleId :
 SSTD005449

Tgt Ion:278 Resp: 24675
 Ion Ratio Lower Upper
 278 100
 139 9.2 12.2 18.4#
 279 19.4 19.5 29.3#



#96
 Benzo(g,h,i)perylene
 Concen: 2.181 ng/ul
 RT: 30.664 min Scan# 4626
 Delta R.T. -0.028 min
 Lab File: BG052214.D
 Acq: 31 Jan 2022 13:32

Tgt Ion:276 Resp: 26045
 Ion Ratio Lower Upper
 276 100
 138 10.5 16.6 24.8#
 277 22.4 17.6 26.4

