

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG013122\
 Data File : BG052219.D
 Acq On : 31 Jan 2022 16:48
 Operator : CG/JU
 Sample : SSTD16054
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD160405

Quant Time: Jan 31 17:52:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG013122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 31 15:44:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.237	152	46003	20.000	ng/ul	0.00
20) Naphthalene-d8	11.080	136	183947	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.886	164	124075	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.741	188	1811324	20.000	ng/ul	0.11
79) Chrysene-d12	21.959	240	246425	20.000	ng/ul	# 0.01
88) Perylene-d12	25.443	264	257426	20.000	ng/ul	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.561	96	69227	53.772	ng/uL	0.00
4) Pyridine-d5	3.989	84	555882	138.899	ng/ul	0.00
7) Phenol-d5	7.397	99	659225	146.414	ng/ul	0.02
9) Bis-(2-Chloroethyl)eth...	7.555	67	385314	134.810	ng/ul	0.01
11) 2-Chlorophenol-d4	7.772	132	467331	156.769	ng/ul	0.00
15) 4-Methylphenol-d8	8.965	113	497604	142.171	ng/ul	0.03
21) Nitrobenzene-d5	9.423	128	239963	165.092	ng/ul	0.01
24) 2-Nitrophenol-d4	10.152	143	272243	169.354	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.704	165	522194	166.190	ng/ul	0.01
31) 4-Chloroaniline-d4	11.215	131	637249	145.629	ng/ul	0.01
46) Dimethylphthalate-d6	14.287	166	1414957	137.967	ng/ul	0.02
49) Acenaphthylene-d8	14.587	160	1760390	142.100	ng/ul	0.00
54) 4-Nitrophenol-d4	15.098	143	248680	145.031	ng/ul	0.04
60) Fluorene-d10	15.879	176	1241615	138.856	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.008	200	299440	26.205	ng/ul	0.03
73) Anthracene-d10	17.741	188	1811629	21.084	ng/ul	0.01
81) Pyrene-d10	20.021	212	1999811	141.008	ng/ul	0.01
92) Benzo(a)pyrene-d12	25.225	264	2120960	156.505	ng/ul	0.04
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.602	88	73709	49.621	ng/uL#	95
5) Pyridine	4.013	79	558410	133.182	ng/ul	99
6) Benzaldehyde	7.361	77	190643	63.652	ng/ul	96
8) Phenol	7.426	94	640623	139.399	ng/ul	92
10) Bis(2-Chloroethyl)ether	7.649	93	471860	140.266	ng/ul	97
12) 2-Chlorophenol	7.808	128	477859	157.575	ng/ul	98
13) 2-Methylphenol	8.689	108	492792	147.460	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.771	45	655060	127.664	ng/ul#	95
16) Acetophenone	9.082	105	738286	133.651	ng/ul	97
17) N-Nitroso-di-n-propyla...	9.088	70	428382	126.938	ng/ul	98
18) 4-Methylphenol	9.041	108	508089	140.356	ng/ul	91
19) Hexachloroethane	9.341	117	196575	150.141	ng/ul#	75
22) Nitrobenzene	9.470	77	632868	140.377	ng/ul#	97
23) Isophorone	10.017	82	1180556	138.507	ng/ul	95
25) 2-Nitrophenol	10.187	139	284657	161.123	ng/ul	96
26) 2,4-Dimethylphenol	10.240	107	575809	150.207	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.475	93	628875	143.302	ng/ul	98
29) 2,4-Dichlorophenol	10.733	162	477529	154.745	ng/ul	96
30) Naphthalene	11.139	128	1487628	149.244	ng/ul	99
32) 4-Chloroaniline	11.244	127	639720	143.453	ng/ul	98
33) Hexachlorobutadiene	11.420	225	387254	157.431	ng/ul	99
34) Caprolactam	12.049	113	96807	80.604	ng/ul	91
35) 4-Chloro-3-methylphenol	12.366	107	520550	146.100	ng/ul	95
36) 2-Methylnaphthalene	12.730	142	1049909	147.871	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG013122\
 Data File : BG052219.D
 Acq On : 31 Jan 2022 16:48
 Operator : CG/JU
 Sample : SSTD16054
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD160405

Quant Time: Jan 31 17:52:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG013122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 31 15:44:22 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.948	142	1054477	144.710	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	13.095	216	679862	158.451	ng/ul	97
40) Hexachlorocyclopentadiene	13.071	237	465689	158.176	ng/ul	97
41) 2,4,6-Trichlorophenol	13.330	196	444108	163.764	ng/ul	95
42) 2,4,5-Trichlorophenol	13.412	196	467382	159.875	ng/ul	99
43) 1,1'-Biphenyl	13.729	154	1388315	144.940	ng/ul#	97
44) 2-Chloronaphthalene	13.776	162	1090363	146.985	ng/ul	96
45) 2-Nitroaniline	13.976	65	404024	142.113	ng/ul	93
47) Dimethylphthalate	14.340	163	1331653	131.683	ng/ul#	98
48) 2,6-Dinitrotoluene	14.463	165	325528	152.674	ng/ul	91
50) Acenaphthylene	14.622	152	1679050	137.638	ng/ul	96
51) 3-Nitroaniline	14.792	138	193229	100.185	ng/ul	94
52) Acenaphthene	14.957	153	1164680	143.591	ng/ul	93
53) 2,4-Dinitrophenol	14.998	184	208910	166.186	ng/ul	91
55) 4-Nitrophenol	15.115	109	261237	132.766	ng/ul	85
56) Dibenzofuran	15.292	168	1597304	135.593	ng/ul	93
57) 2,4-Dinitrotoluene	15.251	165	444396	144.160	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.515	232	426700	157.362	ng/ul#	89
59) Diethylphthalate	15.697	149	1363330	130.672	ng/ul	94
61) Fluorene	15.938	166	1335029	136.000	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.920	204	759236	139.375	ng/ul	96
63) 4-Nitroaniline	15.961	138	155759	78.602	ng/ul	88
66) 4,6-Dinitro-2-methylph...	16.091	198	334	0.031	ng/ul#	81
67) N-Nitrosodiphenylamine	16.132	169	1170887	24.057	ng/ul	98
68) 4-Bromophenyl-phenylether	16.948	248	7836	0.370	ng/ul#	1
69) Hexachlorobenzene	16.948	284	580814	24.697	ng/ul#	90
70) Atrazine	17.083	200	503288	22.288	ng/ul	97
71) Pentachlorophenol	17.289	266	345969	26.342	ng/ul	96
72) Phenanthrene	17.782	178	1956008	20.609	ng/ul#	92
74) Anthracene	17.782	178	1954383	20.571	ng/ul	94
75) 1,2,3,4-Tetrachloroben...	13.700	216	763037	28.049	ng/ul	97
76) Pentachlorobenzene	15.215	250	687352	25.961	ng/ul	100
77) Carbazole	18.041	167	1800039	21.027	ng/ul	95
78) Di-n-butylphthalate	18.576	149	2017359	19.935	ng/ul#	95
80) Fluoranthene	19.686	202	2325269	141.877	ng/ul#	93
82) Pyrene	20.050	202	2170247	134.495	ng/ul#	93
83) Butylbenzylphthalate	20.908	149	926532	164.992	ng/ul#	95
84) 3,3'-Dichlorobenzidine	21.836	252	613515	107.852	ng/ul#	95
85) Benzo(a)anthracene	21.942	228	2354435	145.281	ng/ul	93
86) Bis(2-ethylhexyl)phtha...	21.812	149	1341682	163.367	ng/ul#	93
87) Chrysene	22.018	228	2252331	144.332	ng/ul	93
89) Di-n-octyl phthalate	23.116	149	2166228	152.334	ng/ul	100
90) Benzo(b)fluoranthene	24.344	252	2698551	158.180	ng/ul#	94
91) Benzo(k)fluoranthene	24.421	252	2311231	146.525	ng/ul#	93
93) Benzo(a)pyrene	25.313	252	2463517	152.583	ng/ul#	93
94) Indeno(1,2,3-cd)pyrene	29.508	276	2976715	161.062	ng/ul#	92
95) Dibenzo(a,h)anthracene	29.596	278	2466791	157.798	ng/ul#	93
96) Benzo(g,h,i)perylene	30.788	276	1944638	125.140	ng/ul#	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG013122\
 Data File : BG052219.D
 Acq On : 31 Jan 2022 16:48
 Operator : CG/JU
 Sample : SSTD16054
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD160405

Quant Time: Jan 31 17:52:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG013122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 31 15:44:22 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 49.621 ng/uL
RT: 3.602 min Scan# 19
Delta R.T. 0.003 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion: 88 Resp: 73709
Ion Ratio Lower Upper
88 100
43 35.2 23.0 34.4#
58 76.5 62.4 93.6



#5
Pyridine
Concen: 133.182 ng/uL
RT: 4.013 min Scan# 89
Delta R.T. 0.003 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion: 79 Resp: 558410
Ion Ratio Lower Upper
79 100
52 76.2 60.8 91.2
51 36.8 30.5 45.7









#6
Benzaldehyde
Concen: 63.652 ng/ul
RT: 7.361 min Scan# 61
Delta R.T. 0.003 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion: 77 Resp: 190643
Ion Ratio Lower Upper
77 100
105 87.8 70.4 105.6
106 83.2 61.2 91.8



#8
Phenol
Concen: 139.399 ng/ul
RT: 7.426 min Scan# 670
Delta R.T. 0.020 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion: 94 Resp: 640623
Ion Ratio Lower Upper
94 100
65 34.8 25.1 37.7
66 45.1 31.6 47.4









#10
Bis(2-Chloroethyl)ether
Concen: 140.266 ng/ul
RT: 7.649 min Scan# 706
Delta R.T. 0.014 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion: 93 Resp: 471860
Ion Ratio Lower Upper
93 100
63 82.7 68.6 103.0
95 33.2 26.1 39.1



#12
2-Chlorophenol
Concen: 157.575 ng/ul
RT: 7.808 min Scan# 735
Delta R.T. 0.008 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:128 Resp: 477859
Ion Ratio Lower Upper
128 100
64 54.5 45.3 67.9
130 33.1 27.1 40.7









#13
2-Methylphenol
Concen: 147.460 ng/ul
RT: 8.689 min Scan# 885
Delta R.T. 0.014 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:108 Resp: 492792
Ion Ratio Lower Upper
108 100
107 88.9 71.8 107.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 127.664 ng/ul
RT: 8.771 min Scan# 899
Delta R.T. 0.008 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion: 45 Resp: 655060
Ion Ratio Lower Upper
45 100
77 16.2 10.7 16.1#
79 11.3 8.5 12.7









#16
Acetophenone
Concen: 133.651 ng/ul
RT: 9.082 min Scan# 91
Delta R.T. 0.020 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:105 Resp: 738286
Ion Ratio Lower Upper
105 100
77 86.8 67.3 100.9
51 31.5 24.0 36.0



#17
N-Nitroso-di-n-propylamine
Concen: 126.938 ng/ul
RT: 9.088 min Scan# 953
Delta R.T. 0.050 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion: 70 Resp: 428382
Ion Ratio Lower Upper
70 100
42 58.3 46.0 69.0
101 9.1 7.0 10.4
130 18.4 13.4 20.2









#18
4-Methylphenol
Concen: 140.356 ng/ul
RT: 9.041 min Scan# 945
Delta R.T. 0.038 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:108 Resp: 508089
Ion Ratio Lower Upper
108 100
107 109.0 95.0 142.4



#19
Hexachloroethane
Concen: 150.141 ng/ul
RT: 9.341 min Scan# 996
Delta R.T. 0.003 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:117 Resp: 196575
Ion Ratio Lower Upper
117 100
201 124.2 77.2 115.8#
199 77.8 49.9 74.9#









#22
Nitrobenzene
Concen: 140.377 ng/ul
RT: 9.470 min Scan# 1015
Delta R.T. 0.020 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion: 77 Resp: 632868
Ion Ratio Lower Upper
77 100
123 38.4 30.1 45.1
65 15.2 9.6 14.4#



#23
Isophorone
Concen: 138.507 ng/ul
RT: 10.017 min Scan# 1111
Delta R.T. 0.044 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion: 82 Resp: 1180556
Ion Ratio Lower Upper
82 100
95 8.1 5.6 8.4
138 17.0 11.8 17.8









#25
2-Nitrophenol
Concen: 161.123 ng/ul
RT: 10.187 min Scan# 1138
Delta R.T. 0.014 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:139 Resp: 284657
Ion Ratio Lower Upper
139 100
65 61.4 48.0 72.0
109 39.7 28.6 42.8



#26
2,4-Dimethylphenol
Concen: 150.207 ng/ul
RT: 10.240 min Scan# 1149
Delta R.T. 0.020 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:107 Resp: 575809
Ion Ratio Lower Upper
107 100
121 48.4 39.3 58.9
122 80.0 63.7 95.5





#27

Bis(2-Chloroethoxy)methane

Concen: 143.302 ng/ul

RT: 10.475 min Scan# 1189

Delta R.T. 0.020 min

Lab File: BG052219.D

Acq: 31 Jan 2022 16:48

Instrument :

BNA_G

ClientSampleId :

SSTD160405





Tgt Ion: 93 Resp: 628875

Ion Ratio Lower Upper

93 100

95 34.6 26.5 39.7

123 10.7 8.1 12.1



#29

2,4-Dichlorophenol

Concen: 154.745 ng/ul

RT: 10.733 min Scan# 1233

Delta R.T. 0.014 min

Lab File: BG052219.D

Acq: 31 Jan 2022 16:48




Tgt Ion:162 Resp: 477529


Ion Ratio Lower Upper

162 100

164 66.1 49.4 74.0

98 37.7 30.4 45.6





#30

Naphthalene

Concen: 149.244 ng/ul

RT: 11.139 min Scan# 11

Delta R.T. 0.014 min

Lab File: BG052219.D

Acq: 31 Jan 2022 16:48

Instrument :

BNA_G

ClientSampleId :

SSTD160405





Tgt Ion:128 Resp: 1487628

Ion Ratio Lower Upper

128 100

129 11.8 8.5 12.7

127 14.7 11.8 17.6



#32

4-Chloroaniline

Concen: 143.453 ng/ul

RT: 11.244 min Scan# 1320

Delta R.T. 0.020 min

Lab File: BG052219.D




Acq: 31 Jan 2022 16:48


Tgt Ion:127 Resp: 639720

Ion Ratio Lower Upper

127 100

129 33.9 26.2 39.2









#33
Hexachlorobutadiene
Concen: 157.431 ng/ul
RT: 11.420 min Scan# 11
Delta R.T. 0.008 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:225 Resp: 387254
Ion Ratio Lower Upper
225 100
223 62.7 50.6 75.8
227 64.0 49.8 74.8



#34
Caprolactam
Concen: 80.604 ng/ul
RT: 12.049 min Scan# 1457
Delta R.T. 0.079 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:113 Resp: 96807
Ion Ratio Lower Upper
113 100
55 164.6 147.0 220.6
56 133.3 109.2 163.8









#35
4-Chloro-3-methylphenol
Concen: 146.100 ng/ul
RT: 12.366 min Scan# 11
Delta R.T. 0.032 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:107 Resp: 520550
Ion Ratio Lower Upper
107 100
144 25.2 19.8 29.6
142 75.6 56.2 84.4



#36
2-Methylnaphthalene
Concen: 147.871 ng/ul
RT: 12.730 min Scan# 1573
Delta R.T. 0.008 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:142 Resp: 1049909
Ion Ratio Lower Upper
142 100
141 90.4 73.3 109.9









#37
1-Methylnaphthalene
Concen: 144.710 ng/ul
RT: 12.948 min Scan# 1
Delta R.T. 0.008 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:142 Resp: 1054477
Ion Ratio Lower Upper
142 100
141 94.5 77.8 116.8
116 4.6 3.3 4.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 158.451 ng/ul
RT: 13.095 min Scan# 1635
Delta R.T. 0.008 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:216 Resp: 679862
Ion Ratio Lower Upper
216 100
214 78.7 61.0 91.4
179 20.6 16.2 24.2
108 15.2 13.1 19.7









#40
Hexachlorocyclopentadiene
Concen: 158.176 ng/ul
RT: 13.071 min Scan# 1631
Delta R.T. 0.003 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:237 Resp: 465689
Ion Ratio Lower Upper
237 100
235 64.5 49.4 74.0
272 13.0 11.1 16.7



#41
2,4,6-Trichlorophenol
Concen: 163.764 ng/ul
RT: 13.330 min Scan# 1675
Delta R.T. 0.008 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:196 Resp: 444108
Ion Ratio Lower Upper
196 100
198 97.2 74.3 111.5
200 34.1 23.8 35.6









#42
2,4,5-Trichlorophenol
Concen: 159.875 ng/ul
RT: 13.412 min Scan# 1
Delta R.T. 0.020 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:196 Resp: 467382
Ion Ratio Lower Upper
196 100
198 95.3 77.0 115.4
200 33.0 25.4 38.2



#43
1,1'-Biphenyl
Concen: 144.940 ng/ul
RT: 13.729 min Scan# 1743
Delta R.T. 0.014 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:154 Resp: 1388315
Ion Ratio Lower Upper
154 100
153 44.2 34.5 51.7
76 13.0 13.1 19.7#





#44

2-Chloronaphthalene

Concen: 146.985 ng/ul

RT: 13.776 min Scan# 1

Delta R.T. 0.014 min

Lab File: BG052219.D

Acq: 31 Jan 2022 16:48

Instrument :

BNA_G

ClientSampleId :

SSTD160405





Tgt Ion:162 Resp: 1090363

Ion Ratio Lower Upper

162 100

164 35.5 26.1 39.1

127 40.6 31.0 46.4



#45

2-Nitroaniline

Concen: 142.113 ng/ul

RT: 13.976 min Scan# 1785

Delta R.T. 0.026 min

Lab File: BG052219.D

Acq: 31 Jan 2022 16:48




Tgt Ion: 65 Resp: 404024


Ion Ratio Lower Upper

65 100

92 65.7 53.0 79.4

138 94.9 67.4 101.0









#47
Dimethylphthalate
Concen: 131.683 ng/ul
RT: 14.340 min Scan# 1847
Delta R.T. 0.026 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:163 Resp: 1331653
Ion Ratio Lower Upper
163 100
194 4.6 3.0 4.4#
164 10.9 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 152.674 ng/ul
RT: 14.463 min Scan# 1868
Delta R.T. 0.020 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:165 Resp: 325528
Ion Ratio Lower Upper
165 100
89 63.6 57.4 86.0
121 21.0 18.8 28.2









#50
Acenaphthylene
Concen: 137.638 ng/ul
RT: 14.622 min Scan# 1895
Delta R.T. 0.020 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:152 Resp: 1679050
Ion Ratio Lower Upper
152 100
151 23.0 17.4 26.0
153 15.1 10.3 15.5



#51
3-Nitroaniline
Concen: 100.185 ng/ul
RT: 14.792 min Scan# 1924
Delta R.T. 0.020 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:138 Resp: 193229
Ion Ratio Lower Upper
138 100
108 9.9 8.8 13.2
92 130.0 98.0 147.0









#52
Acenaphthene
Concen: 143.591 ng/ul
RT: 14.957 min Scan# 1952
Delta R.T. 0.014 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:153 Resp: 1164680
Ion Ratio Lower Upper
153 100
152 49.9 36.5 54.7
154 89.3 65.8 98.6



#53
2,4-Dinitrophenol
Concen: 166.186 ng/ul
RT: 14.998 min Scan# 1959
Delta R.T. 0.020 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:184 Resp: 208910
Ion Ratio Lower Upper
184 100
63 69.5 66.2 99.2
154 66.3 53.6 80.4









#55
4-Nitrophenol
Concen: 132.766 ng/ul
RT: 15.115 min Scan# 1979
Delta R.T. 0.038 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:109 Resp: 261237
Ion Ratio Lower Upper
109 100
139 99.0 88.7 133.1
65 120.6 113.6 170.4



#56
Dibenzofuran
Concen: 135.593 ng/ul
RT: 15.292 min Scan# 2009
Delta R.T. 0.014 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:168 Resp: 1597304
Ion Ratio Lower Upper
168 100
139 43.1 31.2 46.8









#57
2,4-Dinitrotoluene
Concen: 144.160 ng/ul
RT: 15.251 min Scan# 2002
Delta R.T. 0.026 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:165 Resp: 444396
Ion Ratio Lower Upper
165 100
63 50.6 42.1 63.1
182 2.6 1.8 2.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 157.362 ng/ul
RT: 15.515 min Scan# 2047
Delta R.T. 0.014 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:232 Resp: 426700
Ion Ratio Lower Upper
232 100
131 39.9 38.9 58.3
130 2.4 2.4 3.6#
166 27.4 25.8 38.8









#59
Diethylphthalate
Concen: 130.672 ng/ul
RT: 15.697 min Scan# 2074
Delta R.T. 0.026 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:149 Resp: 1363330
Ion Ratio Lower Upper
149 100
177 24.8 16.8 25.2
150 13.7 10.1 15.1



#61
Fluorene
Concen: 136.000 ng/ul
RT: 15.938 min Scan# 2119
Delta R.T. 0.014 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:166 Resp: 1335029
Ion Ratio Lower Upper
166 100
165 99.3 81.0 121.6
167 15.0 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 139.375 ng/ul
RT: 15.920 min Scan# 2116
Delta R.T. 0.014 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:204 Resp: 759236
Ion Ratio Lower Upper
204 100
206 35.2 27.4 41.2
141 56.9 48.6 73.0



#63
4-Nitroaniline
Concen: 78.602 ng/ul
RT: 15.961 min Scan# 2123
Delta R.T. 0.032 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:138 Resp: 155759
Ion Ratio Lower Upper
138 100
92 54.7 49.3 73.9
108 104.7 72.6 108.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 0.031 ng/ul
 RT: 16.091 min Scan# 2145
 Delta R.T. 0.097 min
 Lab File: BG052219.D
 Acq: 31 Jan 2022 16:48




Instrument :
 BNA_G
 ClientSampleId :
 SSTD160405


Tgt Ion:198 Resp: 334
 Ion Ratio Lower Upper
 198 100
 182 0.0 3.8 5.8#
 77 39.8 23.2 34.8#



#67
 N-Nitrosodiphenylamine
 Concen: 24.057 ng/ul
 RT: 16.132 min Scan# 2152
 Delta R.T. 0.014 min
 Lab File: BG052219.D
 Acq: 31 Jan 2022 16:48

Tgt Ion:169 Resp: 1170887
 Ion Ratio Lower Upper
 169 100
 168 71.4 58.5 87.7
 167 39.5 31.3 46.9









#68
4-Bromophenyl-phenylether
Concen: 0.370 ng/ul
RT: 16.948 min Scan# 21
Delta R.T. 0.144 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:248 Resp: 7836
Ion Ratio Lower Upper
248 100
250 152.2 75.5 113.3#
141 276.6 67.6 101.4#



#69
Hexachlorobenzene
Concen: 24.697 ng/ul
RT: 16.948 min Scan# 2291
Delta R.T. 0.014 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:284 Resp: 580814
Ion Ratio Lower Upper
284 100
142 28.3 31.9 47.9#
249 32.3 25.8 38.6





#70

Atrazine

Concen: 22.288 ng/ul

RT: 17.083 min Scan# 21

Delta R.T. 0.020 min

Lab File: BG052219.D

Acq: 31 Jan 2022 16:48

Instrument :

BNA_G

ClientSampleId :

SSTD160405





Tgt Ion:200 Resp: 503288

Ion Ratio Lower Upper

200 100

173 25.1 20.2 30.2

215 50.2 38.1 57.1



#71

Pentachlorophenol

Concen: 26.342 ng/ul

RT: 17.289 min Scan# 2349

Delta R.T. 0.008 min

Lab File: BG052219.D

Acq: 31 Jan 2022 16:48




Tgt Ion:266 Resp: 345969


Ion Ratio Lower Upper

266 100

264 64.0 54.3 81.5

268 65.8 51.0 76.6









#72
Phenanthrene
Concen: 20.609 ng/u1
RT: 17.782 min Scan# 2414
Delta R.T. 0.114 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:178 Resp: 1956008
Ion Ratio Lower Upper
178 100
179 19.1 12.3 18.5#
176 22.3 15.5 23.3



#74
Anthracene
Concen: 20.571 ng/u1
RT: 17.782 min Scan# 2433
Delta R.T. 0.020 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:178 Resp: 1954383
Ion Ratio Lower Upper
178 100
179 19.1 13.0 19.6
176 22.3 15.6 23.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 28.049 ng/uL
RT: 13.700 min Scan# 1738
Delta R.T. 0.014 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:216 Resp: 763037
Ion Ratio Lower Upper
216 100
214 78.1 61.3 91.9
218 49.8 37.0 55.6



#76
Pentachlorobenzene
Concen: 25.961 ng/uL
RT: 15.215 min Scan# 1996
Delta R.T. 0.014 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:250 Resp: 687352
Ion Ratio Lower Upper
250 100
248 63.4 50.5 75.7
252 64.4 51.2 76.8









#77
Carbazole
Concen: 21.027 ng/ul
RT: 18.041 min Scan# 2475
Delta R.T. 0.014 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:167 Resp: 1800039
Ion Ratio Lower Upper
167 100
166 24.8 17.3 25.9
139 14.2 10.6 16.0



#78
Di-n-butylphthalate
Concen: 19.935 ng/ul
RT: 18.576 min Scan# 2568
Delta R.T. 0.008 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:149 Resp: 2017359
Ion Ratio Lower Upper
149 100
150 11.5 7.5 11.3#
104 7.5 5.0 7.6









#80
Fluoranthene
Concen: 141.877 ng/ul
RT: 19.686 min Scan# 21
Delta R.T. 0.014 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:202 Resp: 2325269
Ion Ratio Lower Upper
202 100
101 8.4 9.3 13.9#
100 7.1 7.0 10.6



#82
Pyrene
Concen: 134.495 ng/ul
RT: 20.050 min Scan# 2819
Delta R.T. 0.014 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:202 Resp: 2170247
Ion Ratio Lower Upper
202 100
101 10.0 10.7 16.1#
100 9.2 9.0 13.6





#83

Butylbenzylphthalate

Concen: 164.992 ng/ul

RT: 20.908 min Scan# 2965

Delta R.T. 0.003 min

Lab File: BG052219.D

Acq: 31 Jan 2022 16:48

Instrument :

BNA_G

ClientSampleId :

SSTD160405





Tgt Ion:149 Resp: 926532

Ion Ratio Lower Upper

149 100

91 78.6 65.2 97.8

206 25.6 16.2 24.2#



#84

3,3'-Dichlorobenzidine

Concen: 107.852 ng/ul

RT: 21.836 min Scan# 3123

Delta R.T. 0.008 min

Lab File: BG052219.D

Acq: 31 Jan 2022 16:48




Tgt Ion:252 Resp: 613515


Ion Ratio Lower Upper

252 100

254 66.9 50.8 76.2

126 7.1 9.0 13.4#









#85
Benzo(a)anthracene
Concen: 145.281 ng/ul
RT: 21.942 min Scan# 3119
Delta R.T. 0.014 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:228 Resp: 2354435
Ion Ratio Lower Upper
228 100
229 23.3 15.7 23.5
226 31.0 22.3 33.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 163.367 ng/ul
RT: 21.812 min Scan# 3119
Delta R.T. 0.008 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:149 Resp: 1341682
Ion Ratio Lower Upper
149 100
167 29.8 21.1 31.7
279 5.7 2.7 4.1#









#87
Chrysene
Concen: 144.332 ng/ul
RT: 22.018 min Scan# 3151
Delta R.T. 0.020 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:228 Resp: 2252331
Ion Ratio Lower Upper
228 100
226 34.0 24.8 37.2
229 23.6 15.8 23.6



#89
Di-n-octyl phthalate
Concen: 152.334 ng/ul
RT: 23.116 min Scan# 3341
Delta R.T. 0.014 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:149 Resp: 2166228









#90
Benzo(b)fluoranthene
Concen: 158.180 ng/ul
RT: 24.344 min Scan# 3550
Delta R.T. 0.032 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48




Instrument :
BNA_G
ClientSampleId :
SSTD160405


Tgt Ion:252 Resp: 2698551
Ion Ratio Lower Upper
252 100
253 23.7 17.8 26.6
125 6.6 8.5 12.7#



#91
Benzo(k)fluoranthene
Concen: 146.525 ng/ul
RT: 24.421 min Scan# 3563
Delta R.T. 0.038 min
Lab File: BG052219.D
Acq: 31 Jan 2022 16:48

Tgt Ion:252 Resp: 2311231
Ion Ratio Lower Upper
252 100
253 24.8 17.3 25.9
125 6.5 7.8 11.6#





#93

Benzo(a)pyrene

Concen: 152.583 ng/uI

RT: 25.313 min Scan# 31

Delta R.T. 0.061 min

Lab File: BG052219.D

Acq: 31 Jan 2022 16:48

Instrument :

BNA_G

ClientSampleId :

SSTD160405





Tgt Ion:252 Resp: 2463517

Ion Ratio Lower Upper

252 100

253 23.7 17.4 26.2

125 6.8 9.1 13.7#



#94

Indeno(1,2,3-cd)pyrene

Concen: 161.062 ng/uI

RT: 29.508 min Scan# 4429

Delta R.T. 0.085 min

Lab File: BG052219.D

Acq: 31 Jan 2022 16:48




Tgt Ion:276 Resp: 2976715


Ion Ratio Lower Upper

276 100

138 11.7 15.5 23.3#

277 24.8 20.5 30.7





#95

Dibenzo(a,h)anthracene

Concen: 157.798 ng/ul

RT: 29.596 min Scan# 4444

Delta R.T. 0.097 min

Lab File: BG052219.D

Acq: 31 Jan 2022 16:48

Instrument :

BNA_G

ClientSampleId :

SSTD160405





Tgt Ion:278 Resp: 2466791

Ion Ratio Lower Upper

278 100

139 9.4 12.2 18.4#

279 25.8 19.5 29.3



#96

Benzo(g,h,i)perylene

Concen: 125.140 ng/ul

RT: 30.788 min Scan# 4647

Delta R.T. 0.097 min

Lab File: BG052219.D

Acq: 31 Jan 2022 16:48

Tgt Ion:276 Resp: 1944638

Ion Ratio Lower Upper

276 100

138 12.3 16.6 24.8#

277 23.9 17.6 26.4

