

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020124\
 Data File : BG060513.D
 Acq On : 31 Jan 2024 17:54
 Operator : MA/JU
 Sample : P1265-10
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SVOC-GPC2-BLANK

Quant Time: Jan 31 22:08:44 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG012624.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 31 22:04:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	119691	20.000	ng/ul	0.00
20) Naphthalene-d8	10.736	136	557584	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.573	164	510968	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.317	188	1348502	20.000	ng/ul	0.00
79) Chrysene-d12	21.582	240	1286211	20.000	ng/ul	0.00
88) Perylene-d12	24.749	264	1422421	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.398	96	70	0.029	ng/uL	0.03
4) Pyridine-d5	3.768	84	175	0.023	ng/ul	-0.01
7) Phenol-d5	0.000	99	0	0.000	ng/ul	
9) Bis-(2-Chloroethyl)eth...	7.235	67	87	0.012	ng/ul	-0.03
11) 2-Chlorophenol-d4	0.000	132	0	0.000	ng/ul	
15) 4-Methylphenol-d8	0.000	113	0	0.000	ng/ul	
21) Nitrobenzene-d5	0.000	128	0	0.000	ng/ul	
24) 2-Nitrophenol-d4	9.755	143	68	0.014	ng/ul	-0.07
28) 2,4-Dichlorophenol-d3	0.000	165	0	0.000	ng/ul	
31) 4-Chloroaniline-d4	10.742	131	707	0.054	ng/ul	-0.13
46) Dimethylphthalate-d6	0.000	166	0	0.000	ng/ul	
49) Acenaphthylene-d8	0.000	160	0	0.000	ng/ul	
54) 4-Nitrophenol-d4	0.000	143	0	0.000	ng/ul	
60) Fluorene-d10	15.566	176	229	0.006	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	0.000	200	0	0.000	ng/ul	
73) Anthracene-d10	17.317	188	1348502	21.095	ng/ul	-0.10
81) Pyrene-d10	19.702	212	549	0.007	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.514	264	329	0.004	ng/ul	-0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020124\
Data File : BG060513.D
Acq On : 31 Jan 2024 17:54
Operator : MA/JU
Sample : P1265-10
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SVOC-GPC2-BLANK

Quant Time: Jan 31 22:08:44 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG012624.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jan 31 22:04:13 2024
Response via : Initial Calibration

