

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG020217\
 Data File : BG025751.D
 Acq On : 2 Feb 2017 9:49
 Operator : SJ/MA
 Sample : PB96576BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK76

Quant Time: Feb 02 14:26:41 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG011817.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 02 14:22:02 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	85247	20.00	ng/ul	0.00
18) Naphthalene-d8	10.99	136	364216	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.79	164	280315	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.54	188	601940	20.00	ng/ul	0.00
75) Chrysene-d12	21.83	240	700399	20.00	ng/ul	0.00
83) Perylene-d12	25.21	264	707139	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.51	96	11916	6.99	ng/uL	0.00
5) Phenol-d5	7.33	99	256878	35.16	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.48	67	187616	39.76	ng/ul	0.00
9) 2-Chlorophenol-d4	7.70	132	188136	36.51	ng/ul	0.00
13) 4-Methylphenol-d8	8.88	113	216267	34.82	ng/ul	0.00
19) Nitrobenzene-d5	9.35	128	92114	35.68	ng/ul	0.00
22) 2-Nitrophenol-d4	10.08	143	114119	37.05	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.62	165	200828	33.05	ng/ul	0.00
29) 4-Chloroaniline-d4	11.14	131	256894	37.82	ng/ul	0.00
43) Dimethylphthalate-d6	14.20	166	685075	34.61	ng/ul	0.00
46) Acenaphthylene-d8	14.49	160	811098	37.93	ng/ul	0.00
51) 4-Nitrophenol-d4	15.05	143	86889	32.03	ng/ul	0.01
57) Fluorene-d10	15.79	176	610662	32.73	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.95	200	81282	21.89	ng/ul	0.01
70) Anthracene-d10	17.64	188	971926	37.92	ng/ul	0.00
76) Pyrene-d10	19.92	212	1135999	38.98	ng/ul	0.00
87) Benzo(a)pyrene-d12	24.98	264	1104148	37.15	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG020217\
Data File : BG025751.D
Acq On : 2 Feb 2017 9:49
Operator : SJ/MA
Sample : PB96576BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK76

Quant Time: Feb 02 14:26:41 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG011817.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 02 14:22:02 2017
Response via : Initial Calibration

