

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020224\
 Data File : BG060537.D
 Acq On : 1 Feb 2024 17:38
 Operator : MA/JU
 Sample : PB158757BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB158757BL

Quant Time: Feb 02 00:40:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 09 04:07:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.922	152	128410	20.000	ng	0.00
21) Naphthalene-d8	10.725	136	492162	20.000	ng	0.00
39) Acenaphthene-d10	14.556	164	372093	20.000	ng	-0.01
64) Phenanthrene-d10	17.306	188	1059145	20.000	ng	0.00
76) Chrysene-d12	21.571	240	1071735	20.000	ng	0.00
86) Perylene-d12	24.732	264	1221622	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.496	112	993945	127.313	ng	0.00
7) Phenol-d6	7.088	99	1395894	120.747	ng	0.00
23) Nitrobenzene-d5	9.080	82	876570	84.102	ng	-0.01
42) 2,4,6-Tribromophenol	16.048	330	825552	150.798	ng	0.00
45) 2-Fluorobiphenyl	13.181	172	2302152	86.019	ng	0.00
79) Terphenyl-d14	19.926	244	4360244	120.496	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020224\
Data File : BG060537.D
Acq On : 1 Feb 2024 17:38
Operator : MA/JU
Sample : PB158757BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB158757BL

Quant Time: Feb 02 00:40:42 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 09 04:07:12 2024
Response via : Initial Calibration

