

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG020421\
 Data File : BG048114.D
 Acq On : 4 Feb 2021 14:04
 Operator : CG/JU
 Sample : PB134078BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK078

Quant Time: Feb 04 14:40:44 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG012921.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 04 12:55:42 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.11	152	49464	20.00	ng/ul	0.00
20) Naphthalene-d8	10.92	136	188333	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.73	164	135845	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.47	188	346243	20.00	ng/ul	0.00
79) Chrysene-d12	21.75	240	396428	20.00	ng/ul	0.00
88) Perylene-d12	25.01	264	384525	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.52	96	8732	6.64	ng/uL	0.00
4) Pyridine-d5	3.93	84	136789	35.61	ng/ul	0.00
7) Phenol-d5	7.26	99	169221	36.69	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	7.43	67	106939	38.49	ng/ul	0.00
11) 2-Chlorophenol-d4	7.64	132	124772	37.18	ng/ul	0.00
15) 4-Methylphenol-d8	8.80	113	133390	36.64	ng/ul	0.00
21) Nitrobenzene-d5	9.27	128	62565	39.50	ng/ul	0.00
24) 2-Nitrophenol-d4	9.99	143	75753	38.92	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.54	165	131861	35.81	ng/ul	0.00
31) 4-Chloroaniline-d4	11.05	131	168030	35.62	ng/ul	0.00
46) Dimethylphthalate-d6	14.13	166	463348	40.84	ng/ul	0.00
49) Acenaphthylene-d8	14.43	160	487072	37.31	ng/ul	0.00
54) 4-Nitrophenol-d4	14.91	143	80393	39.51	ng/ul	0.00
60) Fluorene-d10	15.72	176	393879	38.74	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.83	200	87876	35.49	ng/ul	0.00
73) Anthracene-d10	17.57	188	644391	38.54	ng/ul	0.00
81) Pyrene-d10	19.85	212	832636	37.67	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.80	264	878901	41.81	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG020421\
Data File : BG048114.D
Acq On : 4 Feb 2021 14:04
Operator : CG/JU
Sample : PB134078BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK078

Quant Time: Feb 04 14:40:44 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG012921.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 04 12:55:42 2021
Response via : Initial Calibration

