


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020421\
 Data File : BG048110.D
 Acq On : 4 Feb 2021 10:57
 Operator : CG/JU
 Sample : DFTPP81
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP081

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG012921.M
 Title : SVOA CALIBRATION
 Last Update : Thu Feb 04 12:55:42 2021



AutoFind: Scans 2478, 2479, 2480; Background Corrected with Scan 2473

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.8	34573	PASS
68	69	0.00	2	0.7	284	PASS
69	198	0.00	100	41.3	41077	PASS
70	69	0.00	2	0.2	96	PASS
127	198	10	80	43.6	43293	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	99349	PASS
199	198	5	9	7.1	7068	PASS
275	198	10	60	31.1	30866	PASS
365	198	1	100	4.3	4297	PASS
441	198	0.01	100	10.8	10736	PASS
442	198	50	100	68.7	68290	PASS
443	442	15	24	18.5	12665	PASS

Scan 2473 (17.867 min): BG048110.D
DFTPP81

m/z	Abundance
39.90	158.0
44.00	743.0
86.90	261.0
116.20	174.0
119.00	211.0
166.90	173.0
182.00	224.0
197.90	165.0
207.00	1151.0
207.80	303.0
221.10	151.0
281.20	252.0
324.20	176.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

m/z	Abundance
37.00	178.0
38.10	371.0
39.10	2903.0
40.00	264.0
41.10	310.0
44.00	847.0
48.20	171.0
49.00	429.0
50.10	9506.0
51.10	34488.0
52.00	1700.0
56.00	1050.0
57.00	2316.0
61.10	610.0
62.00	864.0
63.00	1799.0
64.00	158.0
65.10	888.0
69.00	39672.0
69.80	246.0
73.20	807.0
74.10	3557.0
75.00	6178.0
76.10	2626.0
77.10	44480.0
78.10	2550.0
79.00	2771.0
80.00	2564.0
81.10	3548.0
82.00	423.0
83.00	734.0
83.80	266.0
85.10	522.0
86.00	1086.0
87.00	454.0
88.10	248.0
91.00	863.0
92.00	713.0
93.00	6649.0
94.10	405.0
94.90	184.0
96.10	553.0
98.00	4110.0
99.00	2944.0
99.90	347.0
101.10	1691.0
103.20	975.0
104.00	1314.0
105.00	1069.0
106.10	596.0
107.00	13513.0
108.00	2417.0
108.80	360.0
110.00	20752.0
111.00	3394.0
111.70	345.0
112.20	373.0
113.10	159.0
116.00	1016.0
117.00	10968.0
118.00	542.0
120.00	249.0
120.50	174.0
122.00	917.0
122.90	1062.0
124.00	615.0
124.90	772.0
127.00	42376.0
128.10	3128.0
129.00	17400.0
130.00	1761.0
130.90	269.0
132.00	260.0
132.80	289.0
133.90	595.0
135.00	1459.0
135.90	603.0
137.00	646.0
140.10	212.0
141.00	2495.0
141.90	724.0
143.10	430.0
144.00	150.0
144.80	263.0
146.10	600.0
147.00	1637.0
148.10	2787.0
148.90	617.0
151.00	295.0
152.90	794.0
154.00	680.0
155.00	1761.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

156.10	1849.0
157.10	491.0
158.10	496.0
158.90	314.0
160.00	631.0
161.00	1342.0
162.10	572.0
163.90	291.0
165.00	1008.0
165.90	736.0
167.00	4282.0
168.00	2150.0
169.00	627.0
170.90	275.0
172.00	584.0
172.90	672.0
173.90	1425.0
175.00	1678.0
175.90	388.0
177.00	1160.0
179.00	3552.0
180.00	2683.0
181.00	1128.0
182.00	152.0
183.90	309.0
185.00	1670.0
186.00	12960.0
187.00	3285.0
187.90	534.0
189.10	969.0
189.90	218.0
190.90	513.0
192.00	773.0
193.10	1344.0
194.20	304.0
194.80	170.0
195.90	2663.0
197.90	86872.0
199.00	6342.0
200.10	315.0
201.60	548.0
203.00	912.0
204.00	3752.0
205.00	6619.0
206.00	26304.0
207.00	4825.0
207.90	1200.0
209.00	792.0
210.10	500.0
211.00	1456.0
215.10	320.0
215.80	441.0
217.00	7597.0
218.00	870.0
218.90	267.0
221.00	3265.0
221.80	1083.0
222.90	1952.0
224.00	14844.0
225.00	3362.0
225.90	458.0
227.00	6065.0
228.00	1241.0
229.00	1727.0
231.00	815.0
234.10	429.0
234.80	512.0
236.10	236.0
237.00	629.0
238.80	191.0
239.10	176.0
240.10	225.0
240.90	461.0
242.00	1142.0
243.10	704.0
244.00	10235.0
245.00	1590.0
246.00	2746.0
247.00	477.0
249.00	415.0
251.00	393.0
252.20	241.0
253.00	308.0
255.00	58080.0
256.00	8331.0
257.00	804.0
258.00	4148.0
258.90	668.0
259.90	151.0
265.00	1921.0
271.00	221.0
271.80	240.0
273.00	2053.0
274.00	4025.0
275.00	24680.0
276.00	2701.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

277.00	1901.0
278.00	463.0
281.00	249.0
283.00	282.0
284.00	434.0
285.00	336.0
288.90	180.0
292.80	376.0
295.00	214.0
296.00	6683.0
297.10	1133.0
303.00	485.0
304.00	227.0
308.00	204.0
312.90	150.0
314.00	368.0
314.90	751.0
315.90	375.0
320.90	290.0
323.00	1716.0
323.90	308.0
324.20	266.0
326.90	462.0
328.10	324.0
331.90	173.0
334.00	2203.0
334.90	360.0
341.00	420.0
342.00	162.0
346.00	448.0
352.00	515.0
353.00	579.0
354.00	1025.0
365.00	2697.0
365.90	832.0
372.00	866.0
373.00	247.0
382.80	181.0
390.20	225.0
402.00	455.0
403.00	624.0
404.10	271.0
420.90	371.0
422.10	317.0
423.00	3113.0
424.10	710.0
441.00	6879.0
442.00	45400.0
443.00	9223.0
444.00	907.0
445.10	297.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

m/z	Abundance
36.90	186.0
38.10	394.0
39.10	2951.0
40.10	410.0
44.00	610.0
50.10	10533.0
51.10	40792.0
52.10	1917.0
55.00	398.0
56.00	1053.0
57.10	2695.0
58.00	175.0
61.00	549.0
62.10	1076.0
63.00	1858.0
64.00	228.0
64.90	584.0
67.00	189.0
69.00	49656.0
73.10	939.0
74.00	4549.0
75.00	7632.0
76.10	2865.0
77.10	46752.0
78.10	3550.0
79.10	3597.0
80.00	2719.0
81.00	3808.0
82.00	1000.0
83.00	773.0
85.10	789.0
86.00	972.0
87.00	567.0
87.80	297.0
90.90	641.0
92.00	889.0
93.00	6627.0
94.00	588.0
95.80	367.0
96.80	189.0
98.00	3913.0
99.00	3337.0
100.10	238.0
101.00	2565.0
101.80	152.0
102.90	780.0
104.00	2177.0
105.00	1069.0
107.00	16848.0
108.00	2386.0
109.10	758.0
110.00	25880.0
111.00	4408.0
112.00	421.0
113.00	222.0
114.80	475.0
115.90	865.0
117.00	13141.0
118.10	983.0
118.90	179.0
120.10	262.0
121.10	182.0
122.00	1036.0
123.00	1474.0
124.10	631.0
125.10	768.0
127.00	48976.0
128.00	3959.0
129.00	22376.0
130.00	1751.0
131.10	342.0
133.20	244.0
133.90	599.0
135.00	1631.0
136.00	668.0
137.00	761.0
138.00	363.0
138.80	201.0
139.90	208.0
141.00	3218.0
142.00	1003.0
143.00	682.0
143.90	406.0
146.00	399.0
147.00	1783.0
148.00	3867.0
149.10	679.0
151.10	434.0
152.10	403.0
153.00	1269.0
154.00	779.0
154.90	1561.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

156.00	2116.0
157.00	409.0
157.90	586.0
159.00	682.0
160.00	894.0
161.00	1399.0
161.90	518.0
165.10	931.0
166.00	895.0
167.00	5061.0
168.00	2665.0
168.90	432.0
170.10	280.0
170.80	294.0
172.00	466.0
173.00	731.0
174.00	1604.0
175.00	2364.0
176.10	796.0
177.00	1559.0
177.90	583.0
179.00	5212.0
180.00	3837.0
181.00	1222.0
181.90	196.0
184.00	354.0
185.00	2871.0
186.00	17656.0
187.00	4653.0
187.90	744.0
189.00	1201.0
191.00	767.0
191.90	1622.0
193.00	2311.0
194.00	516.0
195.00	381.0
196.00	3362.0
198.00	115320.0
198.90	8134.0
200.10	651.0
200.90	277.0
202.00	254.0
203.00	1076.0
204.00	5250.0
205.00	8945.0
206.00	37288.0
207.00	5288.0
208.10	1472.0
209.00	854.0
210.30	495.0
211.00	1364.0
212.00	336.0
214.10	195.0
215.00	491.0
215.90	643.0
217.00	10747.0
218.00	1721.0
218.90	168.0
221.00	5374.0
223.00	2220.0
224.00	20568.0
225.00	5761.0
226.10	743.0
227.00	9597.0
228.10	1322.0
229.00	2287.0
231.10	1261.0
233.00	223.0
233.90	745.0
235.00	1020.0
236.00	635.0
236.90	524.0
238.90	308.0
239.80	265.0
241.00	739.0
241.90	1383.0
243.00	1307.0
244.00	16880.0
245.00	2167.0
246.00	4331.0
247.00	697.0
247.70	218.0
248.80	536.0
249.90	175.0
252.80	610.0
254.00	1254.0
255.00	86000.0
256.00	11992.0
256.90	1165.0
258.00	5691.0
258.90	896.0
260.80	232.0
263.80	157.0
265.00	1784.0
265.80	400.0
267.00	192.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

270.80	229.0
271.60	225.0
273.00	2681.0
274.00	6360.0
275.00	36864.0
276.10	4109.0
276.90	3923.0
278.00	631.0
278.90	163.0
280.90	338.0
281.90	165.0
283.00	537.0
284.00	386.0
285.00	740.0
289.20	210.0
289.90	213.0
291.10	164.0
291.90	203.0
293.00	901.0
294.00	231.0
295.00	429.0
296.00	10980.0
297.00	1466.0
298.00	163.0
300.80	265.0
302.10	242.0
303.10	1151.0
304.10	351.0
309.10	219.0
309.80	199.0
312.80	237.0
314.10	490.0
315.00	1313.0
316.00	770.0
321.00	453.0
321.80	195.0
323.00	3573.0
324.00	639.0
324.90	167.0
327.00	824.0
328.10	457.0
333.00	224.0
334.00	2095.0
334.70	345.0
335.00	344.0
339.10	169.0
341.10	417.0
346.00	526.0
352.00	1344.0
352.80	467.0
354.10	849.0
354.80	293.0
365.00	5356.0
365.90	655.0
370.90	327.0
372.00	2009.0
373.10	624.0
382.90	451.0
390.10	167.0
401.90	692.0
403.00	792.0
404.00	225.0
421.00	812.0
422.00	674.0
423.10	4537.0
424.00	818.0
441.00	12781.0
442.00	80368.0
443.00	14308.0
444.00	1268.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

m/z	Abundance
38.10	387.0
39.00	2111.0
40.00	364.0
44.00	786.0
49.20	302.0
50.10	7383.0
51.10	28440.0
52.10	1404.0
56.00	1225.0
57.00	2832.0
58.00	200.0
61.00	358.0
62.20	492.0
63.00	1709.0
64.00	234.0
65.00	601.0
68.00	853.0
69.00	33904.0
70.20	288.0
73.00	418.0
74.10	3285.0
75.00	7289.0
76.10	1687.0
77.10	35520.0
78.00	2108.0
79.00	2337.0
80.00	2040.0
81.00	2626.0
82.00	680.0
83.00	581.0
83.80	259.0
85.10	462.0
86.00	890.0
86.90	530.0
87.90	154.0
91.10	706.0
92.00	729.0
93.00	4940.0
94.00	595.0
96.10	328.0
98.00	3914.0
99.00	2041.0
100.00	455.0
100.90	2098.0
102.10	175.0
103.10	546.0
104.00	1407.0
105.10	1335.0
105.80	268.0
107.00	12520.0
108.00	2063.0
110.00	18240.0
111.00	3004.0
111.90	487.0
113.00	251.0
116.00	990.0
117.00	10819.0
118.10	800.0
122.00	902.0
123.00	1346.0
124.00	798.0
125.00	778.0
127.00	38528.0
128.10	3272.0
129.00	16752.0
130.00	1132.0
131.10	238.0
132.20	236.0
133.90	606.0
135.00	1173.0
136.10	417.0
137.00	663.0
137.90	224.0
140.10	301.0
141.00	2732.0
142.00	546.0
143.00	487.0
145.10	262.0
145.90	951.0
147.00	1520.0
148.00	3017.0
149.00	391.0
151.00	230.0
151.90	259.0
152.90	916.0
154.00	365.0
155.10	1202.0
156.00	1880.0
157.10	413.0
158.10	512.0
158.90	349.0
160.00	667.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

161.00	1525.0
161.60	376.0
161.80	378.0
163.10	184.0
163.70	160.0
165.00	1138.0
166.10	530.0
167.00	4135.0
167.90	1977.0
169.00	559.0
170.10	175.0
172.00	377.0
173.10	710.0
174.00	1108.0
175.00	1703.0
175.90	238.0
176.80	743.0
177.00	734.0
177.90	259.0
179.00	4715.0
180.00	2705.0
181.00	1198.0
184.00	700.0
185.00	2855.0
186.00	16072.0
187.00	4044.0
188.00	595.0
189.00	909.0
191.00	651.0
192.10	1779.0
192.90	1360.0
193.90	366.0
195.10	294.0
196.00	2734.0
198.00	96352.0
199.00	6728.0
200.00	752.0
201.30	394.0
201.60	406.0
203.10	874.0
204.00	3978.0
205.00	7874.0
206.00	31248.0
207.00	5446.0
208.00	1472.0
209.00	512.0
209.90	638.0
210.90	1304.0
214.80	307.0
216.00	791.0
217.00	9403.0
217.90	1212.0
221.00	4340.0
222.90	2151.0
224.00	17352.0
225.00	4408.0
226.20	448.0
227.00	7992.0
228.00	974.0
228.90	1819.0
229.90	434.0
231.10	653.0
232.00	248.0
233.00	221.0
234.00	435.0
234.90	439.0
235.90	326.0
236.90	568.0
239.10	472.0
239.90	344.0
241.10	319.0
242.00	679.0
243.00	1025.0
244.00	12505.0
245.00	1535.0
246.00	2888.0
247.00	531.0
248.00	191.0
248.80	490.0
250.90	158.0
251.80	279.0
253.00	444.0
255.00	73184.0
256.00	10504.0
257.00	778.0
258.00	4292.0
258.90	794.0
261.10	206.0
265.00	1772.0
265.90	223.0
272.00	382.0
273.00	2175.0
274.00	6194.0
275.00	31056.0
276.00	4442.0
277.00	2495.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

277.90	774.0
281.00	417.0
282.90	376.0
284.00	184.0
285.00	809.0
285.90	159.0
289.90	325.0
292.90	889.0
294.80	266.0
296.00	10478.0
297.00	1522.0
301.90	239.0
303.00	1348.0
304.00	308.0
308.00	342.0
314.10	290.0
315.00	1187.0
316.10	656.0
320.90	340.0
321.90	155.0
323.00	2925.0
324.10	425.0
327.00	641.0
327.90	369.0
331.90	240.0
333.00	459.0
334.00	1435.0
334.90	502.0
339.80	157.0
341.00	503.0
342.10	208.0
346.00	857.0
352.00	1362.0
352.90	550.0
353.90	496.0
355.00	203.0
355.30	154.0
365.00	4840.0
365.90	509.0
370.00	172.0
370.90	247.0
372.00	1324.0
372.90	512.0
382.90	707.0
391.80	241.0
401.90	624.0
403.00	854.0
403.90	373.0
421.00	536.0
422.00	537.0
423.00	4741.0
424.00	1362.0
441.00	12549.0
442.00	79104.0
443.00	14465.0
444.00	1298.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081