

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020425\
 Data File : BG063994.D
 Acq On : 4 Feb 2025 17:38
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 05 01:11:22 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 04 02:51:00 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.592	0.571	3.5	98	0.00
3	Pyridine	1.600	1.586	0.9	98	0.00
4	n-Nitrosodimethylamine	0.920	0.937	-1.8	104	0.00
5 S	2-Fluorophenol	1.247	1.291	-3.5	100	0.00
6	Aniline	1.834	1.863	-1.6	102	0.00
7 S	Phenol-d6	1.712	1.821	-6.4	105	0.00
8	2-Chlorophenol	1.271	1.382	-8.7	105	0.00
9	Benzaldehyde	0.935	1.029	-10.1	111	0.00
10 C	Phenol	1.802	1.892	-5.0	105	0.00
11	bis(2-Chloroethyl)ether	1.451	1.460	-0.6	102	0.00
12	1,3-Dichlorobenzene	1.513	1.464	3.2	98	0.00
13 C	1,4-Dichlorobenzene	1.528	1.526	0.1	101	0.00
14	1,2-Dichlorobenzene	1.464	1.465	-0.1	102	0.00
15	Benzyl Alcohol	1.280	1.374	-7.3	107	0.00
16	2,2'-oxybis(1-Chloropropane	2.922	3.064	-4.9	107	0.00
17	2-Methylphenol	1.165	1.249	-7.2	105	0.00
18	Hexachloroethane	0.503	0.544	-8.2	104	0.00
19 P	n-Nitroso-di-n-propylamine	1.185	1.285	-8.4	107	0.00
20	3+4-Methylphenols	1.553	1.681	-8.2	106	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
22	Acetophenone	0.550	0.561	-2.0	106	0.00
23 S	Nitrobenzene-d5	0.329	0.378	-14.9	117	0.00
24	Nitrobenzene	0.330	0.389	-17.9	114	0.00
25	Isophorone	0.713	0.741	-3.9	108	0.00
26 C	2-Nitrophenol	0.069	0.117	-69.6#	160#	0.00
27	2,4-Dimethylphenol	0.205	0.224	-9.3	108	0.00
28	bis(2-Chloroethoxy)methane	0.454	0.463	-2.0	106	0.00
29 C	2,4-Dichlorophenol	0.249	0.288	-15.7	111	0.00
30	1,2,4-Trichlorobenzene	0.337	0.339	-0.6	104	0.00
31	Naphthalene	1.072	1.075	-0.3	103	0.00
32	Benzoic acid	0.123	0.164	-33.3#	151#	0.00
33	4-Chloroaniline	0.411	0.416	-1.2	103	0.00
34 C	Hexachlorobutadiene	0.215	0.218	-1.4	105	0.00
35	Caprolactam	0.095	0.110	-15.8	119	0.00
36 C	4-Chloro-3-methylphenol	0.328	0.357	-8.8	106	0.00
37	2-Methylnaphthalene	0.747	0.763	-2.1	106	0.00
38	1-Methylnaphthalene	0.739	0.747	-1.1	106	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	0.590	0.598	-1.4	106	0.00
41 P	Hexachlorocyclopentadiene	0.149	0.189	-26.8#	121	0.00
42 S	2,4,6-Tribromophenol	0.192	0.236	-22.9	130	0.00
43 C	2,4,6-Trichlorophenol	0.288	0.366	-27.1#	119	0.00
44	2,4,5-Trichlorophenol	0.333	0.407	-22.2	114	0.00
45 S	2-Fluorobiphenyl	1.305	1.327	-1.7	106	0.00
46	1,1'-Biphenyl	1.523	1.549	-1.7	106	0.00
47	2-Chloronaphthalene	1.110	1.135	-2.3	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020425\
 Data File : BG063994.D
 Acq On : 4 Feb 2025 17:38
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 05 01:11:22 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 04 02:51:00 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.249	0.344	-38.2#	134	0.00
49	Acenaphthylene	1.724	1.731	-0.4	104	0.00
50	Dimethylphthalate	1.326	1.441	-8.7	109	0.00
51	2,6-Dinitrotoluene	0.208	0.270	-29.8#	122	0.00
52 C	Acenaphthene	1.156	1.187	-2.7	108	0.00
53	3-Nitroaniline	0.238	0.298	-25.2#	116	0.00
54 P	2,4-Dinitrophenol	0.057	0.076	-33.3#	156#	0.00
55	Dibenzofuran	1.878	1.878	0.0	105	0.00
56 P	4-Nitrophenol	0.203	0.247	-21.7	128	0.00
57	2,4-Dinitrotoluene	0.268	0.362	-35.1#	129	0.00
58	Fluorene	1.451	1.458	-0.5	106	0.00
59	2,3,4,6-Tetrachlorophenol	0.309	0.394	-27.5#	122	0.00
60	Diethylphthalate	1.406	1.542	-9.7	109	0.00
61	4-Chlorophenyl-phenylether	0.728	0.739	-1.5	109	0.00
62	4-Nitroaniline	0.267	0.319	-19.5	115	0.00
63	Azobenzene	1.661	1.670	-0.5	107	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	0.046	0.062	-34.8#	156#	0.00
66 c	n-Nitrosodiphenylamine	0.555	0.570	-2.7	107	0.00
67	4-Bromophenyl-phenylether	0.199	0.210	-5.5	110	0.00
68	Hexachlorobenzene	0.234	0.239	-2.1	106	0.00
69	Atrazine	0.185	0.201	-8.6	108	0.00
70 C	Pentachlorophenol	0.120	0.146	-21.7#	128	0.00
71	Phenanthrene	1.050	1.073	-2.2	108	0.00
72	Anthracene	1.051	1.068	-1.6	107	0.00
73	Carbazole	0.984	1.003	-1.9	108	0.00
74	Di-n-butylphthalate	1.026	1.219	-18.8	114	0.00
75 C	Fluoranthene	1.279	1.302	-1.8	107	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
77	Benzidine	0.426	0.477	-12.0	107	0.00
78	Pyrene	1.235	1.258	-1.9	108	0.00
79 S	Terphenyl-d14	0.958	0.963	-0.5	107	0.00
80	Butylbenzylphthalate	0.346	0.457	-32.1#	125	0.00
81	Benzo(a)anthracene	1.270	1.286	-1.3	107	0.00
82	3,3'-Dichlorobenzidine	0.395	0.443	-12.2	108	0.00
83	Chrysene	1.264	1.245	1.5	106	0.00
84	Bis(2-ethylhexyl)phthalate	0.575	0.738	-28.3#	120	0.00
85 c	Di-n-octyl phthalate	1.001	1.229	-22.8#	128	0.00
86 I	Perylene-d12	1.000	1.000	0.0	105	0.00
87	Indeno(1,2,3-cd)pyrene	1.290	1.342	-4.0	108	0.00
88	Benzo(b)fluoranthene	1.178	1.233	-4.7	108	0.00
89	Benzo(k)fluoranthene	1.209	1.206	0.2	102	0.00
90 C	Benzo(a)pyrene	1.051	1.085	-3.2	107	0.00
91	Dibenzo(a,h)anthracene	1.089	1.135	-4.2	108	0.00
92	Benzo(g,h,i)perylene	1.100	1.126	-2.4	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020425\
Data File : BG063994.D
Acq On : 4 Feb 2025 17:38
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Feb 05 01:11:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 04 02:51:00 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 4