



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020525\
Data File : BG063995.D
Acq On : 5 Feb 2025 8:37
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020325.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Feb 04 02:51:00 2025






AutoFind: Scans 2485, 2486, 2487; Background Corrected with Scan 2478

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	53.8	86925	PASS
68	69	0.00	2	0.5	320	PASS
69	198	0.00	100	38.0	61336	PASS
70	69	0.00	2	0.5	326	PASS
127	198	10	80	46.3	74813	PASS
197	198	0.00	2	0.5	870	PASS
198	198	100	100	100.0	161547	PASS
199	198	5	9	6.8	10957	PASS
275	198	10	60	26.9	43392	PASS
365	198	1	100	3.8	6185	PASS
441	198	0.01	100	13.6	21987	PASS
442	442	50	100	100.0	132672	PASS
443	442	15	24	19.5	25883	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020525\
Data File : BG063995.D
Acq On : 5 Feb 2025 8:37
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Feb 05 11:06:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 04 02:51:00 2025
Response via : Initial Calibration



TIC: BG063995.D\data.ms

(70) Pentachlorophenol (C)

16.889min (-0.000) 0.00 ng




response 94760

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	68.50	61.29
264.00	62.50	64.15
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020525\
Data File : BG063995.D
Acq On : 5 Feb 2025 8:37
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Feb 05 11:06:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 04 02:51:00 2025
Response via : Initial Calibration



TIC: BG063995.D\data.ms

(77) Benzidine

19.474min (-0.001) 25002.34 ng

response 1185140

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.00	14.54
183.00	10.90	11.68
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
2/5/2025	BNA_G	<u>BG063995.D</u>
Compound Name	Response	Retention Time
DDT	364157	20.714
DDD	9933	20.273
DDE	57	19.768
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
9990	374147	2.67

Instrument :
BNA_G
ClientSampleId :
DFTPP