




Data Path : Z:\HPCHEM1\BNA G\DATA\BG020617\
Data File : BG025810.D
Acq On : 6 Feb 2017 15:19
Operator : SJ/MA
Sample : SSTD08008
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Feb 06 15:59:46 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG020617.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Feb 06 15:01:17 2017
Response via : Initial Calibration



TIC: BG025810.D

(32) Caprolactam




11.945min (+0.021) 41.34ng/ul

response 105514

Ion	Exp%	Act%
113.00	100	100
55.00	210.70	224.60
56.00	160.60	166.86
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG020617\
Data File : BG025810.D
Acq On : 6 Feb 2017 15:19
Operator : SJ/MA
Sample : SST08008
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Feb 06 15:59:46 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG020617.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Feb 06 15:01:17 2017
Response via : Initial Calibration



TIC: BG025810.D

(32) Caprolactam




11.945min (+0.021) 77.24ng/ul m

response 197125

Ion	Exp%	Act%
113.00	100	100
55.00	210.70	224.60
56.00	160.60	166.86
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG020617\
Data File : BG025810.D
Acq On : 6 Feb 2017 15:19
Operator : SJ/MA
Sample : SSTD08008
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Feb 06 15:59:46 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG020617.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Feb 06 15:01:17 2017
Response via : Initial Calibration



TIC: BG025810.D

(69) Phenanthrene




17.673min (+0.098) 81.89ng/ul

response 2484888

Ion	Exp%	Act%
178.00	100	100
179.00	15.50	18.59
176.00	20.60	24.06
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG020617\
Data File : BG025810.D
Acq On : 6 Feb 2017 15:19
Operator : SJ/MA
Sample : SSTD08008
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Feb 06 15:59:46 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG020617.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Feb 06 15:01:17 2017
Response via : Initial Calibration



TIC: BG025810.D

(69) Phenanthrene




17.579min (+0.004) 80.59ng/ul m

response 2445478

Ion	Exp%	Act%
178.00	100	100
179.00	15.50	19.16#
176.00	20.60	23.31
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG020617\
Data File : BG025810.D
Acq On : 6 Feb 2017 15:19
Operator : SJ/MA
Sample : SSTD08008
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Feb 06 15:59:46 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG020617.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Feb 06 15:01:17 2017
Response via : Initial Calibration



TIC: BG025810.D

(70) Anthracene-d10 (S)




17.532min (-0.102) 23.49ng/ul

response 624273

Ion	Exp%	Act%
188.00	100	100
94.00	8.80	8.82
80.00	9.50	11.05
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG020617\
Data File : BG025810.D
Acq On : 6 Feb 2017 15:19
Operator : SJ/MA
Sample : SSTD08008
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Feb 06 15:59:46 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG020617.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Feb 06 15:01:17 2017
Response via : Initial Calibration



TIC: BG025810.D

(70) Anthracene-d10 (S)

17.638min (+0.004) 81.22ng/ul m

response 2158765

Ion	Exp%	Act%
188.00	100	100
94.00	8.80	11.99#
80.00	9.50	12.00#
0.00	0.00	0.00