




Data Path : Z:\HPCHEM1\BNA G\DATA\BG020818\
Data File : BG032625.D
Acq On : 9 Feb 2018 2:12
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 09 06:49:51 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG012918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 06 18:57:08 2018
Response via : Initial Calibration



TIC: BG032625.D

(69) Pentachlorophenol (C)




17.706min (-0.012) 29586.43ng

response 65331

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	64.00
264.00	62.00	66.89
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG020818\
Data File : BG032625.D
Acq On : 9 Feb 2018 2:12
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 09 06:49:51 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG012918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 06 18:57:08 2018
Response via : Initial Calibration



TIC: BG032625.D

(76) Benzidine

20.238min (-0.006) 220839.80ng

response 242839

Ion	Exp%	Act%
184.00	100	100
185.00	13.90	15.01
183.00	13.30	12.98
0.00	0.00	0.00