

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020823\
 Data File : BG056580.D
 Acq On : 8 Feb 2023 13:35
 Operator : CG/JU
 Sample : PB150729BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB150729BL

Quant Time: Feb 09 02:06:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG012723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 08 03:02:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.264	152	30263	20.000	ng	0.00
21) Naphthalene-d8	11.096	136	125403	20.000	ng	0.00
39) Acenaphthene-d10	14.892	164	104095	20.000	ng	0.00
64) Phenanthrene-d10	17.630	188	283665	20.000	ng	0.00
76) Chrysene-d12	21.919	240	266554	20.000	ng	0.00
86) Perylene-d12	25.338	264	287119	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.791	112	262338	159.466	ng	0.00
7) Phenol-d6	7.418	99	360331	147.828	ng	0.00
23) Nitrobenzene-d5	9.445	82	183476	83.082	ng	0.00
42) 2,4,6-Tribromophenol	16.372	330	182412	131.812	ng	0.00
45) 2-Fluorobiphenyl	13.511	172	636859	84.822	ng	0.00
79) Terphenyl-d14	20.203	244	1334805	87.952	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020823\
Data File : BG056580.D
Acq On : 8 Feb 2023 13:35
Operator : CG/JU
Sample : PB150729BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB150729BL

Quant Time: Feb 09 02:06:40 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG012723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 08 03:02:11 2023
Response via : Initial Calibration

