

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020823\
 Data File : BG056583.D
 Acq On : 8 Feb 2023 15:51
 Operator : CG/JU
 Sample : 01418-16
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 JPP-22-020223

Quant Time: Feb 09 02:07:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG012723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 08 03:02:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.265	152	40340	20.000	ng	0.00
21) Naphthalene-d8	11.103	136	148654	20.000	ng	# 0.00
39) Acenaphthene-d10	14.893	164	28691	20.000	ng	# 0.00
64) Phenanthrene-d10	17.631	188	205276	20.000	ng	0.00
76) Chrysene-d12	21.926	240	162130	20.000	ng	0.00
86) Perylene-d12	25.351	264	33742	20.000	ng	# 0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.797	112	310769	141.716	ng	0.00
7) Phenol-d6	7.419	99	116872	35.970	ng	0.00
23) Nitrobenzene-d5	9.446	82	259216	99.019	ng	0.00
42) 2,4,6-Tribromophenol	16.379	330	206052	540.210	ng	0.00
45) 2-Fluorobiphenyl	13.518	172	801228	387.175	ng	0.00
79) Terphenyl-d14	20.210	244	1229170	133.157	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020823\
Data File : BG056583.D
Acq On : 8 Feb 2023 15:51
Operator : CG/JU
Sample : 01418-16
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
JPP-22-020223

Quant Time: Feb 09 02:07:33 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG012723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 08 03:02:11 2023
Response via : Initial Calibration

