

Data Path : Z:\HPCHEM1\BNA_G\Data\BG021016\
 Data File : BG020842.D
 Acq On : 10 Feb 2016 16:03
 Operator : SJ/UM
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02025

Quant Time: Feb 11 03:31:18 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG020816.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 11 03:24:50 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	139	0.00
2	1,4-Dioxane	0.499	0.458	8.2	129	0.00
3 S	1,4-Dioxane-d8	0.412	0.403	2.2	127	0.00
4	Benzaldehyde	1.120	1.277	-14.0	136	0.00
5 S	Phenol-d5	2.026	1.975	2.5	137	0.00
6	Phenol	2.145	2.096	2.3	133	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.081	1.058	2.1	131	0.00
8	Bis(2-Chloroethyl)ether	1.644	1.607	2.3	130	0.00
9 S	2-Chlorophenol-d4	1.585	1.601	-1.0	138	0.00
10	2-Chlorophenol	1.630	1.625	0.3	135	0.00
11	2-Methylphenol	1.697	1.657	2.4	133	0.00
12	2,2'-oxybis(1-Chloropropane	1.706	1.679	1.6	130	0.00
13 S	4-Methylphenol-d8	1.761	1.670	5.2	129	0.00
14	Acetophenone	2.663	2.587	2.9	131	0.00
15 P	N-Nitroso-di-n-propylamine	1.347	1.281	4.9	126	0.00
16	4-Methylphenol	1.883	1.833	2.7	132	0.00
17	Hexachloroethane	0.608	0.613	-0.8	138	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	136	0.00
19 S	Nitrobenzene-d5	0.160	0.163	-1.9	131	0.00
20	Nitrobenzene	0.337	0.339	-0.6	128	0.00
21	Isophorone	0.753	0.703	6.6	123	0.00
22 S	2-Nitrophenol-d4	0.160	0.169	-5.6	137	0.00
23 C	2-Nitrophenol	0.187	0.191	-2.1	134	0.00
24	2,4-Dimethylphenol	0.374	0.371	0.8	129	0.00
25	Bis(2-Chloroethoxy)methane	0.478	0.461	3.6	126	0.00
26 S	2,4-Dichlorophenol-d3	0.307	0.321	-4.6	138	0.00
27 C	2,4-Dichlorophenol	0.322	0.329	-2.2	135	0.00
28	Naphthalene	1.111	1.149	-3.4	136	0.00
29 S	4-Chloroaniline-d4	0.409	0.443	-8.3	128	0.00
30	4-Chloroaniline	0.432	0.474	-9.7	129	0.00
31 C	Hexachlorobutadiene	0.154	0.163	-5.8	141	0.00
32	Caprolactam	0.131	0.119	9.2	113	0.00
33 C	4-Chloro-3-methylphenol	0.377	0.359	4.8	123	0.00
34	2-Methylnaphthalene	0.841	0.852	-1.3	132	0.00
35	1-Methylnaphthalene	0.796	0.799	-0.4	131	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	128	0.00
37	1,2,4,5-Tetrachlorobenzene	0.570	0.620	-8.8	133	0.00
38	Hexachlorocyclopentadiene	0.311	0.275	11.6	121	0.00
39 C	2,4,6-Trichlorophenol	0.412	0.434	-5.3	128	0.00
40	2,4,5-Trichlorophenol	0.456	0.466	-2.2	125	0.00
41	1,1'-Biphenyl	1.770	1.839	-3.9	128	0.00
42	2-Chloronaphthalene	1.305	1.369	-4.9	127	0.00
43	2-Nitroaniline	0.355	0.345	2.8	116	0.00
44 S	Dimethylphthalate-d6	1.721	1.630	5.3	115	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG021016\
 Data File : BG020842.D
 Acq On : 10 Feb 2016 16:03
 Operator : SJ/UM
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02025

Quant Time: Feb 11 03:31:18 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG020816.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 11 03:24:50 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.792	1.699	5.2	115	0.00
46	2,6-Dinitrotoluene	0.352	0.350	0.6	120	0.00
47 S	Acenaphthylene-d8	2.102	2.153	-2.4	125	0.00
48	Acenaphthylene	2.306	2.360	-2.3	123	0.00
49	3-Nitroaniline	0.397	0.407	-2.5	118	0.00
50 C	Acenaphthene	1.611	1.634	-1.4	125	0.00
51	2,4-Dinitrophenol	0.138	0.106	23.2#	115	0.00
52 S	4-Nitrophenol-d4	0.345	0.311	9.9	110	0.00
53	4-Nitrophenol	0.196	0.178	9.2	107	0.00
54	Dibenzofuran	2.092	2.080	0.6	119	0.00
55	2,4-Dinitrotoluene	0.489	0.468	4.3	114	0.00
56	2,3,4,6-Tetrachlorophenol	0.396	0.373	5.8	117	0.00
57	Diethylphthalate	1.888	1.753	7.2	113	0.00
58 S	Fluorene-d10	1.464	1.440	1.6	119	0.00
59	Fluorene	1.820	1.782	2.1	119	0.00
60	4-Chlorophenyl-phenylether	0.777	0.770	0.9	122	0.00
61	4-Nitroaniline	0.449	0.432	3.8	112	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.097	0.088	9.3	111	0.00
64	4,6-Dinitro-2-methylphenol	0.109	0.099	9.2	109	0.00
65	N-Nitrosodiphenylamine	0.687	0.725	-5.5	115	0.00
66	4-Bromophenyl-phenylether	0.215	0.225	-4.7	114	0.00
67	Hexachlorobenzene	0.237	0.248	-4.6	115	0.00
68	Atrazine	0.222	0.227	-2.3	113	0.00
69 C	Pentachlorophenol	0.139	0.135	2.9	111	0.00
70	Phenanthrene	1.261	1.294	-2.6	111	0.00
71 S	Anthracene-d10	0.988	0.999	-1.1	109	0.00
72	Anthracene	1.269	1.305	-2.8	110	0.00
73	1,2,3,4-Tetrachlorobenzene	0.248	0.295	-19.0	133	0.00
74	Pentachlorobenzene	0.252	0.282	-11.9	123	0.00
75	Carbazole	1.114	1.159	-4.0	111	0.00
76	Di-n-butylphthalate	1.433	1.470	-2.6	111	0.00
77 C	Fluoranthene	1.304	1.357	-4.1	112	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	112	0.00
79 S	Pyrene-d10	1.028	1.077	-4.8	112	0.00
80	Pyrene	1.444	1.503	-4.1	112	0.00
81	Butylbenzylphthalate	0.670	0.697	-4.0	114	0.00
82	3,3'-Dichlorobenzidine	0.431	0.461	-7.0	109	0.00
83	Benzo(a)anthracene	1.304	1.336	-2.5	111	0.00
84	Bis(2-ethylhexyl)phthalate	1.016	1.039	-2.3	111	0.00
85	Chrysene	1.201	1.232	-2.6	111	0.00
86 I	Perylene-d12	1.000	1.000	0.0	114	0.00
87	Di-n-octyl phthalate	1.802	1.825	-1.3	111	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG021016\
Data File : BG020842.D
Acq On : 10 Feb 2016 16:03
Operator : SJ/UM
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02025

Quant Time: Feb 11 03:31:18 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG020816.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 11 03:24:50 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.362	1.357	0.4	110	0.00
89	Benzo(k)fluoranthene	1.301	1.320	-1.5	112	0.00
90 S	Benzo(a)pyrene-d12	1.108	1.097	1.0	110	0.00
91 C	Benzo(a)pyrene	1.289	1.295	-0.5	112	0.00
92	Indeno(1,2,3-cd)pyrene	1.486	1.516	-2.0	113	0.00
93	Dibenzo(a,h)anthracene	1.218	1.220	-0.2	111	0.00
94	Benzo(g,h,i)perylene	1.216	1.245	-2.4	115	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0