


Data Path : Z:\HPCHEM1\BNA_G\Data\BG021017\
Data File : BG025821.D
Acq On : 10 Feb 2017 9:04
Operator : SJ/MA
Sample : DFTPP03
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP03

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG021017MA.M
Title : SVOA CALIBRATION
Last Update : Fri Feb 10 22:04:56 2017



AutoFind: Scans 2556, 2557, 2558; Background Corrected with Scan 2550

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.8	52232	PASS
68	69	0.00	2	1.6	754	PASS
69	198	0.00	100	31.5	47218	PASS
70	69	0.00	2	0.7	309	PASS
127	198	10	80	44.8	67213	PASS
197	198	0.00	2	0.7	1041	PASS
198	198	100	100	100.0	150072	PASS
199	198	5	9	6.9	10285	PASS
275	198	10	60	24.9	37437	PASS
365	198	1	100	3.3	4904	PASS
441	442	0.01	100	15.7	16542	PASS
442	198	50	100	70.1	105130	PASS
443	442	15	24	20.2	21185	PASS

m/z	Abundance
42.90	175.0
44.10	2127.0
73.10	279.0
96.00	391.0
96.50	181.0
132.90	219.0
147.00	184.0
176.80	171.0
191.00	303.0
193.10	253.0
207.10	4415.0
208.10	893.0
209.10	783.0
281.10	476.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03

m/z	Abundance
37.20	178.0
38.10	502.0
39.10	3238.0
40.10	219.0
41.10	236.0
43.20	257.0
44.10	1635.0
44.90	167.0
49.10	496.0
50.10	11841.0
51.10	50880.0
52.10	2567.0
53.10	223.0
55.10	246.0
56.00	1435.0
57.10	3830.0
61.10	897.0
62.10	955.0
63.10	2139.0
64.10	366.0
65.10	1069.0
69.10	45656.0
70.10	371.0
73.00	708.0
74.10	4742.0
75.10	8400.0
76.10	3164.0
77.10	59008.0
78.10	3532.0
79.10	2691.0
80.00	2549.0
81.10	3470.0
82.10	1209.0
83.10	1012.0
84.00	280.0
85.10	609.0
86.00	904.0
87.20	379.0
87.80	198.0
88.10	236.0
89.00	188.0
91.10	1109.0
92.10	863.0
93.10	5716.0
94.10	447.0
96.10	677.0
97.20	256.0
98.10	3616.0
99.10	4044.0
100.00	180.0
101.10	2624.0
103.10	709.0
104.10	1412.0
105.00	1143.0
106.00	691.0
107.10	15482.0
108.10	2760.0
110.10	33384.0
111.10	5075.0
112.10	404.0
114.80	157.0
116.10	862.0
117.10	10723.0
118.00	972.0
119.00	309.0
120.20	181.0
121.00	165.0
122.00	896.0
123.00	1780.0
124.00	902.0
125.10	840.0
127.10	63056.0
128.10	4993.0
129.10	21872.0
130.10	2186.0
131.00	300.0
133.00	370.0
134.00	734.0
135.10	1980.0
136.10	940.0
137.10	909.0
138.10	259.0
140.00	226.0
141.00	3548.0
142.10	1020.0
143.00	762.0
144.20	185.0
146.00	584.0
147.00	1711.0
148.00	3116.0
149.00	619.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03

150.10	424.0
151.10	558.0
151.90	182.0
153.00	724.0
153.90	819.0
155.00	1890.0
156.10	2250.0
157.00	355.0
158.10	408.0
159.00	461.0
160.10	832.0
161.10	1753.0
162.10	373.0
163.00	248.0
164.10	156.0
165.00	1276.0
166.10	877.0
167.10	5855.0
168.10	3972.0
169.10	484.0
169.90	271.0
170.50	187.0
171.10	293.0
172.00	488.0
173.00	637.0
174.10	1193.0
175.10	2250.0
176.10	368.0
177.10	1370.0
178.20	586.0
179.00	4315.0
180.00	3538.0
181.10	1476.0
183.10	234.0
184.00	680.0
185.10	2375.0
186.10	17256.0
187.10	4225.0
188.10	887.0
189.00	970.0
190.00	304.0
191.00	1048.0
192.00	1724.0
193.10	2145.0
194.20	333.0
194.90	258.0
196.10	3953.0
196.70	1373.0
198.00	131048.0
199.00	9576.0
200.00	665.0
201.00	221.0
201.60	700.0
203.00	872.0
204.10	4036.0
205.10	7681.0
206.10	32032.0
207.10	9629.0
208.10	2063.0
209.10	817.0
211.00	1381.0
211.70	290.0
212.10	371.0
215.00	572.0
216.10	821.0
217.00	7794.0
218.00	851.0
219.00	251.0
219.90	171.0
221.00	4896.0
221.90	1036.0
223.00	2064.0
224.10	18256.0
225.10	4517.0
226.00	429.0
227.00	6996.0
228.00	963.0
229.00	1463.0
229.90	194.0
231.00	769.0
234.10	531.0
235.10	542.0
236.10	355.0
237.00	861.0
238.00	156.0
239.00	408.0
240.00	266.0
241.10	561.0
242.00	1096.0
243.00	1024.0
244.10	13408.0
245.10	1756.0
246.00	2841.0
247.00	386.0
249.00	805.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03

250.90	179.0
252.40	192.0
253.10	484.0
255.10	68144.0
256.10	11185.0
257.00	1019.0
258.00	4135.0
258.90	319.0
265.00	1625.0
266.00	313.0
267.00	182.0
270.10	153.0
271.00	248.0
272.20	287.0
273.10	2210.0
274.10	5517.0
275.10	28648.0
276.10	4223.0
277.10	2594.0
277.90	442.0
281.00	642.0
283.00	414.0
284.20	171.0
285.10	835.0
293.10	613.0
294.00	171.0
296.10	8407.0
297.10	1260.0
303.10	1200.0
304.20	211.0
310.10	167.0
313.10	163.0
314.10	413.0
315.00	999.0
316.10	423.0
321.00	266.0
321.90	155.0
323.10	2719.0
324.00	381.0
327.00	398.0
327.90	398.0
333.10	320.0
334.10	1807.0
335.10	556.0
341.10	266.0
342.20	256.0
346.00	574.0
352.00	566.0
353.10	757.0
354.10	900.0
355.10	175.0
365.00	3976.0
365.90	415.0
372.10	1189.0
373.10	307.0
382.90	340.0
389.90	187.0
402.00	417.0
403.00	710.0
404.10	500.0
421.00	492.0
422.10	587.0
423.10	4805.0
424.00	1084.0
425.10	215.0
438.70	163.0
441.10	11020.0
442.10	72176.0
443.10	14588.0
444.10	1413.0

Instrument :

BNA_G

ClientSampleId :

DFTPP03

m/z	Abundance
36.10	159.0
38.10	508.0
39.10	3892.0
40.00	199.0
44.10	1930.0
45.10	159.0
49.10	504.0
50.10	13821.0
51.10	62016.0
52.10	3265.0
53.10	165.0
55.00	349.0
56.10	1890.0
57.10	4726.0
58.00	231.0
59.10	154.0
61.10	789.0
62.10	980.0
63.10	2409.0
64.00	391.0
65.10	1543.0
68.10	1205.0
69.10	57184.0
70.00	228.0
71.10	255.0
73.10	621.0
74.10	5497.0
75.10	9470.0
76.20	3506.0
77.10	72768.0
78.10	4428.0
79.10	3936.0
80.10	3243.0
81.10	4610.0
82.00	1394.0
83.00	1064.0
84.00	157.0
85.10	958.0
86.10	1426.0
87.10	726.0
91.00	819.0
92.10	1151.0
93.10	6176.0
94.10	412.0
95.10	162.0
96.10	570.0
97.00	266.0
98.10	4227.0
99.10	5051.0
100.10	412.0
101.10	2726.0
102.00	247.0
103.00	1015.0
104.10	2071.0
105.10	1645.0
106.10	706.0
107.10	21184.0
108.00	2916.0
110.10	39728.0
111.10	6555.0
112.10	818.0
115.30	167.0
116.10	1151.0
117.10	13353.0
118.10	1413.0
119.10	305.0
120.00	232.0
122.00	1483.0
123.10	1976.0
124.00	1017.0
125.00	900.0
127.10	79064.0
128.10	6487.0
129.00	29440.0
130.10	2258.0
131.10	636.0
132.40	156.0
133.00	481.0
134.00	1030.0
135.10	2491.0
136.10	1314.0
137.10	1227.0
138.10	183.0
140.10	360.0
141.10	3872.0
142.00	1264.0
143.00	750.0
144.00	159.0
145.50	176.0
146.00	811.0
147.00	2089.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03

148.00	3540.0
149.00	989.0
150.10	354.0
151.00	595.0
152.30	194.0
153.10	1167.0
154.10	614.0
155.10	2699.0
156.10	4009.0
157.10	919.0
158.10	728.0
159.00	608.0
160.10	1303.0
161.10	1798.0
162.00	282.0
163.00	313.0
164.10	195.0
165.00	1421.0
166.00	1008.0
167.10	7249.0
168.10	4647.0
169.00	873.0
169.90	246.0
171.10	267.0
172.00	540.0
173.10	1098.0
174.10	1852.0
175.10	2980.0
176.00	993.0
177.00	1941.0
178.10	415.0
179.10	5915.0
180.10	3805.0
181.10	1884.0
182.00	503.0
183.00	163.0
184.00	640.0
185.10	3016.0
186.10	23816.0
187.10	6954.0
188.00	608.0
189.10	1300.0
190.00	293.0
191.00	862.0
192.00	1712.0
193.10	1977.0
194.10	560.0
195.10	621.0
196.00	5291.0
196.80	1752.0
198.00	178880.0
199.00	11957.0
200.10	1094.0
201.50	1006.0
202.10	390.0
203.10	1222.0
204.10	6445.0
205.10	10466.0
206.10	44840.0
207.10	10300.0
208.10	2092.0
209.10	791.0
210.10	771.0
210.50	805.0
211.10	1809.0
215.10	596.0
216.10	1241.0
217.00	10112.0
218.00	1433.0
219.00	310.0
219.90	239.0
221.10	8185.0
221.80	1433.0
222.00	1434.0
223.00	2996.0
224.10	24536.0
225.10	6062.0
226.20	762.0
227.10	10257.0
228.00	1600.0
229.10	1859.0
230.00	366.0
231.00	689.0
233.10	214.0
234.00	716.0
235.00	942.0
236.00	626.0
237.10	925.0
239.10	361.0
240.00	380.0
241.00	563.0
242.10	1226.0
243.10	1173.0
244.10	19208.0
245.10	3237.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03

246.00	4215.0
247.10	906.0
248.10	163.0
249.00	679.0
252.10	230.0
253.10	642.0
255.10	97176.0
256.10	15269.0
257.10	1641.0
258.00	6153.0
259.10	844.0
260.00	168.0
261.00	248.0
265.10	2169.0
265.90	171.0
270.00	176.0
271.10	150.0
272.00	287.0
273.10	2937.0
274.10	7583.0
275.10	43512.0
276.10	5522.0
277.10	4537.0
278.10	649.0
281.00	344.0
282.00	192.0
283.10	528.0
284.20	311.0
285.10	683.0
286.00	196.0
292.00	240.0
293.00	801.0
294.00	302.0
295.10	361.0
296.00	13120.0
297.10	1213.0
302.00	200.0
303.10	1349.0
304.00	230.0
308.10	320.0
310.00	251.0
313.00	252.0
314.10	442.0
315.00	1753.0
316.10	749.0
321.00	454.0
323.10	4148.0
324.10	837.0
326.90	741.0
328.00	418.0
332.10	452.0
333.10	429.0
334.10	2787.0
335.10	742.0
339.20	162.0
341.10	678.0
346.00	960.0
351.00	178.0
352.10	997.0
353.00	697.0
354.10	1142.0
355.00	191.0
359.00	151.0
365.10	5523.0
366.10	746.0
370.10	175.0
371.10	642.0
372.10	1614.0
373.10	611.0
383.00	524.0
383.90	312.0
390.10	241.0
391.10	171.0
392.10	217.0
401.00	236.0
402.10	761.0
403.00	1002.0
404.10	399.0
421.00	1206.0
422.10	908.0
423.00	6127.0
424.10	1595.0
441.10	18680.0
442.10	120832.0
443.10	25016.0
444.10	1613.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03

m/z	Abundance
38.10	349.0
39.10	2999.0
40.00	245.0
43.20	298.0
44.10	1756.0
49.10	467.0
50.10	10937.0
51.10	43800.0
52.10	2327.0
55.00	266.0
56.00	1614.0
57.10	3047.0
58.10	169.0
61.10	615.0
62.10	670.0
63.10	2026.0
64.00	234.0
65.10	1117.0
68.10	1059.0
69.10	38816.0
70.00	328.0
73.10	875.0
74.10	3832.0
75.10	7114.0
76.10	2298.0
77.10	48856.0
78.10	2903.0
79.10	2827.0
80.00	2392.0
81.10	2565.0
82.00	618.0
83.10	957.0
84.20	250.0
85.10	640.0
86.10	927.0
87.10	574.0
88.10	199.0
89.20	156.0
91.10	825.0
92.10	756.0
93.00	4967.0
94.10	569.0
96.10	687.0
98.10	4537.0
99.00	3432.0
99.90	230.0
101.00	2257.0
102.00	306.0
103.00	924.0
104.10	1205.0
105.10	1227.0
106.20	708.0
107.10	14314.0
108.10	2128.0
109.10	842.0
110.10	30160.0
111.00	4380.0
112.00	455.0
116.10	867.0
117.10	10495.0
118.00	848.0
119.10	371.0
120.00	232.0
122.10	663.0
123.00	1032.0
124.10	623.0
125.00	814.0
127.10	59520.0
128.10	4519.0
129.10	21488.0
130.00	1947.0
131.00	395.0
132.00	262.0
132.70	320.0
134.10	997.0
135.10	2152.0
136.00	897.0
137.00	1110.0
138.00	283.0
140.10	303.0
141.10	2726.0
142.00	971.0
143.00	713.0
143.90	209.0
145.10	152.0
146.10	578.0
147.00	1825.0
148.00	3725.0
149.00	785.0
150.00	223.0
151.00	591.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03

151.30	546.0
152.20	152.0
153.00	907.0
154.10	476.0
155.00	1952.0
156.10	2471.0
157.10	708.0
158.00	475.0
159.00	554.0
160.00	929.0
161.10	1805.0
162.10	557.0
163.00	160.0
164.10	211.0
165.00	1363.0
166.00	1002.0
167.10	5298.0
168.10	2762.0
169.00	757.0
170.00	214.0
171.90	263.0
173.10	962.0
174.10	1017.0
175.10	2172.0
176.10	776.0
177.00	1361.0
178.10	475.0
179.00	5213.0
180.00	3452.0
181.10	1482.0
182.10	462.0
184.10	384.0
185.10	2238.0
186.10	18136.0
187.10	4379.0
188.10	653.0
189.00	1461.0
190.00	315.0
191.00	1039.0
192.00	1587.0
193.10	2253.0
193.90	307.0
194.70	171.0
195.10	369.0
196.10	4205.0
198.00	140288.0
199.00	9323.0
200.00	643.0
200.90	340.0
201.60	658.0
203.00	909.0
204.10	5378.0
205.10	8227.0
206.10	35528.0
207.10	9228.0
208.00	2285.0
209.00	1108.0
211.10	1453.0
213.20	171.0
215.10	490.0
216.00	611.0
217.00	9437.0
218.00	1075.0
219.10	265.0
221.10	6209.0
222.10	1135.0
223.00	1916.0
224.10	20704.0
225.10	5225.0
226.10	641.0
227.10	7512.0
228.00	1190.0
229.00	1594.0
229.90	172.0
231.00	642.0
233.90	499.0
235.00	755.0
236.00	465.0
237.10	902.0
239.10	224.0
241.00	589.0
242.00	952.0
243.10	968.0
244.10	15300.0
245.10	2166.0
246.00	2611.0
247.00	675.0
249.00	603.0
250.90	227.0
252.30	184.0
253.20	614.0
255.00	82896.0
256.00	11507.0
257.10	760.0
258.00	4614.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03

259.00	732.0
261.00	192.0
264.00	193.0
265.00	2162.0
266.10	284.0
267.10	151.0
271.00	211.0
272.20	341.0
273.10	3141.0
274.10	6075.0
275.10	40152.0
276.10	5552.0
277.00	3031.0
278.10	396.0
279.00	150.0
281.10	341.0
283.00	453.0
285.10	829.0
286.00	177.0
292.00	316.0
293.00	921.0
294.10	268.0
295.10	258.0
296.10	10336.0
297.00	1209.0
297.90	207.0
301.10	244.0
302.00	248.0
303.10	1273.0
304.10	367.0
310.00	173.0
313.10	162.0
314.00	484.0
315.00	958.0
316.10	1053.0
321.00	181.0
322.00	343.0
323.10	3571.0
324.10	724.0
327.10	848.0
328.00	255.0
332.00	316.0
333.10	305.0
334.10	2736.0
335.10	585.0
341.10	226.0
342.10	169.0
346.00	607.0
347.20	241.0
352.00	990.0
353.00	725.0
354.10	1348.0
354.90	278.0
365.00	5215.0
366.10	733.0
370.10	155.0
371.20	251.0
372.10	2251.0
373.10	514.0
383.10	488.0
384.10	253.0
390.10	332.0
392.20	169.0
402.00	760.0
403.00	846.0
404.10	430.0
405.30	171.0
421.10	886.0
422.10	838.0
423.10	7732.0
424.10	1426.0
425.10	213.0
439.90	255.0
441.10	19928.0
442.10	122384.0
443.10	23952.0
444.10	2073.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03