


Data Path : Z:\HPCHEM1\BNA\_G\DATA\BG021115\  
Data File : BG015978.D  
Acq On : 11 Feb 2015 9:32  
Operator : TP/IZ  
Sample : DFTPP02  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP02

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA\_G\METHODS\SOM01.2-EPA-BG021015.M  
Title : SVOA CALIBRATION  
Last Update : Wed Feb 11 22:42:45 2015



AutoFind: Scans 2543, 2544, 2545; Background Corrected with Scan 2538

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 39.1      | 25856   | PASS             |
| 68          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 69          | 198          | 0.00         | 100          | 43.6      | 28845   | PASS             |
| 70          | 69           | 0.00         | 2            | 1.3       | 373     | PASS             |
| 127         | 198          | 10           | 80           | 51.5      | 34064   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 66088   | PASS             |
| 199         | 198          | 5            | 9            | 7.0       | 4602    | PASS             |
| 275         | 198          | 10           | 60           | 24.3      | 16087   | PASS             |
| 365         | 198          | 1            | 100          | 3.4       | 2247    | PASS             |
| 441         | 443          | 0.01         | 100          | 78.8      | 10118   | PASS             |
| 442         | 198          | 50           | 100          | 98.3      | 64962   | PASS             |
| 443         | 442          | 15           | 24           | 19.8      | 12848   | PASS             |

| m/z    | Abundance |
|--------|-----------|
| 38.25  | 337.0     |
| 39.15  | 1982.0    |
| 40.15  | 1160.0    |
| 41.15  | 268.0     |
| 42.25  | 166.0     |
| 43.05  | 178.0     |
| 44.15  | 1478.0    |
| 45.15  | 159.0     |
| 49.25  | 367.0     |
| 50.15  | 6024.0    |
| 51.15  | 26024.0   |
| 52.15  | 1723.0    |
| 53.15  | 156.0     |
| 55.15  | 438.0     |
| 56.05  | 852.0     |
| 57.15  | 2334.0    |
| 61.15  | 211.0     |
| 62.15  | 375.0     |
| 63.15  | 1033.0    |
| 64.15  | 155.0     |
| 65.15  | 532.0     |
| 69.15  | 30448.0   |
| 70.15  | 308.0     |
| 73.15  | 287.0     |
| 74.15  | 2806.0    |
| 75.15  | 4380.0    |
| 76.15  | 1695.0    |
| 77.15  | 34816.0   |
| 78.15  | 2133.0    |
| 79.15  | 2099.0    |
| 80.15  | 1320.0    |
| 81.05  | 2284.0    |
| 82.15  | 602.0     |
| 83.15  | 644.0     |
| 85.05  | 398.0     |
| 86.15  | 468.0     |
| 87.15  | 318.0     |
| 89.15  | 159.0     |
| 91.05  | 606.0     |
| 92.15  | 437.0     |
| 93.15  | 3171.0    |
| 94.05  | 192.0     |
| 96.05  | 245.0     |
| 98.15  | 2241.0    |
| 99.15  | 2244.0    |
| 100.15 | 239.0     |
| 101.05 | 1117.0    |
| 103.15 | 429.0     |
| 104.05 | 553.0     |
| 105.05 | 697.0     |
| 106.15 | 409.0     |
| 107.15 | 8349.0    |
| 108.05 | 1578.0    |
| 110.05 | 17784.0   |
| 111.05 | 3100.0    |
| 112.15 | 422.0     |
| 113.05 | 186.0     |
| 116.05 | 760.0     |
| 117.15 | 6311.0    |
| 118.05 | 627.0     |
| 122.15 | 504.0     |
| 123.15 | 919.0     |
| 124.05 | 462.0     |
| 125.05 | 315.0     |
| 127.15 | 33928.0   |
| 128.15 | 2698.0    |
| 129.05 | 11196.0   |
| 130.05 | 972.0     |
| 131.05 | 285.0     |
| 134.25 | 335.0     |
| 135.05 | 1094.0    |
| 136.05 | 371.0     |
| 137.05 | 583.0     |
| 140.15 | 331.0     |
| 141.05 | 1326.0    |
| 142.15 | 639.0     |
| 143.05 | 428.0     |
| 146.15 | 307.0     |
| 147.05 | 983.0     |
| 148.05 | 1440.0    |
| 149.05 | 426.0     |
| 153.05 | 461.0     |
| 154.05 | 322.0     |
| 155.05 | 907.0     |
| 156.05 | 1516.0    |
| 157.05 | 230.0     |
| 157.95 | 290.0     |
| 158.95 | 290.0     |
| 160.05 | 473.0     |
| 161.05 | 549.0     |
| 165.05 | 716.0     |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP02

|        |         |
|--------|---------|
| 166.05 | 491.0   |
| 167.05 | 2826.0  |
| 168.05 | 1556.0  |
| 169.15 | 345.0   |
| 171.05 | 154.0   |
| 172.15 | 338.0   |
| 173.05 | 340.0   |
| 174.05 | 689.0   |
| 175.15 | 815.0   |
| 176.05 | 202.0   |
| 176.35 | 191.0   |
| 177.05 | 518.0   |
| 179.05 | 2211.0  |
| 180.05 | 1502.0  |
| 181.05 | 735.0   |
| 182.05 | 234.0   |
| 184.15 | 196.0   |
| 185.05 | 1267.0  |
| 186.05 | 7487.0  |
| 187.05 | 2325.0  |
| 188.05 | 237.0   |
| 189.05 | 581.0   |
| 191.15 | 243.0   |
| 192.05 | 819.0   |
| 193.05 | 965.0   |
| 195.05 | 156.0   |
| 196.05 | 1919.0  |
| 198.05 | 59968.0 |
| 199.05 | 4253.0  |
| 200.05 | 277.0   |
| 201.25 | 271.0   |
| 203.05 | 427.0   |
| 204.05 | 2002.0  |
| 205.05 | 3411.0  |
| 206.15 | 15590.0 |
| 207.05 | 2243.0  |
| 208.15 | 483.0   |
| 209.05 | 161.0   |
| 210.15 | 225.0   |
| 211.15 | 367.0   |
| 215.15 | 200.0   |
| 216.05 | 352.0   |
| 217.05 | 3638.0  |
| 218.05 | 363.0   |
| 221.05 | 3539.0  |
| 223.05 | 921.0   |
| 224.05 | 7991.0  |
| 225.05 | 1859.0  |
| 226.05 | 211.0   |
| 227.05 | 3199.0  |
| 228.05 | 399.0   |
| 229.05 | 892.0   |
| 231.15 | 302.0   |
| 233.95 | 200.0   |
| 235.05 | 187.0   |
| 237.05 | 325.0   |
| 241.05 | 273.0   |
| 242.05 | 681.0   |
| 243.15 | 659.0   |
| 244.05 | 6908.0  |
| 245.15 | 839.0   |
| 246.05 | 903.0   |
| 247.15 | 259.0   |
| 249.05 | 294.0   |
| 255.05 | 30408.0 |
| 256.05 | 4526.0  |
| 257.05 | 397.0   |
| 258.05 | 1684.0  |
| 258.95 | 325.0   |
| 265.05 | 961.0   |
| 265.95 | 192.0   |
| 273.05 | 961.0   |
| 274.05 | 2133.0  |
| 275.05 | 13519.0 |
| 276.05 | 1984.0  |
| 277.05 | 1042.0  |
| 277.95 | 297.0   |
| 282.95 | 154.0   |
| 289.05 | 151.0   |
| 293.15 | 247.0   |
| 295.15 | 164.0   |
| 296.15 | 3983.0  |
| 297.15 | 576.0   |
| 303.05 | 463.0   |
| 304.05 | 186.0   |
| 314.05 | 157.0   |
| 315.05 | 387.0   |
| 316.05 | 309.0   |
| 323.05 | 1174.0  |
| 324.05 | 276.0   |
| 326.95 | 266.0   |
| 334.05 | 752.0   |
| 335.15 | 152.0   |
| 345.95 | 292.0   |
| 352.05 | 355.0   |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP02

|        |         |
|--------|---------|
| 353.15 | 442.0   |
| 354.15 | 260.0   |
| 365.05 | 1537.0  |
| 366.05 | 347.0   |
| 372.05 | 588.0   |
| 372.95 | 263.0   |
| 383.05 | 197.0   |
| 402.05 | 211.0   |
| 403.05 | 402.0   |
| 404.05 | 179.0   |
| 421.05 | 362.0   |
| 422.05 | 342.0   |
| 423.05 | 2419.0  |
| 424.05 | 510.0   |
| 441.15 | 6430.0  |
| 442.15 | 42600.0 |
| 443.15 | 8473.0  |
| 444.05 | 818.0   |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP02

| m/z    | Abundance |
|--------|-----------|
| 38.15  | 457.0     |
| 39.15  | 2644.0    |
| 40.15  | 1031.0    |
| 41.15  | 287.0     |
| 43.15  | 234.0     |
| 44.15  | 1382.0    |
| 49.15  | 264.0     |
| 50.15  | 7640.0    |
| 51.15  | 31560.0   |
| 52.15  | 1732.0    |
| 55.15  | 406.0     |
| 56.15  | 1061.0    |
| 57.15  | 2514.0    |
| 61.05  | 479.0     |
| 62.15  | 417.0     |
| 63.15  | 1409.0    |
| 64.15  | 278.0     |
| 65.15  | 590.0     |
| 69.15  | 34120.0   |
| 70.15  | 537.0     |
| 73.15  | 341.0     |
| 74.15  | 3399.0    |
| 75.15  | 5501.0    |
| 76.15  | 1766.0    |
| 77.15  | 40264.0   |
| 78.15  | 2566.0    |
| 79.15  | 2332.0    |
| 80.05  | 1882.0    |
| 81.15  | 2749.0    |
| 82.15  | 634.0     |
| 83.05  | 684.0     |
| 85.15  | 383.0     |
| 86.05  | 724.0     |
| 87.15  | 279.0     |
| 91.15  | 507.0     |
| 92.15  | 615.0     |
| 93.15  | 3221.0    |
| 94.05  | 383.0     |
| 95.25  | 190.0     |
| 96.15  | 238.0     |
| 98.15  | 2457.0    |
| 99.05  | 2327.0    |
| 100.15 | 212.0     |
| 101.05 | 1666.0    |
| 103.05 | 846.0     |
| 104.05 | 907.0     |
| 105.05 | 713.0     |
| 106.15 | 502.0     |
| 107.15 | 11132.0   |
| 108.05 | 1787.0    |
| 110.05 | 20392.0   |
| 111.05 | 3767.0    |
| 112.05 | 511.0     |
| 113.15 | 203.0     |
| 116.15 | 706.0     |
| 117.05 | 6739.0    |
| 118.05 | 565.0     |
| 122.05 | 624.0     |
| 123.05 | 1210.0    |
| 124.15 | 495.0     |
| 125.05 | 459.0     |
| 127.15 | 39840.0   |
| 128.05 | 3647.0    |
| 129.05 | 14419.0   |
| 130.05 | 1551.0    |
| 130.95 | 227.0     |
| 132.15 | 166.0     |
| 134.05 | 449.0     |
| 135.15 | 1372.0    |
| 136.05 | 377.0     |
| 137.05 | 686.0     |
| 137.95 | 159.0     |
| 140.05 | 179.0     |
| 141.05 | 1545.0    |
| 142.05 | 673.0     |
| 143.05 | 414.0     |
| 145.05 | 164.0     |
| 146.15 | 381.0     |
| 147.15 | 836.0     |
| 148.05 | 2140.0    |
| 149.05 | 694.0     |
| 149.95 | 198.0     |
| 151.05 | 218.0     |
| 152.05 | 230.0     |
| 153.05 | 586.0     |
| 154.05 | 562.0     |
| 155.05 | 1240.0    |
| 156.15 | 1963.0    |
| 157.15 | 535.0     |
| 158.05 | 338.0     |
| 159.05 | 269.0     |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP02

|        |         |
|--------|---------|
| 160.05 | 654.0   |
| 161.05 | 981.0   |
| 161.85 | 267.0   |
| 164.05 | 165.0   |
| 165.05 | 535.0   |
| 166.05 | 662.0   |
| 167.05 | 3997.0  |
| 168.05 | 2128.0  |
| 169.15 | 343.0   |
| 170.15 | 210.0   |
| 171.95 | 394.0   |
| 172.95 | 374.0   |
| 174.05 | 805.0   |
| 175.15 | 1523.0  |
| 176.15 | 542.0   |
| 177.05 | 553.0   |
| 178.05 | 381.0   |
| 179.05 | 2889.0  |
| 180.15 | 1775.0  |
| 181.05 | 710.0   |
| 182.15 | 161.0   |
| 184.15 | 235.0   |
| 185.05 | 1665.0  |
| 186.15 | 10446.0 |
| 187.15 | 3019.0  |
| 188.05 | 272.0   |
| 188.95 | 719.0   |
| 190.05 | 174.0   |
| 191.05 | 288.0   |
| 192.05 | 718.0   |
| 193.05 | 945.0   |
| 194.05 | 241.0   |
| 195.15 | 167.0   |
| 196.05 | 2771.0  |
| 198.05 | 81320.0 |
| 199.05 | 5717.0  |
| 200.05 | 534.0   |
| 201.55 | 605.0   |
| 203.05 | 554.0   |
| 204.05 | 3122.0  |
| 205.05 | 4610.0  |
| 206.15 | 19328.0 |
| 207.05 | 2679.0  |
| 208.05 | 762.0   |
| 209.05 | 237.0   |
| 210.45 | 553.0   |
| 211.05 | 703.0   |
| 212.05 | 162.0   |
| 215.05 | 248.0   |
| 216.05 | 433.0   |
| 217.05 | 5399.0  |
| 218.05 | 601.0   |
| 221.05 | 4424.0  |
| 223.05 | 1543.0  |
| 224.05 | 11587.0 |
| 225.05 | 2774.0  |
| 226.15 | 200.0   |
| 227.05 | 5054.0  |
| 228.05 | 714.0   |
| 229.05 | 929.0   |
| 229.85 | 232.0   |
| 231.05 | 426.0   |
| 233.95 | 285.0   |
| 235.05 | 380.0   |
| 236.05 | 188.0   |
| 237.05 | 264.0   |
| 241.05 | 370.0   |
| 242.15 | 688.0   |
| 243.05 | 649.0   |
| 244.15 | 8762.0  |
| 245.15 | 1224.0  |
| 246.05 | 2153.0  |
| 247.05 | 425.0   |
| 249.05 | 344.0   |
| 253.05 | 323.0   |
| 255.05 | 44360.0 |
| 256.05 | 6785.0  |
| 257.15 | 622.0   |
| 258.05 | 2965.0  |
| 259.05 | 396.0   |
| 265.05 | 965.0   |
| 266.05 | 209.0   |
| 272.05 | 218.0   |
| 273.05 | 1268.0  |
| 274.05 | 3816.0  |
| 275.05 | 20288.0 |
| 276.05 | 2433.0  |
| 277.05 | 1457.0  |
| 278.15 | 225.0   |
| 283.15 | 157.0   |
| 284.95 | 338.0   |
| 292.95 | 310.0   |
| 296.05 | 5668.0  |
| 297.05 | 1035.0  |
| 303.15 | 712.0   |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP02

|        |         |
|--------|---------|
| 303.95 | 166.0   |
| 314.05 | 393.0   |
| 315.05 | 589.0   |
| 316.05 | 389.0   |
| 321.05 | 353.0   |
| 323.05 | 1914.0  |
| 324.05 | 443.0   |
| 327.05 | 238.0   |
| 328.05 | 230.0   |
| 332.15 | 160.0   |
| 333.05 | 339.0   |
| 334.05 | 1132.0  |
| 335.05 | 274.0   |
| 341.05 | 294.0   |
| 346.05 | 361.0   |
| 352.05 | 587.0   |
| 353.15 | 501.0   |
| 354.05 | 601.0   |
| 365.05 | 2854.0  |
| 366.05 | 515.0   |
| 371.15 | 250.0   |
| 372.15 | 976.0   |
| 373.05 | 285.0   |
| 382.95 | 298.0   |
| 402.05 | 551.0   |
| 403.05 | 766.0   |
| 404.05 | 266.0   |
| 421.15 | 663.0   |
| 422.05 | 406.0   |
| 423.05 | 4517.0  |
| 424.15 | 612.0   |
| 441.15 | 12283.0 |
| 442.15 | 78424.0 |
| 443.15 | 16161.0 |
| 444.15 | 1468.0  |
| 445.15 | 159.0   |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP02

| m/z    | Abundance |
|--------|-----------|
| 38.15  | 298.0     |
| 39.15  | 1408.0    |
| 40.15  | 1409.0    |
| 41.15  | 205.0     |
| 43.15  | 290.0     |
| 44.15  | 1417.0    |
| 49.15  | 283.0     |
| 50.15  | 5082.0    |
| 51.15  | 19984.0   |
| 52.15  | 827.0     |
| 55.15  | 316.0     |
| 56.15  | 592.0     |
| 57.15  | 2119.0    |
| 61.15  | 236.0     |
| 62.25  | 294.0     |
| 63.15  | 887.0     |
| 65.15  | 543.0     |
| 69.15  | 21968.0   |
| 70.15  | 275.0     |
| 73.15  | 230.0     |
| 74.15  | 1973.0    |
| 75.15  | 3402.0    |
| 76.15  | 1199.0    |
| 77.15  | 27160.0   |
| 78.15  | 1903.0    |
| 79.15  | 1710.0    |
| 80.05  | 1423.0    |
| 81.15  | 1548.0    |
| 82.05  | 587.0     |
| 83.15  | 427.0     |
| 83.85  | 166.0     |
| 85.05  | 394.0     |
| 86.05  | 604.0     |
| 87.05  | 218.0     |
| 91.15  | 563.0     |
| 92.15  | 641.0     |
| 93.05  | 1932.0    |
| 94.05  | 191.0     |
| 96.05  | 180.0     |
| 98.15  | 1963.0    |
| 99.15  | 1717.0    |
| 101.15 | 941.0     |
| 103.05 | 305.0     |
| 104.05 | 753.0     |
| 105.15 | 729.0     |
| 107.15 | 6607.0    |
| 108.05 | 1268.0    |
| 110.15 | 14104.0   |
| 111.15 | 2237.0    |
| 112.05 | 386.0     |
| 114.05 | 154.0     |
| 117.05 | 4326.0    |
| 118.15 | 338.0     |
| 122.05 | 504.0     |
| 123.05 | 836.0     |
| 124.05 | 283.0     |
| 125.05 | 315.0     |
| 127.15 | 28424.0   |
| 128.05 | 2405.0    |
| 129.05 | 10760.0   |
| 130.05 | 783.0     |
| 131.05 | 177.0     |
| 134.15 | 402.0     |
| 135.15 | 783.0     |
| 136.15 | 331.0     |
| 136.95 | 400.0     |
| 139.95 | 283.0     |
| 141.05 | 1372.0    |
| 142.15 | 390.0     |
| 143.05 | 247.0     |
| 144.15 | 157.0     |
| 147.05 | 602.0     |
| 148.05 | 1418.0    |
| 149.05 | 293.0     |
| 153.05 | 552.0     |
| 154.15 | 465.0     |
| 155.05 | 807.0     |
| 156.05 | 1211.0    |
| 157.95 | 298.0     |
| 159.15 | 181.0     |
| 160.05 | 439.0     |
| 161.05 | 664.0     |
| 162.15 | 184.0     |
| 165.05 | 500.0     |
| 166.05 | 358.0     |
| 167.05 | 2404.0    |
| 168.05 | 1276.0    |
| 172.05 | 238.0     |
| 173.15 | 364.0     |
| 174.15 | 685.0     |
| 175.15 | 969.0     |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP02



|        |         |
|--------|---------|
| 176.05 | 246.0   |
| 177.05 | 487.0   |
| 177.95 | 180.0   |
| 179.05 | 2054.0  |
| 180.05 | 1246.0  |
| 181.05 | 696.0   |
| 183.95 | 155.0   |
| 185.05 | 878.0   |
| 186.15 | 8247.0  |
| 187.05 | 2287.0  |
| 188.25 | 319.0   |
| 189.05 | 324.0   |
| 191.25 | 153.0   |
| 192.05 | 680.0   |
| 193.05 | 768.0   |
| 193.95 | 180.0   |
| 196.05 | 1897.0  |
| 198.05 | 56976.0 |
| 199.05 | 3836.0  |
| 200.05 | 304.0   |
| 201.25 | 221.0   |
| 201.85 | 186.0   |
| 203.05 | 498.0   |
| 204.05 | 2408.0  |
| 205.05 | 3468.0  |
| 206.15 | 13962.0 |
| 207.15 | 2244.0  |
| 208.05 | 477.0   |
| 209.05 | 232.0   |
| 210.05 | 290.0   |
| 210.95 | 475.0   |
| 215.95 | 285.0   |
| 217.05 | 4712.0  |
| 217.95 | 431.0   |
| 221.05 | 2690.0  |
| 223.05 | 1063.0  |
| 224.05 | 8613.0  |
| 225.05 | 1735.0  |
| 227.05 | 3384.0  |
| 228.05 | 445.0   |
| 229.05 | 637.0   |
| 230.15 | 199.0   |
| 231.05 | 198.0   |
| 234.15 | 247.0   |
| 235.05 | 215.0   |
| 235.95 | 201.0   |
| 236.95 | 254.0   |
| 239.05 | 165.0   |
| 242.15 | 455.0   |
| 243.05 | 564.0   |
| 244.15 | 6781.0  |
| 245.15 | 986.0   |
| 246.05 | 1245.0  |
| 247.05 | 214.0   |
| 248.95 | 356.0   |
| 255.05 | 34568.0 |
| 256.05 | 4706.0  |
| 257.15 | 230.0   |
| 258.05 | 2291.0  |
| 259.05 | 280.0   |
| 265.05 | 749.0   |
| 271.15 | 222.0   |
| 272.15 | 185.0   |
| 272.95 | 1106.0  |
| 274.05 | 2831.0  |
| 275.05 | 14456.0 |
| 276.05 | 2031.0  |
| 277.05 | 1175.0  |
| 278.15 | 331.0   |
| 285.05 | 208.0   |
| 293.05 | 259.0   |
| 294.15 | 163.0   |
| 296.05 | 4395.0  |
| 297.05 | 545.0   |
| 303.05 | 558.0   |
| 304.05 | 167.0   |
| 314.05 | 286.0   |
| 315.05 | 463.0   |
| 316.05 | 362.0   |
| 323.15 | 1739.0  |
| 324.05 | 350.0   |
| 327.05 | 302.0   |
| 328.05 | 189.0   |
| 333.05 | 229.0   |
| 334.05 | 1135.0  |
| 335.05 | 300.0   |
| 341.05 | 264.0   |
| 346.05 | 410.0   |
| 352.05 | 508.0   |
| 353.05 | 387.0   |
| 354.15 | 441.0   |
| 365.05 | 2352.0  |
| 366.05 | 322.0   |
| 370.95 | 207.0   |
| 372.15 | 1081.0  |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP02

|        |         |
|--------|---------|
| 373.05 | 208.0   |
| 383.15 | 302.0   |
| 390.15 | 212.0   |
| 402.05 | 354.0   |
| 403.05 | 714.0   |
| 420.95 | 430.0   |
| 422.05 | 471.0   |
| 423.05 | 3870.0  |
| 424.05 | 863.0   |
| 441.15 | 11643.0 |
| 442.15 | 73864.0 |
| 443.15 | 13911.0 |
| 444.05 | 1303.0  |

**Instrument :**  
BNA\_G  
**ClientSampleId :**  
DFTPP02

Scan 2538 (17.597 min): BG015978.D  
DFTPP02

| m/z    | Abundance |
|--------|-----------|
| 40.15  | 1040.0    |
| 41.25  | 250.0     |
| 44.15  | 1494.0    |
| 55.15  | 173.0     |
| 57.25  | 394.0     |
| 85.95  | 172.0     |
| 207.05 | 482.0     |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP02