

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG021116\
 Data File : BG020860.D
 Acq On : 11 Feb 2016 17:30
 Operator : SJ/UM
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 11 22:42:45 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG021116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 11 15:26:53 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	91	0.00
2	1,4-Dioxane	40.000	39.680	0.8	93	0.00
3	Pyridine	40.000	42.605	-6.5	94	0.00
4	n-Nitrosodimethylamine	40.000	42.306	-5.8	94	0.00
5 S	2-Fluorophenol	80.000	81.867	-2.3	93	0.00
6	Aniline	40.000	41.743	-4.4	94	0.00
7 S	Phenol-d6	80.000	83.530	-4.4	94	0.00
8	2-Chlorophenol	40.000	41.195	-3.0	94	0.00
9	Benzaldehyde	40.000	42.255	-5.6	96	0.00
10 C	Phenol	40.000	41.697	-4.2	95	0.00
11	bis(2-Chloroethyl)ether	40.000	40.702	-1.8	91	0.00
12	1,3-Dichlorobenzene	40.000	41.033	-2.6	94	0.00
13 C	1,4-Dichlorobenzene	40.000	41.426	-3.6	94	0.00
14	1,2-Dichlorobenzene	40.000	41.360	-3.4	93	0.00
15	Benzyl Alcohol	40.000	41.951	-4.9	94	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	41.195	-3.0	93	0.00
17	2-Methylphenol	40.000	41.339	-3.3	94	0.00
18	Hexachloroethane	40.000	41.704	-4.3	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.466	-6.2	97	0.00
20	3+4-Methylphenols	40.000	42.337	-5.8	97	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	95	0.00
22	Acetophenone	40.000	40.422	-1.1	95	0.00
23 S	Nitrobenzene-d5	80.000	81.470	-1.8	96	0.00
24	Nitrobenzene	40.000	40.260	-0.6	96	0.00
25	Isophorone	40.000	41.036	-2.6	96	0.00
26 C	2-Nitrophenol	40.000	41.725	-4.3	94	0.00
27	2,4-Dimethylphenol	40.000	41.090	-2.7	95	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.107	-2.8	97	0.00
29 C	2,4-Dichlorophenol	40.000	41.002	-2.5	98	0.00
30	1,2,4-Trichlorobenzene	40.000	40.813	-2.0	97	0.00
31	Naphthalene	40.000	40.559	-1.4	96	0.00
32	Benzoic acid	40.000	40.362	-0.9	97	0.00
33	4-Chloroaniline	40.000	40.889	-2.2	97	0.00
34 C	Hexachlorobutadiene	40.000	40.943	-2.4	97	0.00
35	Caprolactam	40.000	42.506	-6.3	102	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.668	-4.2	100	0.00
37	2-Methylnaphthalene	40.000	41.280	-3.2	99	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	40.021	-0.1	99	0.00
40 P	Hexachlorocyclopentadiene	40.000	38.944	2.6	94	0.00
41 S	2,4,6-Tribromophenol	80.000	84.731	-5.9	105	0.00
42 C	2,4,6-Trichlorophenol	40.000	40.847	-2.1	99	0.00
43	2,4,5-Trichlorophenol	40.000	41.319	-3.3	102	0.00
44 S	2-Fluorobiphenyl	80.000	80.577	-0.7	99	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG021116\
 Data File : BG020860.D
 Acq On : 11 Feb 2016 17:30
 Operator : SJ/UM
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 11 22:42:45 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG021116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 11 15:26:53 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	40.029	-0.1	99	0.00
46	2-Chloronaphthalene	40.000	39.930	0.2	100	0.00
47	2-Nitroaniline	40.000	41.516	-3.8	102	0.00
48	Acenaphthylene	40.000	40.919	-2.3	102	0.00
49	Dimethylphthalate	40.000	41.187	-3.0	102	0.00
50	2,6-Dinitrotoluene	40.000	41.080	-2.7	101	0.00
51 C	Acenaphthene	40.000	40.469	-1.2	101	0.00
52	3-Nitroaniline	40.000	40.755	-1.9	100	0.00
53 P	2,4-Dinitrophenol	40.000	37.102	7.2	97	0.00
54	Dibenzofuran	40.000	40.514	-1.3	102	0.00
55 P	4-Nitrophenol	40.000	42.055	-5.1	101	0.00
56	2,4-Dinitrotoluene	40.000	42.668	-6.7	103	0.00
57	Fluorene	40.000	41.071	-2.7	103	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	41.129	-2.8	102	0.00
59	Diethylphthalate	40.000	40.867	-2.2	102	0.00
60	4-Chlorophenyl-phenylether	40.000	40.724	-1.8	104	0.00
61	4-Nitroaniline	40.000	41.919	-4.8	103	0.00
62	Azobenzene	40.000	41.021	-2.6	103	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
64	4,6-Dinitro-2-methylphenol	40.000	39.228	1.9	103	0.00
65 c	n-Nitrosodiphenylamine	40.000	41.453	-3.6	103	0.00
66	4-Bromophenyl-phenylether	40.000	41.194	-3.0	103	0.00
67	Hexachlorobenzene	40.000	40.787	-2.0	103	0.00
68	Atrazine	40.000	40.719	-1.8	103	0.00
69 C	Pentachlorophenol	40.000	40.568	-1.4	102	0.00
70	Phenanthrene	40.000	40.751	-1.9	103	0.00
71	Anthracene	40.000	40.673	-1.7	102	0.00
72	Carbazole	40.000	41.138	-2.8	104	0.00
73	Di-n-butylphthalate	40.000	41.145	-2.9	103	0.00
74 C	Fluoranthene	40.000	40.947	-2.4	103	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
76	Benzidine	40.000	39.858	0.4	98	0.00
77	Pyrene	40.000	41.208	-3.0	105	0.00
78 S	Terphenyl-d14	80.000	83.909	-4.9	106	0.00
79	Butylbenzylphthalate	40.000	41.193	-3.0	102	0.00
80	Benzo(a)anthracene	40.000	41.129	-2.8	105	0.00
81	3,3'-Dichlorobenzidine	40.000	41.353	-3.4	104	0.00
82	Chrysene	40.000	41.159	-2.9	104	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	41.388	-3.5	104	0.00
84 c	Di-n-octyl phthalate	40.000	41.504	-3.8	103	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	41.999	-5.0	105	0.00
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Benzo(b)fluoranthene	40.000	42.027	-5.1	106	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG021116\
Data File : BG020860.D
Acq On : 11 Feb 2016 17:30
Operator : SJ/UM
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Feb 11 22:42:45 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG021116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 11 15:26:53 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	40.691	-1.7	104	0.00
89 C	Benzo(a)pyrene	40.000	41.126	-2.8	104	0.00
90	Dibenzo(a,h)anthracene	40.000	41.418	-3.5	106	0.00
91	Benzo(a,h,i)perylene	40.000	41.070	-2.7	104	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0