

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG021116\
 Data File : BG020860.D
 Acq On : 11 Feb 2016 17:30
 Operator : SJ/UM
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 11 22:42:45 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG021116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 11 15:26:53 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2	1,4-Dioxane	0.432	0.428	0.9	93	0.00
3	Pyridine	1.277	1.360	-6.5	94	0.00
4	n-Nitrosodimethylamine	0.334	0.353	-5.7	94	0.00
5 S	2-Fluorophenol	1.188	1.215	-2.3	93	0.00
6	Aniline	2.164	2.259	-4.4	94	0.00
7 S	Phenol-d6	1.736	1.812	-4.4	94	0.00
8	2-Chlorophenol	1.484	1.528	-3.0	94	0.00
9	Benzaldehyde	0.869	0.918	-5.6	96	0.00
10 C	Phenol	1.837	1.915	-4.2	95	0.00
11	bis(2-Chloroethyl)ether	1.467	1.493	-1.8	91	0.00
12	1,3-Dichlorobenzene	1.569	1.609	-2.5	94	0.00
13 C	1,4-Dichlorobenzene	1.590	1.647	-3.6	94	0.00
14	1,2-Dichlorobenzene	1.547	1.600	-3.4	93	0.00
15	Benzyl Alcohol	1.094	1.148	-4.9	94	0.00
16	2,2'-oxybis(1-Chloropropane	1.530	1.576	-3.0	93	0.00
17	2-Methylphenol	1.323	1.368	-3.4	94	0.00
18	Hexachloroethane	0.547	0.570	-4.2	94	0.00
19 P	n-Nitroso-di-n-propylamine	1.120	1.189	-6.2	97	0.00
20	3+4-Methylphenols	1.844	1.952	-5.9	97	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
22	Acetophenone	0.482	0.487	-1.0	95	0.00
23 S	Nitrobenzene-d5	0.304	0.310	-2.0	96	0.00
24	Nitrobenzene	0.317	0.319	-0.6	96	0.00
25	Isophorone	0.641	0.658	-2.7	96	0.00
26 C	2-Nitrophenol	0.167	0.174	-4.2	94	0.00
27	2,4-Dimethylphenol	0.320	0.328	-2.5	95	0.00
28	bis(2-Chloroethoxy)methane	0.395	0.406	-2.8	97	0.00
29 C	2,4-Dichlorophenol	0.287	0.294	-2.4	98	0.00
30	1,2,4-Trichlorobenzene	0.289	0.295	-2.1	97	0.00
31	Naphthalene	1.035	1.049	-1.4	96	0.00
32	Benzoic acid	0.256	0.259	-1.2	97	0.00
33	4-Chloroaniline	0.445	0.455	-2.2	97	0.00
34 C	Hexachlorobutadiene	0.147	0.151	-2.7	97	0.00
35	Caprolactam	0.110	0.117	-6.4	102	0.00
36 C	4-Chloro-3-methylphenol	0.328	0.342	-4.3	100	0.00
37	2-Methylnaphthalene	0.726	0.749	-3.2	99	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
39	1,2,4,5-Tetrachlorobenzene	0.587	0.587	0.0	99	0.00
40 P	Hexachlorocyclopentadiene	0.258	0.251	2.7	94	0.00
41 S	2,4,6-Tribromophenol	0.197	0.208	-5.6	105	0.00
42 C	2,4,6-Trichlorophenol	0.394	0.402	-2.0	99	0.00
43	2,4,5-Trichlorophenol	0.414	0.428	-3.4	102	0.00
44 S	2-Fluorobiphenyl	1.424	1.434	-0.7	99	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG021116\
 Data File : BG020860.D
 Acq On : 11 Feb 2016 17:30
 Operator : SJ/UM
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 11 22:42:45 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG021116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 11 15:26:53 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.764	1.766	-0.1	99	0.00
46	2-Chloronaphthalene	1.263	1.261	0.2	100	0.00
47	2-Nitroaniline	0.303	0.314	-3.6	102	0.00
48	Acenaphthylene	2.196	2.247	-2.3	102	0.00
49	Dimethylphthalate	1.564	1.610	-2.9	102	0.00
50	2,6-Dinitrotoluene	0.335	0.344	-2.7	101	0.00
51 C	Acenaphthene	1.332	1.347	-1.1	101	0.00
52	3-Nitroaniline	0.389	0.397	-2.1	100	0.00
53 P	2,4-Dinitrophenol	0.116	0.114	1.7	97	0.00
54	Dibenzofuran	1.762	1.784	-1.2	102	0.00
55 P	4-Nitrophenol	0.294	0.309	-5.1	101	0.00
56	2,4-Dinitrotoluene	0.433	0.462	-6.7	103	0.00
57	Fluorene	1.622	1.665	-2.7	103	0.00
58	2,3,4,6-Tetrachlorophenol	0.316	0.325	-2.8	102	0.00
59	Diethylphthalate	1.608	1.643	-2.2	102	0.00
60	4-Chlorophenyl-phenylether	0.720	0.734	-1.9	104	0.00
61	4-Nitroaniline	0.391	0.410	-4.9	103	0.00
62	Azobenzene	1.264	1.296	-2.5	103	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
64	4,6-Dinitro-2-methylphenol	0.094	0.097	-3.2	103	0.00
65 c	n-Nitrosodiphenylamine	0.646	0.670	-3.7	103	0.00
66	4-Bromophenyl-phenylether	0.211	0.218	-3.3	103	0.00
67	Hexachlorobenzene	0.225	0.229	-1.8	103	0.00
68	Atrazine	0.196	0.200	-2.0	103	0.00
69 C	Pentachlorophenol	0.129	0.131	-1.6	102	0.00
70	Phenanthrene	1.139	1.160	-1.8	103	0.00
71	Anthracene	1.156	1.175	-1.6	102	0.00
72	Carbazole	1.037	1.067	-2.9	104	0.00
73	Di-n-butylphthalate	1.330	1.368	-2.9	103	0.00
74 C	Fluoranthene	1.230	1.259	-2.4	103	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
76	Benzidine	0.647	0.644	0.5	98	0.00
77	Pyrene	1.319	1.359	-3.0	105	0.00
78 S	Terphenyl-d14	0.857	0.899	-4.9	106	0.00
79	Butylbenzylphthalate	0.632	0.650	-2.8	102	0.00
80	Benzo(a)anthracene	1.194	1.228	-2.8	105	0.00
81	3,3'-Dichlorobenzidine	0.443	0.458	-3.4	104	0.00
82	Chrysene	1.123	1.155	-2.8	104	0.00
83	Bis(2-ethylhexyl)phthalate	0.936	0.969	-3.5	104	0.00
84 c	Di-n-octyl phthalate	1.537	1.595	-3.8	103	0.00
85	Indeno(1,2,3-cd)pyrene	1.291	1.356	-5.0	105	0.00
86 I	Perylene-d12	1.000	1.000	0.0	102	0.00
87	Benzo(b)fluoranthene	1.224	1.286	-5.1	106	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG021116\
Data File : BG020860.D
Acq On : 11 Feb 2016 17:30
Operator : SJ/UM
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Feb 11 22:42:45 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG021116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 11 15:26:53 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.199	1.220	-1.8	104	0.00
89 C	Benzo(a)pyrene	1.166	1.199	-2.8	104	0.00
90	Dibenzo(a,h)anthracene	1.134	1.174	-3.5	106	0.00
91	Benzo(a,h,i)perylene	1.126	1.156	-2.7	104	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0