

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG021122\
 Data File : BG052366.D
 Acq On : 11 Feb 2022 10:35
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 11 11:23:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 08 01:21:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	80	0.00
2	1,4-Dioxane	40.000	34.981	12.5	74	0.00
3	Pyridine	40.000	36.796	8.0	73	0.00
4	n-Nitrosodimethylamine	40.000	34.026	14.9	70	0.00
5 S	2-Fluorophenol	80.000	79.778	0.3	81	0.01
6	Aniline	40.000	38.344	4.1	75	0.00
7 S	Phenol-d6	80.000	77.053	3.7	76	0.01
8	2-Chlorophenol	40.000	38.675	3.3	80	0.00
9	Benzaldehyde	40.000	34.245	14.4	70	0.00
10 C	Phenol	40.000	38.598	3.5	77	0.01
11	bis(2-Chloroethyl)ether	40.000	36.746	8.1	76	0.00
12	1,3-Dichlorobenzene	40.000	39.713	0.7	81	0.00
13 C	1,4-Dichlorobenzene	40.000	39.951	0.1	82	0.00
14	1,2-Dichlorobenzene	40.000	38.712	3.2	80	0.00
15	Benzyl Alcohol	40.000	37.970	5.1	74	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.454	11.4	74	0.02
17	2-Methylphenol	40.000	39.110	2.2	78	0.01
18	Hexachloroethane	40.000	40.149	-0.4	84	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.526	8.7	74	0.00
20	3+4-Methylphenols	40.000	38.615	3.5	77	0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	78	0.00
22	Acetophenone	40.000	37.488	6.3	75	0.00
23 S	Nitrobenzene-d5	80.000	75.317	5.9	75	0.00
24	Nitrobenzene	40.000	36.398	9.0	71	0.00
25	Isophorone	40.000	37.860	5.4	74	0.00
26 C	2-Nitrophenol	40.000	40.876	-2.2	78	0.00
27	2,4-Dimethylphenol	40.000	39.751	0.6	77	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.270	4.3	76	0.00
29 C	2,4-Dichlorophenol	40.000	40.061	-0.2	79	0.00
30	1,2,4-Trichlorobenzene	40.000	38.541	3.6	77	0.00
31	Naphthalene	40.000	38.365	4.1	76	0.01
32	Benzoic acid	40.000	44.632	-11.6	94	0.00
33	4-Chloroaniline	40.000	39.189	2.0	76	0.00
34 C	Hexachlorobutadiene	40.000	37.717	5.7	77	0.00
35	Caprolactam	40.000	41.351	-3.4	80	0.01
36 C	4-Chloro-3-methylphenol	40.000	40.180	-0.4	77	0.00
37	2-Methylnaphthalene	40.000	38.556	3.6	77	0.00
38	1-Methylnaphthalene	40.000	39.182	2.0	77	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	79	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.383	6.5	76	0.00
41 P	Hexachlorocyclopentadiene	40.000	34.016	15.0	69	0.00
42 S	2,4,6-Tribromophenol	80.000	82.350	-2.9	82	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.123	2.2	77	0.00
44	2,4,5-Trichlorophenol	40.000	39.015	2.5	78	0.00
45 S	2-Fluorobiphenyl	80.000	76.019	5.0	77	0.00
46	1,1'-Biphenyl	40.000	37.824	5.4	76	0.00
47	2-Chloronaphthalene	40.000	37.836	5.4	76	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG021122\
 Data File : BG052366.D
 Acq On : 11 Feb 2022 10:35
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 11 11:23:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 08 01:21:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.041	4.9	76	0.00
49	Acenaphthylene	40.000	38.314	4.2	78	0.00
50	Dimethylphthalate	40.000	38.492	3.8	78	0.00
51	2,6-Dinitrotoluene	40.000	39.514	1.2	78	0.00
52 C	Acenaphthene	40.000	38.617	3.5	78	0.00
53	3-Nitroaniline	40.000	41.102	-2.8	81	0.00
54 P	2,4-Dinitrophenol	40.000	39.767	0.6	81	0.00
55	Dibenzofuran	40.000	37.771	5.6	78	0.00
56 P	4-Nitrophenol	40.000	41.626	-4.1	80	0.00
57	2,4-Dinitrotoluene	40.000	40.962	-2.4	81	0.00
58	Fluorene	40.000	38.731	3.2	78	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.177	2.1	79	0.00
60	Diethylphthalate	40.000	39.161	2.1	79	0.00
61	4-Chlorophenyl-phenylether	40.000	38.101	4.7	78	0.00
62	4-Nitroaniline	40.000	42.264	-5.7	83	0.00
63	Azobenzene	40.000	37.165	7.1	76	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	84	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.333	-3.3	82	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.206	4.5	79	0.00
67	4-Bromophenyl-phenylether	40.000	37.113	7.2	78	0.00
68	Hexachlorobenzene	40.000	37.256	6.9	79	0.00
69	Atrazine	40.000	40.543	-1.4	84	0.00
70 C	Pentachlorophenol	40.000	40.015	-0.0	81	0.00
71	Phenanthrene	40.000	37.902	5.2	81	0.00
72	Anthracene	40.000	38.147	4.6	81	0.00
73	Carbazole	40.000	39.421	1.4	83	0.00
74	Di-n-butylphthalate	40.000	40.203	-0.5	84	0.00
75 C	Fluoranthene	40.000	39.049	2.4	84	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	88	0.00
77	Benzydine	40.000	44.590	-11.5	85	0.00
78	Pyrene	40.000	38.882	2.8	84	0.00
79 S	Terphenyl-d14	80.000	78.012	2.5	86	0.00
80	Butylbenzylphthalate	40.000	42.110	-5.3	90	0.00
81	Benzo(a)anthracene	40.000	38.606	3.5	85	0.00
82	3,3'-Dichlorobenzidine	40.000	38.471	3.8	82	0.00
83	Chrysene	40.000	37.383	6.5	83	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.543	-1.4	84	0.00
85 c	Di-n-octyl phthalate	40.000	40.980	-2.4	87	0.00
86 I	Perylene-d12	20.000	20.000	0.0	87	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.728	5.7	82	0.03
88	Benzo(b)fluoranthene	40.000	38.398	4.0	84	0.01
89	Benzo(k)fluoranthene	40.000	37.404	6.5	81	0.01
90 C	Benzo(a)pyrene	40.000	38.014	5.0	82	0.00
91	Dibenzo(a,h)anthracene	40.000	37.318	6.7	82	0.01
92	Benzo(g,h,i)perylene	40.000	37.579	6.1	82	0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG021122\
Data File : BG052366.D
Acq On : 11 Feb 2022 10:35
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Feb 11 11:23:58 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 08 01:21:20 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0