

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG021122\
 Data File : BG052377.D
 Acq On : 11 Feb 2022 19:39
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 12 01:26:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 08 01:21:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	77	0.00
2	1,4-Dioxane	40.000	37.457	6.4	76	0.00
3	Pyridine	40.000	38.326	4.2	73	0.00
4	n-Nitrosodimethylamine	40.000	36.267	9.3	71	0.00
5 S	2-Fluorophenol	80.000	79.839	0.2	78	0.00
6	Aniline	40.000	39.609	1.0	74	0.00
7 S	Phenol-d6	80.000	78.625	1.7	75	0.00
8	2-Chlorophenol	40.000	40.007	-0.0	80	0.00
9	Benzaldehyde	40.000	37.936	5.2	73	0.00
10 C	Phenol	40.000	39.732	0.7	76	0.00
11	bis(2-Chloroethyl)ether	40.000	38.115	4.7	76	0.00
12	1,3-Dichlorobenzene	40.000	39.022	2.4	77	0.00
13 C	1,4-Dichlorobenzene	40.000	39.657	0.9	78	0.00
14	1,2-Dichlorobenzene	40.000	38.232	4.4	75	0.00
15	Benzyl Alcohol	40.000	39.749	0.6	74	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.287	6.8	75	0.01
17	2-Methylphenol	40.000	39.891	0.3	77	0.00
18	Hexachloroethane	40.000	38.971	2.6	79	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.777	0.6	78	0.00
20	3+4-Methylphenols	40.000	39.508	1.2	75	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	80	0.00
22	Acetophenone	40.000	36.901	7.7	76	0.00
23 S	Nitrobenzene-d5	80.000	74.709	6.6	76	0.00
24	Nitrobenzene	40.000	36.682	8.3	73	0.00
25	Isophorone	40.000	38.151	4.6	77	0.00
26 C	2-Nitrophenol	40.000	39.279	1.8	77	0.00
27	2,4-Dimethylphenol	40.000	39.088	2.3	78	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.520	3.7	78	0.00
29 C	2,4-Dichlorophenol	40.000	38.563	3.6	78	0.00
30	1,2,4-Trichlorobenzene	40.000	37.302	6.7	77	0.00
31	Naphthalene	40.000	37.672	5.8	77	0.00
32	Benzoic acid	40.000	39.849	0.4	84	0.00
33	4-Chloroaniline	40.000	39.469	1.3	78	0.00
34 C	Hexachlorobutadiene	40.000	37.682	5.8	79	0.00
35	Caprolactam	40.000	42.499	-6.2	84	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.906	0.2	79	0.00
37	2-Methylnaphthalene	40.000	38.805	3.0	79	0.00
38	1-Methylnaphthalene	40.000	38.713	3.2	78	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	82	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.498	8.8	77	0.00
41 P	Hexachlorocyclopentadiene	40.000	33.216	17.0	70	0.00
42 S	2,4,6-Tribromophenol	80.000	81.686	-2.1	85	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.989	2.5	80	0.00
44	2,4,5-Trichlorophenol	40.000	39.261	1.8	82	0.00
45 S	2-Fluorobiphenyl	80.000	74.302	7.1	78	0.00
46	1,1'-Biphenyl	40.000	37.327	6.7	79	0.00
47	2-Chloronaphthalene	40.000	37.500	6.3	79	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG021122\
 Data File : BG052377.D
 Acq On : 11 Feb 2022 19:39
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 12 01:26:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 08 01:21:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.069	2.3	81	0.00
49	Acenaphthylene	40.000	38.142	4.6	81	0.00
50	Dimethylphthalate	40.000	39.153	2.1	83	0.00
51	2,6-Dinitrotoluene	40.000	39.436	1.4	82	0.00
52 C	Acenaphthene	40.000	38.318	4.2	80	0.00
53	3-Nitroaniline	40.000	41.343	-3.4	85	0.00
54 P	2,4-Dinitrophenol	40.000	37.406	6.5	79	0.00
55	Dibenzofuran	40.000	38.184	4.5	82	0.00
56 P	4-Nitrophenol	40.000	39.771	0.6	80	0.00
57	2,4-Dinitrotoluene	40.000	40.654	-1.6	83	0.00
58	Fluorene	40.000	39.352	1.6	83	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.872	0.3	84	0.00
60	Diethylphthalate	40.000	39.463	1.3	84	0.00
61	4-Chlorophenyl-phenylether	40.000	38.513	3.7	83	0.00
62	4-Nitroaniline	40.000	41.964	-4.9	87	0.00
63	Azobenzene	40.000	38.265	4.3	82	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	88	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.535	-1.3	84	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.628	5.9	82	0.00
67	4-Bromophenyl-phenylether	40.000	37.688	5.8	83	0.00
68	Hexachlorobenzene	40.000	37.424	6.4	83	0.00
69	Atrazine	40.000	38.753	3.1	84	0.00
70 C	Pentachlorophenol	40.000	38.510	3.7	81	0.00
71	Phenanthrene	40.000	38.070	4.8	85	0.00
72	Anthracene	40.000	38.064	4.8	84	0.00
73	Carbazole	40.000	39.246	1.9	87	0.00
74	Di-n-butylphthalate	40.000	40.183	-0.5	88	0.00
75 C	Fluoranthene	40.000	38.837	2.9	88	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	93	0.00
77	Benzydine	40.000	38.566	3.6	78	0.00
78	Pyrene	40.000	38.349	4.1	87	0.00
79 S	Terphenyl-d14	80.000	76.304	4.6	89	0.00
80	Butylbenzylphthalate	40.000	41.110	-2.8	92	0.00
81	Benzo(a)anthracene	40.000	38.566	3.6	89	0.00
82	3,3'-Dichlorobenzidine	40.000	38.598	3.5	87	0.00
83	Chrysene	40.000	37.677	5.8	88	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.331	-0.8	88	0.00
85 c	Di-n-octyl phthalate	40.000	40.748	-1.9	91	0.00
86 I	Perylene-d12	20.000	20.000	0.0	92	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.940	5.2	87	0.02
88	Benzo(b)fluoranthene	40.000	37.858	5.4	87	0.00
89	Benzo(k)fluoranthene	40.000	38.129	4.7	87	0.00
90 C	Benzo(a)pyrene	40.000	38.284	4.3	88	0.00
91	Dibenzo(a,h)anthracene	40.000	37.594	6.0	87	0.01
92	Benzo(g,h,i)perylene	40.000	37.803	5.5	87	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG021122\
Data File : BG052377.D
Acq On : 11 Feb 2022 19:39
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Feb 12 01:26:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 08 01:21:20 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0