

Data Path : U:\HPCHEM1\BNA G\DATA\BG021216\
 Data File : BG020898.D
 Acq On : 12 Feb 2016 18:31
 Operator : SJ/UM
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG021216

Quant Time: Feb 13 03:41:01 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\625-TERPINEOL-BG021216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 12 18:22:09 2016
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.27	152	16589	20.00	ng	0.00
4) Naphthalene-d8	11.09	136	89589	20.00	ng	0.00
7) Acenaphthene-d10	14.88	164	54356	20.00	ng	0.00
10) Phenanthrene-d10	17.62	188	111022	20.00	ng	0.00
11) Chrysene-d12	21.90	240	106952	20.00	ng	0.00
13) Perylene-d12	25.28	264	98437	20.00	ng	0.00
System Monitoring Compounds						
2) 2-Fluorophenol	5.80	112	87731	82.19	ng	0.00
3) Phenol-d6	7.41	99	131447	82.24	ng	0.00
5) Nitrobenzene-d5	9.44	82	114371	82.12	ng	0.00
8) 2,4,6-Tribromophenol	16.36	330	48014	80.90	ng	0.00
9) 2-Fluorobiphenyl	13.51	172	309810	80.78	ng	0.00
12) Terphenyl-d14	20.20	244	410066	80.82	ng	0.00
Target Compounds						
6) Terpineol	11.14	59	41532	41.07	ng	Qvalue # 33


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : U:\HPCHEM1\BNA G\DATA\BG021216\
Data File : BG020898.D
Acq On : 12 Feb 2016 18:31
Operator : SJ/UM
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG021216

Quant Time: Feb 13 03:41:01 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\625-TERPINEOL-BG021216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 12 18:22:09 2016
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 8.27 min Scan# 924
Delta R.T. 0.01 min
Lab File: BG020898.D
Acq: 12 Feb 2016 18:31

Instrument :
BNA_G
ClientSampleId :
ICVBG021216

Tot Ion:152 Resp: 16589
Ion Ratio Lower Upper
152 100
150 160.9 119.8 179.6
115 58.4 49.6 74.4




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


2-Fluorophenol
Concen: 82.19 ng
RT: 5.80 min Scan# 504
Delta R.T. -0.00 min
Lab File: BG020898.D
Acq: 12 Feb 2016 18:31

Tgt Ion:112 Resp: 87731
Ion Ratio Lower Upper
112 100
64 46.7 43.5 65.3
63 23.3 32.2 48.2#



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BGC
Ion 63.00 (62.70 to 63.70): BGC





#3

Phenol-d6

Concen: 82.24 ng

RT: 7.41 min Scan# 778

Delta R.T. -0.00 min

Lab File: BG020898.D

Acq: 12 Feb 2016 18:31

Instrument :

BNA_G

ClientSampleId :

ICVBG021216


Tot Ion: 99 Resp: 131447

Ion Ratio Lower Upper




99 100

42 10.7 18.8 28.2#

71 25.0 53.8 80.8#



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#4

Naphthalene-d8

Concen: 20.00 ng

RT: 11.09 min Scan# 1405

Delta R.T. -0.00 min

Lab File: BG020898.D

Acq: 12 Feb 2016 18:31

Tgt Ion: 136 Resp: 89589


Ion Ratio Lower Upper

136 100



137 11.3 9.0 13.4


54 6.7 4.4 6.6#

68 6.3 5.0 7.6



Abundance Ion 136.00 (135.70 to 136.70): BGC
Ion 137.00 (136.70 to 137.70): BGC
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC





#5

Nitrobenzene-d5

Concen: 82.12 ng

RT: 9.44 min Scan# 1123

Delta R.T. -0.00 min

Lab File: BG020898.D

Acq: 12 Feb 2016 18:31

Instrument :

BNA_G

ClientSampleId :

ICVBG021216


Tot Ion: 82 Resp: 114371

Ion Ratio Lower Upper

82 100

128 51.1 29.0 43.4#




54 37.6 34.5 51.7



Abundance Ion 82.00 (81.70 to 82.70): BGC

Ion 128.00 (127.70 to 128.70): BGC

Ion 54.00 (53.70 to 54.70): BGC



#6

Terpineol

Concen: 41.07 ng

RT: 11.14 min Scan# 1413

Delta R.T. 0.01 min

Lab File: BG020898.D

Acq: 12 Feb 2016 18:31


Tgt Ion: 59 Resp: 41532

Ion Ratio Lower Upper

59 100

93 62.8 118.9 178.3#



121 54.7 104.3 156.5#




Abundance Ion 59.00 (58.70 to 59.70): BGC

Ion 93.00 (92.70 to 93.70): BGC

Ion 121.00 (120.70 to 121.70): BGC





#7

Acenaphthene-d10

Concen: 20.00 ng

RT: 14.88 min Scan# 2050

Delta R.T. -0.00 min

Lab File: BG020898.D

Acq: 12 Feb 2016 18:31

Instrument :

BNA_G

ClientSampleId :

ICVBG021216


Tot Ion:164 Resp: 54356

Ion Ratio Lower Upper

164 100

162 105.3 74.5 111.7




160 47.1 32.7 49.1



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#8

2,4,6-Tribromophenol

Concen: 80.90 ng

RT: 16.36 min Scan# 2302

Delta R.T. 0.01 min

Lab File: BG020898.D

Acq: 12 Feb 2016 18:31


Tgt Ion:330 Resp: 48014

Ion Ratio Lower Upper

330 100

332 96.2 77.7 116.5



141 39.2 37.0 55.4




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E





#9

2-Fluorobiphenyl

Concen: 80.78 ng

RT: 13.51 min Scan# 1816

Delta R.T. -0.00 min

Lab File: BG020898.D

Acq: 12 Feb 2016 18:31

Instrument :

BNA_G

ClientSampleId :

ICVBG021216


Tot Ion:172 Resp: 309810

Ion Ratio Lower Upper

172 100

171 38.2 31.1 46.7




170 25.3 19.0 28.4



Abundance Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E



#10

Phenanthrene-d10

Concen: 20.00 ng

RT: 17.62 min Scan# 2516

Delta R.T. 0.01 min

Lab File: BG020898.D

Acq: 12 Feb 2016 18:31


Tgt Ion:188 Resp: 111022

Ion Ratio Lower Upper

188 100

94 10.6 4.6 7.0#



80 11.2 5.6 8.4#




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BGC

Ion 80.00 (79.70 to 80.70): BGC





#11

Chrysene-d12

Concen: 20.00 ng

RT: 21.90 min Scan# 3244

Delta R.T. -0.00 min

Lab File: BG020898.D

Acq: 12 Feb 2016 18:31

Instrument :

BNA_G

ClientSampleId :

ICVBG021216


Tot Ion:240 Resp: 106952

Ion Ratio Lower Upper

240 100

120 15.2 2.4 3.6#




236 27.6 19.5 29.3



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#12

Terphenyl-d14

Concen: 80.82 ng

RT: 20.20 min Scan# 2955

Delta R.T. -0.00 min

Lab File: BG020898.D

Acq: 12 Feb 2016 18:31


Tgt Ion:244 Resp: 410066

Ion Ratio Lower Upper

244 100

212 7.1 6.7 10.1



122 14.0 4.3 6.5#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#13

Pervlene-d12

Concen: 20.00 ng

RT: 25.28 min Scan# 3820

Delta R.T. -0.00 min

Lab File: BG020898.D

Acq: 12 Feb 2016 18:31

Instrument :

BNA_G

ClientSampleId :

ICVBG021216


Tot Ion:264 Resp: 98437

Ion Ratio Lower Upper

264 100

260 24.9 17.3 25.9

265 22.8 17.9 26.9



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

