


Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG021221\
 Data File : BG048144.D
 Acq On : 12 Feb 2021 10:14
 Operator : CG/JU
 Sample : DFTPP81
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP081

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG012921.M
 Title : SVOA CALIBRATION
 Last Update : Fri Feb 12 11:47:26 2021



AutoFind: Scans 2476, 2477, 2478; Background Corrected with Scan 2471

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.8	40544	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	41.0	45237	PASS
70	69	0.00	2	0.2	95	PASS
127	198	10	80	43.9	48368	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	110248	PASS
199	198	5	9	7.1	7780	PASS
275	198	10	60	29.8	32837	PASS
365	198	1	100	3.9	4352	PASS
441	198	0.01	100	9.8	10822	PASS
442	198	50	100	63.5	70058	PASS
443	442	15	24	19.7	13782	PASS

Scan 2471 (17.857 min): BG048144.D
DFTPP81

m/z	Abundance
39.90	271.0
44.00	742.0
96.00	154.0
133.00	185.0
190.70	160.0
207.10	1456.0
208.00	419.0
209.00	265.0
266.80	171.0
281.00	310.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

m/z	Abundance
37.00	240.0
38.00	292.0
39.10	2898.0
40.10	467.0
44.00	856.0
49.20	317.0
50.10	8929.0
51.10	32976.0
52.00	1689.0
55.10	284.0
56.00	913.0
57.00	2345.0
61.00	325.0
62.20	357.0
63.00	1506.0
64.00	160.0
65.00	1015.0
69.00	36432.0
74.00	4234.0
75.00	5645.0
76.10	2240.0
77.10	39488.0
78.10	2777.0
79.00	3187.0
80.00	2177.0
81.10	3109.0
82.00	1034.0
82.90	532.0
85.00	446.0
86.10	862.0
87.10	579.0
87.80	215.0
91.10	796.0
92.10	944.0
93.00	4162.0
93.90	332.0
96.00	225.0
98.00	3906.0
99.00	3127.0
101.00	1917.0
103.00	270.0
104.00	1167.0
105.00	1127.0
106.00	411.0
107.00	12171.0
107.90	2033.0
110.00	19872.0
111.00	3464.0
112.00	326.0
114.90	174.0
116.00	1203.0
117.00	9481.0
118.00	829.0
118.90	192.0
119.20	193.0
122.00	502.0
123.10	1018.0
124.10	520.0
125.00	665.0
125.70	291.0
127.00	37912.0
128.10	3130.0
129.00	14441.0
130.10	1244.0
130.90	383.0
133.90	730.0
135.00	1643.0
136.20	633.0
136.90	1069.0
139.70	247.0
140.10	242.0
141.00	1979.0
142.00	547.0
143.00	490.0
144.10	283.0
146.00	343.0
147.00	1304.0
148.00	2297.0
149.00	429.0
150.10	160.0
151.10	250.0
151.60	171.0
153.00	691.0
154.00	700.0
155.00	1343.0
156.10	2107.0
156.80	305.0
158.00	372.0
159.90	351.0
161.00	808.0
161.90	380.0
164.20	206.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

165.10	1014.0
166.00	702.0
167.00	3840.0
168.10	1861.0
168.90	449.0
169.90	185.0
172.00	474.0
173.10	332.0
174.00	910.0
175.10	1894.0
175.80	376.0
176.20	337.0
176.90	802.0
177.90	325.0
179.00	3483.0
180.00	2252.0
180.90	997.0
183.90	366.0
185.00	1595.0
186.00	10880.0
187.00	2676.0
188.00	483.0
189.10	799.0
189.80	184.0
191.00	545.0
191.90	1042.0
193.10	1661.0
193.90	167.0
194.80	391.0
196.00	2908.0
198.00	82176.0
198.90	6547.0
200.10	573.0
200.90	289.0
201.40	217.0
201.90	213.0
203.00	456.0
204.00	4212.0
205.00	6339.0
206.00	22936.0
207.10	4100.0
208.10	924.0
209.00	727.0
210.00	602.0
211.00	774.0
215.00	578.0
216.00	553.0
217.00	7738.0
218.00	1228.0
219.10	182.0
221.00	2794.0
223.00	1376.0
224.00	13932.0
225.00	3600.0
226.00	342.0
227.00	6312.0
228.10	975.0
229.00	1272.0
229.80	269.0
231.00	536.0
234.10	643.0
234.90	418.0
236.00	233.0
237.10	510.0
237.80	203.0
238.90	159.0
240.00	200.0
241.00	501.0
242.00	1081.0
243.10	795.0
244.00	9735.0
245.10	1627.0
246.00	2056.0
247.00	555.0
248.10	233.0
249.00	303.0
251.20	234.0
252.10	258.0
253.00	314.0
255.00	52552.0
256.00	9084.0
257.00	941.0
257.90	2983.0
259.10	607.0
259.80	158.0
260.80	261.0
265.00	1379.0
273.00	1651.0
274.00	4504.0
275.00	21424.0
276.00	3400.0
277.00	1940.0
278.00	477.0
281.90	242.0
282.90	312.0
283.80	275.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

285.10	427.0
293.00	515.0
294.90	162.0
296.00	6362.0
297.00	1174.0
300.90	155.0
301.90	204.0
303.00	685.0
304.00	272.0
312.80	165.0
314.00	313.0
314.90	635.0
316.00	532.0
318.10	196.0
321.00	224.0
322.00	174.0
323.00	2051.0
324.10	163.0
327.80	303.0
331.90	218.0
333.00	215.0
334.00	1189.0
335.10	320.0
340.80	205.0
345.90	506.0
352.00	535.0
353.00	247.0
354.00	969.0
354.80	207.0
358.80	157.0
364.90	2375.0
365.90	315.0
371.00	258.0
372.20	748.0
372.90	283.0
382.90	424.0
390.00	156.0
391.10	162.0
402.20	216.0
403.00	370.0
403.80	215.0
421.00	366.0
422.00	324.0
423.00	2654.0
424.00	804.0
441.00	6019.0
442.00	38632.0
443.00	7415.0
444.00	776.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

m/z	Abundance
38.00	731.0
39.10	3327.0
40.10	399.0
44.10	840.0
44.90	175.0
49.20	634.0
50.10	11925.0
51.10	49656.0
52.10	2280.0
53.10	168.0
54.80	217.0
56.10	1459.0
57.00	3058.0
58.10	207.0
61.00	750.0
62.00	922.0
63.00	1795.0
64.00	288.0
65.10	972.0
67.10	156.0
69.00	53280.0
69.90	201.0
73.10	601.0
74.10	5687.0
75.00	8848.0
76.10	3923.0
77.10	56104.0
78.10	3470.0
79.00	4545.0
80.10	3253.0
81.00	4458.0
82.00	1084.0
83.00	804.0
83.90	165.0
85.10	812.0
86.00	1170.0
87.10	803.0
88.00	240.0
91.00	900.0
92.10	1370.0
93.00	6418.0
93.90	295.0
94.20	316.0
95.00	209.0
96.10	583.0
98.00	4781.0
99.00	4337.0
100.00	361.0
101.00	2260.0
102.10	340.0
103.00	752.0
104.10	1690.0
105.00	1441.0
107.00	17864.0
108.00	2221.0
110.00	27088.0
111.00	4490.0
112.20	893.0
113.00	330.0
116.00	1480.0
117.00	14641.0
118.10	1150.0
118.90	154.0
120.00	359.0
122.10	1151.0
123.00	1696.0
124.00	1164.0
125.00	949.0
127.00	56912.0
128.00	5183.0
129.00	25832.0
130.00	1776.0
131.10	691.0
133.90	783.0
135.00	1986.0
136.00	1111.0
137.00	849.0
137.90	420.0
139.10	255.0
139.90	294.0
141.00	3229.0
141.90	709.0
142.90	601.0
144.00	293.0
145.20	375.0
146.00	669.0
147.00	1531.0
148.00	4230.0
149.10	753.0
150.10	286.0
151.20	478.0
151.70	433.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

153.00	1043.0
154.00	636.0
155.00	1886.0
156.00	2872.0
156.90	670.0
158.10	933.0
159.10	1008.0
160.10	1326.0
161.00	1484.0
161.90	568.0
163.10	368.0
164.00	214.0
165.00	1496.0
166.10	1043.0
167.00	5679.0
168.00	2565.0
169.00	941.0
170.00	423.0
170.80	227.0
172.00	1000.0
173.00	493.0
174.00	1910.0
175.00	3430.0
176.10	920.0
176.90	1704.0
178.20	379.0
178.90	4965.0
180.00	3386.0
181.00	1536.0
182.00	520.0
183.90	364.0
185.00	3407.0
186.00	20528.0
187.00	5783.0
188.00	521.0
189.00	1904.0
190.00	458.0
191.20	597.0
192.00	2764.0
193.00	2234.0
193.80	398.0
196.00	3703.0
198.00	127480.0
198.90	9049.0
200.00	1002.0
201.00	555.0
201.60	410.0
203.00	935.0
204.00	6023.0
205.00	10540.0
206.00	37424.0
207.10	6759.0
208.00	1657.0
209.00	829.0
210.10	610.0
211.00	2036.0
212.20	329.0
215.00	670.0
216.00	720.0
217.00	10974.0
218.00	1323.0
219.10	189.0
220.90	4106.0
221.80	1396.0
223.00	2915.0
224.00	22392.0
225.00	5891.0
226.00	1055.0
227.00	11150.0
228.00	1935.0
229.00	2374.0
231.00	770.0
233.00	262.0
234.00	731.0
235.10	822.0
235.80	554.0
237.10	631.0
238.00	319.0
238.70	223.0
240.00	155.0
241.00	834.0
242.00	1319.0
243.00	1552.0
244.00	15685.0
245.00	2308.0
246.00	3569.0
247.00	663.0
248.10	200.0
248.90	571.0
252.10	268.0
252.90	528.0
255.00	89208.0
256.00	13292.0
257.10	801.0
258.00	4891.0
259.00	829.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

259.80	218.0
261.20	170.0
265.00	2437.0
270.90	309.0
272.00	154.0
273.00	3307.0
274.00	6550.0
275.00	39544.0
276.00	4970.0
277.00	4128.0
278.00	758.0
283.00	594.0
283.80	213.0
284.90	616.0
285.80	217.0
286.00	207.0
291.90	328.0
293.00	413.0
293.90	239.0
295.90	12331.0
297.00	1712.0
302.00	432.0
303.10	1280.0
304.00	474.0
309.90	163.0
314.10	471.0
314.90	1211.0
316.00	558.0
321.00	356.0
323.10	3527.0
324.20	722.0
327.00	790.0
327.90	522.0
332.00	319.0
332.90	419.0
334.00	2199.0
334.90	350.0
336.00	210.0
341.10	398.0
345.80	701.0
352.00	1150.0
353.10	843.0
354.00	796.0
354.90	225.0
355.90	189.0
364.90	5202.0
366.00	783.0
370.90	543.0
372.00	1524.0
372.90	582.0
382.90	275.0
390.10	291.0
391.10	263.0
402.00	829.0
403.10	799.0
404.10	527.0
421.00	837.0
421.90	619.0
423.00	4510.0
424.00	1176.0
431.20	177.0
441.10	12357.0
442.00	78688.0
443.00	15979.0
444.10	1507.0
444.80	198.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

m/z	Abundance
38.10	384.0
39.10	2499.0
40.10	471.0
41.20	220.0
44.00	725.0
44.90	232.0
49.00	430.0
50.10	9026.0
51.10	39000.0
52.10	2108.0
54.10	151.0
55.10	152.0
55.90	1357.0
57.10	2834.0
58.10	163.0
61.00	628.0
62.00	645.0
63.10	2082.0
64.00	255.0
65.10	962.0
67.10	162.0
69.00	46000.0
70.10	285.0
71.10	207.0
73.10	454.0
74.00	4872.0
75.10	7071.0
76.10	2385.0
77.10	47896.0
78.10	2966.0
79.00	3657.0
80.00	2801.0
81.00	3640.0
82.00	877.0
82.90	954.0
85.00	386.0
85.90	1003.0
87.10	646.0
88.00	238.0
91.10	753.0
92.10	1144.0
93.00	6896.0
93.90	383.0
95.10	326.0
95.90	504.0
98.00	4395.0
99.00	3402.0
99.90	380.0
101.00	1776.0
103.00	672.0
104.00	1034.0
105.00	1001.0
106.00	392.0
107.00	14499.0
108.00	1896.0
110.00	24600.0
111.10	4376.0
111.90	686.0
112.80	202.0
115.10	206.0
116.00	1071.0
117.00	14074.0
118.10	868.0
120.10	323.0
121.90	1250.0
123.00	1920.0
124.00	725.0
125.00	582.0
125.80	290.0
127.00	50280.0
128.10	3817.0
129.00	22368.0
129.90	1797.0
131.10	379.0
132.70	266.0
134.00	835.0
135.00	1811.0
135.90	588.0
137.00	802.0
138.20	472.0
140.10	180.0
141.00	2719.0
142.00	893.0
143.10	368.0
144.90	244.0
146.00	586.0
147.10	1728.0
148.00	3033.0
149.00	898.0
150.90	488.0
151.50	260.0
152.20	342.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

153.00	844.0
154.00	566.0
155.00	1806.0
156.00	2882.0
156.80	440.0
158.00	860.0
159.10	386.0
159.90	930.0
161.00	1766.0
162.10	657.0
163.90	242.0
165.00	1276.0
166.00	739.0
167.00	5732.0
168.00	2350.0
169.00	851.0
171.00	197.0
171.90	585.0
173.00	586.0
173.90	1206.0
175.10	3167.0
175.90	520.0
177.00	1281.0
178.90	4567.0
180.00	3610.0
181.00	1733.0
182.00	362.0
184.00	522.0
185.10	3026.0
186.00	19760.0
187.00	4961.0
188.20	324.0
189.10	1454.0
190.20	176.0
190.90	687.0
192.00	1590.0
193.00	1723.0
194.00	672.0
195.10	324.0
196.00	3040.0
198.00	121088.0
198.90	7745.0
200.00	539.0
200.70	161.0
201.40	721.0
203.00	784.0
204.00	6148.0
205.00	9786.0
206.00	38352.0
207.00	5417.0
208.00	2052.0
208.90	767.0
210.00	799.0
211.10	1243.0
212.20	271.0
215.00	735.0
215.90	916.0
217.00	9918.0
217.90	1242.0
219.20	326.0
220.00	285.0
221.00	5861.0
222.00	1335.0
223.00	2488.0
224.00	21200.0
225.00	5184.0
226.20	678.0
227.00	8678.0
227.90	1462.0
229.00	1836.0
229.90	434.0
230.90	773.0
233.80	461.0
234.90	971.0
236.10	466.0
237.00	438.0
238.90	293.0
239.80	296.0
241.10	466.0
242.00	1187.0
243.00	1128.0
244.00	16496.0
245.00	2197.0
246.00	3473.0
246.90	514.0
247.80	203.0
248.10	232.0
248.90	816.0
250.90	341.0
252.10	294.0
252.90	398.0
255.00	88464.0
256.00	13210.0
257.00	1430.0
258.00	6024.0
259.00	679.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081

261.00	267.0
265.00	2803.0
265.90	334.0
267.00	174.0
271.00	196.0
272.00	355.0
273.00	2876.0
274.00	7400.0
275.00	37544.0
276.00	5266.0
277.00	3395.0
278.00	610.0
278.90	159.0
281.00	391.0
283.10	410.0
284.00	445.0
285.10	734.0
285.90	186.0
288.80	171.0
289.90	194.0
292.00	348.0
293.00	628.0
295.10	808.0
296.00	11576.0
297.00	1910.0
301.00	159.0
302.00	197.0
302.90	1734.0
303.90	466.0
308.00	248.0
313.10	246.0
314.10	368.0
315.00	1752.0
316.00	846.0
321.00	356.0
322.00	179.0
323.00	3807.0
324.00	561.0
326.90	829.0
328.00	219.0
329.10	212.0
331.80	196.0
333.10	292.0
334.00	2370.0
335.10	635.0
339.80	151.0
340.90	356.0
342.00	178.0
346.00	608.0
350.90	237.0
352.00	1240.0
352.90	826.0
354.10	829.0
365.00	5481.0
365.90	1059.0
370.90	423.0
372.10	1838.0
373.00	629.0
377.20	188.0
382.90	513.0
384.10	152.0
390.10	298.0
390.80	225.0
402.00	884.0
402.90	897.0
404.00	470.0
421.00	878.0
422.00	1033.0
423.00	6416.0
424.00	1170.0
425.20	152.0
441.00	14090.0
442.00	92856.0
443.00	17952.0
444.10	1483.0

Instrument :
BNA_G
ClientSampleId :
DFTPP081