


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG021315\
Data File : BG015986.D
Acq On : 13 Feb 2015 7:31
Operator : TP/IZ
Sample : DFTPP03
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP03

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG021015.M
Title : SVOA CALIBRATION
Last Update : Wed Feb 11 12:50:03 2015



AutoFind: Scans 2542, 2543, 2544; Background Corrected with Scan 2537

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.4	35256	PASS
68	69	0.00	2	0.5	197	PASS
69	198	0.00	100	43.2	36800	PASS
70	69	0.00	2	0.6	226	PASS
127	198	10	80	53.3	45416	PASS
197	198	0.00	2	0.4	300	PASS
198	198	100	100	100.0	85162	PASS
199	198	5	9	6.8	5800	PASS
275	198	10	60	25.5	21751	PASS
365	198	1	100	3.2	2753	PASS
441	443	0.01	100	78.1	12403	PASS
442	198	50	100	91.0	77469	PASS
443	442	15	24	20.5	15889	PASS

m/z	Abundance
37.15	154.0
38.05	280.0
39.15	2074.0
40.15	1458.0
41.15	501.0
43.15	257.0
44.15	987.0
49.05	298.0
50.15	7003.0
51.15	29448.0
52.15	1585.0
55.15	386.0
56.15	820.0
57.15	2735.0
58.05	170.0
61.05	597.0
62.05	405.0
63.15	1151.0
64.15	202.0
65.15	742.0
68.15	591.0
69.15	29072.0
70.15	337.0
73.15	261.0
74.15	2779.0
75.15	4822.0
76.15	1238.0
77.15	36720.0
78.15	2127.0
79.05	1947.0
80.15	1526.0
81.15	1901.0
82.05	551.0
83.15	675.0
84.05	155.0
85.15	401.0
86.15	519.0
87.05	400.0
91.15	548.0
92.15	629.0
93.05	3500.0
94.25	254.0
96.25	152.0
97.15	257.0
98.05	2622.0
99.05	2162.0
100.05	173.0
101.05	1277.0
103.05	586.0
104.05	849.0
105.15	743.0
106.15	317.0
107.15	8812.0
108.05	1874.0
110.05	17520.0
111.15	2888.0
112.05	488.0
113.15	182.0
116.05	601.0
117.05	5798.0
118.15	526.0
118.95	174.0
120.05	240.0
122.05	726.0
123.15	1352.0
124.15	618.0
125.15	524.0
127.05	33984.0
128.05	2281.0
129.05	12614.0
130.05	1179.0
131.15	324.0
134.15	432.0
135.05	855.0
136.05	399.0
137.05	435.0
138.15	152.0
140.05	227.0
141.15	1685.0
142.05	437.0
142.95	416.0
145.05	177.0
145.95	227.0
147.05	1087.0
148.05	2032.0
149.05	1246.0
150.15	183.0
151.05	272.0
153.05	604.0
154.05	490.0
155.15	1094.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03

156.15	1502.0
157.15	377.0
158.05	263.0
159.05	325.0
160.05	464.0
161.05	717.0
163.05	201.0
165.05	441.0
166.15	511.0
167.05	3606.0
168.15	1344.0
169.05	334.0
171.15	206.0
172.05	403.0
173.05	381.0
174.05	502.0
175.05	1253.0
176.05	369.0
177.05	695.0
177.95	227.0
179.05	2490.0
180.15	1342.0
181.05	641.0
182.15	204.0
185.05	1238.0
186.05	8496.0
187.15	2243.0
187.95	279.0
189.05	555.0
191.05	313.0
192.05	738.0
193.15	675.0
194.05	229.0
194.95	157.0
196.05	1445.0
198.05	58800.0
199.05	3752.0
199.95	309.0
201.35	270.0
201.85	220.0
203.05	458.0
204.15	2104.0
205.05	3460.0
206.05	13439.0
207.15	2637.0
208.05	511.0
209.85	165.0
211.05	689.0
211.85	259.0
215.05	198.0
216.15	278.0
217.05	3266.0
218.05	527.0
221.05	3175.0
223.05	824.0
224.05	7192.0
225.05	1822.0
225.95	323.0
227.05	3247.0
228.05	477.0
229.05	723.0
230.05	156.0
230.95	202.0
234.05	155.0
235.05	180.0
237.05	177.0
241.15	224.0
242.05	433.0
243.05	540.0
244.05	5891.0
245.15	904.0
245.95	1103.0
247.05	414.0
249.05	395.0
253.15	305.0
255.05	27760.0
256.05	4313.0
257.05	289.0
258.05	1752.0
259.05	379.0
265.05	717.0
265.85	162.0
273.05	1041.0
274.05	2540.0
275.05	12310.0
276.05	1738.0
277.05	1085.0
284.85	172.0
292.05	190.0
292.95	398.0
296.05	3353.0
297.05	547.0
303.05	335.0
313.95	229.0
315.05	392.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03

316.15	265.0
323.05	1279.0
326.95	233.0
334.05	767.0
335.05	171.0
341.05	190.0
346.05	249.0
352.05	387.0
353.05	272.0
354.15	434.0
355.25	197.0
365.05	1499.0
366.05	168.0
372.05	626.0
373.05	172.0
402.05	169.0
403.15	359.0
404.15	213.0
421.05	258.0
422.05	381.0
423.15	2060.0
424.15	292.0
441.15	5954.0
442.15	36552.0
443.05	7044.0
444.05	710.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03

m/z	Abundance
37.15	164.0
38.15	589.0
39.15	3204.0
40.05	1465.0
41.25	557.0
43.15	320.0
44.15	731.0
45.05	188.0
49.15	375.0
50.15	10382.0
51.15	41720.0
52.15	2034.0
53.15	351.0
55.15	461.0
56.05	1255.0
57.15	3737.0
61.15	504.0
62.15	551.0
63.15	1684.0
64.05	278.0
65.15	1180.0
67.05	210.0
69.15	45120.0
70.15	414.0
73.15	491.0
74.15	4226.0
75.15	7060.0
76.15	2191.0
77.15	54448.0
78.15	3460.0
79.15	3446.0
80.05	2067.0
81.15	3202.0
82.05	975.0
83.15	760.0
84.05	186.0
85.05	543.0
86.15	949.0
87.15	378.0
88.15	267.0
91.05	793.0
92.15	851.0
93.05	4940.0
94.15	405.0
96.15	397.0
97.25	239.0
98.15	3768.0
99.05	2932.0
100.05	427.0
101.15	1870.0
103.05	613.0
104.15	1344.0
105.05	1558.0
107.15	14698.0
108.05	1977.0
109.15	617.0
110.05	28624.0
111.05	4482.0
112.05	409.0
113.05	219.0
116.05	756.0
117.05	9335.0
118.05	1025.0
120.05	172.0
122.05	757.0
123.05	2082.0
124.05	930.0
125.05	726.0
127.05	54968.0
128.05	4022.0
129.05	20336.0
130.05	2064.0
131.05	291.0
133.15	191.0
134.05	673.0
135.05	1786.0
136.15	789.0
137.15	816.0
138.15	238.0
139.05	232.0
140.15	325.0
141.05	2454.0
142.05	796.0
143.05	454.0
144.15	219.0
145.05	213.0
146.15	395.0
147.05	1178.0
148.05	2615.0
149.15	1512.0
150.05	313.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03

151.45	312.0
151.95	280.0
153.05	728.0
154.05	648.0
155.15	1609.0
156.05	2069.0
156.95	448.0
158.05	462.0
159.05	420.0
160.05	747.0
161.05	1230.0
162.05	549.0
163.15	317.0
164.05	205.0
165.05	1123.0
166.05	881.0
167.05	5404.0
168.05	2272.0
168.95	415.0
170.95	301.0
172.05	510.0
173.05	707.0
174.15	1491.0
175.15	1997.0
176.05	653.0
177.05	912.0
178.05	327.0
179.05	3961.0
180.05	2692.0
181.05	1172.0
182.05	242.0
184.05	301.0
185.05	1758.0
186.05	14299.0
187.15	4178.0
188.15	440.0
189.05	866.0
191.05	444.0
192.05	1153.0
193.15	1393.0
194.05	261.0
194.95	266.0
196.05	3160.0
196.85	902.0
198.05	100928.0
199.05	7084.0
200.05	489.0
201.55	505.0
203.05	629.0
204.05	3319.0
205.15	5728.0
206.05	24592.0
207.05	3937.0
208.05	697.0
209.05	371.0
210.25	410.0
211.05	809.0
215.05	248.0
216.05	444.0
217.05	6516.0
218.05	1155.0
221.05	5504.0
223.05	1915.0
224.05	14378.0
225.05	3621.0
226.05	393.0
227.05	5404.0
228.05	755.0
229.05	1195.0
230.05	192.0
231.05	409.0
232.15	157.0
234.15	287.0
235.15	493.0
236.05	325.0
237.05	519.0
238.95	231.0
240.05	195.0
241.05	380.0
242.05	657.0
243.05	1090.0
244.05	11158.0
245.05	1395.0
246.05	1795.0
247.05	370.0
249.05	333.0
251.05	189.0
251.95	199.0
253.05	296.0
255.05	55728.0
256.05	8611.0
257.05	579.0
258.05	3085.0
259.05	606.0
261.05	151.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03

265.05	1142.0
270.95	235.0
272.05	207.0
273.05	1988.0
274.05	4054.0
275.05	26016.0
276.05	3115.0
277.05	2014.0
278.15	360.0
283.15	389.0
285.05	358.0
293.05	476.0
294.95	157.0
296.05	7288.0
297.05	960.0
303.15	836.0
304.15	225.0
308.05	160.0
308.95	174.0
313.95	336.0
315.05	811.0
316.05	616.0
321.05	280.0
323.15	2416.0
324.05	452.0
327.05	616.0
328.05	247.0
334.05	1685.0
335.05	343.0
340.95	250.0
346.05	555.0
352.05	733.0
353.05	653.0
354.05	651.0
365.05	3284.0
366.05	551.0
371.05	175.0
372.05	1288.0
373.05	446.0
382.95	223.0
383.95	170.0
402.05	537.0
403.05	797.0
404.15	154.0
421.05	687.0
422.15	449.0
423.05	5334.0
424.05	1354.0
425.05	181.0
439.95	167.0
441.15	13879.0
442.05	88848.0
443.05	18616.0
444.15	1774.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03

m/z	Abundance
38.15	521.0
39.15	2632.0
40.15	1113.0
41.15	609.0
43.15	324.0
44.15	700.0
45.15	168.0
49.25	227.0
50.15	8395.0
51.15	34600.0
52.15	1775.0
53.15	203.0
55.15	454.0
56.05	1306.0
57.15	2591.0
58.05	163.0
61.15	465.0
62.15	371.0
63.15	1308.0
64.15	331.0
65.15	929.0
67.15	249.0
69.05	36208.0
70.15	393.0
71.15	151.0
73.15	336.0
74.15	3527.0
75.15	5974.0
76.15	2255.0
77.15	43592.0
78.15	3239.0
79.15	2411.0
80.15	2149.0
81.05	2492.0
82.05	828.0
83.15	886.0
84.05	233.0
85.05	545.0
86.05	661.0
87.05	408.0
88.15	276.0
89.15	157.0
91.15	895.0
92.05	821.0
93.15	4461.0
94.05	382.0
96.15	391.0
97.15	185.0
98.05	3120.0
99.15	2368.0
100.15	300.0
101.05	2001.0
103.05	688.0
104.05	1096.0
105.15	1204.0
106.05	371.0
107.05	12344.0
108.05	2193.0
110.05	22872.0
111.05	3699.0
112.05	476.0
113.15	257.0
116.05	707.0
117.15	8747.0
118.05	689.0
120.05	337.0
122.05	816.0
123.05	1932.0
124.05	696.0
125.05	580.0
127.05	47296.0
128.15	3574.0
129.05	17640.0
130.05	1265.0
131.05	301.0
132.05	170.0
133.05	203.0
134.05	548.0
135.15	1598.0
136.05	504.0
137.05	834.0
138.05	219.0
140.15	371.0
141.05	2420.0
142.15	751.0
142.95	570.0
145.15	172.0
146.15	348.0
147.05	1083.0
148.05	2606.0
149.05	1342.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03

150.05	255.0
150.95	275.0
151.45	300.0
151.95	182.0
153.15	715.0
154.05	522.0
155.15	1512.0
156.15	1771.0
157.05	402.0
158.05	605.0
159.05	401.0
160.05	726.0
161.05	861.0
162.15	415.0
163.05	217.0
164.15	178.0
165.15	841.0
166.15	657.0
167.05	4674.0
168.05	2165.0
169.05	370.0
171.05	207.0
172.05	339.0
173.05	454.0
174.05	1176.0
175.05	1824.0
176.15	748.0
177.05	1083.0
178.05	418.0
179.05	3336.0
180.05	2688.0
181.05	1129.0
182.05	170.0
184.15	273.0
185.05	1790.0
186.05	13364.0
187.15	4196.0
188.15	509.0
189.05	980.0
191.05	393.0
192.05	1164.0
193.05	1145.0
194.05	244.0
195.05	185.0
196.05	2719.0
198.05	95760.0
199.05	6565.0
200.05	842.0
201.55	560.0
203.05	782.0
204.05	3002.0
205.05	5924.0
206.05	24680.0
207.05	3420.0
208.05	897.0
209.05	523.0
210.05	628.0
211.15	916.0
211.75	232.0
214.95	329.0
215.95	535.0
217.05	5792.0
218.05	871.0
221.05	5700.0
223.05	1725.0
224.05	13389.0
225.05	3364.0
226.05	382.0
227.05	6326.0
228.05	776.0
228.95	1052.0
231.05	561.0
233.95	365.0
235.05	550.0
235.95	326.0
236.95	506.0
238.95	275.0
240.05	193.0
241.05	206.0
242.05	740.0
243.05	772.0
244.05	11200.0
245.15	1727.0
246.05	2042.0
246.95	465.0
249.05	624.0
252.05	169.0
253.05	237.0
255.05	55760.0
256.05	7607.0
257.05	747.0
258.05	2974.0
259.05	535.0
265.05	1587.0
266.05	158.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03

272.05	194.0
273.05	1987.0
274.05	4845.0
275.05	26928.0
276.05	3498.0
277.05	1831.0
278.05	380.0
281.15	162.0
283.05	294.0
284.15	208.0
285.15	484.0
292.15	150.0
293.05	462.0
295.05	219.0
296.05	7691.0
297.05	1201.0
301.95	175.0
303.15	802.0
304.05	248.0
313.95	298.0
315.05	1332.0
316.05	422.0
321.05	355.0
322.05	181.0
323.05	2473.0
324.05	579.0
327.05	549.0
327.95	286.0
332.05	250.0
333.05	219.0
334.05	1828.0
335.05	312.0
341.05	305.0
345.95	522.0
352.05	671.0
353.15	585.0
354.05	864.0
355.05	165.0
365.05	3476.0
366.05	557.0
370.85	155.0
372.05	1278.0
373.05	380.0
383.05	408.0
390.05	328.0
390.95	156.0
402.05	699.0
403.05	969.0
404.05	472.0
421.05	697.0
422.15	596.0
423.05	5708.0
424.05	1219.0
438.65	160.0
441.15	17376.0
442.15	107008.0
443.05	22008.0
444.05	1923.0
445.05	193.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03

m/z	Abundance
40.15	833.0
41.15	626.0
43.15	204.0
44.15	929.0
55.25	321.0
57.25	678.0
70.15	155.0
77.15	181.0
79.15	249.0
84.05	196.0
86.05	154.0
91.05	184.0
93.15	206.0
105.15	243.0
123.15	526.0
149.05	1110.0
206.95	265.0

Instrument :
BNA_G
ClientSampleId :
DFTPP03