

Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG021422\  
 Data File : BG052386.D  
 Acq On : 14 Feb 2022 16:38  
 Operator : CG/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 LabSampleId :  
 SSTDCCC040

Quant Time: Feb 14 18:00:05 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_G\Methods\8270-BG020722.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Feb 14 14:37:39 2022  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | Amount | Calc.  | %Dev | Area% | Dev(min) |
|------|-----------------------------|--------|--------|------|-------|----------|
| 1 I  | 1,4-Dichlorobenzene-d4      | 20.000 | 20.000 | 0.0  | 86    | 0.00     |
| 2    | 1,4-Dioxane                 | 40.000 | 37.662 | 5.8  | 86    | 0.00     |
| 3    | Pyridine                    | 40.000 | 37.600 | 6.0  | 80    | 0.00     |
| 4    | n-Nitrosodimethylamine      | 40.000 | 35.709 | 10.7 | 79    | 0.00     |
| 5 S  | 2-Fluorophenol              | 80.000 | 77.221 | 3.5  | 84    | 0.00     |
| 6    | Aniline                     | 40.000 | 36.758 | 8.1  | 78    | 0.00     |
| 7 S  | Phenol-d6                   | 80.000 | 74.616 | 6.7  | 80    | 0.00     |
| 8    | 2-Chlorophenol              | 40.000 | 37.438 | 6.4  | 84    | 0.00     |
| 9    | Benzaldehyde                | 40.000 | 34.394 | 14.0 | 78    | 0.00     |
| 10 C | Phenol                      | 40.000 | 37.680 | 5.8  | 81    | 0.00     |
| 11   | bis(2-Chloroethyl)ether     | 40.000 | 36.077 | 9.8  | 81    | 0.00     |
| 12   | 1,3-Dichlorobenzene         | 40.000 | 37.334 | 6.7  | 82    | 0.00     |
| 13 C | 1,4-Dichlorobenzene         | 40.000 | 36.833 | 7.9  | 81    | 0.00     |
| 14   | 1,2-Dichlorobenzene         | 40.000 | 35.977 | 10.1 | 80    | 0.00     |
| 15   | Benzyl Alcohol              | 40.000 | 36.679 | 8.3  | 76    | 0.00     |
| 16   | 2,2'-oxybis(1-Chloropropane | 40.000 | 35.438 | 11.4 | 80    | 0.00     |
| 17   | 2-Methylphenol              | 40.000 | 37.602 | 6.0  | 81    | 0.00     |
| 18   | Hexachloroethane            | 40.000 | 37.274 | 6.8  | 84    | 0.00     |
| 19 P | n-Nitroso-di-n-propylamine  | 40.000 | 34.981 | 12.5 | 77    | 0.00     |
| 20   | 3+4-Methylphenols           | 40.000 | 36.542 | 8.6  | 78    | 0.00     |
| 21 I | Naphthalene-d8              | 20.000 | 20.000 | 0.0  | 84    | 0.00     |
| 22   | Acetophenone                | 40.000 | 37.000 | 7.5  | 80    | 0.00     |
| 23 S | Nitrobenzene-d5             | 80.000 | 73.184 | 8.5  | 78    | 0.00     |
| 24   | Nitrobenzene                | 40.000 | 36.018 | 10.0 | 75    | 0.00     |
| 25   | Isophorone                  | 40.000 | 36.504 | 8.7  | 77    | 0.00     |
| 26 C | 2-Nitrophenol               | 40.000 | 38.210 | 4.5  | 78    | 0.00     |
| 27   | 2,4-Dimethylphenol          | 40.000 | 38.219 | 4.5  | 80    | 0.00     |
| 28   | bis(2-Chloroethoxy)methane  | 40.000 | 36.927 | 7.7  | 78    | 0.00     |
| 29 C | 2,4-Dichlorophenol          | 40.000 | 37.267 | 6.8  | 79    | 0.00     |
| 30   | 1,2,4-Trichlorobenzene      | 40.000 | 37.656 | 5.9  | 81    | 0.00     |
| 31   | Naphthalene                 | 40.000 | 36.730 | 8.2  | 79    | 0.00     |
| 32   | Benzoic acid                | 40.000 | 36.811 | 8.0  | 76    | 0.00     |
| 33   | 4-Chloroaniline             | 40.000 | 37.871 | 5.3  | 79    | 0.00     |
| 34 C | Hexachlorobutadiene         | 40.000 | 37.049 | 7.4  | 81    | 0.00     |
| 35   | Caprolactam                 | 40.000 | 38.771 | 3.1  | 81    | 0.00     |
| 36 C | 4-Chloro-3-methylphenol     | 40.000 | 38.277 | 4.3  | 79    | 0.00     |
| 37   | 2-Methylnaphthalene         | 40.000 | 37.001 | 7.5  | 79    | 0.00     |
| 38   | 1-Methylnaphthalene         | 40.000 | 36.939 | 7.7  | 78    | 0.00     |
| 39 I | Acenaphthene-d10            | 20.000 | 20.000 | 0.0  | 86    | 0.00     |
| 40   | 1,2,4,5-Tetrachlorobenzene  | 40.000 | 35.683 | 10.8 | 79    | 0.00     |
| 41 P | Hexachlorocyclopentadiene   | 40.000 | 42.719 | -6.8 | 92    | 0.00     |
| 42 S | 2,4,6-Tribromophenol        | 80.000 | 73.785 | 7.8  | 80    | 0.00     |
| 43 C | 2,4,6-Trichlorophenol       | 40.000 | 37.435 | 6.4  | 81    | 0.00     |
| 44   | 2,4,5-Trichlorophenol       | 40.000 | 36.396 | 9.0  | 80    | 0.00     |
| 45 S | 2-Fluorobiphenyl            | 80.000 | 71.923 | 10.1 | 79    | 0.00     |
| 46   | 1,1'-Biphenyl               | 40.000 | 36.031 | 9.9  | 79    | 0.00     |
| 47   | 2-Chloronaphthalene         | 40.000 | 35.934 | 10.2 | 79    | 0.00     |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG021422\  
 Data File : BG052386.D  
 Acq On : 14 Feb 2022 16:38  
 Operator : CG/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 LabSampleId :  
 SSTDCCC040

Quant Time: Feb 14 18:00:05 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_G\Methods\8270-BG020722.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Feb 14 14:37:39 2022  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                   | Amount | Calc.  | %Dev | Area% | Dev(min) |
|------|----------------------------|--------|--------|------|-------|----------|
| 48   | 2-Nitroaniline             | 40.000 | 36.920 | 7.7  | 80    | 0.00     |
| 49   | Acenaphthylene             | 40.000 | 35.991 | 10.0 | 80    | 0.00     |
| 50   | Dimethylphthalate          | 40.000 | 36.146 | 9.6  | 80    | 0.00     |
| 51   | 2,6-Dinitrotoluene         | 40.000 | 37.115 | 7.2  | 80    | 0.00     |
| 52 C | Acenaphthene               | 40.000 | 35.953 | 10.1 | 79    | 0.00     |
| 53   | 3-Nitroaniline             | 40.000 | 37.773 | 5.6  | 81    | 0.00     |
| 54 P | 2,4-Dinitrophenol          | 40.000 | 34.534 | 13.7 | 72    | 0.00     |
| 55   | Dibenzofuran               | 40.000 | 35.918 | 10.2 | 81    | 0.00     |
| 56 P | 4-Nitrophenol              | 40.000 | 37.284 | 6.8  | 78    | 0.00     |
| 57   | 2,4-Dinitrotoluene         | 40.000 | 37.965 | 5.1  | 81    | 0.00     |
| 58   | Fluorene                   | 40.000 | 36.381 | 9.0  | 80    | 0.00     |
| 59   | 2,3,4,6-Tetrachlorophenol  | 40.000 | 36.018 | 10.0 | 79    | 0.00     |
| 60   | Diethylphthalate           | 40.000 | 36.917 | 7.7  | 82    | 0.00     |
| 61   | 4-Chlorophenyl-phenylether | 40.000 | 35.912 | 10.2 | 80    | 0.00     |
| 62   | 4-Nitroaniline             | 40.000 | 37.988 | 5.0  | 82    | 0.00     |
| 63   | Azobenzene                 | 40.000 | 35.714 | 10.7 | 80    | 0.00     |
| 64 I | Phenanthrene-d10           | 20.000 | 20.000 | 0.0  | 87    | 0.00     |
| 65   | 4,6-Dinitro-2-methylphenol | 40.000 | 38.706 | 3.2  | 80    | 0.00     |
| 66 c | n-Nitrosodiphenylamine     | 40.000 | 36.912 | 7.7  | 80    | 0.00     |
| 67   | 4-Bromophenyl-phenylether  | 40.000 | 36.939 | 7.7  | 81    | 0.00     |
| 68   | Hexachlorobenzene          | 40.000 | 36.877 | 7.8  | 82    | 0.00     |
| 69   | Atrazine                   | 40.000 | 37.486 | 6.3  | 81    | 0.00     |
| 70 C | Pentachlorophenol          | 40.000 | 36.384 | 9.0  | 77    | 0.00     |
| 71   | Phenanthrene               | 40.000 | 36.847 | 7.9  | 82    | 0.00     |
| 72   | Anthracene                 | 40.000 | 36.838 | 7.9  | 81    | 0.00     |
| 73   | Carbazole                  | 40.000 | 37.965 | 5.1  | 84    | 0.00     |
| 74   | Di-n-butylphthalate        | 40.000 | 38.873 | 2.8  | 85    | 0.00     |
| 75 C | Fluoranthene               | 40.000 | 37.561 | 6.1  | 85    | 0.00     |
| 76 I | Chrysene-d12               | 20.000 | 20.000 | 0.0  | 95    | 0.00     |
| 77   | Benzydine                  | 40.000 | 43.674 | -9.2 | 90    | 0.00     |
| 78   | Pyrene                     | 40.000 | 36.439 | 8.9  | 84    | 0.00     |
| 79 S | Terphenyl-d14              | 80.000 | 73.614 | 8.0  | 87    | 0.00     |
| 80   | Butylbenzylphthalate       | 40.000 | 39.969 | 0.1  | 91    | 0.00     |
| 81   | Benzo(a)anthracene         | 40.000 | 37.672 | 5.8  | 88    | 0.00     |
| 82   | 3,3'-Dichlorobenzidine     | 40.000 | 37.554 | 6.1  | 86    | 0.00     |
| 83   | Chrysene                   | 40.000 | 37.320 | 6.7  | 89    | 0.00     |
| 84   | Bis(2-ethylhexyl)phthalate | 40.000 | 39.686 | 0.8  | 89    | 0.00     |
| 85 c | Di-n-octyl phthalate       | 40.000 | 39.650 | 0.9  | 90    | 0.00     |
| 86 I | Perylene-d12               | 20.000 | 20.000 | 0.0  | 94    | 0.00     |
| 87   | Indeno(1,2,3-cd)pyrene     | 40.000 | 37.801 | 5.5  | 89    | 0.01     |
| 88   | Benzo(b)fluoranthene       | 40.000 | 37.856 | 5.4  | 89    | 0.00     |
| 89   | Benzo(k)fluoranthene       | 40.000 | 37.748 | 5.6  | 88    | 0.00     |
| 90 C | Benzo(a)pyrene             | 40.000 | 37.611 | 6.0  | 88    | 0.00     |
| 91   | Dibenzo(a,h)anthracene     | 40.000 | 37.687 | 5.8  | 89    | 0.00     |
| 92   | Benzo(g,h,i)perylene       | 40.000 | 37.535 | 6.2  | 88    | 0.00     |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG021422\  
Data File : BG052386.D  
Acq On : 14 Feb 2022 16:38  
Operator : CG/JU  
Sample : SSTDCCC040  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_G  
LabSampleId :  
SSTDCCC040

Quant Time: Feb 14 18:00:05 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_G\Methods\8270-BG020722.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Feb 14 14:37:39 2022  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|----------|--------|-------|------|-------|----------|
|----------|--------|-------|------|-------|----------|

(#) = Out of Range                      SPCC's out = 0    CCC's out = 0