


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG021425\
Data File : BG064023.D
Acq On : 14 Feb 2025 8:13
Operator : RC/JU
Sample : DFTPP452
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP452

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG021425.MA.M
Title : SVOA CALIBRATION
Last Update : Tue Feb 18 15:56:04 2025



AutoFind: Scans 2484, 2485, 2486; Background Corrected with Scan 2478

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	58.5	66837	PASS
68	69	0.00	2	1.6	726	PASS
69	198	0.00	100	39.1	44627	PASS
70	69	0.00	2	0.6	248	PASS
127	198	10	80	47.8	54605	PASS
197	198	0.00	2	0.4	445	PASS
198	198	100	100	100.0	114168	PASS
199	198	5	9	6.7	7623	PASS
275	198	10	60	27.3	31224	PASS
365	198	1	100	4.4	4977	PASS
441	198	0.01	100	13.4	15349	PASS
442	198	50	100	84.7	96667	PASS
443	442	15	24	20.1	19414	PASS

m/z	Abundance
39.90	217.0
44.00	3451.0
167.00	183.0
207.00	335.0
487.50	188.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

m/z	Abundance
36.90	310.0
38.10	646.0
39.00	4691.0
40.00	417.0
43.10	341.0
44.00	4217.0
45.20	376.0
49.10	489.0
50.00	18168.0
51.00	67448.0
52.00	3588.0
53.10	159.0
54.90	397.0
56.00	1916.0
57.00	4212.0
61.00	821.0
61.90	811.0
63.00	2366.0
63.90	269.0
65.00	1283.0
68.00	905.0
69.00	45728.0
70.00	159.0
73.10	442.0
74.00	5135.0
75.00	7196.0
76.00	2721.0
77.00	55904.0
78.10	3589.0
79.00	3764.0
80.00	2446.0
81.00	4303.0
82.00	1049.0
83.00	813.0
83.90	320.0
85.00	834.0
86.00	1397.0
87.10	476.0
88.00	162.0
91.10	948.0
92.00	1258.0
93.00	6617.0
94.00	487.0
96.00	174.0
98.00	5226.0
99.00	4245.0
100.00	244.0
100.90	1924.0
101.80	186.0
103.00	800.0
104.00	1409.0
105.00	1032.0
106.00	628.0
107.00	15701.0
107.90	2497.0
110.00	27160.0
111.00	4513.0
111.90	790.0
113.00	331.0
116.00	859.0
117.00	15878.0
118.00	1686.0
119.90	157.0
121.90	860.0
123.10	1530.0
123.90	527.0
125.00	744.0
127.00	50752.0
128.10	3716.0
129.00	20424.0
130.00	1519.0
130.80	546.0
131.80	192.0
133.00	154.0
133.90	685.0
135.00	1478.0
136.00	628.0
137.00	1022.0
137.80	363.0
139.90	441.0
141.00	2368.0
142.10	937.0
143.00	555.0
146.10	548.0
147.00	1071.0
148.00	3342.0
149.00	587.0
150.00	225.0
151.00	356.0
151.40	288.0
152.00	262.0
152.90	877.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

154.00	598.0
155.00	1479.0
156.00	2403.0
157.10	253.0
158.00	495.0
158.90	438.0
159.80	847.0
161.00	1102.0
162.10	303.0
164.20	354.0
165.00	1401.0
166.00	1089.0
167.00	5897.0
168.00	3098.0
169.00	559.0
170.10	232.0
171.20	312.0
172.00	693.0
173.00	537.0
174.00	1381.0
175.00	2179.0
176.00	421.0
177.10	998.0
177.80	361.0
179.00	3989.0
180.00	2407.0
181.00	1529.0
182.00	306.0
183.10	171.0
183.90	366.0
185.10	2272.0
186.00	15104.0
187.00	3750.0
187.90	533.0
189.00	1033.0
190.00	243.0
190.90	602.0
192.00	1637.0
193.00	1575.0
193.80	447.0
195.00	328.0
196.00	3003.0
198.00	102464.0
199.00	6076.0
199.90	435.0
201.20	306.0
202.10	272.0
203.00	861.0
204.00	4248.0
205.00	6707.0
206.00	27568.0
207.10	3975.0
208.10	1440.0
208.90	210.0
210.80	835.0
214.90	235.0
215.90	686.0
217.00	7128.0
218.00	1077.0
219.10	172.0
220.10	242.0
221.00	3127.0
222.00	925.0
223.10	1661.0
224.00	16544.0
225.00	4396.0
226.00	568.0
227.00	8092.0
228.00	1183.0
229.00	1443.0
230.00	209.0
231.00	515.0
234.00	384.0
235.00	624.0
237.00	380.0
239.10	370.0
239.80	373.0
240.70	403.0
242.00	878.0
243.00	841.0
244.00	13194.0
245.00	1626.0
245.90	2604.0
246.90	726.0
248.10	223.0
248.90	446.0
251.00	318.0
253.00	477.0
255.00	59376.0
256.00	8487.0
257.10	842.0
258.00	4296.0
259.10	720.0
259.90	221.0
264.10	204.0
265.00	1757.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

266.00	171.0
270.90	335.0
272.10	396.0
273.00	2061.0
273.90	3996.0
275.00	26144.0
276.00	3231.0
277.00	2447.0
278.10	194.0
279.00	182.0
282.90	373.0
283.80	237.0
285.00	595.0
292.10	233.0
293.00	757.0
294.00	317.0
296.00	8237.0
296.90	1260.0
300.90	186.0
303.10	849.0
304.20	199.0
313.90	354.0
314.90	1019.0
316.00	464.0
321.00	247.0
321.70	193.0
323.00	2854.0
323.90	488.0
327.10	441.0
327.90	392.0
332.00	169.0
332.80	281.0
334.00	1835.0
335.00	713.0
341.00	495.0
345.90	598.0
352.00	677.0
353.00	409.0
353.90	893.0
364.90	3717.0
366.00	471.0
370.80	171.0
372.00	1248.0
373.20	345.0
382.90	353.0
402.00	649.0
403.10	601.0
403.90	327.0
420.90	517.0
422.00	696.0
423.00	4028.0
424.10	839.0
428.70	267.0
436.80	250.0
441.00	11152.0
442.00	66976.0
443.00	12945.0
444.00	859.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

m/z	Abundance
36.90	335.0
37.90	1120.0
39.00	6672.0
40.00	839.0
44.00	3527.0
50.00	20472.0
51.00	79416.0
52.00	4116.0
53.00	161.0
55.00	371.0
56.00	2118.0
57.00	5181.0
57.90	215.0
61.00	963.0
62.00	770.0
63.00	2639.0
64.00	468.0
65.00	1243.0
67.10	280.0
68.00	1273.0
69.00	52712.0
70.00	240.0
73.00	523.0
74.10	6417.0
75.00	8941.0
76.00	3573.0
77.00	67584.0
78.00	4325.0
79.00	4676.0
80.00	3751.0
81.00	4803.0
82.00	1326.0
83.00	1335.0
85.00	938.0
86.00	1471.0
87.00	491.0
88.00	313.0
91.00	1050.0
92.00	1181.0
93.00	8843.0
93.90	576.0
95.00	216.0
95.80	205.0
96.80	241.0
98.00	6784.0
99.00	5233.0
100.00	548.0
101.00	2679.0
101.90	376.0
103.00	713.0
103.90	1443.0
105.00	1253.0
107.00	21512.0
108.00	3362.0
110.00	34448.0
111.00	5103.0
112.10	744.0
112.90	245.0
116.00	1371.0
117.00	20720.0
117.90	1529.0
119.90	254.0
122.00	1436.0
122.90	2347.0
124.00	974.0
125.00	790.0
127.00	68104.0
128.00	4585.0
129.00	25904.0
129.90	2391.0
130.90	340.0
132.10	343.0
132.90	150.0
133.90	594.0
135.00	1822.0
136.00	769.0
137.00	1446.0
137.90	386.0
138.90	196.0
140.00	353.0
141.00	3744.0
142.00	1177.0
143.00	731.0
143.90	155.0
145.90	756.0
147.00	1924.0
148.00	4492.0
149.00	1196.0
150.10	467.0
151.10	602.0
152.00	371.0
152.90	1037.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

153.90	1056.0
154.90	1956.0
156.00	3491.0
157.10	766.0
158.00	720.0
159.00	575.0
160.00	1087.0
161.00	1629.0
162.10	426.0
163.00	161.0
165.00	1891.0
165.90	834.0
167.00	7965.0
168.00	3688.0
168.90	638.0
170.00	153.0
170.90	334.0
172.00	518.0
173.00	863.0
174.10	1395.0
175.00	2725.0
176.10	823.0
177.00	1508.0
177.90	430.0
179.00	5565.0
180.00	3627.0
181.00	1769.0
182.00	292.0
182.80	339.0
184.10	475.0
185.00	2725.0
186.00	20976.0
187.00	6900.0
188.00	541.0
188.90	1295.0
189.90	344.0
190.90	704.0
192.00	2123.0
193.10	2068.0
194.00	786.0
194.90	344.0
196.00	3229.0
196.70	1336.0
198.00	135744.0
199.00	9467.0
199.90	755.0
201.50	734.0
203.00	1143.0
204.00	5604.0
205.00	8547.0
206.00	37288.0
207.00	5088.0
208.00	1545.0
208.90	510.0
210.90	1701.0
211.70	516.0
213.00	157.0
215.00	366.0
216.10	826.0
217.00	11296.0
217.90	1306.0
219.70	257.0
221.10	4204.0
221.90	1502.0
223.00	2395.0
224.00	21296.0
225.00	6033.0
226.00	522.0
227.00	11390.0
227.90	1454.0
229.00	1777.0
230.10	310.0
231.00	733.0
233.90	888.0
235.00	522.0
235.90	644.0
236.90	832.0
239.00	494.0
239.90	324.0
241.00	570.0
242.10	1164.0
243.00	1315.0
244.00	16584.0
245.00	1922.0
246.00	3810.0
247.00	721.0
248.00	170.0
248.90	572.0
250.10	219.0
250.80	245.0
251.20	171.0
251.80	227.0
252.30	305.0
253.10	697.0
253.70	830.0
255.00	89232.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

256.00	12764.0
257.00	1091.0
258.00	5758.0
259.10	880.0
260.00	219.0
263.90	201.0
265.00	1818.0
265.80	486.0
270.90	510.0
272.00	288.0
272.90	2578.0
274.00	6776.0
275.00	37008.0
276.00	4810.0
277.00	3832.0
278.00	429.0
283.10	628.0
284.00	212.0
285.10	390.0
285.80	159.0
288.90	176.0
290.80	158.0
292.00	239.0
293.00	871.0
296.00	12217.0
297.00	2229.0
301.00	233.0
302.10	203.0
303.00	1991.0
304.20	300.0
307.90	283.0
314.10	496.0
314.90	1132.0
316.10	596.0
317.20	181.0
321.00	503.0
322.10	253.0
323.00	3666.0
323.90	782.0
327.00	808.0
327.90	463.0
332.10	254.0
332.80	782.0
334.00	2509.0
335.00	825.0
340.90	582.0
345.90	1372.0
347.00	175.0
352.00	1261.0
352.90	838.0
354.00	1014.0
355.10	231.0
364.90	5870.0
365.90	904.0
369.90	351.0
371.00	297.0
372.00	2401.0
373.00	908.0
382.90	627.0
384.60	172.0
389.90	166.0
391.00	186.0
392.00	228.0
402.00	594.0
403.00	960.0
403.80	338.0
420.80	1001.0
421.90	913.0
423.00	5986.0
423.90	1399.0
429.40	167.0
431.20	175.0
433.40	167.0
435.00	222.0
435.50	223.0
438.90	160.0
441.00	16472.0
442.00	112216.0
443.00	22928.0
444.00	1917.0
444.70	312.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

m/z	Abundance
38.00	940.0
39.00	3935.0
40.00	458.0
42.90	226.0
44.00	3361.0
49.10	432.0
50.00	12549.0
51.00	53648.0
52.00	2617.0
54.90	363.0
56.00	1712.0
56.90	3232.0
61.00	622.0
62.00	534.0
63.00	1667.0
64.00	378.0
65.00	852.0
69.00	35440.0
70.10	346.0
73.10	599.0
74.00	4879.0
75.00	5706.0
76.10	1653.0
77.00	45456.0
78.10	3242.0
79.00	2845.0
79.90	2291.0
81.00	3815.0
82.00	603.0
83.00	675.0
84.00	210.0
84.90	455.0
86.00	711.0
87.00	434.0
91.00	808.0
91.90	697.0
93.00	5289.0
94.10	307.0
96.10	175.0
98.00	4345.0
99.00	3200.0
100.00	368.0
101.00	1732.0
102.00	162.0
103.10	477.0
104.00	1312.0
105.00	1122.0
107.00	14319.0
108.00	2163.0
110.00	24448.0
111.00	3560.0
112.00	578.0
113.00	277.0
116.10	975.0
117.00	14217.0
118.00	1006.0
120.00	176.0
122.00	719.0
122.90	1906.0
124.00	375.0
124.90	740.0
125.60	255.0
127.00	44960.0
128.00	3126.0
129.00	19096.0
130.00	1934.0
131.00	195.0
133.00	247.0
133.90	771.0
135.00	1270.0
136.00	720.0
137.00	726.0
137.90	338.0
138.70	171.0
140.00	177.0
141.00	3170.0
142.10	860.0
142.90	539.0
143.90	270.0
145.80	366.0
147.00	1317.0
147.90	3175.0
149.00	800.0
151.10	452.0
151.80	279.0
152.90	734.0
154.00	610.0
155.00	1443.0
156.00	1983.0
157.00	543.0
157.80	650.0
158.90	361.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

159.90	1001.0
161.00	1008.0
161.70	425.0
162.10	299.0
163.00	152.0
164.10	246.0
165.00	1149.0
166.00	824.0
167.00	6871.0
168.00	3231.0
169.00	544.0
169.80	312.0
170.90	304.0
172.00	707.0
172.90	527.0
173.90	1188.0
175.00	2282.0
175.90	854.0
177.00	1142.0
177.90	451.0
179.00	4126.0
179.90	2617.0
181.00	1383.0
181.90	161.0
182.70	200.0
184.00	542.0
184.90	2290.0
186.00	14554.0
187.00	4027.0
188.00	625.0
189.00	1168.0
190.00	337.0
191.00	811.0
192.00	1492.0
193.00	1569.0
194.10	413.0
195.90	3078.0
198.00	104296.0
199.00	7325.0
200.00	1033.0
201.60	884.0
202.10	213.0
203.00	1242.0
204.00	5480.0
205.00	7642.0
206.00	29832.0
207.00	3578.0
207.90	1031.0
209.10	351.0
210.00	444.0
211.00	1193.0
211.80	412.0
214.90	646.0
216.00	674.0
217.00	9162.0
218.10	1267.0
221.00	4232.0
221.80	1048.0
222.90	1905.0
224.00	18760.0
225.00	4555.0
226.10	733.0
227.00	9068.0
228.10	777.0
229.00	2023.0
230.10	288.0
231.00	603.0
232.00	218.0
233.10	320.0
233.90	618.0
234.90	750.0
236.10	539.0
237.10	666.0
237.80	179.0
239.10	417.0
239.90	339.0
241.00	389.0
242.00	1031.0
242.90	928.0
244.00	13509.0
245.10	1812.0
246.00	3142.0
247.00	698.0
247.80	220.0
248.90	530.0
251.10	211.0
251.80	203.0
253.00	498.0
255.00	74024.0
256.00	10476.0
257.00	899.0
258.00	4792.0
258.90	897.0
259.70	150.0
265.10	1978.0
266.20	201.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

270.30	223.0
270.90	305.0
272.00	368.0
272.90	2178.0
274.00	5905.0
275.00	30520.0
276.00	4551.0
277.00	3021.0
278.00	586.0
281.80	189.0
283.00	347.0
284.10	263.0
285.10	530.0
291.00	153.0
292.40	207.0
293.10	645.0
293.90	154.0
296.00	11255.0
297.10	1588.0
303.00	1178.0
304.10	413.0
313.30	203.0
314.00	385.0
315.00	1392.0
316.10	532.0
320.90	557.0
321.90	286.0
323.00	3034.0
324.00	362.0
327.00	517.0
328.00	536.0
332.00	260.0
333.00	371.0
334.00	2682.0
335.00	811.0
341.00	328.0
342.00	159.0
346.10	571.0
346.80	163.0
351.00	168.0
352.10	995.0
353.00	932.0
354.00	1140.0
354.80	199.0
365.00	5345.0
365.90	538.0
371.00	239.0
372.00	1340.0
373.00	609.0
383.00	426.0
384.00	315.0
390.00	242.0
391.10	150.0
392.10	185.0
401.00	174.0
402.00	806.0
403.00	1027.0
403.90	354.0
412.30	155.0
421.00	775.0
422.00	855.0
423.00	5758.0
424.00	1550.0
425.10	385.0
433.40	154.0
434.60	292.0
436.50	249.0
437.30	212.0
441.00	18424.0
442.00	110808.0
443.00	22368.0
444.00	2064.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452