


Data Path : Z:\HPCHEM1\BNA G\Data\BG021518\
 Data File : BG032732.D
 Acq On : 16 Feb 2018 2:43
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG021218.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Feb 12 18:08:01 2018



AutoFind: Scans 2569, 2570, 2571; Background Corrected with Scan 2563

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	22.0	5276	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	22.1	5305	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	33.8	8087	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	23951	PASS
199	198	5	9	5.8	1392	PASS
275	198	10	60	34.3	8216	PASS
365	198	1	100	4.8	1157	PASS
441	442	0.01	24	15.8	3774	PASS
442	198	50	100	99.7	23890	PASS
443	442	15	24	19.8	4733	PASS