

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG021615\
 Data File : BG015993.D
 Acq On : 16 Feb 2015 13:01
 Operator : TP/IZ
 Sample : PB81744BL
 Misc : SBLK19
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK19

Quant Time: Feb 17 18:04:04 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG021015.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 17 13:09:30 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.81	152	53209	20.00	ng/ul	-0.02
18) Naphthalene-d8	10.60	136	240394	20.00	ng/ul	-0.02
35) Acenaphthene-d10	14.44	164	148413	20.00	ng/ul	-0.01
61) Phenanthrene-d10	17.18	188	325650	20.00	ng/ul	0.00
75) Chrysene-d12	21.36	240	358742	20.00	ng/ul	0.00
83) Perylene-d12	23.69	264	344107	20.00	ng/ul	0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.28	96	10338	7.96	ng/uL	-0.01
5) Phenol-d5	6.97	99	193643	36.68	ng/ul	-0.01
7) Bis-(2-Chloroethyl)ether-d	7.14	67	119627	39.34	ng/ul	-0.01
9) 2-Chlorophenol-d4	7.34	132	144775	36.99	ng/ul	-0.02
13) 4-Methylphenol-d8	8.51	113	148131	36.10	ng/ul	-0.01
19) Nitrobenzene-d5	8.96	128	72478	36.21	ng/ul	-0.01
22) 2-Nitrophenol-d4	9.68	143	70926	34.68	ng/ul	-0.02
26) 2,4-Dichlorophenol-d3	10.22	165	126216	34.95	ng/ul	-0.01
29) 4-Chloroaniline-d4	10.73	131	227019	56.55	ng/ul	-0.01
43) Dimethylphthalate-d6	13.85	166	426161	42.75	ng/ul	-0.01
46) Acenaphthylene-d8	14.13	160	555846	40.91	ng/ul	-0.01
51) 4-Nitrophenol-d4	14.63	143	84571	33.99	ng/ul	0.00
57) Fluorene-d10	15.43	176	403603	41.53	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.54	200	43655	22.66	ng/ul	0.00
70) Anthracene-d10	17.28	188	620735	41.50	ng/ul	0.00
76) Pyrene-d10	19.58	212	684051	42.85	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.53	264	630376	41.37	ng/ul	0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG021615\
Data File : BG015993.D
Acq On : 16 Feb 2015 13:01
Operator : TP/IZ
Sample : PB81744BL
Misc : SBLK19
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK19

Quant Time: Feb 17 18:04:04 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG021015.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Feb 17 13:09:30 2015
Response via : Initial Calibration

