


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG021616\
 Data File : BG020908.D
 Acq On : 16 Feb 2016 10:38
 Operator : SJ/UM
 Sample : DFTPP32
 Misc : LOQ 20PPM
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP32

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG020816.M
 Title : SVOA CALIBRATION
 Last Update : Wed Feb 17 12:49:19 2016



AutoFind: Scans 2585, 2586, 2587; Background Corrected with Scan 2580

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.3	4655	PASS
68	69	0.00	2	0.9	51	PASS
69	198	0.00	100	34.1	5820	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	51.3	8763	PASS
197	198	0.00	2	0.5	92	PASS
198	198	100	100	100.0	17073	PASS
199	198	5	9	6.6	1120	PASS
275	198	10	60	25.0	4272	PASS
365	198	1	100	3.7	631	PASS
441	442	0.01	100	16.5	2400	PASS
442	198	50	100	85.1	14536	PASS
443	442	15	24	19.0	2761	PASS

m/z	Abundance
39.25	446.0
50.15	1328.0
51.25	5524.0
52.25	280.0
56.15	158.0
57.15	501.0
63.15	278.0
68.25	154.0
69.15	6441.0
74.15	722.0
75.15	1112.0
76.25	454.0
77.25	8437.0
78.25	649.0
79.15	459.0
80.15	388.0
81.15	621.0
82.15	186.0
83.15	154.0
86.15	153.0
91.15	161.0
92.15	163.0
93.15	788.0
98.15	601.0
99.15	630.0
101.15	392.0
104.15	188.0
105.25	235.0
107.15	2548.0
108.15	350.0
110.15	4617.0
111.15	871.0
116.15	157.0
117.15	1682.0
118.15	160.0
122.15	176.0
123.15	353.0
127.25	9803.0
128.25	822.0
129.15	3720.0
130.25	295.0
135.15	342.0
137.25	211.0
141.15	383.0
143.25	155.0
147.15	251.0
148.15	565.0
155.15	242.0
156.25	398.0
161.25	274.0
165.15	157.0
167.25	881.0
168.25	441.0
173.25	182.0
174.25	228.0
175.25	393.0
177.25	172.0
179.25	577.0
180.25	506.0
181.25	178.0
185.25	294.0
186.25	2381.0
187.25	703.0
192.15	228.0
193.25	198.0
196.25	549.0
197.25	277.0
198.25	17496.0
199.25	1208.0
204.25	682.0
205.25	1366.0
206.25	4652.0
207.25	700.0
208.15	231.0
211.25	221.0
217.25	1257.0
218.15	244.0
221.25	930.0
223.15	284.0
224.25	2815.0
225.25	484.0
227.25	1066.0
229.25	190.0
231.25	174.0
242.35	153.0
243.25	177.0
244.35	1969.0
245.25	325.0
246.25	329.0
255.25	10035.0
256.25	1297.0

Instrument :
BNA_G
ClientSampleId :
DFTPP32

258.25	430.0
265.25	184.0
273.25	375.0
274.25	610.0
275.35	4036.0
276.35	669.0
277.35	307.0
296.35	1207.0
297.25	185.0
315.35	159.0
323.35	377.0
334.35	269.0
365.35	445.0
372.45	197.0
423.45	755.0
441.45	2036.0
442.45	12315.0
443.55	2524.0
444.45	232.0

Instrument :
BNA_G
ClientSampleId :
DFTPP32

m/z	Abundance
39.15	312.0
50.15	1244.0
51.25	5319.0
52.15	313.0
56.15	166.0
57.15	409.0
63.15	225.0
69.15	6743.0
74.15	765.0
75.15	1030.0
76.25	459.0
77.25	8538.0
78.15	555.0
79.15	328.0
80.15	506.0
81.15	646.0
91.15	158.0
92.15	175.0
93.15	784.0
98.15	686.0
99.15	547.0
101.15	366.0
103.15	157.0
104.05	241.0
107.25	2517.0
108.15	339.0
110.15	4627.0
111.15	650.0
117.15	1759.0
118.15	175.0
123.15	238.0
127.25	10209.0
128.25	810.0
129.25	3411.0
130.15	329.0
135.15	266.0
141.15	432.0
142.15	191.0
143.15	152.0
147.15	176.0
148.15	560.0
153.15	152.0
155.15	274.0
156.25	409.0
160.15	176.0
161.15	236.0
165.25	190.0
167.25	837.0
168.25	418.0
174.25	189.0
175.15	307.0
177.15	153.0
179.25	696.0
180.25	608.0
181.25	265.0
185.25	368.0
186.25	2766.0
187.25	696.0
192.25	226.0
193.25	297.0
196.25	828.0
198.25	20064.0
199.15	1381.0
204.25	701.0
205.25	1403.0
206.25	5669.0
207.25	746.0
211.35	201.0
217.25	1395.0
218.15	237.0
221.25	881.0
222.25	270.0
223.25	330.0
224.25	2924.0
225.25	744.0
227.25	1252.0
228.25	175.0
229.25	269.0
242.25	226.0
243.25	171.0
244.35	2136.0
245.35	268.0
246.25	500.0
255.25	11748.0
256.25	1685.0
258.25	647.0
265.25	272.0
273.35	428.0
274.35	906.0
275.35	5123.0
276.35	655.0

Instrument :
BNA_G
ClientSampleId :
DFTPP32

277.25	354.0
296.35	1381.0
297.35	165.0
303.35	167.0
323.35	541.0
334.35	328.0
354.35	152.0
365.35	850.0
372.45	205.0
403.45	207.0
421.45	184.0
423.45	974.0
424.45	168.0
441.45	2820.0
442.45	17408.0
443.45	3083.0
444.45	333.0

Instrument :
BNA_G
ClientSampleId :
DFTPP32

m/z	Abundance
39.15	258.0
50.15	782.0
51.25	3124.0
52.15	165.0
57.15	212.0
69.15	4276.0
74.15	370.0
75.15	678.0
76.15	258.0
77.25	5181.0
78.25	391.0
79.15	340.0
80.15	264.0
81.15	337.0
93.15	451.0
98.15	404.0
99.15	288.0
101.15	240.0
105.15	173.0
107.15	1592.0
108.25	222.0
110.15	3097.0
111.15	481.0
117.15	1161.0
123.15	165.0
127.25	6278.0
128.15	496.0
129.15	2223.0
130.25	163.0
135.25	229.0
141.25	393.0
142.25	155.0
147.15	156.0
148.15	363.0
155.15	181.0
156.25	191.0
167.25	492.0
168.15	384.0
175.25	202.0
179.15	457.0
180.25	367.0
181.25	153.0
185.25	276.0
186.25	1942.0
187.25	448.0
189.15	151.0
193.25	182.0
196.25	426.0
198.25	13659.0
199.25	772.0
204.25	342.0
205.25	830.0
206.25	3505.0
207.25	623.0
211.15	199.0
217.25	876.0
218.15	162.0
221.25	767.0
223.25	210.0
224.25	1833.0
225.25	453.0
227.25	786.0
228.25	160.0
229.25	179.0
243.25	169.0
244.25	1681.0
245.25	235.0
246.25	330.0
255.25	8039.0
256.25	1134.0
258.25	381.0
273.25	247.0
274.25	555.0
275.35	3659.0
276.25	450.0
277.25	302.0
296.35	920.0
323.35	331.0
334.35	229.0
354.45	152.0
365.35	599.0
372.35	198.0
421.45	157.0
423.45	641.0
424.45	228.0
441.45	2345.0
442.45	13887.0
443.45	2677.0
444.45	219.0

Instrument :
BNA_G
ClientSampleId :
DFTPP32

m/z	Abundance
39.15	258.0
50.15	782.0
51.25	3124.0
52.15	165.0
57.15	212.0
69.15	4276.0
74.15	370.0
75.15	678.0
76.15	258.0
77.25	5181.0
78.25	391.0
79.15	340.0
80.15	264.0
81.15	337.0
93.15	451.0
98.15	404.0
99.15	288.0
101.15	240.0
105.15	173.0
107.15	1592.0
108.25	222.0
110.15	3097.0
111.15	481.0
117.15	1161.0
123.15	165.0
127.25	6278.0
128.15	496.0
129.15	2223.0
130.25	163.0
135.25	229.0
141.25	393.0
142.25	155.0
147.15	156.0
148.15	363.0
155.15	181.0
156.25	191.0
167.25	492.0
168.15	384.0
175.25	202.0
179.15	457.0
180.25	367.0
181.25	153.0
185.25	276.0
186.25	1942.0
187.25	448.0
189.15	151.0
193.25	182.0
196.25	426.0
198.25	13659.0
199.25	772.0
204.25	342.0
205.25	830.0
206.25	3505.0
207.25	623.0
211.15	199.0
217.25	876.0
218.15	162.0
221.25	767.0
223.25	210.0
224.25	1833.0
225.25	453.0
227.25	786.0
228.25	160.0
229.25	179.0
243.25	169.0
244.25	1681.0
245.25	235.0
246.25	330.0
255.25	8039.0
256.25	1134.0
258.25	381.0
273.25	247.0
274.25	555.0
275.35	3659.0
276.25	450.0
277.25	302.0
296.35	920.0
323.35	331.0
334.35	229.0
354.45	152.0
365.35	599.0
372.35	198.0
421.45	157.0
423.45	641.0
424.45	228.0
441.45	2345.0
442.45	13887.0
443.45	2677.0
444.45	219.0

Instrument :
BNA_G
ClientSampleId :
DFTPP32

m/z Abundance

Instrument :
BNA_G
ClientSampleId :
DFTPP32