

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG021617\
 Data File : BG025851.D
 Acq On : 16 Feb 2017 13:32
 Operator : SJ/MA
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SICV21

Quant Time: Feb 16 14:09:17 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.3-EPA-BG021617MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 16 13:32:00 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	110	0.00
3	Naphthalene	0.986	1.012	-2.6	110	0.00
4	2-Methylnaphthalene	0.777	0.808	-4.0	110	0.00
5	1-Methylnaphthalene	0.758	0.785	-3.6	110	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
7 S	Acenaphthylene-d8	1.656	1.764	-6.5	110	0.00
8	Acenaphthylene	1.697	1.796	-5.8	110	0.00
9 C	Acenaphthene	1.212	1.272	-5.0	111	0.00
10 S	Fluorene-d10	1.272	1.338	-5.2	111	0.00
11	Fluorene	1.423	1.495	-5.1	110	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
13	Phenanthrene	1.067	1.128	-5.7	111	0.00
14 S	Anthracene-d10	0.904	0.954	-5.5	110	0.00
15	Anthracene	1.084	1.144	-5.5	109	0.00
16 C	Fluoranthene	1.172	1.243	-6.1	109	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
18 S	Pyrene-d10	0.964	1.013	-5.1	109	0.00
19	Pyrene	1.210	1.274	-5.3	108	0.00
20	Benzo(a)anthracene	1.125	1.183	-5.2	109	0.00
21	Chrysene	1.075	1.115	-3.7	108	0.00
22 I	Perylene-d12	1.000	1.000	0.0	111	0.00
23	Benzo(b)fluoranthene	1.156	1.191	-3.0	107	0.00
24	Benzo(k)fluoranthene	1.116	1.157	-3.7	111	0.00
25 S	Benzo(a)pyrene-d12	0.885	0.922	-4.2	109	0.00
26 C	Benzo(a)pyrene	1.118	1.165	-4.2	109	0.00
27	Indeno(1,2,3-cd)pyrene	1.310	1.350	-3.1	111	0.00
28	Dibenzo(a,h)anthracene	1.085	1.124	-3.6	110	0.00
29	Benzo(g,h,i)perylene	1.081	1.108	-2.5	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0