


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG021621\
 Data File : BG048158.D
 Acq On : 16 Feb 2021 9:47
 Operator : CG/JU
 Sample : DFTPP83
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP83

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG021321MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu Feb 18 12:52:13 2021



AutoFind: Scans 2477, 2478, 2479; Background Corrected with Scan 2471

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.1	34744	PASS
68	69	0.00	2	0.8	292	PASS
69	198	0.00	100	37.8	38528	PASS
70	69	0.00	2	0.4	164	PASS
127	198	10	80	40.4	41133	PASS
197	198	0.00	2	0.3	320	PASS
198	198	100	100	100.0	101890	PASS
199	198	5	9	7.4	7534	PASS
275	198	10	60	30.0	30549	PASS
365	198	1	100	4.1	4173	PASS
441	198	0.01	100	10.7	10881	PASS
442	198	50	100	68.9	70160	PASS
443	442	15	24	18.9	13283	PASS

Scan 2471 (17.856 min): BG048158.D
DFTPP83

m/z	Abundance
44.00	1524.0
207.00	1952.0
208.10	439.0
208.90	242.0
280.90	156.0
333.70	208.0

Instrument :
BNA_G
ClientSampleId :
DFTPP83

m/z	Abundance
38.00	420.0
39.10	3182.0
39.90	301.0
41.10	184.0
44.00	1561.0
49.10	328.0
50.10	10395.0
51.10	40096.0
52.10	1893.0
52.90	157.0
56.00	1499.0
57.00	2728.0
61.10	543.0
62.00	595.0
63.00	1576.0
64.00	194.0
65.00	841.0
65.90	153.0
69.00	45832.0
70.00	205.0
71.90	204.0
72.90	491.0
73.10	491.0
74.00	4505.0
75.00	7567.0
76.00	3099.0
77.10	48056.0
78.10	3021.0
79.00	3136.0
80.00	2691.0
81.00	3809.0
82.00	1235.0
83.00	886.0
85.00	853.0
86.00	1538.0
87.00	438.0
88.00	151.0
88.90	256.0
90.90	816.0
92.00	1220.0
93.00	6065.0
94.10	354.0
95.10	201.0
96.10	209.0
96.90	229.0
98.00	4919.0
99.00	3252.0
100.00	197.0
101.00	1545.0
103.00	717.0
104.00	1261.0
105.00	1328.0
105.90	408.0
107.00	14723.0
108.00	2776.0
109.00	600.0
110.00	23592.0
111.10	3374.0
112.00	570.0
113.20	232.0
116.00	930.0
117.00	13298.0
118.00	1029.0
119.00	179.0
119.70	220.0
122.00	1010.0
123.00	1786.0
123.90	858.0
125.00	664.0
127.00	47584.0
128.00	4382.0
129.00	21184.0
130.00	1722.0
131.00	325.0
134.00	686.0
135.00	1772.0
136.00	1019.0
137.00	1001.0
137.90	289.0
139.80	234.0
141.00	2616.0
142.00	817.0
143.00	562.0
144.10	247.0
145.90	424.0
147.00	1319.0
148.00	2960.0
149.20	511.0
149.90	177.0
151.00	364.0
152.10	202.0
153.00	805.0

Instrument :
BNA_G
ClientSampleId :
DFTPP83

154.00	850.0
155.00	1367.0
156.00	1930.0
157.10	476.0
157.90	850.0
158.90	630.0
160.00	1069.0
161.00	1488.0
161.90	373.0
163.90	271.0
164.90	1109.0
167.00	4975.0
168.00	2518.0
169.00	531.0
170.00	224.0
170.60	170.0
171.20	202.0
171.80	426.0
173.00	461.0
174.00	1357.0
175.10	2221.0
176.00	565.0
177.10	1406.0
178.00	396.0
179.00	5376.0
180.00	3044.0
181.00	1220.0
182.00	332.0
184.10	384.0
184.90	1999.0
186.00	15742.0
187.00	4771.0
188.10	736.0
189.00	1036.0
189.80	212.0
191.00	631.0
192.00	1212.0
192.90	1549.0
193.90	286.0
195.00	251.0
196.00	3514.0
196.70	962.0
198.00	108872.0
198.90	8742.0
200.00	714.0
200.90	200.0
201.60	469.0
203.00	1163.0
204.10	4669.0
205.00	8175.0
206.00	31072.0
207.00	6025.0
208.00	1727.0
209.00	510.0
210.00	442.0
211.00	1455.0
211.90	379.0
214.90	287.0
216.00	772.0
217.00	9302.0
218.00	1202.0
219.20	245.0
221.00	3931.0
221.80	1808.0
223.00	2026.0
224.00	18368.0
225.00	4140.0
225.80	555.0
227.00	8601.0
227.90	1090.0
228.90	1552.0
229.90	211.0
231.10	386.0
233.10	290.0
234.00	859.0
234.90	543.0
236.00	308.0
237.00	573.0
238.90	239.0
240.10	151.0
241.00	459.0
242.10	839.0
243.10	970.0
244.00	12908.0
245.00	2041.0
246.00	2770.0
247.00	605.0
247.90	244.0
248.90	681.0
250.00	254.0
251.70	196.0
253.20	394.0
255.00	71504.0
256.00	10407.0
257.00	796.0
258.00	4684.0

Instrument :
BNA_G
ClientSampleId :
DFTPP83

259.10	486.0
264.00	205.0
265.00	1961.0
266.00	449.0
269.40	195.0
269.90	274.0
272.20	164.0
273.00	2201.0
274.00	5939.0
275.00	29952.0
276.00	4483.0
277.00	2450.0
277.90	381.0
279.10	284.0
281.10	232.0
281.80	274.0
283.00	392.0
284.00	150.0
285.10	554.0
293.00	649.0
294.00	171.0
295.00	262.0
296.00	9384.0
297.00	1451.0
301.90	202.0
303.00	1120.0
304.00	180.0
308.10	221.0
310.00	252.0
313.00	193.0
314.10	359.0
315.00	1322.0
316.10	432.0
320.90	566.0
322.00	366.0
323.10	2839.0
324.00	840.0
326.00	189.0
327.00	855.0
332.10	173.0
333.00	190.0
334.00	2045.0
335.00	488.0
339.00	197.0
341.00	399.0
342.10	193.0
346.00	408.0
352.00	734.0
352.90	763.0
354.00	989.0
355.20	397.0
365.00	3931.0
366.00	808.0
371.10	223.0
372.00	1175.0
373.00	359.0
383.00	411.0
390.10	442.0
400.90	236.0
402.00	712.0
402.90	890.0
421.00	528.0
422.00	579.0
423.00	3088.0
424.00	1329.0
425.00	160.0
439.10	178.0
441.00	8598.0
442.00	57880.0
443.00	12173.0
444.00	1249.0
445.20	263.0

Instrument :
BNA_G
ClientSampleId :
DFTPP83

m/z	Abundance
38.10	310.0
39.10	2737.0
40.00	347.0
44.10	1460.0
44.90	171.0
49.00	367.0
50.10	11577.0
51.10	40352.0
52.00	1898.0
55.00	302.0
56.00	1173.0
57.00	2737.0
61.00	584.0
62.10	893.0
63.00	2081.0
64.20	354.0
65.10	1075.0
68.00	877.0
69.00	42824.0
69.90	288.0
73.10	292.0
74.10	4978.0
75.10	6505.0
76.10	2776.0
77.10	47896.0
78.10	3172.0
79.00	3736.0
80.00	3198.0
81.10	3992.0
82.00	880.0
82.90	642.0
84.00	177.0
85.20	483.0
86.00	1142.0
86.90	411.0
88.30	195.0
89.20	153.0
91.10	773.0
92.00	773.0
93.00	5010.0
94.00	632.0
96.10	230.0
96.90	167.0
98.00	3854.0
99.00	3563.0
100.00	322.0
101.00	1791.0
103.10	732.0
104.00	1467.0
105.10	1433.0
107.00	15281.0
108.00	2251.0
109.00	677.0
110.00	23200.0
111.00	4055.0
111.90	290.0
116.00	1032.0
117.00	14165.0
118.00	803.0
119.90	226.0
122.00	1260.0
123.00	1379.0
123.90	597.0
125.10	641.0
125.80	204.0
127.00	46384.0
128.10	3953.0
129.00	19376.0
129.90	1307.0
130.90	443.0
132.00	189.0
134.10	603.0
135.10	1495.0
136.10	707.0
137.10	646.0
137.80	244.0
139.80	297.0
141.00	3417.0
142.00	929.0
142.90	466.0
144.10	166.0
146.20	372.0
147.00	1603.0
148.00	3838.0
148.80	615.0
149.00	592.0
149.90	196.0
151.10	581.0
152.90	716.0
154.00	553.0
155.00	1902.0
156.10	1624.0

Instrument :
BNA_G
ClientSampleId :
DFTPP83

156.90	418.0
158.00	570.0
159.00	258.0
160.00	669.0
161.00	1310.0
162.00	332.0
165.00	1289.0
166.10	1256.0
167.00	5631.0
168.10	2283.0
169.10	477.0
170.10	228.0
171.10	260.0
171.80	438.0
173.00	572.0
173.90	944.0
175.00	2263.0
175.90	288.0
177.10	1100.0
178.00	635.0
179.00	4999.0
180.10	3763.0
180.90	959.0
181.90	271.0
184.00	473.0
185.00	2488.0
186.00	18624.0
187.10	5049.0
188.10	477.0
189.00	1300.0
191.10	531.0
192.00	1646.0
193.00	1491.0
194.00	528.0
195.90	3021.0
197.90	117352.0
198.90	8970.0
199.90	593.0
201.50	585.0
203.00	1089.0
204.00	4911.0
205.00	9209.0
206.00	35264.0
207.00	5697.0
208.00	1841.0
209.10	771.0
210.10	714.0
211.00	1116.0
214.90	308.0
216.00	1016.0
217.00	10702.0
218.00	1691.0
221.00	4996.0
221.90	1153.0
223.00	2846.0
224.00	19480.0
225.00	6077.0
227.00	8487.0
228.00	1527.0
229.00	1891.0
230.00	256.0
230.90	833.0
232.00	190.0
234.00	525.0
235.00	677.0
236.00	363.0
236.90	606.0
238.90	527.0
240.00	191.0
241.10	666.0
242.00	1327.0
243.00	1111.0
244.00	15065.0
245.00	1758.0
246.00	4107.0
247.00	954.0
248.00	167.0
248.90	817.0
249.90	235.0
251.60	195.0
252.20	206.0
252.90	536.0
255.00	84816.0
256.00	11731.0
257.00	896.0
258.00	6031.0
258.90	978.0
260.00	321.0
264.90	2047.0
266.00	505.0
271.10	398.0
272.00	277.0
273.00	2623.0
274.00	6977.0
275.00	37456.0
276.00	4990.0

Instrument :
BNA_G
ClientSampleId :
DFTPP83

277.00	3314.0
278.10	578.0
279.10	271.0
281.00	275.0
282.10	154.0
283.00	473.0
283.90	408.0
285.00	866.0
288.80	180.0
293.00	573.0
294.10	229.0
294.80	345.0
296.00	10600.0
297.00	1413.0
303.00	1172.0
303.90	348.0
307.90	301.0
309.90	190.0
313.20	162.0
314.00	637.0
314.90	1622.0
316.10	634.0
321.00	335.0
321.80	156.0
323.00	3493.0
324.10	670.0
326.90	754.0
327.90	265.0
329.10	253.0
331.90	332.0
333.10	428.0
334.00	1912.0
335.00	562.0
341.00	509.0
342.20	249.0
346.00	693.0
347.00	250.0
352.00	790.0
353.00	599.0
354.00	955.0
355.00	241.0
365.00	5011.0
365.80	454.0
370.90	186.0
372.00	1420.0
372.90	309.0
383.00	460.0
384.00	184.0
389.80	348.0
391.80	190.0
394.70	166.0
402.00	453.0
403.00	811.0
403.90	425.0
421.00	900.0
422.00	725.0
423.00	5317.0
424.00	1158.0
425.20	259.0
441.00	12823.0
442.00	83136.0
443.00	14854.0
444.10	1301.0

Instrument :
BNA_G
ClientSampleId :
DFTPP83

m/z	Abundance
38.30	174.0
39.10	1415.0
40.10	305.0
44.00	1159.0
50.00	6628.0
51.10	23784.0
52.00	883.0
55.90	648.0
57.00	1840.0
58.10	227.0
61.00	287.0
62.00	297.0
62.90	872.0
65.10	504.0
69.00	26928.0
73.00	356.0
74.00	2765.0
75.00	4737.0
76.10	1442.0
77.10	27272.0
78.10	1731.0
79.00	2359.0
80.00	2073.0
81.10	2475.0
82.00	748.0
83.00	278.0
84.80	349.0
86.10	801.0
87.00	389.0
89.20	152.0
90.90	524.0
92.00	812.0
93.00	3230.0
93.90	358.0
96.00	286.0
98.00	2983.0
99.10	2184.0
99.90	228.0
101.00	1179.0
103.10	457.0
104.10	645.0
105.00	875.0
106.00	532.0
107.00	10650.0
108.10	1361.0
110.00	16021.0
110.90	1823.0
112.10	547.0
117.00	8481.0
118.10	442.0
121.80	488.0
123.10	1133.0
124.10	474.0
125.00	431.0
127.00	29432.0
128.00	2485.0
129.00	13103.0
129.90	903.0
131.20	327.0
132.80	170.0
134.00	429.0
135.00	875.0
136.00	605.0
137.00	447.0
140.00	183.0
141.00	1627.0
142.10	904.0
143.00	321.0
145.10	164.0
145.90	317.0
147.10	893.0
148.00	2193.0
148.80	340.0
149.80	183.0
150.90	238.0
152.90	626.0
154.00	508.0
155.10	1116.0
156.00	1484.0
157.10	203.0
158.00	621.0
159.10	205.0
159.90	882.0
161.00	1209.0
162.10	377.0
163.70	200.0
165.00	1038.0
165.90	750.0
167.00	2751.0
168.00	1487.0
169.00	381.0
170.00	285.0

Instrument :
BNA_G
ClientSampleId :
DFTPP83

170.80	200.0
172.10	470.0
172.90	518.0
174.00	1368.0
175.10	1855.0
175.90	455.0
176.90	922.0
178.10	385.0
178.90	3149.0
180.00	1573.0
181.00	1005.0
182.00	200.0
184.10	179.0
184.90	1396.0
186.00	10881.0
187.00	3625.0
188.00	254.0
188.90	1030.0
189.80	183.0
190.80	511.0
192.00	1198.0
193.10	1021.0
193.80	175.0
194.10	153.0
196.00	2075.0
198.00	79448.0
199.00	4892.0
200.00	245.0
201.50	498.0
203.10	905.0
204.00	3346.0
205.00	5981.0
206.00	22928.0
207.00	4906.0
208.00	1434.0
209.00	433.0
210.10	574.0
210.80	1132.0
212.20	332.0
215.00	422.0
215.80	338.0
217.00	7929.0
217.90	1088.0
221.00	3067.0
222.00	1240.0
222.90	1123.0
224.00	14806.0
225.00	4356.0
226.10	441.0
227.00	6116.0
228.00	818.0
229.00	1467.0
230.10	196.0
230.90	353.0
232.90	170.0
234.10	461.0
234.90	416.0
235.90	336.0
237.00	433.0
240.00	363.0
241.00	342.0
242.00	487.0
243.00	836.0
244.00	11796.0
245.00	1482.0
246.00	2738.0
246.90	797.0
247.90	246.0
249.00	676.0
250.80	153.0
252.10	174.0
253.00	297.0
255.00	60856.0
256.00	8668.0
257.00	617.0
258.00	3636.0
259.00	663.0
260.80	203.0
264.90	1957.0
265.90	421.0
271.70	194.0
272.10	151.0
273.00	1815.0
274.00	3830.0
275.00	24240.0
276.00	3395.0
277.00	2523.0
277.90	663.0
281.00	471.0
281.90	164.0
282.80	389.0
285.00	525.0
291.90	288.0
292.90	481.0
294.10	243.0
296.00	8190.0

Instrument :
BNA_G
ClientSampleId :
DFTPP83

297.00	956.0
302.00	229.0
303.00	810.0
304.10	255.0
313.90	495.0
314.90	927.0
316.00	684.0
322.00	211.0
323.00	2501.0
324.00	535.0
327.00	502.0
328.00	285.0
333.00	239.0
334.00	1697.0
335.00	482.0
340.80	229.0
341.90	177.0
346.00	207.0
352.10	974.0
353.00	526.0
353.90	900.0
355.10	214.0
359.10	179.0
364.90	3578.0
365.70	531.0
365.90	516.0
371.10	184.0
372.00	1150.0
373.00	566.0
383.00	387.0
390.00	232.0
392.10	288.0
401.80	647.0
403.10	502.0
404.00	207.0
421.00	495.0
421.90	791.0
423.00	4534.0
424.00	809.0
441.00	11223.0
442.00	69464.0
443.00	12822.0
444.10	1327.0
451.10	200.0

Instrument :
BNA_G
ClientSampleId :
DFTPP83